

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
Lab Code: ACE SDG NO.: Q3495
Lab File ID: VX048445.D Date Analyzed: 11/03/2025
Instrument ID: MSVOA_X Time Analyzed: 09:14
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	182398	5.53	256609	6.74	229985	10.04
UPPER LIMIT	364796	6.032	513218	7.239	459970	10.537
LOWER LIMIT	91199	5.032	128305	6.239	114993	9.537
EPA SAMPLE NO.						
TW-1025-1RE	146009	5.54	263069	6.75	207013	10.04
TB	163243	5.54	253289	6.75	218816	10.04
VX1103WBL01	160467	5.53	268071	6.74	290044	10.04
VX1103WBS01	172640	5.53	288363	6.74	217740	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.