

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
 Lab Code: ACE SDG NO.: Q3495
 Lab File ID: VY023620.D Date Analyzed: 10/30/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:09
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	476273	7.71	677359	8.62	596160	11.42
UPPER LIMIT	952546	8.213	1354720	9.116	1192320	11.92
LOWER LIMIT	238137	7.213	338680	8.116	298080	10.92
EPA SAMPLE NO.						
SB-1025-2	765269	7.71	1282589	8.62	1076446	11.42
VY1030SBL01	742027	7.71	1269076	8.62	1061841	11.42
VY1030SBS01	459671	7.71	679981	8.62	588601	11.41
VY1030SBSD01	451310	7.71	657731	8.62	578799	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.