

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJEC	CT NUMBER:	VX110325					
SequenceID:	<u>VX110325</u>				NA	NO	YES
1. Chromatograms Labe	led/Compounds Identi	fied. (Field samples and Method	Blanks)				_
2. GC/MS Tuning Spec (NOTE THAT THERE		Criteria ITERIA FOR NY ASP CLP, CLP	AND NJ)				_
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series							
	calibration performed	formed Calibration performed bef within 24 hours of sample analys					
5. GC/MS Calibration R	equirements						
5a. Initial Calibration M If not met, list those con		veries which fall outside the accep	otable ranges.		<u> </u>		
5b. Continous Calibratic If not met, list those con		veries which fall outside the accep	otable ranges.				✓
6. Blank Contamination	- If yes, list compound	ls and concentrations in each blan	k:				
7. Surrogate Recoveries If not met, list those con		veries which fall outside the accep	otable ranges.				✓
8a. Matrix Spike/Matrix If not met, list those con		veries Meet Criteria veries which fall outside the accep	otable range.				
8b. Blank Spike/Blank s If not met, list those con		eries Meet Criteria veries which fall outside the accep	otable range.				
9. Internal Standard Are	a/Retention Time Shift	Meet Criteria					✓
Comments:							
MMDadoda				11/04/2025			
Analyst]	Date			



284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900, Fax: 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER:		VX110325				
SequenceID:	VX110325		1	NΑ	NO	YES
10. Analysis Holding Time Met If not met, list number of days exceeded for each sample:		ach sample:	_			
ADDITIONAL COM	MENTS:					
O3505-04 SAMPLE	has been screened at hi	gh dilution and then analyzed at lower dilution due to n	matrix of the sample			

q3500 samples initially screened as sample has sediment and looks turbid, after initial dilution lab has analyzed them straight and we have standard failure as lab used both vials lab will report analysis with failure

MMDadoda 11/04/2025 Date

Analyst



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER:		VX110325				
SequenceID:	VX110325			NA	NO	YES

MMDadoda 11/04/2025

Analyst Date