

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3506
 Lab File ID: VX048416.D Date Analyzed: 10/31/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:18
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	111736	12				
UPPER LIMIT	223472	12.5				
LOWER LIMIT	55868	11.5				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	120863	12.01				
STORAGE-BLANK-WATER-REF-5	129205	12.01				
STORAGE-BLANK-WATER-REF-4	130748	12.01				
STORAGE-BLANK-SAMPLE-MANAGEMENT	136155	12.01				
VX1031WBL01	119652	12.01				
VX1031WBS01	146834	12.01				
VX1031WBSD01	120989	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.