

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3506</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>11/03/2025 09:14</u>
Lab File ID:	<u>VX048445.D</u>	Init. Calib. Date(s):	<u>10/29/2025 10/29/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>11:12 13:03</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.510	0.440		-13.73	20
Chloromethane	0.288	0.260	0.1	-9.72	20
Vinyl Chloride	0.665	0.635		-4.51	20
Ethyl Acetate	0.408	0.407		-0.25	20
Bromomethane	0.240	0.250		4.17	20
Chloroethane	0.395	0.378		-4.3	20
Trichlorofluoromethane	1.019	0.953		-6.48	20
1,1,2-Trichlorotrifluoroethane	0.573	0.569		-0.7	20
Tert butyl alcohol	0.055	0.046		-16.36	20
1,1-Dichloroethene	0.582	0.593		1.89	20
Acrolein	0.115	0.117		1.74	20
Acrylonitrile	0.251	0.207		-17.53	20
Acetone	0.163	0.156		-4.29	20
Carbon Disulfide	1.669	1.411		-15.46	20
Methyl tert-butyl Ether	1.864	1.651		-11.43	20
Methyl Acetate	0.545	0.458		-15.96	20
Methylene Chloride	0.628	0.543		-13.53	20
trans-1,2-Dichloroethene	0.618	0.522		-15.53	20
Vinyl Acetate	1.373	1.234		-10.12	20
1,1-Dichloroethane	1.145	0.990	0.1	-13.54	20
Cyclohexane	0.994	0.888		-10.66	20
2-Butanone	0.260	0.218		-16.15	20
Carbon Tetrachloride	0.553	0.609		10.13	20
2,2-Dichloropropane	0.986	0.883		-10.45	20
cis-1,2-Dichloroethene	0.738	0.611		-17.21	20
Bromochloromethane	0.519	0.550		5.97	20
Chloroform	1.178	1.091		-7.39	20
1,1,1-Trichloroethane	1.040	1.003		-3.56	20
Methylcyclohexane	0.616	0.548		-11.04	20
1,1-Dichloropropene	0.479	0.531		10.86	20
Benzene	1.462	1.691		15.66	20
1,2-Dichloroethane	0.515	0.594		15.34	20
Trichloroethene	0.383	0.366		-4.44	20
1,2-Dichloropropane	0.365	0.354		-3.01	20
Dibromomethane	0.256	0.266		3.91	20
Bromodichloromethane	0.543	0.581		7	20
4-Methyl-2-Pentanone	0.375	0.363		-3.2	20
Toluene	0.914	0.880		-3.72	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG No.: Q3506

Instrument ID: MSVOA_X

Calibration Date/Time: 11/03/2025 09:14

Lab File ID: VX048445.D

Init. Calib. Date(s): 10/29/2025 10/29/2025

Heated Purge: (Y/N) N

Init. Calib. Time(s): 11:12 13:03

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
t-1,3-Dichloropropene	0.492	0.517		5.08	20
cis-1,3-Dichloropropene	0.531	0.542		2.07	20
1,1,2-Trichloroethane	0.331	0.322		-2.72	20
1,3-Dichloropropane	0.573	0.560		-2.27	20
2-Chloroethyl Vinyl ether	0.251	0.285		13.55	20
2-Hexanone	0.244	0.235		-3.69	20
Dibromochloromethane	0.409	0.419		2.44	20
1,2-Dibromoethane	0.342	0.342		0	20
Tetrachloroethene	0.403	0.374		-7.2	20
Chlorobenzene	1.152	1.076	0.3	-6.6	20
1,1,1,2-Tetrachloroethane	0.392	0.394		0.51	20
Hexachloroethane	0.571	0.572		0.17	20
Ethyl Benzene	1.940	1.879		-3.14	20
m/p-Xylenes	0.734	0.704		-4.09	20
o-Xylene	0.702	0.676		-3.7	20
Styrene	1.189	1.172		-1.43	20
Bromoform	0.283	0.286	0.1	1.06	20
Isopropylbenzene	3.698	4.029		8.95	20
1,1,2,2-Tetrachloroethane	0.937	1.000	0.3	6.72	20
1,2,3-Trichloropropane	0.719	0.817		13.63	20
Bromobenzene	0.914	1.028		12.47	20
n-propylbenzene	4.371	4.732		8.26	20
2-Chlorotoluene	2.624	2.585		-1.49	20
1,3,5-Trimethylbenzene	3.032	3.010		-0.73	20
4-Chlorotoluene	3.085	2.941		-4.67	20
tert-Butylbenzene	3.086	2.818		-8.68	20
1,2,4-Trimethylbenzene	3.002	2.907		-3.16	20
sec-Butylbenzene	3.846	3.538		-8.01	20
p-Isopropyltoluene	3.251	3.087		-5.05	20
1,3-Dichlorobenzene	1.745	1.647		-5.62	20
1,4-Dichlorobenzene	1.777	1.651		-7.09	20
n-Butylbenzene	3.052	2.723		-10.78	20
1,2-Dichlorobenzene	1.617	1.555		-3.83	20
1,2-Dibromo-3-Chloropropane	0.183	0.179		-2.19	20
1,2,4-Trichlorobenzene	1.116	1.006		-9.86	20
Hexachlorobutadiene	0.468	0.390		-16.67	20
Naphthalene	2.818	2.580		-8.45	20
1,2,3-Trichlorobenzene	1.013	0.928		-8.39	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3506</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>11/03/2025 09:14</u>
Lab File ID:	<u>VX048445.D</u>	Init. Calib. Date(s):	<u>10/29/2025 10/29/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>11:12 13:03</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
1,2-Dichloroethane-d4	0.719	0.649		-9.74	20
Dibromofluoromethane	0.285	0.346		21.4	20
Toluene-d8	1.258	1.255		-0.24	20
4-Bromofluorobenzene	0.471	0.531		12.74	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.