

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102925

Review By	Amit Patel	Review On	10/30/2025 2:04:38 PM
Supervise By	Maresh Dadoda	Supervise On	10/30/2025 2:06:52 PM
SubDirectory	VX102925	HP Acquire Method	HP Processing Method 82X102925W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135960 VP135996,VP135997,VP135998,VP135999,VP136000,VP136001 VP135002

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048403.D	29 Oct 2025 08:39		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048404.D	29 Oct 2025 10:02		JC/MD	Ok
3	IBLK	IBLK	VX048405.D	29 Oct 2025 10:31		JC/MD	Ok
4	IBLK	IBLK	VX048406.D	29 Oct 2025 10:51		JC/MD	Ok
5	VSTDICC001	VSTDICC001	VX048407.D	29 Oct 2025 11:12	LR-58	JC/MD	Ok,M
6	VSTDICC005	VSTDICC005	VX048408.D	29 Oct 2025 11:41		JC/MD	Ok,M
7	VSTDICC020	VSTDICC020	VX048409.D	29 Oct 2025 12:01		JC/MD	Ok,M
8	VSTDICCC050	VSTDICCC050	VX048410.D	29 Oct 2025 12:22		JC/MD	Ok,M
9	VSTDICC100	VSTDICC100	VX048411.D	29 Oct 2025 12:43		JC/MD	Ok,M
10	VSTDICC150	VSTDICC150	VX048412.D	29 Oct 2025 13:03		JC/MD	Ok,M
11	IBLK	IBLK	VX048413.D	29 Oct 2025 13:24		JC/MD	Ok
12	VSTDICV050	ICVVX102925	VX048414.D	29 Oct 2025 14:27		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX103125

Review By	Amit Patel	Review On	11/3/2025 10:48:48 AM
Supervise By	Semsettin Yesilyurt	Supervise On	11/3/2025 11:01:57 AM
SubDirectory	VX103125	HP Acquire Method	HP Processing Method 82X102925W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135989
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135991,VP135992

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048415.D	31 Oct 2025 08:06		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048416.D	31 Oct 2025 09:18	pH#lot#v14222	JC/MD	Ok,M
3	VX1031MBL01	VX1031MBL01	VX048417.D	31 Oct 2025 09:46		JC/MD	Ok
4	VX1031WBL01	VX1031WBL01	VX048418.D	31 Oct 2025 10:06		JC/MD	Ok
5	VX1031WBS01	VX1031WBS01	VX048419.D	31 Oct 2025 10:34	BS Failed Low For Com.#87,88	JC/MD	Ok,M
6	VX1031WBSD01	VX1031WBSD01	VX048420.D	31 Oct 2025 11:00	vial B pH#5.0	JC/MD	Ok,M
7	Q3480-02	RT3449	VX048421.D	31 Oct 2025 11:21	vial B pH#5.0	JC/MD	Ok,M
8	Q3500-02	462	VX048422.D	31 Oct 2025 11:41	vial A pH<2.0 ,Need Straight Run	JC/MD	Not Ok
9	Q3500-03	463	VX048423.D	31 Oct 2025 12:02	vial A pH<2.0 Need Straight Run	JC/MD	Not Ok
10	Q3500-05	461-LIQUID	VX048424.D	31 Oct 2025 12:23	vial A pH<2.0 ,Need Straight Run	JC/MD	Not Ok
11	Q3494-01	MW3	VX048425.D	31 Oct 2025 12:43	vial A pH<2.0	JC/MD	Ok
12	Q3494-02	Field	VX048426.D	31 Oct 2025 13:04	vial A pH<2.0 ,FB;hit of com.#16	JC/MD	Not Ok
13	Q3495-04	TB	VX048427.D	31 Oct 2025 13:25	vial A pH<2.0 ,TB;BS Failed Low	JC/MD	ReRun
14	Q3495-03	TW-1025-1	VX048428.D	31 Oct 2025 13:46	BS failed Low	JC/MD	ReRun
15	Q3505-06	10-17-25-WATER	VX048429.D	31 Oct 2025 14:06		JC/MD	Ok
16	Q3505-05	10-16-25-WATER	VX048430.D	31 Oct 2025 14:27	Surrogate fail	JC/MD	ReRun
17	Q3505-04	10-10-25-WATER	VX048431.D	31 Oct 2025 14:48	Need Straight Run	JC/MD	Not Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX103125

Review By	Amit Patel	Review On	11/3/2025 10:48:48 AM
Supervise By	Semsettin Yesilyurt	Supervise On	11/3/2025 11:01:57 AM
SubDirectory	VX103125	HP Acquire Method	HP Processing Method 82X102925W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135989		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135991,VP135992		

18	IBLK	IBLK	VX048432.D	31 Oct 2025 15:08		JC/MD	Ok
19	Q3506-01	STORAGE-BLANK-SO	VX048433.D	31 Oct 2025 15:29		JC/MD	Ok
20	Q3506-02	STORAGE-BLANK-WA	VX048434.D	31 Oct 2025 15:50	Surrogate fail	JC/MD	ReRun
21	Q3506-03	STORAGE-BLANK-WA	VX048435.D	31 Oct 2025 16:10	Surrogate fail	JC/MD	ReRun
22	Q3506-04	STORAGE-BLANK-SAM	VX048436.D	31 Oct 2025 16:31		JC/MD	Ok
23	Q3507-02	TB	VX048437.D	31 Oct 2025 16:52	TB;Surrogate fail	JC/MD	ReRun
24	Q3508-02	TB	VX048438.D	31 Oct 2025 17:12	TB;Surrogate fail	JC/MD	ReRun
25	Q3507-01	SW-1	VX048439.D	31 Oct 2025 17:33		JC/MD	Ok
26	Q3508-01	SW-1	VX048440.D	31 Oct 2025 17:54		JC/MD	Ok
27	Q3510-01	OUTFALL-002	VX048441.D	31 Oct 2025 18:14	Surrogate fail	JC/MD	ReRun
28	Q3511-01	OUTFALL-001	VX048442.D	31 Oct 2025 18:35	Surrogate fail	JC/MD	ReRun
29	VSTDCCC050	VSTDCCC050EC	VX048443.D	31 Oct 2025 18:56		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX110325

Review By	Maresh Dadoda	Review On	11/4/2025 3:28:58 AM
Supervise By	mohammad ahmed	Supervise On	11/4/2025 3:29:09 AM
SubDirectory	VX110325	HP Acquire Method	MSVOA_X HP Processing Method 82X102925W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136024
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136025,VP136023 VP133935

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048444.D	03 Nov 2025 08:20		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048445.D	03 Nov 2025 09:14	pH#lot#v14222	JC/MD	Ok,M
3	VX1103MBL01	VX1103MBL01	VX048446.D	03 Nov 2025 09:40		JC/MD	Ok
4	VX1103WBL01	VX1103WBL01	VX048447.D	03 Nov 2025 11:25		JC/MD	Ok
5	VX1103MBS01	VX1103WBS01	VX048448.D	03 Nov 2025 11:52		JC/MD	Not Ok
6	VX1103WBS01	VX1103WBS01	VX048449.D	03 Nov 2025 12:21		JC/MD	Ok,M
7	VX1103WBSD01	VX1103WBSD01	VX048450.D	03 Nov 2025 12:44		JC/MD	Ok,M
8	VX1103MBS02	VX1103MBS02	VX048451.D	03 Nov 2025 13:06		JC/MD	Not Ok
9	Q3506-02	STORAGE-BLANK-WA	VX048452.D	03 Nov 2025 13:37	pH<2.0 B	JC/MD	Ok
10	Q3506-03RE	STORAGE-BLANK-WA	VX048453.D	03 Nov 2025 13:58	pH<2.0 B Surrogate fail	JC/MD	Confirms
11	Q3510-01	OUTFALL-002	VX048454.D	03 Nov 2025 14:18	pH<2.0 B	JC/MD	Ok
12	Q3511-01	OUTFALL-001	VX048455.D	03 Nov 2025 14:39	pH<2.0 B	JC/MD	Ok
13	Q3495-03RE	TW-1025-1RE	VX048456.D	03 Nov 2025 14:59	pH<2.0 B Surrogate fail	JC/MD	Confirms
14	Q3495-04	TB	VX048457.D	03 Nov 2025 15:20	pH<2.0 B TB	JC/MD	Ok
15	Q3422-02	PR132-FB2-20251021	VX048458.D	03 Nov 2025 15:41	vial B,pH#5.0 FB	JC/MD	Ok
16	Q3518-02	50831	VX048459.D	03 Nov 2025 16:01	pH<2.0 A Surrogate fail	JC/MD	ReRun
17	Q3494-02	Field	VX048460.D	03 Nov 2025 16:22	pH<2.0 B FB	JC/MD	Ok
18	Q3505-05	10-16-25-WATER	VX048461.D	03 Nov 2025 16:42	pH<2.0 B	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX110325

Review By	Mahesh Dadoda	Review On	11/4/2025 3:28:58 AM		
Supervise By	mohammad ahmed	Supervise On	11/4/2025 3:29:09 AM		
SubDirectory	VX110325	HP Acquire Method	MSVOA_X	HP Processing Method	82X102925W.M
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP136024			
CCC		VP136025,VP136023			
Internal Standard/PEM		VP133935			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3500-02	10-17-25-WATER	VX048462.D	03 Nov 2025 17:03	vial B pH<2.0 Surrogate fail	JC/MD	Ok
20	Q3500-03	463	VX048463.D	03 Nov 2025 17:24	vial B pH<2 Internal standard fail; Surrogate fail	JC/MD	Ok
21	Q3500-05	461-LIQUID	VX048464.D	03 Nov 2025 17:44	vial B pH<2.0 Surrogate fail	JC/MD	Ok
22	Q3505-04	10-10-25-WATER	VX048465.D	03 Nov 2025 18:05	vial B pH<2	JC/MD	Ok
23	Q3521-03	TANK-2025	VX048466.D	03 Nov 2025 18:25	Need Straight Run	JC/MD	Not Ok
24	VIBLK	VIBLK	VX048467.D	03 Nov 2025 18:46		JC/MD	Ok
25	VIBLK	VIBLK	VX048468.D	03 Nov 2025 19:06		JC/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VX048469.D	03 Nov 2025 19:27		JC/MD	Ok,M

M : Manual Integration