

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3506</u>
Lab File ID:	<u>VX048416.D</u>	Date Analyzed:	<u>10/31/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:18</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	111736	12				
UPPER LIMIT	223472	12.5				
LOWER LIMIT	55868	11.5				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	120863	12.01				
STORAGE-BLANK-WATER-REF-4	130748	12.01				
STORAGE-BLANK-SAMPLE-MANAGEMENT	136155	12.01				
VX1031WBL01	119652	12.01				
VX1031WBS01	146834	12.01				
VX1031WBSD01	120989	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.