

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ATGG01
 Lab Code: ACE SDG NO.: Q3507
 Lab File ID: VX048416.D Date Analyzed: 10/31/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:18
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	152790	5.53	255523	6.74	229228	10.04
UPPER LIMIT	305580	6.031	511046	7.239	458456	10.537
LOWER LIMIT	76395	5.031	127762	6.239	114614	9.537
EPA SAMPLE NO.						
SW-1	104739	5.54	226642	6.75	210025	10.04
TB	110576	5.54	187149	6.75	202287	10.04
VX1031WBL01	109923	5.54	221003	6.75	229713	10.04
VX1031WBS01	179964	5.53	315014	6.74	292819	10.04
VX1031WBSD01	161570	5.53	274506	6.74	244643	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.