
From: Staci Bruce <staci.bruce@alliancetg.com>
Sent: Friday, October 31, 2025 11:57 AM
Subject: RE: sample not mentioned on COC.

For Chester and Goshen. Just the two you received yesterday for project AEC-2025-0013.



Staci Bruce
Director, Compliance
Mobile: 423-902-9441
www.alliancetg.com

From: Deepak Parmar <Deepak.Parmar@alliancetg.com>
Sent: Friday, October 31, 2025 11:46 AM
To: Staci Bruce <staci.bruce@alliancetg.com>
Subject: Re: sample not mentioned on COC.

which project? we have only two.

Thanks & Regards,



Deepak Parmar
Sr. Project Manager
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3154
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com



From: Staci Bruce <staci.bruce@alliancetg.com>
Sent: Friday, October 31, 2025 11:27 AM
To: Deepak Parmar <Deepak.Parmar@AllianceTG.com>
Subject: RE: sample not mentioned on COC.

After research it looks like a TB is required for BTEX so please analyze.

Staci Bruce
Director, Compliance
Mobile: 423-902-9441
www.alliancetg.com

From: Deepak Parmar <Deepak.Parmar@alliancetg.com>
Sent: Friday, October 31, 2025 11:25 AM
To: Staci Bruce <staci.bruce@alliancetg.com>
Subject: Re: sample not mentioned on COC.

ok

Thanks & Regards,



Deepak Parmar
Sr. Project Manager
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3154
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com



From: Staci Bruce <staci.bruce@alliancetg.com>
Sent: Friday, October 31, 2025 11:22 AM
To: Deepak Parmar <Deepak.Parmar@AllianceTG.com>
Subject: RE: sample not mentioned on COC.

Are the TB necessary to validate the data? Otherwise, I'm ok with not analyzing the TB.

Thank you,

Staci Bruce
Director, Compliance
Mobile: 423-902-9441
www.alliancetg.com

From: Deepak Parmar <Deepak.Parmar@alliancetg.com>
Sent: Friday, October 31, 2025 11:19 AM

To: Staci Bruce <staci.bruce@alliancetg.com>

Subject: sample not mentioned on COC.

hi,

The lab received two TB samples but not mentioned on COC, please let us know how to proceed with analysis .
please see attachment

Thanks & Regards,



Deepak Parmar

Sr. Project Manager

An Alliance Technical Group Company

Main: 908-789-8900

Direct: 908-728-3154

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com

