

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3508
Project:	AEC-2025-0013-CSC 4133 Goshen	Sample ID:	Q3507-01
Client ID:	SW-1DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	19.8		19.8		1	0		11/04/2025

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3508
Project:	AEC-2025-0013-CSC 4133 Goshen	Sample ID:	Q3507-01
Client ID:	SW-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	61.3		63.3		1	3.21		11/04/2025

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3508
Project:	AEC-2025-0013-CSC 4133 Goshen	Sample ID:	Q3551-01
Client ID:	MANHOLEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	25.9		26.0		1	0.39		11/07/2025
Oil and Grease	mg/L	+/-18	38.6		38.8		1	0.52		11/07/2025