

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>ATGG01</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3508</u>                 |
| Instrument ID:      | <u>MSVOA_X</u>    | Calibration Date/Time: | <u>10/31/2025 09:18</u>      |
| Lab File ID:        | <u>VX048416.D</u> | Init. Calib. Date(s):  | <u>10/29/2025 10/29/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>11:12 13:03</u>           |
| GC Column:          | <u>DB-624UI</u>   | ID:                    | <u>0.18</u> (mm)             |

| COMPOUND              | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|-----------------------|-------|--------|------------|-------|-------|
| Benzene               | 1.462 | 1.472  |            | 0.68  | 20    |
| Toluene               | 0.914 | 0.927  |            | 1.42  | 20    |
| Ethyl Benzene         | 1.940 | 1.943  |            | 0.16  | 20    |
| m/p-Xylenes           | 0.734 | 0.741  |            | 0.95  | 20    |
| o-Xylene              | 0.702 | 0.709  |            | 1     | 20    |
| 1,2-Dichloroethane-d4 | 0.719 | 0.743  |            | 3.34  | 20    |
| Dibromofluoromethane  | 0.285 | 0.296  |            | 3.86  | 20    |
| Toluene-d8            | 1.258 | 1.277  |            | 1.51  | 20    |
| 4-Bromofluorobenzene  | 0.471 | 0.470  |            | -0.21 | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.