

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3508</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ATG-GREENVILLE AEC</u>	Datafile :	<u>VX048419.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1031WBS01	Benzene	20	17.7	ug/L	89			82	109	
	Toluene	20	17.8	ug/L	89			82	110	
	Ethyl Benzene	20	16.8	ug/L	84			83	109	
	m/p-Xylenes	40	34.2	ug/L	86			82	110	
	o-Xylene	20	17.0	ug/L	85			83	109	