



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3509

Project: Linden Generating Station - LRSA

RunNo.: LB137755

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.02	7	100	90-110	10/31/2025
Sample ID:	CCV1						
Field pH		pH	7.01	7	100	90-110	10/31/2025
Sample ID:	CCV2						
Field pH		pH	7.01	7	100	90-110	10/31/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3509

Project: Linden Generating Station - LRSA

RunNo.: LB137760

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	51.184	50	102	95-105	11/04/2025
Sample ID: CCV2 COD	mg/L	49.161	50	98	95-105	11/04/2025
Sample ID: CCV3 COD	mg/L	50.173	50	100	95-105	11/04/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3509

Project: Linden Generating Station - LRSA

RunNo.: LB137760

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------