

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3509

Project: Linden Generating Station - LRSA

Run No.: LB137730

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137730BS							
BOD5	mg/L	198	197		100	1	84.6-115.4	10/31/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3509

Project: Linden Generating Station - LRSA

Run No.: LB137737

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137737BS							
TSS	mg/L	550	531		96	1	90-110	11/03/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3509

Project: Linden Generating Station - LRSA

Run No.: LB137760

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137760BS							
COD	mg/L	50	49.2		98	1	90-110	11/04/2025