

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC
Project: AEC-2025-0013- CSC 4153

SDG No.: Q3510
RunNo.: LB137760

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	51.184	50	102	95-105	11/04/2025
Sample ID: CCV2 COD	mg/L	49.161	50	98	95-105	11/04/2025
Sample ID: CCV3 COD	mg/L	50.173	50	100	95-105	11/04/2025