

VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>ATGG01</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3510</u>                 |
| Instrument ID:      | <u>MSVOA_X</u>    | Calibration Date/Time: | <u>11/03/2025 09:14</u>      |
| Lab File ID:        | <u>VX048445.D</u> | Init. Calib. Date(s):  | <u>10/29/2025 10/29/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>11:12 13:03</u>           |
| GC Column:          | <u>DB-624UI</u>   | ID:                    | <u>0.18</u> (mm)             |

| COMPOUND              | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|-----------------------|-------|--------|------------|-------|-------|
| Benzene               | 1.462 | 1.691  |            | 15.66 | 20    |
| Toluene               | 0.914 | 0.880  |            | -3.72 | 20    |
| Ethyl Benzene         | 1.940 | 1.879  |            | -3.14 | 20    |
| m/p-Xylenes           | 0.734 | 0.704  |            | -4.09 | 20    |
| o-Xylene              | 0.702 | 0.676  |            | -3.7  | 20    |
| 1,2-Dichloroethane-d4 | 0.719 | 0.649  |            | -9.74 | 20    |
| Dibromofluoromethane  | 0.285 | 0.346  |            | 21.4  | 20    |
| Toluene-d8            | 1.258 | 1.255  |            | -0.24 | 20    |
| 4-Bromofluorobenzene  | 0.471 | 0.531  |            | 12.74 | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.