

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3510</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ATG-GREENVILLE AEC</u>	Datafile :	<u>VX048449.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1103WBS01	Benzene	20	19.2	ug/L	96			82	109	
	Toluene	20	17.6	ug/L	88			82	110	
	Ethyl Benzene	20	20.5	ug/L	103			83	109	
	m/p-Xylenes	40	42.2	ug/L	106			82	110	
	o-Xylene	20	20.7	ug/L	104			83	109	