

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3510</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ATG-GREENVILLE AEC</u>	Datafile :	<u>VX048450.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1103WBSD01	Benzene	20	22.8	ug/L	114	17	* *	82	109	15
	Toluene	20	19.7	ug/L	99	12		82	110	16
	Ethyl Benzene	20	20.3	ug/L	102	1		83	109	16
	m/p-Xylenes	40	40.7	ug/L	102	4		82	110	15
	o-Xylene	20	19.8	ug/L	99	5		83	109	20