



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX110425\
Data File : VX048470.D
Acq On : 04 Nov 2025 08:38
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Title : SW846 8260
Last Update : Thu Oct 30 02:37:17 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.0	8161	PASS
75	95	30	60	49.4	22400	PASS
95	95	100	100	100.0	45373	PASS
96	95	5	9	6.7	3046	PASS
173	174	0.00	2	1.0	345	PASS
174	95	50	100	75.7	34341	PASS
175	174	5	9	7.3	2510	PASS
176	174	95	101	95.6	32813	PASS
177	176	5	9	6.8	2243	PASS