

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: VX110425

SequenceID : VX110425

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	<u>✓</u>
2. GC/MS Tuning Specifications. BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	<u>✓</u>
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	<u>✓</u>
4. GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	<u>✓</u>
5. GC/MS Calibration Requirements	_____	_____	<u>✓</u>
5a. Initial Calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	<u>✓</u>	_____	_____
5b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	_____	_____	<u>✓</u>
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	<u>✓</u>	_____
7. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	_____	_____	<u>✓</u>
8a. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u>✓</u>	_____	_____
8b. Blank Spike/Blank spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	<u>✓</u>
9. Internal Standard Area/Retention Time Shift Meet Criteria	_____	_____	<u>✓</u>
Comments:			

MMDadoda

Analyst

11/04/2025

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: VX110425

SequenceID : VX110425

NA NO YES

10. Analysis Holding Time Met

\_\_\_\_

If not met, list number of days exceeded for each sample:

\_\_\_\_

\_\_\_\_ ✓

ADDITIONAL COMMENTS:

Q3502-04 sample is oil ,lab has analyzed 1 gm of sample in to 10ml methanol and loaded straight 100x for analysis,sample has high hits of heavy analytes

q3488 samples are wipe samples and lab has analyzed them as medium level analysis

MMDadoda  
Analyst

11/04/2025  
Date



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: VX110425

SequenceID : VX110425 NA NO YES

MMDadoda  
Analyst

11/04/2025  
Date