

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3511</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ATG-GREENVILLE AEC</u>	Datafile :	<u>VX048485.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1104WBSD02	Benzene	20	18.6	ug/L	93	10		82	109	15
	Toluene	20	16.8	ug/L	84	4		82	110	16
	Ethyl Benzene	20	18.1	ug/L	91	8		83	109	16
	m/p-Xylenes	40	36.4	ug/L	91	16	*	82	110	15
	o-Xylene	20	18.5	ug/L	93	16		83	109	20