

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ATGG01
 Lab Code: ACE SDG NO.: Q3511
 Lab File ID: VX048471.D Date Analyzed: 11/04/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:10
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	141843	5.53	232065	6.74	217919	10.04
UPPER LIMIT	283686	6.025	464130	7.239	435838	10.537
LOWER LIMIT	70921.5	5.025	116033	6.239	108960	9.537
EPA SAMPLE NO.						
OUTFALL-002	142347	5.54	286906	6.75	258868	10.04
VX1104WBL01	132762	5.54	255401	6.75	268346	10.04
VX1104WBS02	167270	5.54	242904	6.75	206293	10.04
VX1104WBSD02	159074	5.53	271466	6.74	244989	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.