



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

CASE NARRATIVE

NYU Langone Health

Project Name: NYU Clinical Lab Water Testing 2025 - H252243895

Project # N/A

Order ID # Q3524

Test Name: Metals Group3,pH

A. Number of Samples and Date of Receipt:

4 Water samples were received on 10/31/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Metals Group3,pH. This data package contains results for Metals Group3(200.7), pH(9040C).

C. Analytical Techniques:

Metals Group3 : The analysis and digestion of Metals Group3 was based on method 200.7.

Wetchem : The analysis of pH was based on method 9040C.

D. QA/ QC Samples:

The Holding Times were met for all analysis except following Wetchem : TH-404-DI-4 of pH, for TH-430-DI-1 of pH, for TH-430-DI-2 of pH and for TH-430-DI-3 of pH as samples were receive out of holding time.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the acceptable requirements.

The RPD recoveries met criteria.

The Blank Spike met requirements for all compounds

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Duplicate analysis met criteria for all samples.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____