

Report of Analysis

Client:	NYU Langone Health	Date Collected:	10/31/25
Project:	NYU Clinical Lab Water Testing 2025 - H252243895	Date Received:	10/31/25
Client Sample ID:	TH-430-DI-2	SDG No.:	Q3524
Lab Sample ID:	Q3524-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
Dissolved Silica	Dissolved Silica	40.0	U	1	40.0	428	ug/L	11/03/25 10:40	11/05/25 13:45	EPA 200.7	M200.7

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits