

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137798

Review By	Janvi	Review On	11/6/2025 3:46:12 PM
Supervise By	mohan	Supervise On	11/10/2025 8:18:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/05/25 11:07		Jaswal	OK
2	S1	S1	CAL2	11/05/25 11:11		Jaswal	OK
3	S2	S2	CAL3	11/05/25 11:15		Jaswal	OK
4	S3	S3	CAL4	11/05/25 11:19		Jaswal	OK
5	S4	S4	CAL5	11/05/25 11:23		Jaswal	OK
6	S5	S5	CAL6	11/05/25 11:27		Jaswal	OK
7	ICV01	ICV01	ICV	11/05/25 11:32		Jaswal	OK
8	LLICV01	LLICV01	LLICV	11/05/25 11:36		Jaswal	OK
9	ICB01	ICB01	ICB	11/05/25 11:41		Jaswal	OK
10	CRI01	CRI01	CRDL	11/05/25 11:49		Jaswal	OK
11	ICSA01	ICSA01	ICSA	11/05/25 11:53		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	11/05/25 12:11		Jaswal	OK
13	ICSADL	ICSADL	ICSA	11/05/25 12:15		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	11/05/25 12:19		Jaswal	OK
15	CCV01	CCV01	CCV	11/05/25 12:23		Jaswal	OK
16	CCB01	CCB01	CCB	11/05/25 12:33		Jaswal	OK
17	LR-1	LR-1	HIGH STD	11/05/25 12:39		Jaswal	OK
18	LR-2	LR-2	HIGH STD	11/05/25 12:44		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137798

Review By	Janvi	Review On	11/6/2025 3:46:12 PM
Supervise By	mohan	Supervise On	11/10/2025 8:18:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

19	LR-3	LR-3	HIGH STD	11/05/25 12:48		Jaswal	OK
20	Q3512-01DL	10200DL	SAM	11/05/25 12:53		Jaswal	OK
21	Q3495-01	SB-1025-1	SAM	11/05/25 12:57		Jaswal	OK
22	Q3495-01DUP	SB-1025-1DUP	DUP	11/05/25 13:01		Jaswal	OK
23	Q3495-01L	SB-1025-1L	SD	11/05/25 13:05		Jaswal	OK
24	Q3495-01MS	SB-1025-1MS	MS	11/05/25 13:09		Jaswal	OK
25	Q3495-01MSD	SB-1025-1MSD	MSD	11/05/25 13:13		Jaswal	OK
26	Q3495-01A	SB-1025-1A	PS	11/05/25 13:17	0.1 ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
27	CCV02	CCV02	CCV	11/05/25 13:21		Jaswal	OK
28	CCB02	CCB02	CCB	11/05/25 13:25		Jaswal	OK
29	PB170376BL	PB170376BL	MB	11/05/25 13:32		Jaswal	OK
30	PB170376BS	PB170376BS	LCS	11/05/25 13:36		Jaswal	OK
31	Q3524-01	TH-430-DI-1	SAM	11/05/25 13:40		Jaswal	OK
32	Q3524-02	TH-430-DI-2	SAM	11/05/25 13:45		Jaswal	OK
33	Q3524-03	TH-430-DI-3	SAM	11/05/25 13:49		Jaswal	OK
34	Q3524-04	TH-404-DI-4	SAM	11/05/25 13:53		Jaswal	OK
35	Q3524-04DUP	TH-404-DI-4DUP	DUP	11/05/25 13:58		Jaswal	OK
36	Q3524-04L	TH-404-DI-4L	SD	11/05/25 14:02		Jaswal	OK
37	Q3524-04MS	TH-404-DI-4MS	MS	11/05/25 14:06		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137798

Review By	Janvi	Review On	11/6/2025 3:46:12 PM
Supervise By	mohan	Supervise On	11/10/2025 8:18:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

38	Q3524-04MSD	TH-404-DI-4MSD	MSD	11/05/25 14:10		Jaswal	OK
39	CCV03	CCV03	CCV	11/05/25 14:14		Jaswal	OK
40	CCB03	CCB03	CCB	11/05/25 14:18		Jaswal	OK
41	Q3524-04A	TH-404-DI-4A	PS	11/05/25 14:23	0.1 ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
42	PBS	PBS	LCS	11/05/25 14:27	Not Use	Jaswal	Not Ok
43	LCSS	LCSS	SAM	11/05/25 14:31	Not Use	Jaswal	Not Ok
44	Q3483-11	HW1025-PT-BNA-SO	SAM	11/05/25 14:38	Not Use	Jaswal	Not Ok
45	Q3483-12	HW1025-PT-TRIAZIN	SAM	11/05/25 14:42	Not Use	Jaswal	Not Ok
46	Q3483-11DL	HW1025-PT-BNA-SO	SAM	11/05/25 14:48	Not Use	Jaswal	Not Ok
47	Q3483-12DL	HW1025-PT-TRIAZIN	SAM	11/05/25 14:52	Not Use	Jaswal	Not Ok
48	CCV04	CCV04	CCV	11/05/25 15:16		Jaswal	OK
49	CCB04	CCB04	CCB	11/05/25 15:21		Jaswal	OK
50	PB170392BL	PB170392BL	MB	11/05/25 15:25		Jaswal	OK
51	PB170392BS	PB170392BS	LCS	11/05/25 15:29		Jaswal	OK
52	Q3534-01	MW-6	SAM	11/05/25 15:33		Jaswal	OK
53	Q3534-02	MW-1	SAM	11/05/25 15:38	Na high, need dilution	Jaswal	Dilution
54	Q3534-03	MW-11	SAM	11/05/25 15:42	Na high, need dilution	Jaswal	Dilution
55	Q3534-04	MW-5	SAM	11/05/25 15:47		Jaswal	OK
56	Q3534-05	MW-EB-1	SAM	11/05/25 15:51		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137798

Review By	Janvi	Review On	11/6/2025 3:46:12 PM
Supervise By	mohan	Supervise On	11/10/2025 8:18:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

57	Q3534-06	MW-EB-2	SAM	11/05/25 15:55	Na high, need dilution	Jaswal	Dilution
58	Q3534-07	MW-EB-3	SAM	11/05/25 16:00	Na high, need dilution	Jaswal	Dilution
59	Q3534-09	FB-1	SAM	11/05/25 16:04		Jaswal	OK
60	CCV05	CCV05	CCV	11/05/25 16:17		Jaswal	OK
61	CCB05	CCB05	CCB	11/05/25 16:25		Jaswal	OK
62	Q3537-03	MW-17-PURGE-WAT	SAM	11/05/25 16:30	CCB,CCV Fail for Na	Jaswal	Not Ok
63	Q3541-01	N48969	SAM	11/05/25 16:34	CCB,CCV Fail for Na	Jaswal	Not Ok
64	Q3542-01	AUD-25-182-194	SAM	11/05/25 16:38	CCB,CCV Fail for Na	Jaswal	Not Ok
65	Q3542-01DUP	AUD-25-182-194DUP	DUP	11/05/25 16:42	CCB,CCV Fail for Na	Jaswal	Not Ok
66	Q3542-01L	AUD-25-182-194L	SD	11/05/25 16:47	CCB,CCV Fail for Na	Jaswal	Not Ok
67	Q3542-01MS	AUD-25-182-194MS	MS	11/05/25 16:51	CCB,CCV Fail for Na	Jaswal	Not Ok
68	Q3542-01MSD	AUD-25-182-194MSD	MSD	11/05/25 16:55	CCB,CCV Fail for Na	Jaswal	Not Ok
69	Q3542-01A	AUD-25-182-194A	PS	11/05/25 17:00	CCB,CCV Fail for Na (0.1 ml of each m6008 and m6017 in 10ml sample.)	Jaswal	Not Ok
70	CCV06	CCV06	CCV	11/05/25 17:04		Jaswal	OK
71	CCB06	CCB06	CCB	11/05/25 17:08		Jaswal	OK