



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Hit Summary Sheet**  
**SW-846**

<b>SDG No.:</b>	Q3524	<b>Order ID:</b>	Q3524
<b>Client:</b>	NYU Langone Health	<b>Project ID:</b>	NYU Clinical Lab Water Testing 2025 - H2

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								



# SAMPLE DATA

A

B

C

D

## Report of Analysis

Client:	NYU Langone Health	Date Collected:	10/31/25
Project:	NYU Clinical Lab Water Testing 2025 - H252243895	Date Received:	10/31/25
Client Sample ID:	TH-430-DI-1	SDG No.:	Q3524
Lab Sample ID:	Q3524-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
Dissolved Silica	Dissolved Silica	40.0	U	1	40.0	428	ug/L	11/03/25 10:40	11/05/25 13:40	EPA 200.7	M200.7

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

## Report of Analysis

Client:	NYU Langone Health	Date Collected:	10/31/25
Project:	NYU Clinical Lab Water Testing 2025 - H252243895	Date Received:	10/31/25
Client Sample ID:	TH-430-DI-2	SDG No.:	Q3524
Lab Sample ID:	Q3524-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
Dissolved Silica	Dissolved Silica	40.0	U	1	40.0	428	ug/L	11/03/25 10:40	11/05/25 13:45	EPA 200.7	M200.7

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N =Spiked sample recovery not within control limits

## Report of Analysis

Client:	NYU Langone Health	Date Collected:	10/31/25
Project:	NYU Clinical Lab Water Testing 2025 - H252243895	Date Received:	10/31/25
Client Sample ID:	TH-430-DI-3	SDG No.:	Q3524
Lab Sample ID:	Q3524-03	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
Dissolved Silica	Dissolved Silica	40.0	U	1	40.0	428	ug/L	11/03/25 10:40	11/05/25 13:49	EPA 200.7	M200.7

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N =Spiked sample recovery not within control limits

## Report of Analysis

Client:	NYU Langone Health	Date Collected:	10/31/25
Project:	NYU Clinical Lab Water Testing 2025 - H252243895	Date Received:	10/31/25
Client Sample ID:	TH-404-DI-4	SDG No.:	Q3524
Lab Sample ID:	Q3524-04	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
Dissolved Silica	Dissolved Silica	40.0	U	1	40.0	428	ug/L	11/03/25 10:40	11/05/25 13:53	EPA 200.7	M200.7

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q3524	OrderDate:	10/31/2025 3:09:00 PM
Client:	NYU Langone Health	Project:	NYU Clinical Lab Water Testing 2025 - H252243895
Contact:	Marie-Ange Exilhomme	Location:	D41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3524-01	TH-430-DI-1	Water	Metals Group3	200.7	10/31/25	11/03/25	11/05/25	10/31/25
Q3524-02	TH-430-DI-2	Water	Metals Group3	200.7	10/31/25	11/03/25	11/05/25	10/31/25
Q3524-03	TH-430-DI-3	Water	Metals Group3	200.7	10/31/25	11/03/25	11/05/25	10/31/25
Q3524-04	TH-404-DI-4	Water	Metals Group3	200.7	10/31/25	11/03/25	11/05/25	10/31/25