

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038118.D
 Acq On : 28 Oct 2025 17:51
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102825

Quant Time: Oct 29 17:07:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.00
2	1,4-Dioxane	0.400	0.440	-10.0	90	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	82	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	76	0.00
5 S	Phenol-d6	0.400	0.351	12.3	74	0.00
6	bis(2-Chloroethyl)ether	0.400	0.427	-6.7	89	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	84	0.01
8 S	Nitrobenzene-d5	0.400	0.353	11.8	76	0.00
9	Naphthalene	0.400	0.407	-1.7	87	0.00
10	Hexachlorobutadiene	0.400	0.435	-8.7	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	89	0.00
12	2-Methylnaphthalene	0.400	0.384	4.0	84	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	0.01
14 S	2,4,6-Tribromophenol	0.400	0.340	15.0	81	0.00
15 S	2-Fluorobiphenyl	0.400	0.411	-2.7	90	0.01
16	Acenaphthylene	0.400	0.455	-13.7	103	0.00
17	Acenaphthene	0.400	0.398	0.5	89	0.01
18	Fluorene	0.400	0.414	-3.5	92	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.411	-2.7	86	0.00
21	4-Bromophenyl-phenylether	0.400	0.400	0.0	93	0.00
22	Hexachlorobenzene	0.400	0.426	-6.5	95	0.00
23	Atrazine	0.400	0.370	7.5	90	0.00
24	Pentachlorophenol	0.400	0.351	12.3	93	0.00
25	Phenanthrene	0.400	0.408	-2.0	94	0.00
26	Anthracene	0.400	0.406	-1.5	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.389	2.8	91	0.00
28	Fluoranthene	0.400	0.380	5.0	89	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	87	0.00
30	Pyrene	0.400	0.422	-5.5	90	0.00
31 S	Terphenyl-d14	0.400	0.428	-7.0	93	0.00
32	Benzo(a)anthracene	0.400	0.402	-0.5	92	0.00
33	Chrysene	0.400	0.428	-7.0	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.358	10.5	81	0.00
35 I	Perylene-d12	0.400	0.400	0.0	87	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.406	-1.5	91	0.00
37	Benzo(b)fluoranthene	0.400	0.405	-1.3	93	0.00
38	Benzo(k)fluoranthene	0.400	0.439	-9.7	102	0.00
39 C	Benzo(a)pyrene	0.400	0.437	-9.2	99	0.00
40	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	93	0.00
41	Benzo(g,h,i)perylene	0.400	0.392	2.0	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN102825