

Surrogate Summary

SW-846

SDG No.: Q3525

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170381BL	PB170381BL	2-Methylnaphthalene-d10	0.4	0.33	83		30	150
		Fluoranthene-d10	0.4	0.35	88		30	150
		Nitrobenzene-d5	0.4	0.35	88		55	111
		2-Fluorobiphenyl	0.4	0.39	97		53	106
		Terphenyl-d14	0.4	0.46	114		58	132
PB170381BS	PB170381BS	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.30	76		30	150
		Nitrobenzene-d5	0.4	0.37	92		55	111
		2-Fluorobiphenyl	0.4	0.42	104		53	106
		Terphenyl-d14	0.4	0.48	121		58	132
PB170381BSD	PB170381BSD	2-Methylnaphthalene-d10	0.4	0.29	73		30	150
		Fluoranthene-d10	0.4	0.25	63		30	150
		Nitrobenzene-d5	0.4	0.31	76		55	111
		2-Fluorobiphenyl	0.4	0.32	81		53	106
		Terphenyl-d14	0.4	0.36	91		58	132
Q3525-01	RW5-SP100-20251030	2-Methylnaphthalene-d10	0.4	0.26	66		30	150
		Fluoranthene-d10	0.4	0.34	85		30	150
		Nitrobenzene-d5	0.4	0.31	77		55	111
		2-Fluorobiphenyl	0.4	0.31	76		53	106
		Terphenyl-d14	0.4	0.47	118		58	132
Q3525-01DL	RW5-SP100-20251030DL	2-Methylnaphthalene-d10	0.4	0.27	68		30	150
		Fluoranthene-d10	0.4	0.34	84		30	150
		Nitrobenzene-d5	0.4	0.31	79		55	111
		2-Fluorobiphenyl	0.4	0.31	77		53	106
		Terphenyl-d14	0.4	0.55	137	*	58	132
Q3525-02	RW5-SP201-20251030	2-Methylnaphthalene-d10	0.4	0.28	69		30	150
		Fluoranthene-d10	0.4	0.34	86		30	150
		Nitrobenzene-d5	0.4	0.29	72		55	111
		2-Fluorobiphenyl	0.4	0.31	77		53	106
		Terphenyl-d14	0.4	0.57	143	*	58	132
Q3525-03	RW5-SP303-20251030	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.33	83		30	150
		Nitrobenzene-d5	0.4	0.31	77		55	111
		2-Fluorobiphenyl	0.4	0.35	86		53	106
		Terphenyl-d14	0.4	0.56	140	*	58	132