

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3525

Client ID : SSTDCCC0.4

Date Analyzed: 11/04/2025

Lab File ID: BN038130.D

Time Analyzed: 10:40

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8599	7.775	24973	10.57	13947	14.44
UPPER LIMIT	17198	8.275	49946	11.073	27894	14.941
LOWER LIMIT	4299.5	7.275	12486.5	10.073	6973.5	13.941
EPA SAMPLE NO.						
01 PB170381BL	8221	7.79	22221	10.59	11601	14.44
02 PB170381BS	11523	7.78	31801	10.57	16154	14.44
03 PB170381BSD	10823	7.78	30534	10.57	15616	14.43
04 RW5-SP100-20251030	8620	7.78	25357	10.57	13158	14.43
05 RW5-SP201-20251030	8491	7.78	25672	10.58	14049	14.44
06 RW5-SP303-20251030	8154	7.78	24363	10.58	13426	14.43
07 RW5-SP100-20251030DL	8331	7.78	24076	10.57	13164	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.