

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038129.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3525
DFTPP Injection Date: 11/04/2025
DFTPP Injection Time: 10:00

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	77.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 (20) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038130.D	11/04/2025	10:40
PB170381BL	PB170381BL	BN038131.D	11/04/2025	11:43
PB170381BS	PB170381BS	BN038132.D	11/04/2025	12:19
PB170381BSD	PB170381BSD	BN038133.D	11/04/2025	12:56
RW5-SP100-20251030	Q3525-01	BN038135.D	11/04/2025	14:08
RW5-SP201-20251030	Q3525-02	BN038136.D	11/04/2025	14:45
RW5-SP303-20251030	Q3525-03	BN038137.D	11/04/2025	15:21
RW5-SP100-20251030DL	Q3525-01DL	BN038144.D	11/04/2025	19:35
SSTDCCC0.4EC	SSTDCCC0.4	BN038145.D	11/04/2025	20:12