

DATA PACKAGE
SEMI-VOLATILE ORGANICS

PROJECT NAME : NWIRP BETHPAGE 112G08005-WE13

TETRA TECH NUS, INC.
661 Andersen Drive
Suite 200
Pittsburgh, PA - 15220-2745
Phone No: 412-921-7090

ORDER ID : Q3526
ATTENTION : Ernie Wu



Laboratory Certification ID # 20012



1) SEMI-VOLATILE DATA	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For SVOC-SIMGroup1	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) LCS/LCSD Summary	15
9.3) Method Blank Summary	17
9.4) GS/MS Tune Summary	18
9.5) Internal Standard Area and RT Summary	20
10) Sample Data	22
10.1) RW7-SP100-20251030	23
10.2) RW7-SP201-20251030	35
10.3) RW7-SP302-20251030	46
10.4) RW7-SP303-20251030	57
11) Calibration Data Summary	68
11.1) Initial Calibration Data	3
11.1.1) BN102825	69
11.2) Continued Calibration Data	253
11.2.1) BN038130.D	253
11.2.2) BN038145.D	279
12) QC Sample Data	305
12.1) Tune Raw Data	306
12.2) Method Blank Data	314
12.3) LCS Data	324
12.4) LCSD Data	348
13) Manual Integration	372
14) Analytical Runlogs	374
15) Extraction Logs	378
15.1) PB170381.pdf	378
15.2) PB170381IC.pdf	380

Table Of Contents for Q3526

16) Standard Prep Logs	381
17) Shipping Document	434
17.1) Chain Of Custody	435
17.2) Lab Certificate	436
18) Not Reviewed Data	437

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Cover Page

Order ID : Q3526

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q3526-01
Q3526-02
Q3526-03
Q3526-04

Client Sample Number

RW7-SP100-20251030
RW7-SP201-20251030
RW7-SP302-20251030
RW7-SP303-20251030

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:33 pm, Nov 11, 2025

Date: 11/10/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager : Ernie Wu

Order ID # Q3526

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

4 Water samples were received on 11/03/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-Semi Volatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for RW7-SP201-20251030 [Terphenyl-d14 - 140%], RW7-SP302-20251030 [Terphenyl-d14 - 148%] and RW7-SP303-20251030 [Terphenyl-d14 - 163%], The failure surrogates not associated with the client parameters list, therefore no corrective action was taken.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD for met criteria except for except for PB170381BSD for 1,4-Dioxane. %RPD was affected due to marginal difference in recoveries.

The Blank Spike Duplicate for {PB170381BSD} with File ID: BN038133.D met requirements for all compounds except for 1,4-Dioxane[65%], Failed marginally biased low, however, it is meeting QC criteria in the continues calibration.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).”

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:34 pm, Nov 11, 2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3526

MATRIX: Water

METHOD: 8270-Modified/3510

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. The Surrogate recoveries were met for all analysis except for PB170381BL [2,4,6- Tribromophenol - 29%], RW7-SP201-20251030 [Terphenyl-d14 - 140%], RW7- SP302-20251030 [Terphenyl-d14 - 148%] and RW7-SP303-20251030 [Terphenyl-d14 - 163%], The failure surrogates not associated with the client parameters list, therefore no corrective action was taken.		✓	
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria The RPD for met criteria except for except for PB170381BSD for 1,4-Dioxane. %RPD was affected due to marginal difference in recoveries. If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike Duplicate for {PB170381BSD} with File ID: BN038133.D met requirements for all compounds except for 1,4-Dioxane[65%], Failed marginally biased low, however, it is meeting QC criteria in the continues calibration..		✓	
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY
(CONTINUED)

		NA	NO	YES
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).”
The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.
The not QT review data is reported in the Miscellaneous.

QA REVIEW

REVIEWED

Sohil Jodhani, QA/QC Director , 11/11/2025, 9:19:37 AM

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3526

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 11/10/2025

LAB CHRONICLE

OrderID:	Q3526	OrderDate:	11/3/2025 9:28:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	A12

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3526-01	RW7-SP100-2025103 0	Water			10/30/25			11/03/25
			SVOC-SIMGroup1	8270-Modified		11/03/25	11/04/25	
Q3526-02	RW7-SP201-2025103 0	Water			10/30/25			11/03/25
			SVOC-SIMGroup1	8270-Modified		11/03/25	11/04/25	
Q3526-03	RW7-SP302-2025103 0	Water			10/30/25			11/03/25
			SVOC-SIMGroup1	8270-Modified		11/03/25	11/04/25	
Q3526-04	RW7-SP303-2025103 0	Water			10/30/25			11/03/25
			SVOC-SIMGroup1	8270-Modified		11/03/25	11/04/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q3526

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :	RW7-SP100-20251030							
Q3526-01	RW7-SP100-20251030	WATER	1,4-Dioxane	2.900	Q	0.07	0.2	0.2 ug/L
Total Svoc :				2.90				
Total Concentration:				2.90				



QC SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Surrogate Summary

SW-846

SDG No.: Q3526

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170381BL	PB170381BL	2-Methylnaphthalene-d10	0.4	0.33	83		30	150
		Fluoranthene-d10	0.4	0.35	88		30	150
		Nitrobenzene-d5	0.4	0.35	88		55	111
		2-Fluorobiphenyl	0.4	0.39	97		53	106
		Terphenyl-d14	0.4	0.46	114		58	132
PB170381BS	PB170381BS	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.30	76		30	150
		Nitrobenzene-d5	0.4	0.37	92		55	111
		2-Fluorobiphenyl	0.4	0.42	104		53	106
		Terphenyl-d14	0.4	0.48	121		58	132
PB170381BSD	PB170381BSD	2-Methylnaphthalene-d10	0.4	0.29	73		30	150
		Fluoranthene-d10	0.4	0.25	63		30	150
		Nitrobenzene-d5	0.4	0.31	76		55	111
		2-Fluorobiphenyl	0.4	0.32	81		53	106
		Terphenyl-d14	0.4	0.36	91		58	132
Q3526-01	RW7-SP100-20251030	2-Methylnaphthalene-d10	0.4	0.28	71		30	150
		Fluoranthene-d10	0.4	0.43	107		30	150
		Nitrobenzene-d5	0.4	0.28	70		55	111
		2-Fluorobiphenyl	0.4	0.31	77		53	106
		Terphenyl-d14	0.4	0.51	128		58	132
Q3526-02	RW7-SP201-20251030	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.31	76		55	111
		2-Fluorobiphenyl	0.4	0.31	77		53	106
		Terphenyl-d14	0.4	0.56	140	*	58	132
Q3526-03	RW7-SP302-20251030	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.33	81		55	111
		2-Fluorobiphenyl	0.4	0.36	90		53	106
		Terphenyl-d14	0.4	0.59	148	*	58	132
Q3526-04	RW7-SP303-20251030	2-Methylnaphthalene-d10	0.4	0.27	67		30	150
		Fluoranthene-d10	0.4	0.32	81		30	150
		Nitrobenzene-d5	0.4	0.27	68		55	111
		2-Fluorobiphenyl	0.4	0.28	69		53	106
		Terphenyl-d14	0.4	0.65	163	*	58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3526

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038132.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	
								Qual		High	RPD
PB170381BS	1,4-Dioxane	0.4	0.33	ug/L	83				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3526</u>	Analytical Method:	<u>8270-Modified</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	DataFile:	<u>BN038133.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB170381BSD	1,4-Dioxane	0.4	0.26	ug/L	65	24	*	*	70	130		20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB170381BL

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3526

Lab File ID: BN038131.D

Lab Sample ID: PB170381BL

Instrument ID: BNA_N

Date Extracted: 11/03/2025

Matrix: (soil/water) Water

Date Analyzed: 11/04/2025

Level: (low/med) LOW

Time Analyzed: 11:43

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB170381BS	PB170381BS	BN038132.D	11/04/2025
PB170381BSD	PB170381BSD	BN038133.D	11/04/2025
RW7-SP100-20251030	Q3526-01	BN038138.D	11/04/2025
RW7-SP201-20251030	Q3526-02	BN038139.D	11/04/2025
RW7-SP302-20251030	Q3526-03	BN038140.D	11/04/2025
RW7-SP303-20251030	Q3526-04	BN038141.D	11/04/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038110.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3526
DFTPP Injection Date: 10/28/2025
DFTPP Injection Time: 12:05

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	83.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN038111.D	10/28/2025	13:00
SSTDICC0.2	SSTDICC0.2	BN038112.D	10/28/2025	13:36
SSTDICCC0.4	SSTDICCC0.4	BN038113.D	10/28/2025	14:13
SSTDICC0.8	SSTDICC0.8	BN038114.D	10/28/2025	14:49
SSTDICC1.6	SSTDICC1.6	BN038115.D	10/28/2025	15:25
SSTDICC3.2	SSTDICC3.2	BN038116.D	10/28/2025	16:02
SSTDICC5.0	SSTDICC5.0	BN038117.D	10/28/2025	16:38

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038129.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3526
DFTPP Injection Date: 11/04/2025
DFTPP Injection Time: 10:00

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	77.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 (20) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038130.D	11/04/2025	10:40
PB170381BL	PB170381BL	BN038131.D	11/04/2025	11:43
PB170381BS	PB170381BS	BN038132.D	11/04/2025	12:19
PB170381BSD	PB170381BSD	BN038133.D	11/04/2025	12:56
RW7-SP100-20251030	Q3526-01	BN038138.D	11/04/2025	15:57
RW7-SP201-20251030	Q3526-02	BN038139.D	11/04/2025	16:34
RW7-SP302-20251030	Q3526-03	BN038140.D	11/04/2025	17:10
RW7-SP303-20251030	Q3526-04	BN038141.D	11/04/2025	17:46
SSTDCCC0.4EC	SSTDCCC0.4	BN038145.D	11/04/2025	20:12

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3526

Client ID : SSTDCCC0.4

Date Analyzed: 11/04/2025

Lab File ID: BN038130.D

Time Analyzed: 10:40

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8599	7.775	24973	10.57	13947	14.44
UPPER LIMIT	17198	8.275	49946	11.073	27894	14.941
LOWER LIMIT	4299.5	7.275	12486.5	10.073	6973.5	13.941
EPA SAMPLE NO.						
01 PB170381BL	8221	7.79	22221	10.59	11601	14.44
02 PB170381BS	11523	7.78	31801	10.57	16154	14.44
03 PB170381BSD	10823	7.78	30534	10.57	15616	14.43
04 RW7-SP100-20251030	8530	7.78	25854	10.57	14999	14.43
05 RW7-SP201-20251030	8635	7.78	25969	10.57	15192	14.43
06 RW7-SP302-20251030	8568	7.78	25578	10.57	14324	14.43
07 RW7-SP303-20251030	8535	7.78	25212	10.57	14257	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3526

Client ID: SSTDCCC0.4

Date Analyzed: 11/04/2025

Lab File ID: BN038130.D

Time Analyzed: 10:40

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	27056	17.186	17059	21.375	13607	23.686
UPPER LIMIT	54112	17.686	34118	21.875	27214	24.186
LOWER LIMIT	13528	16.686	8529.5	20.875	6803.5	23.186
EPA SAMPLE NO.						
01 PB170381BL	21781	17.20	15202	21.38	13241	23.70
02 PB170381BS	30417	17.19	16510	21.38	13194	23.69
03 PB170381BSD	28756	17.19	18053	21.38	14856	23.68
04 RW7-SP100-20251030	31102	17.19	25933	21.38	21372	23.68
05 RW7-SP201-20251030	28331	17.19	19063	21.38	14967	23.69
06 RW7-SP302-20251030	27505	17.19	16355	21.38	12980	23.68
07 RW7-SP303-20251030	26274	17.19	14750	21.38	12357	23.68

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	10/30/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	11/03/25
Client Sample ID:	RW7-SP100-20251030	SDG No.:	Q3526
Lab Sample ID:	Q3526-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 mL	Test:	SVOC-SIMGroup1
Prep Method :	3510C		
	Level: LOW		
	Final Vol: 1000 uL		
	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	2.90	Q	1	0.070	0.20	0.20	ug/L	11/04/25 15:57	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.28			30 - 150		71%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.43			30 - 150		107%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.28			55 - 111		70%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.31			53 - 106		77%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.51			58 - 132		128%	SPK: 0.4		
INTERNAL STANDARDS										
		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	8530								
1146-65-2	Naphthalene-d8	25900								
15067-26-2	Acenaphthene-d10	15000								
1517-22-2	Phenanthrene-d10	31100								
1719-03-5	Chrysene-d12	25900								
1520-96-3	Perylene-d12	21400								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038138.D
 Acq On : 04 Nov 2025 15:57
 Operator : RC/JU
 Sample : Q3526-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP100-20251030

Quant Time: Nov 04 16:25:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8530	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	25854	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	14999	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	31102	0.400	ng	0.00
29) Chrysene-d12	21.375	240	25933	0.400	ng	# 0.00
35) Perylene-d12	23.680	264	21372	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2490	0.124	ng	0.00
5) Phenol-d6	6.937	99	1781	0.073	ng	0.00
8) Nitrobenzene-d5	8.929	82	5875	0.282	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	10503	0.284	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	2316	0.375	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	18588	0.309	ng	0.00
27) Fluoranthene-d10	19.220	212	33654	0.429	ng	0.00
31) Terphenyl-d14	19.819	244	28902	0.513	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	25922	2.941	ng	97
6) bis(2-Chloroethyl)ether	7.197	93	524	0.022	ng	98
34) Bis(2-ethylhexyl)phtha...	21.295	149	22629	0.464	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038138.D
Acq On : 04 Nov 2025 15:57
Operator : RC/JU
Sample : Q3526-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Quant Time: Nov 04 16:25:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.775 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN038138.D
 Acq: 04 Nov 2025 15:57


Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP100-20251030





Tgt Ion: 152 Resp: 8530
 Ion Ratio Lower Upper
 152 100
 150 155.7 122.3 183.5
 115 59.4 47.4 71.0




#2
 1,4-Dioxane
 Concen: 2.941 ng
 RT: 3.247 min Scan# 23
 Delta R.T. -0.007 min
 Lab File: BN038138.D
 Acq: 04 Nov 2025 15:57



Tgt Ion: 88 Resp: 25922
 Ion Ratio Lower Upper
 88 100
 43 35.4 24.3 36.5
 58 75.5 60.4 90.6









#4
2-Fluorophenol
Concen: 0.124 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion:	112	Resp:	2490
Ion Ratio	Lower	Upper	
112	100		
64	58.2	45.4	68.0
63	31.3	23.2	34.8



#5
Phenol-d6
Concen: 0.073 ng
RT: 6.937 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion:	99	Resp:	1781
Ion Ratio	Lower	Upper	
99	100		
42	19.5	16.6	24.8
71	40.9	25.4	38.2









#6
bis(2-Chloroethyl)ether
Concen: 0.022 ng
RT: 7.197 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion: 93	Resp:	524	
Ion Ratio	Lower	Upper	
93	100		
63	71.9	59.3	88.9
95	30.7	25.2	37.8



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 936
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt	Ion:136	Resp:	25854
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.4
54	9.9	5.2	7.8#
68	7.3	4.1	6.1#









#8
 Nitrobenzene-d5
 Concen: 0.282 ng
 RT: 8.929 min Scan# 782
 Delta R.T. -0.000 min
 Lab File: BN038138.D
 Acq: 04 Nov 2025 15:57




Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP100-20251030


Tgt Ion:	82	Resp:	5875
Ion	Ratio	Lower	Upper
82	100		
128	45.7	34.1	51.1
54	53.9	43.5	65.3



#11
 2-Methylnaphthalene-d10
 Concen: 0.284 ng
 RT: 12.172 min Scan# 1154
 Delta R.T. -0.005 min
 Lab File: BN038138.D
 Acq: 04 Nov 2025 15:57

Tgt Ion:	152	Resp:	10503
Ion	Ratio	Lower	Upper
152	100		
151	21.0	16.8	25.2





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.430 min Scan# 14

Delta R.T. -0.000 min

Lab File: BN038138.D

Acq: 04 Nov 2025 15:57

Instrument :

BNA_N

ClientSampleId :

RW7-SP100-20251030





Tgt Ion:164 Resp: 14999

Ion Ratio Lower Upper

164 100

162 106.3 83.0 124.4

160 54.8 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.375 ng

RT: 15.932 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BN038138.D

Acq: 04 Nov 2025 15:57




Tgt Ion:330 Resp: 2316


Ion Ratio Lower Upper

330 100

332 95.3 76.6 114.8

141 50.2 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.309 ng
RT: 13.051 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion:	172	Resp:	18588
Ion Ratio	Lower	Upper	
172	100		
171	36.0	27.8	41.6
170	24.2	18.1	27.1



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion:	188	Resp:	31102
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	11.4	8.6	13.0









#27
Fluoranthene-d10
Concen: 0.429 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion:212 Resp: 33654
Ion Ratio Lower Upper
212 100
106 18.8 12.6 19.0
104 9.7 7.1 10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion:240 Resp: 25933
Ion Ratio Lower Upper
240 100
120 23.9 10.7 16.1#
236 30.4 23.1 34.7









#31
 Terphenyl-d14
 Concen: 0.513 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 04 Nov 2025 15:57




Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP100-20251030


Tgt Ion:244 Resp: 28902
 Ion Ratio Lower Upper
 244 100
 212 16.1 7.8 11.6#
 122 39.1 15.2 22.8#



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.464 ng
 RT: 21.295 min Scan# 2260
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 04 Nov 2025 15:57


Tgt Ion:149 Resp: 22629
 Ion Ratio Lower Upper
 149 100
 167 28.4 21.1 31.7
 279 3.1 2.9 4.3







#35
Perylene-d12
Concen: 0.400 ng
RT: 23.680 min Scan# 2821
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030



Tgt Ion:	264	Resp:	21372
Ion Ratio	Lower	Upper	
264	100		
260	27.6	21.8	32.8
265	72.2	68.6	102.8



Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/30/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	11/03/25
Client Sample ID:	RW7-SP201-20251030		SDG No.:	Q3526
Lab Sample ID:	Q3526-02		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.20	UQ	1	0.070	0.20	0.20	ug/L	11/04/25 16:34	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.30			30 - 150		76%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.38			30 - 150		94%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.31			55 - 111		76%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.31			53 - 106		77%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.56	*		58 - 132		140%	SPK: 0.4		
INTERNAL STANDARDS										
		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	8640								
1146-65-2	Naphthalene-d8	26000								
15067-26-2	Acenaphthene-d10	15200								
1517-22-2	Phenanthrene-d10	28300								
1719-03-5	Chrysene-d12	19100								
1520-96-3	Perylene-d12	15000								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038139.D
Acq On : 04 Nov 2025 16:34
Operator : RC/JU
Sample : Q3526-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030

Quant Time: Nov 04 16:59:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8635	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	25969	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	15192	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	28331	0.400	ng	0.00
29) Chrysene-d12	21.375	240	19063	0.400	ng	0.00
35) Perylene-d12	23.686	264	14967	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2921	0.144	ng	0.00
5) Phenol-d6	6.937	99	2056	0.083	ng	0.00
8) Nitrobenzene-d5	8.929	82	6392	0.305	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	11234	0.303	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1744	0.279	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	18738	0.308	ng	0.00
27) Fluoranthene-d10	19.220	212	26931	0.377	ng	0.00
31) Terphenyl-d14	19.819	244	23204	0.560	ng	0.00
Target Compounds						
9) Naphthalene	10.626	128	1557	0.022	ng	# 76
34) Bis(2-ethylhexyl)phtha...	21.295	149	3536	0.099	ng	# 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038139.D
Acq On : 04 Nov 2025 16:34
Operator : RC/JU
Sample : Q3526-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030


Quant Time: Nov 04 16:59:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 651
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030







Tgt Ion:152 Resp: 8635
Ion Ratio Lower Upper
152 100
150 152.2 122.3 183.5
115 60.6 47.4 71.0



#4
2-Fluorophenol
Concen: 0.144 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Tgt Ion:112 Resp: 2921
Ion Ratio Lower Upper
112 100
64 57.8 45.4 68.0
63 31.5 23.2 34.8









#5
Phenol-d6
Concen: 0.083 ng
RT: 6.937 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030


Tgt Ion: 99 Resp: 2056
Ion Ratio Lower Upper
99 100
42 45.8 16.6 24.8#
71 203.1 25.4 38.2#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 936
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Tgt Ion: 136 Resp: 25969
Ion Ratio Lower Upper
136 100
137 11.5 9.0 13.4
54 10.6 5.2 7.8#
68 7.8 4.1 6.1#









#8
 Nitrobenzene-d5
 Concen: 0.305 ng
 RT: 8.929 min Scan# 782
 Delta R.T. -0.000 min
 Lab File: BN038139.D
 Acq: 04 Nov 2025 16:34




Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP201-20251030


Tgt Ion: 82	Resp: 6392
Ion Ratio	Lower Upper
82	100
128	44.1 34.1 51.1
54	55.0 43.5 65.3



#9
 Naphthalene
 Concen: 0.022 ng
 RT: 10.626 min Scan# 941
 Delta R.T. -0.000 min
 Lab File: BN038139.D
 Acq: 04 Nov 2025 16:34

Tgt Ion: 128	Resp: 1557
Ion Ratio	Lower Upper
128	100
129	23.8 9.1 13.7#
127	19.6 10.5 15.7#









#11
2-Methylnaphthalene-d10
Concen: 0.303 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030


Tgt Ion:152 Resp: 11234
Ion Ratio Lower Upper
152 100
151 20.7 16.8 25.2



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.430 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34


Tgt Ion:164 Resp: 15192
Ion Ratio Lower Upper
164 100
162 106.5 83.0 124.4
160 55.0 40.7 61.1








#14
2,4,6-Tribromophenol
Concen: 0.279 ng
RT: 15.932 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34


Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030





Tgt Ion:330 Resp: 1744
Ion Ratio Lower Upper
330 100
332 95.9 76.6 114.8
141 43.8 34.1 51.1




#15
2-Fluorobiphenyl
Concen: 0.308 ng
RT: 13.051 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34



Tgt Ion:172 Resp: 18738
Ion Ratio Lower Upper
172 100
171 36.2 27.8 41.6
170 24.4 18.1 27.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030


Tgt Ion:188	Resp:	28331
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	9.9	13.0



#27
Fluoranthene-d10
Concen: 0.377 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Tgt Ion:212	Resp:	26931
Ion Ratio	Lower	Upper
212	100	
106	15.5	12.6
104	9.7	7.1





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038139.D

Acq: 04 Nov 2025 16:34

Instrument :

BNA_N

ClientSampleId :

RW7-SP201-20251030





Tgt Ion:240 Resp: 19063

Ion Ratio Lower Upper

240 100

120 14.2 10.7 16.1

236 28.6 23.1 34.7



#31

Terphenyl-d14

Concen: 0.560 ng

RT: 19.819 min Scan# 2076

Delta R.T. -0.000 min

Lab File: BN038139.D

Acq: 04 Nov 2025 16:34




Tgt Ion:244 Resp: 23204


Ion Ratio Lower Upper

244 100

212 9.3 7.8 11.6

122 17.5 15.2 22.8









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.099 ng
 RT: 21.295 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN038139.D
 Acq: 04 Nov 2025 16:34




Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP201-20251030

Tgt Ion:	149	Resp:	3536
Ion Ratio	Lower	Upper	
149	100		
167	26.5	21.1	31.7
279	5.5	2.9	4.3



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.686 min Scan# 2822
 Delta R.T. 0.006 min
 Lab File: BN038139.D
 Acq: 04 Nov 2025 16:34

Tgt Ion:	264	Resp:	14967
Ion Ratio	Lower	Upper	
264	100		
260	27.3	21.8	32.8
265	96.7	68.6	102.8



Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	10/30/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	11/03/25
Client Sample ID:	RW7-SP302-20251030	SDG No.:	Q3526
Lab Sample ID:	Q3526-03	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	940 mL	Test:	SVOC-SIMGroup1
Prep Method :	3510C		
	Level: LOW		
	Final Vol: 1000 uL		
	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.21	UQ	1	0.070	0.21	0.21	ug/L	11/04/25 17:10	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.32			30 - 150		79%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.36			30 - 150		90%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.33			55 - 111		81%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.36			53 - 106		90%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.59	*		58 - 132		148%	SPK: 0.4		
INTERNAL STANDARDS										
		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	8570								
1146-65-2	Naphthalene-d8	25600								
15067-26-2	Acenaphthene-d10	14300								
1517-22-2	Phenanthrene-d10	27500								
1719-03-5	Chrysene-d12	16400								
1520-96-3	Perylene-d12	13000								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038140.D
Acq On : 04 Nov 2025 17:10
Operator : RC/JU
Sample : Q3526-03
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030

Quant Time: Nov 04 17:34:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8568	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	25578	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	14324	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	27505	0.400	ng	0.00
29) Chrysene-d12	21.375	240	16355	0.400	ng	0.00
35) Perylene-d12	23.683	264	12980	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2964	0.147	ng	0.00
5) Phenol-d6	6.937	99	2079	0.085	ng	0.00
8) Nitrobenzene-d5	8.929	82	6714	0.325	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	11602	0.318	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1849	0.313	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	20625	0.360	ng	0.00
27) Fluoranthene-d10	19.220	212	25076	0.362	ng	0.00
31) Terphenyl-d14	19.819	244	21101	0.594	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.295	149	3579	0.116	ng	# 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038140.D
Acq On : 04 Nov 2025 17:10
Operator : RC/JU
Sample : Q3526-03
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030


Quant Time: Nov 04 17:34:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10

Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030







Tgt Ion:152 Resp: 8568
Ion Ratio Lower Upper
152 100
150 155.3 122.3 183.5
115 59.9 47.4 71.0



#4
2-Fluorophenol
Concen: 0.147 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10


Tgt Ion:112 Resp: 2964
Ion Ratio Lower Upper
112 100
64 57.7 45.4 68.0
63 30.7 23.2 34.8








#5
Phenol-d6
Concen: 0.085 ng
RT: 6.937 min Scan# 534
Delta R.T. 0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10


Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030





Tgt Ion: 99 Resp: 2079
Ion Ratio Lower Upper
99 100
42 24.7 16.6 24.8
71 39.5 25.4 38.2#




#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 936
Delta R.T. -0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10



Tgt Ion: 136 Resp: 25578
Ion Ratio Lower Upper
136 100
137 11.5 9.0 13.4
54 10.0 5.2 7.8#
68 7.3 4.1 6.1#









#8
 Nitrobenzene-d5
 Concen: 0.325 ng
 RT: 8.929 min Scan# 782
 Delta R.T. 0.000 min
 Lab File: BN038140.D
 Acq: 04 Nov 2025 17:10




Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP302-20251030


Tgt Ion:	82	Resp:	6714
Ion	Ratio	Lower	Upper
82	100		
128	43.5	34.1	51.1
54	54.0	43.5	65.3



#11
 2-Methylnaphthalene-d10
 Concen: 0.318 ng
 RT: 12.172 min Scan# 1154
 Delta R.T. -0.005 min
 Lab File: BN038140.D
 Acq: 04 Nov 2025 17:10

Tgt Ion:	152	Resp:	11602
Ion	Ratio	Lower	Upper
152	100		
151	21.5	16.8	25.2





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.430 min Scan# 14

Delta R.T. -0.000 min

Lab File: BN038140.D

Acq: 04 Nov 2025 17:10

Instrument :

BNA_N

ClientSampleId :

RW7-SP302-20251030





Tgt Ion:164 Resp: 14324

Ion Ratio Lower Upper

164 100

162 107.2 83.0 124.4

160 56.2 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.313 ng

RT: 15.932 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BN038140.D

Acq: 04 Nov 2025 17:10




Tgt Ion:330 Resp: 1849


Ion Ratio Lower Upper

330 100

332 97.9 76.6 114.8

141 42.3 34.1 51.1









#15
 2-Fluorobiphenyl
 Concen: 0.360 ng
 RT: 13.051 min Scan# 1323
 Delta R.T. 0.000 min
 Lab File: BN038140.D
 Acq: 04 Nov 2025 17:10




Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP302-20251030


Tgt Ion:	172	Resp:	20625
Ion Ratio	Lower	Upper	
172	100		
171	35.6	27.8	41.6
170	24.6	18.1	27.1



#19
 Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.186 min Scan# 1688
 Delta R.T. -0.000 min
 Lab File: BN038140.D
 Acq: 04 Nov 2025 17:10


Tgt Ion:	188	Resp:	27505
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	10.0	8.6	13.0








#27
Fluoranthene-d10
Concen: 0.362 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10

Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030







Tgt Ion:212 Resp: 25076
Ion Ratio Lower Upper
212 100
106 16.5 12.6 19.0
104 9.2 7.1 10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10

Tgt Ion:240 Resp: 16355
Ion Ratio Lower Upper
240 100
120 13.8 10.7 16.1
236 28.8 23.1 34.7









#31
 Terphenyl-d14
 Concen: 0.594 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. 0.000 min
 Lab File: BN038140.D
 Acq: 04 Nov 2025 17:10




Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP302-20251030


Tgt Ion:244	Resp:	21101
Ion Ratio	Lower	Upper
244	100	
212	9.3	7.8 11.6
122	17.6	15.2 22.8



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.116 ng
 RT: 21.295 min Scan# 2260
 Delta R.T. 0.000 min
 Lab File: BN038140.D
 Acq: 04 Nov 2025 17:10


Tgt Ion:149	Resp:	3579
Ion Ratio	Lower	Upper
149	100	
167	24.6	21.1 31.7
279	4.6	2.9 4.3#







#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.683 min Scan# 2821
 Delta R.T. 0.003 min
 Lab File: BN038140.D
 Acq: 04 Nov 2025 17:10

Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP302-20251030



Tgt Ion:264 Resp: 12980

Ion	Ratio	Lower	Upper
264	100		
260	27.6	21.8	32.8
265	103.2	68.6	102.8#



Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	10/30/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	11/03/25
Client Sample ID:	RW7-SP303-20251030	SDG No.:	Q3526
Lab Sample ID:	Q3526-04	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 mL	Test:	SVOC-SIMGroup1
Prep Method :	3510C		
	Level: LOW		
	Final Vol: 1000 uL		
	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.20	UQ	1	0.070	0.20	0.20	ug/L	11/04/25 17:46	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.27			30 - 150		67%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.32			30 - 150		81%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.27			55 - 111		68%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.28			53 - 106		69%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.65	*		58 - 132		163%	SPK: 0.4		
INTERNAL STANDARDS										
		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	8540								
1146-65-2	Naphthalene-d8	25200								
15067-26-2	Acenaphthene-d10	14300								
1517-22-2	Phenanthrene-d10	26300								
1719-03-5	Chrysene-d12	14800								
1520-96-3	Perylene-d12	12400								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038141.D
Acq On : 04 Nov 2025 17:46
Operator : RC/JU
Sample : Q3526-04
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030

Quant Time: Nov 04 18:13:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8535	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	25212	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	14257	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	26274	0.400	ng	0.00
29) Chrysene-d12	21.375	240	14750	0.400	ng	0.00
35) Perylene-d12	23.680	264	12357	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2450	0.122	ng	0.00
5) Phenol-d6	6.937	99	1641	0.067	ng	0.00
8) Nitrobenzene-d5	8.929	82	5492	0.270	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	9669	0.269	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1587	0.270	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	15752	0.276	ng	0.00
27) Fluoranthene-d10	19.220	212	21496	0.324	ng	0.00
31) Terphenyl-d14	19.819	244	20961	0.654	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.295	149	3107	0.112	ng	Qvalue 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038141.D
Acq On : 04 Nov 2025 17:46
Operator : RC/JU
Sample : Q3526-04
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030


Quant Time: Nov 04 18:13:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 651
Delta R.T. -0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030







Tgt Ion:152 Resp: 8535
Ion Ratio Lower Upper
152 100
150 156.1 122.3 183.5
115 59.7 47.4 71.0



#4
2-Fluorophenol
Concen: 0.122 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Tgt Ion:112 Resp: 2450
Ion Ratio Lower Upper
112 100
64 56.8 45.4 68.0
63 29.6 23.2 34.8









#5
Phenol-d6
Concen: 0.067 ng
RT: 6.937 min Scan# 534
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46




Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030


Tgt Ion:	99	Resp:	1641
Ion	Ratio	Lower	Upper
99	100		
42	20.5	16.6	24.8
71	41.1	25.4	38.2



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 936
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46


Tgt Ion:	136	Resp:	25212
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	9.2	5.2	7.8
68	7.0	4.1	6.1








#8
Nitrobenzene-d5
Concen: 0.270 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030







Tgt Ion:	82	Resp:	5492
Ion Ratio	Lower	Upper	
82	100		
128	45.6	34.1	51.1
54	55.1	43.5	65.3



#11
2-Methylnaphthalene-d10
Concen: 0.269 ng
RT: 12.172 min Scan# 1154
Delta R.T. -0.005 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Tgt Ion:	152	Resp:	9669
Ion Ratio	Lower	Upper	
152	100		
151	20.9	16.8	25.2





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.430 min Scan# 14

Delta R.T. 0.000 min

Lab File: BN038141.D

Acq: 04 Nov 2025 17:46

Instrument :

BNA_N

ClientSampleId :

RW7-SP303-20251030





Tgt Ion:164 Resp: 14257

Ion Ratio Lower Upper

164 100

162 107.4 83.0 124.4

160 55.2 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.270 ng

RT: 15.932 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BN038141.D

Acq: 04 Nov 2025 17:46




Tgt Ion:330 Resp: 1587


Ion Ratio Lower Upper

330 100

332 94.8 76.6 114.8

141 44.3 34.1 51.1









#15
 2-Fluorobiphenyl
 Concen: 0.276 ng
 RT: 13.051 min Scan# 13
 Delta R.T. 0.000 min
 Lab File: BN038141.D
 Acq: 04 Nov 2025 17:46




Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP303-20251030


Tgt Ion:	172	Resp:	15752
Ion Ratio	Lower	Upper	
172	100		
171	36.0	27.8	41.6
170	24.4	18.1	27.1



#19
 Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.186 min Scan# 1688
 Delta R.T. 0.000 min
 Lab File: BN038141.D
 Acq: 04 Nov 2025 17:46


Tgt Ion:	188	Resp:	26274
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	10.0	8.6	13.0








#27
Fluoranthene-d10
Concen: 0.324 ng
RT: 19.220 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030







Tgt Ion:212 Resp: 21496
Ion Ratio Lower Upper
212 100
106 16.5 12.6 19.0
104 9.1 7.1 10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Tgt Ion:240 Resp: 14750
Ion Ratio Lower Upper
240 100
120 13.0 10.7 16.1
236 28.8 23.1 34.7





#31

Terphenyl-d14

Concen: 0.654 ng

RT: 19.819 min Scan# 2076

Delta R.T. 0.000 min

Lab File: BN038141.D


Acq: 04 Nov 2025 17:46

Instrument :



BNA_N

ClientSampleId :

RW7-SP303-20251030



Tgt Ion:	244	Resp:	20961
Ion Ratio	Lower	Upper	
244	100		
212	9.2	7.8	11.6
122	17.8	15.2	22.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.112 ng




RT: 21.295 min Scan# 2260


Delta R.T. 0.000 min

Lab File: BN038141.D

Acq: 04 Nov 2025 17:46


Tgt Ion:	149	Resp:	3107
Ion Ratio	Lower	Upper	
149	100		
167	25.4	21.1	31.7
279	3.9	2.9	4.3







#35
Perylene-d12
Concen: 0.400 ng
RT: 23.680 min Scan# 2821
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030



Tgt Ion:264 Resp: 12357

Ion	Ratio	Lower	Upper
264	100		
260	27.3	21.8	32.8
265	103.3	68.6	102.8#





CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN102825.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Oct 29 13:09:46 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN038111.D 0.2 =BN038112.D 0.4 =BN038113.D 0.8 =BN038114.D 1.6 =BN038115.D 3.2 =BN038116.D 5 =BN038117.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.446	0.424	0.405	0.416	0.400	0.389	0.413	4.87	
3)	n-Nitrosodimet...	0.528	0.537	0.527	0.547	0.534	0.518	0.532	1.85	
4) S	2-Fluorophenol	1.066	0.973	0.954	0.907	0.918	0.895	0.882	0.942	6.73
5) S	Phenol-d6	1.266	1.134	1.134	1.097	1.131	1.135	1.137	1.148	4.71
6)	bis(2-Chloroet...	1.164	1.140	1.126	1.122	1.148	1.114	1.079	1.128	2.42
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.403	0.312	0.313	0.301	0.316	0.307	0.307	0.323	11.11
9)	Naphthalene	1.139	1.097	1.081	1.080	1.125	1.107	1.084	1.102	2.08
10)	Hexachlorobuta...	0.198	0.187	0.187	0.182	0.188	0.183	0.177	0.186	3.49
11) SURR2	Methylnaphth...	0.581	0.569	0.551	0.549	0.586	0.584	0.579	0.571	2.69
12)	2-Methylnaphth...	0.765	0.712	0.709	0.713	0.757	0.749	0.738	0.735	3.22
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.156	0.151	0.149	0.152	0.171	0.182	0.194	0.165	10.70
15) S	2-Fluorobiphenyl	1.794	1.649	1.577	1.570	1.615	1.514	1.495	1.602	6.24
16)	Acenaphthylene	1.768	1.728	1.711	1.764	1.891	1.912	1.908	1.811	4.88
17)	Acenaphthene	1.260	1.242	1.211	1.234	1.318	1.310	1.297	1.267	3.25
18)	Fluorene	1.549	1.618	1.596	1.635	1.768	1.731	1.729	1.661	4.93
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.041	0.045	0.051	0.064	0.073		0.055	24.55	
21)	4-Bromophenyl-...	0.260	0.251	0.250	0.256	0.274	0.271	0.275	0.262	4.06
22)	Hexachlorobenzene	0.304	0.292	0.298	0.293	0.306	0.297	0.298	0.298	1.76
23)	Atrazine	0.190	0.176	0.176	0.179	0.203	0.209	0.211	0.192	8.19
24)	Pentachlorophenol	0.112	0.108	0.113	0.132	0.146	0.159	0.128	16.16	
25)	Phenanthrene	1.281	1.231	1.224	1.236	1.327	1.293	1.298	1.270	3.12
26)	Anthracene	1.076	1.029	1.047	1.057	1.182	1.183	1.202	1.111	6.72
27) SURR	Fluoranthene-d10	1.040	0.969	0.957	0.981	1.050	1.055	1.009	1.009	4.01
28)	Fluoranthene	1.448	1.349	1.345	1.394	1.492	1.491	1.414	1.419	4.30
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.834	1.853	1.842	1.664	1.835	1.766	1.846	1.806	3.82
31) S	Terphenyl-d14	0.878	0.881	0.872	0.807	0.892	0.862	0.891	0.869	3.37
32)	Benzo(a)anthra...	1.491	1.379	1.358	1.363	1.469	1.474	1.456	1.427	4.06
33)	Chrysene	1.623	1.586	1.543	1.523	1.558	1.518	1.490	1.549	2.91
34)	Bis(2-ethylhex...	0.910	0.722	0.700	0.721	0.754	0.703	0.752	10.60	
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN102825.M

36)	Indeno(1,2,3-c...	1.521	1.512	1.604	1.620	1.730	1.739	1.778	1.643	6.53
37)	Benzo(b)fluora...	1.603	1.543	1.630	1.729	1.806	1.839	1.822	1.710	6.94
38)	Benzo(k)fluora...	1.652	1.590	1.619	1.669	1.831	1.835	1.855	1.721	6.62
39) C	Benzo(a)pyrene	1.344	1.276	1.316	1.314	1.420	1.451	1.462	1.369	5.44
40)	Dibenzo(a,h)an...	1.149	1.150	1.196	1.240	1.352	1.363	1.402	1.265	8.42
41)	Benzo(g,h,i)pe...	1.402	1.365	1.426	1.432	1.499	1.477	1.509	1.444	3.66

(#) = Out of Range

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038111.D
 Acq On : 28 Oct 2025 13:00
 Operator : RC/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Oct 28 17:31:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.789	152	8239	0.400	ng	0.01
7) Naphthalene-d8	10.584	136	24762	0.400	ng	# 0.01
13) Acenaphthene-d10	14.441	164	14233	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	28114	0.400	ng	# 0.00
29) Chrysene-d12	21.384	240	22622	0.400	ng	0.00
35) Perylene-d12	23.692	264	20615	0.400	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.356	112	2196	0.113	ng	0.00
5) Phenol-d6	6.952	99	2608	0.110	ng	0.01
8) Nitrobenzene-d5	8.939	82	2497	0.125	ng	0.01
11) 2-Methylnaphthalene-d10	12.182	152	3595	0.102	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	558	0.095	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	6383	0.112	ng	0.01
27) Fluoranthene-d10	19.225	212	7307	0.103	ng	0.00
31) Terphenyl-d14	19.829	244	4967	0.101	ng	0.00
Target Compounds						
6) bis(2-Chloroethyl)ether	7.204	93	2398	0.103	ng	Qvalue 95
9) Naphthalene	10.637	128	7049	0.103	ng	95
10) Hexachlorobutadiene	10.936	225	1223	0.106	ng	# 98
12) 2-Methylnaphthalene	12.258	142	4736	0.104	ng	96
16) Acenaphthylene	14.163	152	6290	0.098	ng	100
17) Acenaphthene	14.505	154	4482	0.099	ng	98
18) Fluorene	15.499	166	5513	0.093	ng	97
21) 4-Bromophenyl-phenylether	16.391	248	1824	0.099	ng	# 82
22) Hexachlorobenzene	16.503	284	2137	0.102	ng	98
23) Atrazine	16.652	200	1338	0.099	ng	# 90
25) Phenanthrene	17.236	178	9006	0.101	ng	99
26) Anthracene	17.322	178	7562	0.097	ng	98
28) Fluoranthene	19.257	202	10174	0.102	ng	99
30) Pyrene	19.620	202	10370	0.102	ng	99
32) Benzo(a)anthracene	21.367	228	8433	0.104	ng	99
33) Chrysene	21.420	228	9181	0.105	ng	98
36) Indeno(1,2,3-cd)pyrene	26.048	276	7839	0.093	ng	98
37) Benzo(b)fluoranthene	22.993	252	8261	0.094	ng	# 67
38) Benzo(k)fluoranthene	23.040	252	8515	0.096	ng	# 67
39) Benzo(a)pyrene	23.590	252	6927	0.098	ng	# 58
40) Dibenzo(a,h)anthracene	26.072	278	5922	0.091	ng	# 71
41) Benzo(g,h,i)perylene	26.765	276	7227	0.097	ng	# 88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038111.D
Acq On : 28 Oct 2025 13:00
Operator : RC/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Oct 28 17:31:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.789 min Scan# 61
Delta R.T. 0.014 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:152 Resp: 8239
Ion Ratio Lower Upper
152 100
150 151.7 122.3 183.5
115 60.8 47.4 71.0



#4
2-Fluorophenol
Concen: 0.113 ng
RT: 5.356 min Scan# 315
Delta R.T. 0.007 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:112 Resp: 2196
Ion Ratio Lower Upper
112 100
64 57.2 45.4 68.0
63 32.8 23.2 34.8









#5
Phenol-d6
Concen: 0.110 ng
RT: 6.952 min Scan# 511
Delta R.T. 0.015 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Tgt Ion: 99 Resp: 2608




Ion	Ratio	Lower	Upper
99	100		
42	23.4	16.6	24.8
71	34.0	25.4	38.2




#6
bis(2-Chloroethyl)ether
Concen: 0.103 ng
RT: 7.204 min Scan# 571
Delta R.T. 0.007 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion: 93 Resp: 2398

Ion	Ratio	Lower	Upper
93	100		
63	78.2	59.3	88.9
95	34.1	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	24762
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.0	13.4
54	10.0	5.2	7.8#
68	7.5	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 0.125 ng
RT: 8.939 min Scan# 783
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:	82	Resp:	2497
Ion Ratio	Lower	Upper	
82	100		
128	46.2	34.1	51.1
54	57.2	43.5	65.3









#9
Naphthalene
Concen: 0.103 ng
RT: 10.637 min Scan# 941
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	128	Resp:	7049
Ion Ratio	Lower	Upper	
128	100		
129	13.5	9.1	13.7
127	14.7	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.106 ng
RT: 10.936 min Scan# 970
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:	225	Resp:	1223
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.3	50.3	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.102 ng
RT: 12.182 min Scan# 1156
Delta R.T. 0.005 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 3595
Ion Ratio Lower Upper
152 100
151 21.4 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.104 ng
RT: 12.258 min Scan# 1171
Delta R.T. 0.010 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:142 Resp: 4736
Ion Ratio Lower Upper
142 100
141 90.0 70.3 105.5
115 39.2 27.8 41.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.441 min Scan# 14

Delta R.T. 0.011 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:164 Resp: 14233

Ion Ratio Lower Upper

164 100

162 106.3 83.0 124.4

160 52.7 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.095 ng

RT: 15.932 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00




Tgt Ion:330 Resp: 558


Ion Ratio Lower Upper

330 100

332 95.9 76.6 114.8

141 41.0 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.112 ng
RT: 13.062 min Scan# 11
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	172	Resp:	6383
Ion Ratio	Lower	Upper	
172	100		
171	36.2	27.8	41.6
170	25.2	18.1	27.1



#16
Acenaphthylene
Concen: 0.098 ng
RT: 14.163 min Scan# 1427
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:	152	Resp:	6290
Ion Ratio	Lower	Upper	
152	100		
151	19.9	15.8	23.8
153	12.9	10.2	15.2









#17
Acenaphthene
Concen: 0.099 ng
RT: 14.505 min Scan# 1459
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:154 Resp: 4482
Ion Ratio Lower Upper
154 100
153 114.1 90.0 135.0
152 60.7 47.3 70.9



#18
Fluorene
Concen: 0.093 ng
RT: 15.499 min Scan# 1552
Delta R.T. 0.010 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:166 Resp: 5513
Ion Ratio Lower Upper
166 100
165 102.4 79.0 118.4
167 13.3 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:188 Resp: 28114
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.8 8.6 13.0#



#21
4-Bromophenyl-phenylether
Concen: 0.099 ng
RT: 16.391 min Scan# 1624
Delta R.T. 0.012 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00


Tgt Ion:248 Resp: 1824
Ion Ratio Lower Upper
248 100
250 103.9 76.2 114.2
141 48.0 57.6 86.4#








#22
Hexachlorobenzene
Concen: 0.102 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1





Tgt Ion: 284 Resp: 2137
Ion Ratio Lower Upper
284 100
142 38.9 30.5 45.7
249 31.2 23.5 35.3




#23
Atrazine
Concen: 0.099 ng
RT: 16.652 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00



Tgt Ion: 200 Resp: 1338
Ion Ratio Lower Upper
200 100
173 38.1 21.4 32.2#
215 47.8 36.7 55.1









#25
Phenanthrene
Concen: 0.101 ng
RT: 17.236 min Scan# 1692
Delta R.T. 0.012 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	178	Resp:	9006
Ion Ratio	Lower	Upper	
178	100		
176	19.0	15.5	23.3
179	15.8	12.2	18.4



#26
Anthracene
Concen: 0.097 ng
RT: 17.322 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:	178	Resp:	7562
Ion Ratio	Lower	Upper	
178	100		
176	18.4	14.9	22.3
179	16.0	11.8	17.8





#27

Fluoranthene-d10

Concen: 0.103 ng

RT: 19.225 min Scan# 1947

Delta R.T. 0.005 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:212 Resp: 7307

Ion Ratio Lower Upper

212 100

106 15.9 12.6 19.0

104 9.3 7.1 10.7



#28

Fluoranthene

Concen: 0.102 ng

RT: 19.257 min Scan# 1955

Delta R.T. 0.005 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00




Tgt Ion:202 Resp: 10174


Ion Ratio Lower Upper

202 100

101 11.4 9.4 14.2

203 17.2 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.384 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:240 Resp: 22622

Ion Ratio Lower Upper

240 100

120 12.1 10.7 16.1

236 28.6 23.1 34.7



#30

Pyrene

Concen: 0.102 ng

RT: 19.620 min Scan# 2033

Delta R.T. 0.009 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00




Tgt Ion:202 Resp: 10370


Ion Ratio Lower Upper

202 100

200 21.3 16.7 25.1

203 18.0 14.2 21.2









#31
Terphenyl-d14
Concen: 0.101 ng
RT: 19.829 min Scan# 2076
Delta R.T. 0.009 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:244 Resp: 4967
Ion Ratio Lower Upper
244 100
212 10.8 7.8 11.6
122 18.9 15.2 22.8



#32
Benzo(a)anthracene
Concen: 0.104 ng
RT: 21.367 min Scan# 2268
Delta R.T. 0.009 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:228 Resp: 8433
Ion Ratio Lower Upper
228 100
226 27.7 22.2 33.4
229 20.5 15.8 23.8









#33
 Chrysene
 Concen: 0.105 ng
 RT: 21.420 min Scan# 21
 Delta R.T. 0.009 min
 Lab File: BN038111.D
 Acq: 28 Oct 2025 13:00




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	228	Resp:	9181
Ion Ratio	Lower	Upper	
228	100		
226	30.5	24.3	36.5
229	21.3	15.7	23.5



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.692 min Scan# 2824
 Delta R.T. 0.012 min
 Lab File: BN038111.D
 Acq: 28 Oct 2025 13:00


Tgt Ion:	264	Resp:	20615
Ion Ratio	Lower	Upper	
264	100		
260	27.0	21.8	32.8
265	75.0	68.6	102.8








#36
Indeno(1,2,3-cd)pyrene
Concen: 0.093 ng
RT: 26.048 min Scan# 3630
Delta R.T. 0.015 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1







Tgt Ion:276 Resp: 7839
Ion Ratio Lower Upper
276 100
138 25.8 21.4 32.0
277 24.8 19.2 28.8



#37
Benzo(b)fluoranthene
Concen: 0.094 ng
RT: 22.993 min Scan# 2585
Delta R.T. 0.012 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00


Tgt Ion:252 Resp: 8261
Ion Ratio Lower Upper
252 100
253 37.3 20.9 31.3#
125 39.7 14.1 21.1#








#38
Benzo(k)fluoranthene
Concen: 0.096 ng
RT: 23.040 min Scan# 20
Delta R.T. 0.012 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1







Tgt Ion:252 Resp: 8515
Ion Ratio Lower Upper
252 100
253 38.3 21.2 31.8#
125 39.7 14.6 22.0#



#39
Benzo(a)pyrene
Concen: 0.098 ng
RT: 23.590 min Scan# 2789
Delta R.T. 0.012 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:252 Resp: 6927
Ion Ratio Lower Upper
252 100
253 44.2 23.1 34.7#
125 51.6 18.6 28.0#





#40

Dibenzo(a,h)anthracene

Concen: 0.091 ng

RT: 26.072 min Scan# 3

Delta R.T. 0.023 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:278 Resp: 5922

Ion Ratio Lower Upper

278 100

139 38.9 17.6 26.4#

279 41.0 22.6 33.8#



#41

Benzo(g,h,i)perylene

Concen: 0.097 ng

RT: 26.765 min Scan# 3875

Delta R.T. 0.015 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00




Tgt Ion:276 Resp: 7227

Ion Ratio Lower Upper

276 100

277 27.6 19.7 29.5

138 34.5 20.2 30.4#



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038112.D
 Acq On : 28 Oct 2025 13:36
 Operator : RC/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Oct 28 17:31:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	9322	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	28183	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	15941	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	31512	0.400	ng	0.00
29) Chrysene-d12	21.375	240	23080	0.400	ng	0.00
35) Perylene-d12	23.683	264	20598	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	4536	0.207	ng	0.00
5) Phenol-d6	6.937	99	5284	0.198	ng	0.00
8) Nitrobenzene-d5	8.929	82	4403	0.194	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	8012	0.199	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1201	0.183	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	13145	0.206	ng	0.01
27) Fluoranthene-d10	19.225	212	15247	0.192	ng	0.00
31) Terphenyl-d14	19.824	244	10170	0.203	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	2078	0.216	ng	95
3) n-Nitrosodimethylamine	3.564	42	2463	0.199	ng	# 98
6) bis(2-Chloroethyl)ether	7.197	93	5312	0.202	ng	99
9) Naphthalene	10.626	128	15455	0.199	ng	98
10) Hexachlorobutadiene	10.925	225	2632	0.201	ng	# 97
12) 2-Methylnaphthalene	12.253	142	10027	0.194	ng	99
16) Acenaphthylene	14.152	152	13787	0.191	ng	99
17) Acenaphthene	14.505	154	9898	0.196	ng	100
18) Fluorene	15.489	166	13019	0.197	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	639	0.238	ng	# 40
21) 4-Bromophenyl-phenylether	16.379	248	3948	0.191	ng	98
22) Hexachlorobenzene	16.503	284	4594	0.196	ng	99
23) Atrazine	16.652	200	2774	0.183	ng	98
24) Pentachlorophenol	16.838	266	1769	0.175	ng	98
25) Phenanthrene	17.223	178	19374	0.194	ng	100
26) Anthracene	17.322	178	16201	0.185	ng	99
28) Fluoranthene	19.253	202	21232	0.190	ng	99
30) Pyrene	19.615	202	21387	0.205	ng	100
32) Benzo(a)anthracene	21.358	228	15917	0.193	ng	99
33) Chrysene	21.411	228	18301	0.205	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	10496	0.242	ng	100
36) Indeno(1,2,3-cd)pyrene	26.039	276	15575	0.184	ng	99
37) Benzo(b)fluoranthene	22.984	252	15888	0.180	ng	# 90
38) Benzo(k)fluoranthene	23.031	252	16371	0.185	ng	# 91
39) Benzo(a)pyrene	23.584	252	13137	0.186	ng	# 84
40) Dibenzo(a,h)anthracene	26.057	278	11845	0.182	ng	95
41) Benzo(g,h,i)perylene	26.753	276	14055	0.189	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038112.D
Acq On : 28 Oct 2025 13:36
Operator : RC/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Quant Time: Oct 28 17:31:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.782 min Scan# 61
 Delta R.T. 0.007 min
 Lab File: BN038112.D
 Acq: 28 Oct 2025 13:36


Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2





Tgt Ion:152 Resp: 9322
 Ion Ratio Lower Upper
 152 100
 150 150.4 122.3 183.5
 115 59.1 47.4 71.0




#2
 1,4-Dioxane
 Concen: 0.216 ng
 RT: 3.254 min Scan# 24
 Delta R.T. 0.000 min
 Lab File: BN038112.D
 Acq: 28 Oct 2025 13:36



Tgt Ion: 88 Resp: 2078
 Ion Ratio Lower Upper
 88 100
 43 36.3 24.3 36.5
 58 73.3 60.4 90.6





#3

n-Nitrosodimethylamine

Concen: 0.199 ng

RT: 3.564 min Scan# 61

Delta R.T. 0.007 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion: 42 Resp: 2463

Ion Ratio Lower Upper

42 100

74 121.8 96.8 145.2

44 12.3 5.3 7.9#



#4

2-Fluorophenol

Concen: 0.207 ng

RT: 5.348 min Scan# 314

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion: 112 Resp: 4536


Ion Ratio Lower Upper

112 100

64 56.6 45.4 68.0

63 30.2 23.2 34.8









#5
Phenol-d6
Concen: 0.198 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	99	Resp:	5284
Ion	Ratio	Lower	Upper
99	100		
42	22.1	16.6	24.8
71	32.7	25.4	38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.202 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:	93	Resp:	5312
Ion	Ratio	Lower	Upper
93	100		
63	74.3	59.3	88.9
95	32.2	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	136	Resp:	28183
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.6	5.2	7.8
68	5.2	4.1	6.1



#8
Nitrobenzene-d5
Concen: 0.194 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:	82	Resp:	4403
Ion	Ratio	Lower	Upper
82	100		
128	44.4	34.1	51.1
54	57.3	43.5	65.3









#9
Naphthalene
Concen: 0.199 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	128	Resp:	15455
Ion Ratio	Lower	Upper	
128	100		
129	12.2	9.1	13.7
127	13.8	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.201 ng
RT: 10.925 min Scan# 969
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:	225	Resp:	2632
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	65.1	50.3	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.199 ng
RT: 12.177 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 8012
Ion Ratio Lower Upper
152 100
151 20.7 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.194 ng
RT: 12.253 min Scan# 1170
Delta R.T. 0.005 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:142 Resp: 10027
Ion Ratio Lower Upper
142 100
141 88.4 70.3 105.5
115 36.8 27.8 41.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.441 min Scan# 1453

Delta R.T. 0.011 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:164 Resp: 15941

Ion Ratio Lower Upper

164 100

162 103.1 83.0 124.4

160 49.6 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.183 ng

RT: 15.932 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion:330 Resp: 1201


Ion Ratio Lower Upper

330 100

332 96.6 76.6 114.8

141 41.5 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.206 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	172	Resp:	13145
Ion Ratio	Lower	Upper	
172	100		
171	35.0	27.8	41.6
170	23.6	18.1	27.1



#16
Acenaphthylene
Concen: 0.191 ng
RT: 14.152 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:	152	Resp:	13787
Ion Ratio	Lower	Upper	
152	100		
151	19.6	15.8	23.8
153	13.0	10.2	15.2









#17
Acenaphthene
Concen: 0.196 ng
RT: 14.505 min Scan# 1459
Delta R.T. 0.011 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:154 Resp: 9898
Ion Ratio Lower Upper
154 100
153 112.7 90.0 135.0
152 58.9 47.3 70.9



#18
Fluorene
Concen: 0.197 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:166 Resp: 13019
Ion Ratio Lower Upper
166 100
165 98.6 79.0 118.4
167 13.4 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:188 Resp: 31512

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.0 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.238 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion:198 Resp: 639


Ion Ratio Lower Upper

198 100

51 192.9 88.9 133.3#

105 84.5 49.2 73.8#









#21
4-Bromophenyl-phenylether
Concen: 0.191 ng
RT: 16.379 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	248	Resp:	3948
Ion Ratio	Lower	Upper	
248	100		
250	92.8	76.2	114.2
141	73.9	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.196 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:	284	Resp:	4594
Ion Ratio	Lower	Upper	
284	100		
142	38.0	30.5	45.7
249	30.1	23.5	35.3









#23
 Atrazine
 Concen: 0.183 ng
 RT: 16.652 min Scan# 1645
 Delta R.T. 0.000 min
 Lab File: BN038112.D
 Acq: 28 Oct 2025 13:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:	200	Resp:	2774
Ion Ratio	Lower	Upper	
200	100		
173	28.7	21.4	32.2
215	45.0	36.7	55.1



#24
 Pentachlorophenol
 Concen: 0.175 ng
 RT: 16.838 min Scan# 1660
 Delta R.T. 0.000 min
 Lab File: BN038112.D
 Acq: 28 Oct 2025 13:36

Tgt Ion:	266	Resp:	1769
Ion Ratio	Lower	Upper	
266	100		
264	61.0	49.3	73.9
268	65.9	50.6	75.8









#25
 Phenanthrene
 Concen: 0.194 ng
 RT: 17.223 min Scan# 1691
 Delta R.T. 0.000 min
 Lab File: BN038112.D
 Acq: 28 Oct 2025 13:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:178	Resp:	19374
Ion	Ratio	Lower Upper
178	100	
176	19.2	15.5 23.3
179	15.3	12.2 18.4



#26
 Anthracene
 Concen: 0.185 ng
 RT: 17.322 min Scan# 1699
 Delta R.T. 0.000 min
 Lab File: BN038112.D
 Acq: 28 Oct 2025 13:36

Tgt Ion:178	Resp:	16201
Ion	Ratio	Lower Upper
178	100	
176	18.8	14.9 22.3
179	15.5	11.8 17.8









#27
Fluoranthene-d10
Concen: 0.192 ng
RT: 19.225 min Scan# 1948
Delta R.T. 0.005 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:212 Resp: 15247
Ion Ratio Lower Upper
212 100
106 15.8 12.6 19.0
104 9.0 7.1 10.7



#28
Fluoranthene
Concen: 0.190 ng
RT: 19.253 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:202 Resp: 21232
Ion Ratio Lower Upper
202 100
101 11.6 9.4 14.2
203 17.2 13.4 20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:240 Resp: 23080
Ion Ratio Lower Upper
240 100
120 14.9 10.7 16.1
236 28.8 23.1 34.7



#30
Pyrene
Concen: 0.205 ng
RT: 19.615 min Scan# 2032
Delta R.T. 0.005 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:202 Resp: 21387
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.7 14.2 21.2





#31

Terphenyl-d14

Concen: 0.203 ng

RT: 19.824 min Scan# 2076

Delta R.T. 0.005 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:244 Resp: 10170

Ion Ratio Lower Upper

244 100

212 9.6 7.8 11.6

122 17.6 15.2 22.8



#32

Benzo(a)anthracene

Concen: 0.193 ng

RT: 21.358 min Scan# 2267

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion:228 Resp: 15917


Ion Ratio Lower Upper

228 100

226 28.5 22.2 33.4

229 19.5 15.8 23.8









#33
Chrysene
Concen: 0.205 ng
RT: 21.411 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	228	Resp:	18301
Ion Ratio	Lower	Upper	
228	100		
226	31.1	24.3	36.5
229	19.9	15.7	23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.242 ng
RT: 21.295 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:	149	Resp:	10496
Ion Ratio	Lower	Upper	
149	100		
167	26.3	21.1	31.7
279	3.1	2.9	4.3









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.683 min Scan# 21
 Delta R.T. 0.003 min
 Lab File: BN038112.D
 Acq: 28 Oct 2025 13:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:	264	Resp:	20598
Ion Ratio	Lower	Upper	
264	100		
260	26.6	21.8	32.8
265	75.6	68.6	102.8



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.184 ng
 RT: 26.039 min Scan# 3627
 Delta R.T. 0.006 min
 Lab File: BN038112.D
 Acq: 28 Oct 2025 13:36

Tgt Ion:	276	Resp:	15575
Ion Ratio	Lower	Upper	
276	100		
138	25.6	21.4	32.0
277	24.2	19.2	28.8









#37
Benzo(b)fluoranthene
Concen: 0.180 ng
RT: 22.984 min Scan# 2582
Delta R.T. 0.003 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:252 Resp: 15888
Ion Ratio Lower Upper
252 100
253 29.8 20.9 31.3
125 24.3 14.1 21.1#



#38
Benzo(k)fluoranthene
Concen: 0.185 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.003 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36


Tgt Ion:252 Resp: 16371
Ion Ratio Lower Upper
252 100
253 29.7 21.2 31.8
125 23.6 14.6 22.0#








#39
Benzo(a)pyrene
Concen: 0.186 ng
RT: 23.584 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2







Tgt Ion:252 Resp: 13137
Ion Ratio Lower Upper
252 100
253 35.6 23.1 34.7#
125 33.1 18.6 28.0#



#40
Dibenzo(a,h)anthracene
Concen: 0.182 ng
RT: 26.057 min Scan# 3633
Delta R.T. 0.009 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36


Tgt Ion:278 Resp: 11845
Ion Ratio Lower Upper
278 100
139 24.7 17.6 26.4
279 30.8 22.6 33.8







#41
Benzo(g,h,i)perylene
Concen: 0.189 ng
RT: 26.753 min Scan# 31
Delta R.T. 0.003 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2



Tgt Ion:276 Resp: 14055
Ion Ratio Lower Upper
276 100
277 26.2 19.7 29.5
138 28.5 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038113.D
 Acq On : 28 Oct 2025 14:13
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Oct 28 17:31:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	9189	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	27629	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	15496	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	29547	0.400	ng	0.00
29) Chrysene-d12	21.375	240	21529	0.400	ng	0.00
35) Perylene-d12	23.683	264	17966	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	8768	0.405	ng	0.00
5) Phenol-d6	6.937	99	10420	0.395	ng	0.00
8) Nitrobenzene-d5	8.929	82	8652	0.388	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	15231	0.386	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	2303	0.361	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	24443	0.394	ng	0.01
27) Fluoranthene-d10	19.220	212	28227	0.379	ng	0.00
31) Terphenyl-d14	19.819	244	18767	0.401	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	3895	0.410	ng	Qvalue 100
3) n-Nitrosodimethylamine	3.564	42	4910	0.402	ng	100
6) bis(2-Chloroethyl)ether	7.197	93	10402	0.402	ng	100
9) Naphthalene	10.626	128	29858	0.392	ng	100
10) Hexachlorobutadiene	10.925	225	5174	0.403	ng	# 100
12) 2-Methylnaphthalene	12.253	142	19577	0.386	ng	100
16) Acenaphthylene	14.152	152	26507	0.378	ng	100
17) Acenaphthene	14.494	154	18774	0.382	ng	100
18) Fluorene	15.489	166	24733	0.384	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	1329	0.375	ng	100
21) 4-Bromophenyl-phenylether	16.379	248	7383	0.381	ng	100
22) Hexachlorobenzene	16.503	284	8778	0.398	ng	100
23) Atrazine	16.652	200	5175	0.365	ng	100
24) Pentachlorophenol	16.838	266	3182	0.335	ng	100
25) Phenanthrene	17.223	178	36093	0.385	ng	100
26) Anthracene	17.322	178	30869	0.376	ng	100
28) Fluoranthene	19.253	202	39664	0.378	ng	100
30) Pyrene	19.615	202	39644	0.408	ng	100
32) Benzo(a)anthracene	21.357	228	29230	0.381	ng	100
33) Chrysene	21.411	228	33214	0.398	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	15540	0.384	ng	100
36) Indeno(1,2,3-cd)pyrene	26.037	276	28812	0.390	ng	100
37) Benzo(b)fluoranthene	22.984	252	29283	0.381	ng	100
38) Benzo(k)fluoranthene	23.028	252	29088	0.376	ng	100
39) Benzo(a)pyrene	23.581	252	23635	0.384	ng	100
40) Dibenzo(a,h)anthracene	26.054	278	21487	0.378	ng	100
41) Benzo(g,h,i)perylene	26.747	276	25622	0.395	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038113.D
Acq On : 28 Oct 2025 14:13
Operator : RC/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Quant Time: Oct 28 17:31:42 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.782 min Scan# 61
 Delta R.T. 0.007 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4







Tgt Ion:152 Resp: 9189
 Ion Ratio Lower Upper
 152 100
 150 152.9 122.3 183.5
 115 59.2 47.4 71.0



#2
 1,4-Dioxane
 Concen: 0.410 ng
 RT: 3.254 min Scan# 24
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13

Tgt Ion: 88 Resp: 3895
 Ion Ratio Lower Upper
 88 100
 43 30.4 24.3 36.5
 58 75.5 60.4 90.6









#3
 n-Nitrosodimethylamine
 Concen: 0.402 ng
 RT: 3.564 min Scan# 61
 Delta R.T. 0.007 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4


Tgt Ion: 42 Resp: 4910
 Ion Ratio Lower Upper
 42 100
 74 121.1 96.8 145.2
 44 6.7 5.3 7.9



#4
 2-Fluorophenol
 Concen: 0.405 ng
 RT: 5.348 min Scan# 314
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13

Tgt Ion: 112 Resp: 8768
 Ion Ratio Lower Upper
 112 100
 64 56.7 45.4 68.0
 63 29.0 23.2 34.8









#5
Phenol-d6
Concen: 0.395 ng
RT: 6.937 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 10420
Ion Ratio Lower Upper
99 100
42 20.7 16.6 24.8
71 31.8 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.402 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion: 93 Resp: 10402
Ion Ratio Lower Upper
93 100
63 73.7 59.3 88.9
95 31.4 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	27629
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.5	5.2	7.8
68	5.1	4.1	6.1



#8
Nitrobenzene-d5
Concen: 0.388 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:	82	Resp:	8652
Ion	Ratio	Lower	Upper
82	100		
128	42.6	34.1	51.1
54	54.4	43.5	65.3









#9
Naphthalene
Concen: 0.392 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:128 Resp: 29858
Ion Ratio Lower Upper
128 100
129 11.4 9.1 13.7
127 13.1 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.403 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:225 Resp: 5174
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.9 50.3 75.5









#11
 2-Methylnaphthalene-d10
 Concen: 0.386 ng
 RT: 12.177 min Scan# 1155
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13




Instrument : BNA_N
 ClientSampleId : SSTDICCC0.4


Tgt Ion:	152	Resp:	15231
Ion Ratio	Lower	Upper	
152	100		
151	21.0	16.8	25.2



#12
 2-Methylnaphthalene
 Concen: 0.386 ng
 RT: 12.253 min Scan# 1170
 Delta R.T. 0.005 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13

Tgt Ion:	142	Resp:	19577
Ion Ratio	Lower	Upper	
142	100		
141	87.9	70.3	105.5
115	34.7	27.8	41.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.441 min Scan# 1453

Delta R.T. 0.011 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:164 Resp: 15496

Ion Ratio Lower Upper

164 100

162 103.7 83.0 124.4

160 50.9 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.361 ng

RT: 15.932 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13




Tgt Ion:330 Resp: 2303


Ion Ratio Lower Upper

330 100

332 96.1 76.6 114.8

141 42.6 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.394 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	172	Resp:	24443
Ion Ratio	Lower	Upper	
172	100		
171	34.7	27.8	41.6
170	22.6	18.1	27.1



#16
Acenaphthylene
Concen: 0.378 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:	152	Resp:	26507
Ion Ratio	Lower	Upper	
152	100		
151	19.8	15.8	23.8
153	12.7	10.2	15.2









#17
 Acenaphthene
 Concen: 0.382 ng
 RT: 14.494 min Scan# 1458
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13

Instrument : BNA_N
 ClientSampleId : SSTDICCC0.4

Tgt Ion:154 Resp: 18774




Ion	Ratio	Lower	Upper
154	100		
153	112.5	90.0	135.0
152	59.1	47.3	70.9




#18
 Fluorene
 Concen: 0.384 ng
 RT: 15.489 min Scan# 1551
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13

Tgt Ion:166 Resp: 24733

Ion	Ratio	Lower	Upper
166	100		
165	98.7	79.0	118.4
167	13.4	10.7	16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:188 Resp: 29547
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.8 8.6 13.0



#20
4,6-Dinitro-2-methylphenol
Concen: 0.375 ng
RT: 15.560 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:198 Resp: 1329
Ion Ratio Lower Upper
198 100
51 111.1 88.9 133.3
105 61.5 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.381 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	248	Resp:	7383
Ion Ratio	Lower	Upper	
248	100		
250	95.2	76.2	114.2
141	72.0	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.398 ng
RT: 16.503 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:	284	Resp:	8778
Ion Ratio	Lower	Upper	
284	100		
142	38.1	30.5	45.7
249	29.4	23.5	35.3









#23
Atrazine
Concen: 0.365 ng
RT: 16.652 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:200 Resp: 5175
Ion Ratio Lower Upper
200 100
173 26.8 21.4 32.2
215 45.9 36.7 55.1



#24
Pentachlorophenol
Concen: 0.335 ng
RT: 16.838 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:266 Resp: 3182
Ion Ratio Lower Upper
266 100
264 61.6 49.3 73.9
268 63.2 50.6 75.8









#25
Phenanthrene
Concen: 0.385 ng
RT: 17.223 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 36093
Ion Ratio Lower Upper
178 100
176 19.4 15.5 23.3
179 15.3 12.2 18.4



#26
Anthracene
Concen: 0.376 ng
RT: 17.322 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:178 Resp: 30869
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 14.8 11.8 17.8









#27
Fluoranthene-d10
Concen: 0.379 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:212 Resp: 28227
Ion Ratio Lower Upper
212 100
106 15.8 12.6 19.0
104 8.9 7.1 10.7



#28
Fluoranthene
Concen: 0.378 ng
RT: 19.253 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:202 Resp: 39664
Ion Ratio Lower Upper
202 100
101 11.8 9.4 14.2
203 16.8 13.4 20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:240 Resp: 21529
Ion Ratio Lower Upper
240 100
120 13.4 10.7 16.1
236 28.9 23.1 34.7



#30
Pyrene
Concen: 0.408 ng
RT: 19.615 min Scan# 2032
Delta R.T. 0.005 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:202 Resp: 39644
Ion Ratio Lower Upper
202 100
200 20.9 16.7 25.1
203 17.7 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.401 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:244 Resp: 18767
 Ion Ratio Lower Upper
 244 100
 212 9.7 7.8 11.6
 122 19.0 15.2 22.8



#32
 Benzo(a)anthracene
 Concen: 0.381 ng
 RT: 21.357 min Scan# 2267
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13

Tgt Ion:228 Resp: 29230
 Ion Ratio Lower Upper
 228 100
 226 27.8 22.2 33.4
 229 19.8 15.8 23.8









#33
 Chrysene
 Concen: 0.398 ng
 RT: 21.411 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:	228	Resp:	33214
Ion Ratio	Lower	Upper	
228	100		
226	30.4	24.3	36.5
229	19.6	15.7	23.5



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.384 ng
 RT: 21.295 min Scan# 2260
 Delta R.T. -0.000 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13

Tgt Ion:	149	Resp:	15540
Ion Ratio	Lower	Upper	
149	100		
167	26.4	21.1	31.7
279	3.6	2.9	4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.683 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:264 Resp: 17966
Ion Ratio Lower Upper
264 100
260 27.3 21.8 32.8
265 85.7 68.6 102.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.390 ng
RT: 26.037 min Scan# 3626
Delta R.T. 0.003 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:276 Resp: 28812
Ion Ratio Lower Upper
276 100
138 26.7 21.4 32.0
277 24.0 19.2 28.8





#37

Benzo(b)fluoranthene

Concen: 0.381 ng

RT: 22.984 min Scan# 2582

Delta R.T. 0.003 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 29283

Ion Ratio Lower Upper

252 100

253 26.1 20.9 31.3

125 17.6 14.1 21.1



#38

Benzo(k)fluoranthene

Concen: 0.376 ng

RT: 23.028 min Scan# 2597

Delta R.T. -0.000 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13




Tgt Ion:252 Resp: 29088


Ion Ratio Lower Upper

252 100

253 26.5 21.2 31.8


125 18.3 14.6 22.0








#39
Benzo(a)pyrene
Concen: 0.384 ng
RT: 23.581 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4





Tgt Ion:252 Resp: 23635
Ion Ratio Lower Upper
252 100
253 28.9 23.1 34.7
125 23.3 18.6 28.0




#40
Dibenzo(a,h)anthracene
Concen: 0.378 ng
RT: 26.054 min Scan# 3632
Delta R.T. 0.006 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13



Tgt Ion:278 Resp: 21487
Ion Ratio Lower Upper
278 100
139 22.0 17.6 26.4
279 28.2 22.6 33.8








#41
 Benzo(g,h,i)perylene
 Concen: 0.395 ng
 RT: 26.747 min Scan# 31
 Delta R.T. -0.003 min
 Lab File: BN038113.D
 Acq: 28 Oct 2025 14:13

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4

Tgt Ion:	276	Resp:	25622
Ion Ratio	Lower	Upper	
276	100		
277	24.6	19.7	29.5
138	25.3	20.2	30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038114.D
 Acq On : 28 Oct 2025 14:49
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Oct 28 17:31:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/29/2025
 Supervised By :mohammad ahmed 10/30/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	9110	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	27131	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	14920	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	28506	0.400	ng	0.00
29) Chrysene-d12	21.375	240	23535	0.400	ng	0.00
35) Perylene-d12	23.683	264	20307	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	16552	0.771	ng	0.00
5) Phenol-d6	6.937	99	19989	0.765	ng	0.00
8) Nitrobenzene-d5	8.929	82	16332	0.746	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	29806	0.769	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	4525	0.736	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	46843	0.784	ng	0.01
27) Fluoranthene-d10	19.220	212	55912	0.778	ng	0.00
31) Terphenyl-d14	19.819	244	37975	0.743	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	7707	0.819	ng	99
3) n-Nitrosodimethylamine	3.564	42	9598	0.792	ng	# 100
6) bis(2-Chloroethyl)ether	7.197	93	20444	0.796	ng	99
9) Naphthalene	10.626	128	58627	0.784	ng	99
10) Hexachlorobutadiene	10.925	225	9875	0.783	ng	# 98
12) 2-Methylnaphthalene	12.248	142	38674	0.776	ng	98
16) Acenaphthylene	14.152	152	52624	0.779	ng	99
17) Acenaphthene	14.494	154	36845	0.779	ng	100
18) Fluorene	15.489	166	49204	0.794	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	2898	0.689	ng	# 69
21) 4-Bromophenyl-phenylether	16.379	248	14571	0.780	ng	97
22) Hexachlorobenzene	16.503	284	16700	0.786	ng	99
23) Atrazine	16.652	200	10204	0.745	ng	99
24) Pentachlorophenol	16.838	266	6470	0.707	ng	99
25) Phenanthrene	17.223	178	70455	0.779	ng	100
26) Anthracene	17.322	178	60270	0.761	ng	100
28) Fluoranthene	19.253	202	79478	0.786	ng	100
30) Pyrene	19.615	202	78307	0.737	ng	100
32) Benzo(a)anthracene	21.357	228	64140	0.764	ng	100
33) Chrysene	21.411	228	71675	0.787	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	32963	0.745	ng	99
36) Indeno(1,2,3-cd)pyrene	26.039	276	65796	0.789	ng	99
37) Benzo(b)fluoranthene	22.984	252	69912m	0.805	ng	
38) Benzo(k)fluoranthene	23.031	252	67771	0.776	ng	# 93
39) Benzo(a)pyrene	23.578	252	53367	0.768	ng	# 90
40) Dibenzo(a,h)anthracene	26.054	278	50344	0.784	ng	95
41) Benzo(g,h,i)perylene	26.753	276	58155	0.793	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038114.D
Acq On : 28 Oct 2025 14:49
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Oct 28 17:31:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8






Tgt Ion:152 Resp: 9110
Ion Ratio Lower Upper
152 100
150 157.1 122.3 183.5
115 63.4 47.4 71.0

Manual Integrations

APPROVED




Reviewed By :Rahul Chavli 10/29/2025


Supervised By :mohammad ahmed 10/30/2025



#2
1,4-Dioxane
Concen: 0.819 ng
RT: 3.254 min Scan# 24
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion: 88 Resp: 7707
Ion Ratio Lower Upper
88 100
43 30.6 24.3 36.5
58 74.3 60.4 90.6





#3
n-Nitrosodimethylamine
Concen: 0.792 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8




Tgt Ion: 42 Resp: 959
Ion Ratio Lower Upper
42 100
74 120.8 96.8 145.2
44 8.8 5.3 7.9

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025




#4
2-Fluorophenol
Concen: 0.771 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Tgt Ion: 112 Resp: 16552
Ion Ratio Lower Upper
112 100
64 55.7 45.4 68.0
63 30.2 23.2 34.8





#5
Phenol-d6
Concen: 0.765 ng
RT: 6.937 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8




Tgt Ion: 99 Resp: 19989
Ion Ratio Lower Upper
99 100
42 20.7 16.6 24.8
71 33.2 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025




#6
bis(2-Chloroethyl)ether
Concen: 0.796 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Tgt Ion: 93 Resp: 20444
Ion Ratio Lower Upper
93 100
63 75.0 59.3 88.9
95 31.6 25.2 37.8





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8



Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	6.3	5.2	7.8
68	5.2	4.1	6.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#8
Nitrobenzene-d5
Concen: 0.746 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion	Ratio	Lower	Upper
82	100		
128	41.9	34.1	51.1
54	53.4	43.5	65.3





#9

Naphthalene

Concen: 0.784 ng

RT: 10.626 min Scan# 941

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :





SSTDICC0.8

Tgt Ion:	128	Resp:	5862
Ion Ratio	Lower	Upper	
128	100		
129	11.0	9.1	13.7
127	13.2	10.5	15.7

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 10/29/2025

Supervised By :mohammad ahmed 10/30/2025



#10

Hexachlorobutadiene

Concen: 0.783 ng




RT: 10.925 min Scan# 969


Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49


Tgt Ion:	225	Resp:	9875
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.8	50.3	75.5





#11
2-Methylnaphthalene-d10
Concen: 0.769 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion:152 Resp: 29800
Ion Ratio Lower Upper
152 100
151 21.2 16.8 25.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025




#12
2-Methylnaphthalene
Concen: 0.776 ng
RT: 12.248 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Tgt Ion:142 Resp: 38674
Ion Ratio Lower Upper
142 100
141 90.1 70.3 105.5
115 35.9 27.8 41.6





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion:164 Resp: 14920
Ion Ratio Lower Upper
164 100
162 104.0 83.0 124.4
160 51.9 40.7 61.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025




#14
2,4,6-Tribromophenol
Concen: 0.736 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Tgt Ion:330 Resp: 4525
Ion Ratio Lower Upper
330 100
332 95.3 76.6 114.8
141 41.9 34.1 51.1





#15
2-Fluorobiphenyl
Concen: 0.784 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:172 Resp: 4684
Ion Ratio Lower Upper
172 100
171 34.3 27.8 41.6
170 22.4 18.1 27.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#16
Acenaphthylene
Concen: 0.779 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:152 Resp: 52624
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.8
153 12.9 10.2 15.2





#17
Acenaphthene
Concen: 0.779 ng
RT: 14.494 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:154 Resp: 3684
Ion Ratio Lower Upper
154 100
153 112.3 90.0 135.0
152 59.3 47.3 70.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#18
Fluorene
Concen: 0.794 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:166 Resp: 49204
Ion Ratio Lower Upper
166 100
165 98.7 79.0 118.4
167 13.3 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:188 Resp: 28500
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.1 8.6 13.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#20
4,6-Dinitro-2-methylphenol
Concen: 0.689 ng
RT: 15.560 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:198 Resp: 2898
Ion Ratio Lower Upper
198 100
51 69.5 88.9 133.3#
105 49.4 49.2 73.8





#21
4-Bromophenyl-phenylether
Concen: 0.780 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8




Tgt Ion:248 Resp: 1457
Ion Ratio Lower Upper
248 100
250 93.7 76.2 114.2
141 68.4 57.6 86.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025




#22
Hexachlorobenzene
Concen: 0.786 ng
RT: 16.503 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Tgt Ion:284 Resp: 16700
Ion Ratio Lower Upper
284 100
142 38.0 30.5 45.7
249 30.3 23.5 35.3





#23
Atrazine
Concen: 0.745 ng
RT: 16.652 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:200 Resp: 10204
Ion Ratio Lower Upper
200 100
173 26.2 21.4 32.2
215 46.0 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#24
Pentachlorophenol
Concen: 0.707 ng
RT: 16.838 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:266 Resp: 6470
Ion Ratio Lower Upper
266 100
264 62.0 49.3 73.9
268 64.3 50.6 75.8





#25
Phenanthrene
Concen: 0.779 ng
RT: 17.223 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion:178 Resp: 7045
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.3 12.2 18.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025




#26
Anthracene
Concen: 0.761 ng
RT: 17.322 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49



Tgt Ion:178 Resp: 60270
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.2 11.8 17.8









#27
Fluoranthene-d10
Concen: 0.778 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:212 Resp: 55912
Ion Ratio Lower Upper
212 100
106 16.0 12.6 19.0
104 9.0 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#28
Fluoranthene
Concen: 0.786 ng
RT: 19.253 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:202 Resp: 79478
Ion Ratio Lower Upper
202 100
101 11.8 9.4 14.2
203 17.0 13.4 20.2





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8




Tgt Ion:240 Resp: 2353
Ion Ratio Lower Upper
240 100
120 13.1 10.7 16.1
236 28.4 23.1 34.7

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025




#30
Pyrene
Concen: 0.737 ng
RT: 19.615 min Scan# 2032
Delta R.T. 0.005 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49



Tgt Ion:202 Resp: 78307
Ion Ratio Lower Upper
202 100
200 21.2 16.7 25.1
203 17.7 14.2 21.2





#31
 Terphenyl-d14
 Concen: 0.743 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Instrument :

BNA_N





ClientSampleId :

SSTDICC0.8

Tgt Ion:244 Resp: 37975
 Ion Ratio Lower Upper
 244 100
 212 9.3 7.8 11.6
 122 18.4 15.2 22.8




Manual Integrations
 APPROVED


Reviewed By :Rahul Chavli 10/29/2025
 Supervised By :mohammad ahmed 10/30/2025



#32
 Benzo(a)anthracene
 Concen: 0.764 ng
 RT: 21.357 min Scan# 2267
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion:228 Resp: 64140
 Ion Ratio Lower Upper
 228 100
 226 27.7 22.2 33.4
 229 19.6 15.8 23.8









#33
 Chrysene
 Concen: 0.787 ng
 RT: 21.411 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion: 228 Resp: 7167

Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.3	36.5
229	19.6	15.7	23.5




Manual Integrations
APPROVED
 Reviewed By :Rahul Chavli 10/29/2025
 Supervised By :mohammad ahmed 10/30/2025




#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.745 ng
 RT: 21.295 min Scan# 2260
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion: 149 Resp: 32963


Ion	Ratio	Lower	Upper
149	100		
167	26.8	21.1	31.7
279	3.1	2.9	4.3





#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.683 min Scan# 2821
 Delta R.T. 0.003 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8




Tgt Ion:264 Resp: 20307
 Ion Ratio Lower Upper
 264 100
 260 26.8 21.8 32.8
 265 76.5 68.6 102.8

Manual Integrations
 APPROVED



Reviewed By :Rahul Chavli 10/29/2025
 Supervised By :mohammad ahmed 10/30/2025




#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.789 ng
 RT: 26.039 min Scan# 3627
 Delta R.T. 0.006 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49




Tgt Ion:276 Resp: 65796
 Ion Ratio Lower Upper
 276 100
 138 27.3 21.4 32.0
 277 24.7 19.2 28.8





#37
Benzo(b)fluoranthene
Concen: 0.805 ng m
RT: 22.984 min Scan# 2582
Delta R.T. 0.003 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8




Tgt Ion:252 Resp: 6991
Ion Ratio Lower Upper
252 100
253 23.4 20.9 31.3
125 14.0 14.1 21.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025




#38
Benzo(k)fluoranthene
Concen: 0.776 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.003 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49



Tgt Ion:252 Resp: 67771
Ion Ratio Lower Upper
252 100
253 23.4 21.2 31.8
125 14.2 14.6 22.0






#39
Benzo(a)pyrene
Concen: 0.768 ng
RT: 23.578 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :




SSTDIC0.8




Tgt Ion:252 Resp: 5336
Ion Ratio Lower Upper
252 100
253 24.4 23.1 34.7
125 17.2 18.6 28.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025




#40
Dibenzo(a,h)anthracene
Concen: 0.784 ng
RT: 26.054 min Scan# 3632
Delta R.T. 0.006 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Tgt Ion:278 Resp: 50344
Ion Ratio Lower Upper
278 100
139 19.9 17.6 26.4
279 25.0 22.6 33.8





#41
Benzo(g,h,i)perylene
Concen: 0.793 ng
RT: 26.753 min Scan# 3869
Delta R.T. 0.003 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49



Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:276 Resp: 5815
Ion Ratio Lower Upper
276 100
277 24.0 19.7 29.5
138 24.9 20.2 30.4

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038115.D
 Acq On : 28 Oct 2025 15:25
 Operator : RC/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Oct 28 17:31:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	9872	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	29279	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	16624	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	31469	0.400	ng	0.00
29) Chrysene-d12	21.375	240	25173	0.400	ng	0.00
35) Perylene-d12	23.680	264	19529	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	36286	1.560	ng	0.00
5) Phenol-d6	6.937	99	44651	1.576	ng	0.00
8) Nitrobenzene-d5	8.929	82	36991	1.566	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	68673	1.642	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	11390	1.663	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	107386	1.613	ng	0.00
27) Fluoranthene-d10	19.220	212	132146	1.665	ng	0.00
31) Terphenyl-d14	19.819	244	89776	1.642	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	16443	1.612	ng	99
3) n-Nitrosodimethylamine	3.557	42	21596	1.645	ng	96
6) bis(2-Chloroethyl)ether	7.197	93	45315	1.628	ng	99
9) Naphthalene	10.626	128	131755	1.634	ng	99
10) Hexachlorobutadiene	10.925	225	22006	1.617	ng	# 99
12) 2-Methylnaphthalene	12.248	142	88687	1.649	ng	99
16) Acenaphthylene	14.152	152	125798	1.671	ng	99
17) Acenaphthene	14.495	154	87659	1.664	ng	98
18) Fluorene	15.489	166	117574	1.703	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	7997	1.534	ng	# 54
21) 4-Bromophenyl-phenylether	16.379	248	34434	1.669	ng	95
22) Hexachlorobenzene	16.503	284	38518	1.642	ng	99
23) Atrazine	16.652	200	25585	1.693	ng	97
24) Pentachlorophenol	16.838	266	16638	1.646	ng	99
25) Phenanthrene	17.223	178	166989	1.671	ng	100
26) Anthracene	17.322	178	148754	1.702	ng	100
28) Fluoranthene	19.253	202	187774	1.682	ng	99
30) Pyrene	19.610	202	184735	1.626	ng	100
32) Benzo(a)anthracene	21.358	228	147889	1.647	ng	99
33) Chrysene	21.411	228	156927	1.610	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	72607	1.535	ng	99
36) Indeno(1,2,3-cd)pyrene	26.031	276	135169	1.685	ng	99
37) Benzo(b)fluoranthene	22.981	252	141041	1.689	ng	# 91
38) Benzo(k)fluoranthene	23.028	252	143019	1.702	ng	# 90
39) Benzo(a)pyrene	23.578	252	110949	1.660	ng	# 85
40) Dibenzo(a,h)anthracene	26.048	278	105589	1.710	ng	92
41) Benzo(g,h,i)perylene	26.747	276	117071	1.660	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038115.D
Acq On : 28 Oct 2025 15:25
Operator : RC/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Oct 28 17:31:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.775 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:152 Resp: 9872
 Ion Ratio Lower Upper
 152 100
 150 157.2 122.3 183.5
 115 60.8 47.4 71.0



#2
 1,4-Dioxane
 Concen: 1.612 ng
 RT: 3.254 min Scan# 24
 Delta R.T. 0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25

Tgt Ion: 88 Resp: 16443
 Ion Ratio Lower Upper
 88 100
 43 30.7 24.3 36.5
 58 76.6 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 1.645 ng
RT: 3.557 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 21596
Ion Ratio Lower Upper
42 100
74 115.8 96.8 145.2
44 6.3 5.3 7.9



#4
2-Fluorophenol
Concen: 1.560 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion: 112 Resp: 36286
Ion Ratio Lower Upper
112 100
64 56.2 45.4 68.0
63 29.9 23.2 34.8









#5
Phenol-d6
Concen: 1.576 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	99	Resp:	44651
Ion	Ratio	Lower	Upper
99	100		
42	20.8	16.6	24.8
71	32.5	25.4	38.2



#6
bis(2-Chloroethyl)ether
Concen: 1.628 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:	93	Resp:	45315
Ion	Ratio	Lower	Upper
93	100		
63	74.8	59.3	88.9
95	31.9	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	29279
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.4
54	6.1	5.2	7.8
68	5.0	4.1	6.1



#8
Nitrobenzene-d5
Concen: 1.566 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:	82	Resp:	36991
Ion	Ratio	Lower	Upper
82	100		
128	43.5	34.1	51.1
54	51.3	43.5	65.3









#9
Naphthalene
Concen: 1.634 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 131755
Ion Ratio Lower Upper
128 100
129 10.8 9.1 13.7
127 13.1 10.5 15.7



#10
Hexachlorobutadiene
Concen: 1.617 ng
RT: 10.925 min Scan# 969
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:225 Resp: 22006
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 1.642 ng
RT: 12.177 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 68673
Ion Ratio Lower Upper
152 100
151 20.9 16.8 25.2



#12
2-Methylnaphthalene
Concen: 1.649 ng
RT: 12.248 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25


Tgt Ion:142 Resp: 88687
Ion Ratio Lower Upper
142 100
141 88.5 70.3 105.5
115 34.6 27.8 41.6








#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 14
Delta R.T. 0.011 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6





Tgt Ion:164 Resp: 16624
Ion Ratio Lower Upper
164 100
162 103.5 83.0 124.4
160 50.5 40.7 61.1




#14
2,4,6-Tribromophenol
Concen: 1.663 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25



Tgt Ion:330 Resp: 11390
Ion Ratio Lower Upper
330 100
332 97.4 76.6 114.8
141 41.9 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 1.613 ng
RT: 13.051 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 107386
Ion Ratio Lower Upper
172 100
171 35.9 27.8 41.6
170 24.3 18.1 27.1



#16
Acenaphthylene
Concen: 1.671 ng
RT: 14.152 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:152 Resp: 125798
Ion Ratio Lower Upper
152 100
151 19.4 15.8 23.8
153 13.0 10.2 15.2









#17
Acenaphthene
Concen: 1.664 ng
RT: 14.495 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 87659
Ion Ratio Lower Upper
154 100
153 110.3 90.0 135.0
152 57.5 47.3 70.9



#18
Fluorene
Concen: 1.703 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:166 Resp: 117574
Ion Ratio Lower Upper
166 100
165 98.4 79.0 118.4
167 13.2 10.7 16.1









#19
 Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.186 min Scan# 1688
 Delta R.T. 0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25




Instrument : BNA_N
 ClientSampleId : SSTDICC1.6


Tgt Ion:	188	Resp:	31469
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	10.6	8.6	13.0



#20
 4,6-Dinitro-2-methylphenol
 Concen: 1.534 ng
 RT: 15.560 min Scan# 1557
 Delta R.T. -0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25

Tgt Ion:	198	Resp:	7997
Ion Ratio	Lower	Upper	
198	100		
51	49.4	88.9	133.3#
105	43.5	49.2	73.8#









#21
4-Bromophenyl-phenylether
Concen: 1.669 ng
RT: 16.379 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 34434
Ion Ratio Lower Upper
248 100
250 95.6 76.2 114.2
141 62.3 57.6 86.4



#22
Hexachlorobenzene
Concen: 1.642 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:284 Resp: 38518
Ion Ratio Lower Upper
284 100
142 38.3 30.5 45.7
249 30.0 23.5 35.3









#23
Atrazine
 Concen: 1.693 ng
 RT: 16.652 min Scan# 1645
 Delta R.T. 0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25




Instrument : BNA_N
ClientSampleId : SSTDICC1.6


Tgt Ion:	200	Resp:	25585
Ion Ratio	Lower	Upper	
200	100		
173	23.2	21.4	32.2
215	45.1	36.7	55.1



#24
Pentachlorophenol
 Concen: 1.646 ng
 RT: 16.838 min Scan# 1660
 Delta R.T. 0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25

Tgt Ion:	266	Resp:	16638
Ion Ratio	Lower	Upper	
266	100		
264	62.5	49.3	73.9
268	63.5	50.6	75.8









#25
 Phenanthrene
 Concen: 1.671 ng
 RT: 17.223 min Scan# 1691
 Delta R.T. 0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25

Instrument : BNA_N
 ClientSampleId : SSTDICC1.6

Tgt Ion:178 Resp: 166989




Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.5	23.3
179	15.2	12.2	18.4




#26
 Anthracene
 Concen: 1.702 ng
 RT: 17.322 min Scan# 1699
 Delta R.T. 0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25

Tgt Ion:178 Resp: 148754

Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.9	22.3
179	15.2	11.8	17.8









#27
Fluoranthene-d10
Concen: 1.665 ng
RT: 19.220 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 132146
Ion Ratio Lower Upper
212 100
106 16.0 12.6 19.0
104 8.9 7.1 10.7



#28
Fluoranthene
Concen: 1.682 ng
RT: 19.253 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:202 Resp: 187774
Ion Ratio Lower Upper
202 100
101 11.9 9.4 14.2
203 17.2 13.4 20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 25173
Ion Ratio Lower Upper
240 100
120 11.5 10.7 16.1
236 28.0 23.1 34.7



#30
Pyrene
Concen: 1.626 ng
RT: 19.610 min Scan# 2031
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:202 Resp: 184735
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.7 14.2 21.2









#31
 Terphenyl-d14
 Concen: 1.642 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. 0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:	244	Resp:	89776
Ion Ratio	Lower	Upper	
244	100		
212	9.1	7.8	11.6
122	17.8	15.2	22.8



#32
 Benzo(a)anthracene
 Concen: 1.647 ng
 RT: 21.358 min Scan# 2267
 Delta R.T. 0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25

Tgt Ion:	228	Resp:	147889
Ion Ratio	Lower	Upper	
228	100		
226	26.8	22.2	33.4
229	19.6	15.8	23.8









#33
Chrysene
Concen: 1.610 ng
RT: 21.411 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 156927
Ion Ratio Lower Upper
228 100
226 29.8 24.3 36.5
229 19.7 15.7 23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.535 ng
RT: 21.295 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:149 Resp: 72607
Ion Ratio Lower Upper
149 100
167 26.6 21.1 31.7
279 3.0 2.9 4.3









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.680 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:264 Resp: 19529
 Ion Ratio Lower Upper
 264 100
 260 27.3 21.8 32.8
 265 78.3 68.6 102.8



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 1.685 ng
 RT: 26.031 min Scan# 3624
 Delta R.T. -0.003 min
 Lab File: BN038115.D
 Acq: 28 Oct 2025 15:25

Tgt Ion:276 Resp: 135169
 Ion Ratio Lower Upper
 276 100
 138 27.6 21.4 32.0
 277 24.5 19.2 28.8





#37

Benzo(b)fluoranthene

Concen: 1.689 ng

RT: 22.981 min Scan# 2581

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 141041

Ion Ratio Lower Upper

252 100

253 22.5 20.9 31.3

125 12.6 14.1 21.1#



#38

Benzo(k)fluoranthene

Concen: 1.702 ng

RT: 23.028 min Scan# 2597

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25




Tgt Ion:252 Resp: 143019


Ion Ratio Lower Upper

252 100

253 22.2 21.2 31.8


125 12.5 14.6 22.0#








#39
Benzo(a)pyrene
Concen: 1.660 ng
RT: 23.578 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:252 Resp: 110949
Ion Ratio Lower Upper
252 100
253 22.7 23.1 34.7#
125 14.4 18.6 28.0#



#40
Dibenzo(a,h)anthracene
Concen: 1.710 ng
RT: 26.048 min Scan# 3630
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25


Tgt Ion:278 Resp: 105589
Ion Ratio Lower Upper
278 100
139 18.7 17.6 26.4
279 23.9 22.6 33.8







#41
Benzo(g,h,i)perylene
Concen: 1.660 ng
RT: 26.747 min Scan# 3869
Delta R.T. -0.003 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:276 Resp: 117071
Ion Ratio Lower Upper
276 100
277 23.9 19.7 29.5
138 24.2 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038116.D
 Acq On : 28 Oct 2025 16:02
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Oct 28 17:31:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	9743	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	28864	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	16547	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	31206	0.400	ng	0.00
29) Chrysene-d12	21.375	240	26115	0.400	ng	0.00
35) Perylene-d12	23.683	264	19904	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	69778	3.040	ng	0.00
5) Phenol-d6	6.937	99	88646	3.171	ng	0.00
8) Nitrobenzene-d5	8.929	82	70830	3.041	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	134773	3.270	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	24045	3.527	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	200437	3.024	ng	0.01
27) Fluoranthene-d10	19.220	212	263418	3.348	ng	0.00
31) Terphenyl-d14	19.819	244	180175	3.176	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	31207	3.100	ng	99
3) n-Nitrosodimethylamine	3.557	42	41593	3.211	ng	96
6) bis(2-Chloroethyl)ether	7.197	93	86904	3.164	ng	99
9) Naphthalene	10.626	128	255606	3.215	ng	99
10) Hexachlorobutadiene	10.925	225	42325	3.155	ng	# 99
12) 2-Methylnaphthalene	12.248	142	172992	3.263	ng	98
16) Acenaphthylene	14.152	152	253121	3.378	ng	99
17) Acenaphthene	14.495	154	173433	3.308	ng	97
18) Fluorene	15.489	166	230067	3.348	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	18221	3.363	ng	# 47
21) 4-Bromophenyl-phenylether	16.379	248	67616	3.305	ng	99
22) Hexachlorobenzene	16.503	284	74101	3.185	ng	99
23) Atrazine	16.652	200	52181	3.482	ng	97
24) Pentachlorophenol	16.838	266	36424	3.634	ng	99
25) Phenanthrene	17.223	178	322719	3.257	ng	100
26) Anthracene	17.322	178	295333	3.408	ng	99
28) Fluoranthene	19.253	202	372238	3.363	ng	99
30) Pyrene	19.615	202	368946	3.130	ng	100
32) Benzo(a)anthracene	21.358	228	308037	3.306	ng	99
33) Chrysene	21.411	228	317065	3.136	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	157480	3.209	ng	99
36) Indeno(1,2,3-cd)pyrene	26.031	276	276870	3.386	ng	98
37) Benzo(b)fluoranthene	22.984	252	292856	3.441	ng	# 90
38) Benzo(k)fluoranthene	23.031	252	292143	3.411	ng	# 89
39) Benzo(a)pyrene	23.578	252	231067	3.392	ng	# 84
40) Dibenzo(a,h)anthracene	26.051	278	217022	3.449	ng	92
41) Benzo(g,h,i)perylene	26.750	276	235264	3.274	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038116.D
Acq On : 28 Oct 2025 16:02
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Oct 28 17:31:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 9743
Ion Ratio Lower Upper
152 100
150 157.0 122.3 183.5
115 61.3 47.4 71.0



#2
1,4-Dioxane
Concen: 3.100 ng
RT: 3.254 min Scan# 24
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion: 88 Resp: 31207
Ion Ratio Lower Upper
88 100
43 30.6 24.3 36.5
58 76.1 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 3.211 ng
RT: 3.557 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 41593
Ion Ratio Lower Upper
42 100
74 116.1 96.8 145.2
44 6.8 5.3 7.9



#4
2-Fluorophenol
Concen: 3.040 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion: 112 Resp: 69778
Ion Ratio Lower Upper
112 100
64 56.1 45.4 68.0
63 30.0 23.2 34.8









#5
Phenol-d6
Concen: 3.171 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 88646
Ion Ratio Lower Upper
99 100
42 20.7 16.6 24.8
71 32.4 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 3.164 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion: 93 Resp: 86904
Ion Ratio Lower Upper
93 100
63 75.1 59.3 88.9
95 32.0 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	28864
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.5	5.2	7.8
68	4.9	4.1	6.1



#8
Nitrobenzene-d5
Concen: 3.041 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:	82	Resp:	70830
Ion	Ratio	Lower	Upper
82	100		
128	42.8	34.1	51.1
54	51.5	43.5	65.3









#9
Naphthalene
Concen: 3.215 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 255606
Ion Ratio Lower Upper
128 100
129 10.9 9.1 13.7
127 13.0 10.5 15.7



#10
Hexachlorobutadiene
Concen: 3.155 ng
RT: 10.925 min Scan# 969
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:225 Resp: 42325
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 3.270 ng
RT: 12.177 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 134773
Ion Ratio Lower Upper
152 100
151 20.9 16.8 25.2



#12
2-Methylnaphthalene
Concen: 3.263 ng
RT: 12.248 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02


Tgt Ion:142 Resp: 172992
Ion Ratio Lower Upper
142 100
141 90.0 70.3 105.5
115 35.2 27.8 41.6








#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:164 Resp: 16547
Ion Ratio Lower Upper
164 100
162 102.2 83.0 124.4
160 50.8 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 3.527 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:330 Resp: 24045
Ion Ratio Lower Upper
330 100
332 97.1 76.6 114.8
141 42.2 34.1 51.1









#15
 2-Fluorobiphenyl
 Concen: 3.024 ng
 RT: 13.062 min Scan# 1324
 Delta R.T. 0.011 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	172	Resp:	200437
Ion Ratio	Lower	Upper	
172	100		
171	34.2	27.8	41.6
170	22.3	18.1	27.1



#16
 Acenaphthylene
 Concen: 3.378 ng
 RT: 14.152 min Scan# 1426
 Delta R.T. 0.000 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02

Tgt Ion:	152	Resp:	253121
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.8	23.8
153	12.9	10.2	15.2









#17
Acenaphthene
Concen: 3.308 ng
RT: 14.495 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 173433
Ion Ratio Lower Upper
154 100
153 109.4 90.0 135.0
152 56.0 47.3 70.9



#18
Fluorene
Concen: 3.348 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:166 Resp: 230067
Ion Ratio Lower Upper
166 100
165 99.1 79.0 118.4
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 31206
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.6 8.6 13.0



#20
4,6-Dinitro-2-methylphenol
Concen: 3.363 ng
RT: 15.560 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:198 Resp: 18221
Ion Ratio Lower Upper
198 100
51 40.8 88.9 133.3#
105 38.5 49.2 73.8#









#21
4-Bromophenyl-phenylether
Concen: 3.305 ng
RT: 16.379 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 67616
Ion Ratio Lower Upper
248 100
250 95.3 76.2 114.2
141 69.9 57.6 86.4



#22
Hexachlorobenzene
Concen: 3.185 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:284 Resp: 74101
Ion Ratio Lower Upper
284 100
142 38.7 30.5 45.7
249 29.9 23.5 35.3









#23
 Atrazine
 Concen: 3.482 ng
 RT: 16.652 min Scan# 1645
 Delta R.T. 0.000 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	200	Resp:	52181
Ion Ratio	Lower	Upper	
200	100		
173	24.0	21.4	32.2
215	45.3	36.7	55.1



#24
 Pentachlorophenol
 Concen: 3.634 ng
 RT: 16.838 min Scan# 1660
 Delta R.T. 0.000 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02

Tgt Ion:	266	Resp:	36424
Ion Ratio	Lower	Upper	
266	100		
264	62.5	49.3	73.9
268	63.4	50.6	75.8









#25
 Phenanthrene
 Concen: 3.257 ng
 RT: 17.223 min Scan# 1691
 Delta R.T. 0.000 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02

Instrument : BNA_N
 ClientSampleId : SSTDICC3.2

Tgt Ion:178 Resp: 322719




Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.5	23.3
179	15.3	12.2	18.4




#26
 Anthracene
 Concen: 3.408 ng
 RT: 17.322 min Scan# 1699
 Delta R.T. 0.000 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02

Tgt Ion:178 Resp: 295333

Ion	Ratio	Lower	Upper
178	100		
176	18.8	14.9	22.3
179	15.2	11.8	17.8









#27
Fluoranthene-d10
Concen: 3.348 ng
RT: 19.220 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 263418
Ion Ratio Lower Upper
212 100
106 15.8 12.6 19.0
104 8.9 7.1 10.7



#28
Fluoranthene
Concen: 3.363 ng
RT: 19.253 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:202 Resp: 372238
Ion Ratio Lower Upper
202 100
101 11.8 9.4 14.2
203 17.2 13.4 20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 26115
Ion Ratio Lower Upper
240 100
120 12.3 10.7 16.1
236 28.4 23.1 34.7



#30
Pyrene
Concen: 3.130 ng
RT: 19.615 min Scan# 2032
Delta R.T. 0.005 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:202 Resp: 368946
Ion Ratio Lower Upper
202 100
200 21.1 16.7 25.1
203 17.7 14.2 21.2









#31
 Terphenyl-d14
 Concen: 3.176 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. 0.000 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:244 Resp: 180175
 Ion Ratio Lower Upper
 244 100
 212 9.0 7.8 11.6
 122 17.3 15.2 22.8



#32
 Benzo(a)anthracene
 Concen: 3.306 ng
 RT: 21.358 min Scan# 2267
 Delta R.T. 0.000 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02

Tgt Ion:228 Resp: 308037
 Ion Ratio Lower Upper
 228 100
 226 27.3 22.2 33.4
 229 19.5 15.8 23.8









#33
 Chrysene
 Concen: 3.136 ng
 RT: 21.411 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:228 Resp: 317065




Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.3	36.5
229	19.5	15.7	23.5




#34
 Bis(2-ethylhexyl)phthalate
 Concen: 3.209 ng
 RT: 21.295 min Scan# 2260
 Delta R.T. 0.000 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02

Tgt Ion:149 Resp: 157480

Ion	Ratio	Lower	Upper
149	100		
167	26.9	21.1	31.7
279	3.0	2.9	4.3









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.683 min Scan# 21
 Delta R.T. 0.003 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	264	Resp:	19904
Ion Ratio	Lower	Upper	
264	100		
260	26.5	21.8	32.8
265	76.0	68.6	102.8



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 3.386 ng
 RT: 26.031 min Scan# 3624
 Delta R.T. -0.003 min
 Lab File: BN038116.D
 Acq: 28 Oct 2025 16:02

Tgt Ion:	276	Resp:	276870
Ion Ratio	Lower	Upper	
276	100		
138	27.4	21.4	32.0
277	24.8	19.2	28.8









#37
Benzo(b)fluoranthene
Concen: 3.441 ng
RT: 22.984 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 292856
Ion Ratio Lower Upper
252 100
253 22.3 20.9 31.3
125 11.7 14.1 21.1#



#38
Benzo(k)fluoranthene
Concen: 3.411 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.003 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02


Tgt Ion:252 Resp: 292143
Ion Ratio Lower Upper
252 100
253 22.1 21.2 31.8
125 11.6 14.6 22.0#







#39
Benzo(a)pyrene
Concen: 3.392 ng
RT: 23.578 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:252 Resp: 231067
Ion Ratio Lower Upper
252 100
253 22.1 23.1 34.7#
125 13.4 18.6 28.0#



#40
Dibenzo(a,h)anthracene
Concen: 3.449 ng
RT: 26.051 min Scan# 3631
Delta R.T. 0.003 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02


Tgt Ion:278 Resp: 217022
Ion Ratio Lower Upper
278 100
139 18.4 17.6 26.4
279 23.7 22.6 33.8







#41
Benzo(g,h,i)perylene
Concen: 3.274 ng
RT: 26.750 min Scan# 31
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:276 Resp: 235264
Ion Ratio Lower Upper
276 100
277 23.7 19.7 29.5
138 24.1 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038117.D
 Acq On : 28 Oct 2025 16:38
 Operator : RC/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Oct 28 17:31:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	10579	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	31197	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	17960	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	33242	0.400	ng	0.00
29) Chrysene-d12	21.375	240	25416	0.400	ng	0.00
35) Perylene-d12	23.680	264	18172	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	116608	4.679	ng	0.00
5) Phenol-d6	6.937	99	150343	4.953	ng	0.00
8) Nitrobenzene-d5	8.929	82	119764	4.757	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	225769	5.068	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	43520	5.881	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	335695	4.667	ng	0.00
27) Fluoranthene-d10	19.220	212	419371	5.003	ng	0.00
31) Terphenyl-d14	19.819	244	283126	5.127	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	51426	4.705	ng	Qvalue 100
3) n-Nitrosodimethylamine	3.557	42	68522	4.871	ng	96
6) bis(2-Chloroethyl)ether	7.197	93	142662	4.784	ng	99
9) Naphthalene	10.626	128	422794	4.920	ng	99
10) Hexachlorobutadiene	10.925	225	68839	4.748	ng	# 99
12) 2-Methylnaphthalene	12.248	142	287955	5.025	ng	99
16) Acenaphthylene	14.152	152	428380	5.267	ng	99
17) Acenaphthene	14.494	154	291166	5.117	ng	95
18) Fluorene	15.489	166	389813	5.227	ng	98
21) 4-Bromophenyl-phenylether	16.379	248	114170	5.239	ng	# 92
22) Hexachlorobenzene	16.503	284	123999	5.003	ng	99
23) Atrazine	16.652	200	87823	5.501	ng	97
24) Pentachlorophenol	16.838	266	66071	6.188	ng	99
25) Phenanthrene	17.223	178	539294	5.110	ng	100
26) Anthracene	17.322	178	499584	5.411	ng	99
28) Fluoranthene	19.253	202	587710	4.984	ng	99
30) Pyrene	19.610	202	586526	5.112	ng	100
32) Benzo(a)anthracene	21.358	228	462496	5.100	ng	100
33) Chrysene	21.411	228	473389	4.811	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	223481	4.679	ng	99
36) Indeno(1,2,3-cd)pyrene	26.034	276	403825	5.409	ng	99
37) Benzo(b)fluoranthene	22.981	252	413811	5.326	ng	# 90
38) Benzo(k)fluoranthene	23.028	252	421255	5.387	ng	# 88
39) Benzo(a)pyrene	23.578	252	332107	5.340	ng	# 83
40) Dibenzo(a,h)anthracene	26.048	278	318538	5.545	ng	91
41) Benzo(g,h,i)perylene	26.750	276	342713	5.223	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038117.D
Acq On : 28 Oct 2025 16:38
Operator : RC/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Quant Time: Oct 28 17:31:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.775 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN038117.D
 Acq: 28 Oct 2025 16:38


Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0





Tgt Ion:152 Resp: 10579
 Ion Ratio Lower Upper
 152 100
 150 155.6 122.3 183.5
 115 60.1 47.4 71.0




#2
 1,4-Dioxane
 Concen: 4.705 ng
 RT: 3.254 min Scan# 24
 Delta R.T. 0.000 min
 Lab File: BN038117.D
 Acq: 28 Oct 2025 16:38



Tgt Ion: 88 Resp: 51426
 Ion Ratio Lower Upper
 88 100
 43 30.5 24.3 36.5
 58 75.4 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 4.871 ng
RT: 3.557 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 68522
Ion Ratio Lower Upper
42 100
74 116.1 96.8 145.2
44 6.5 5.3 7.9



#4
2-Fluorophenol
Concen: 4.679 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion: 112 Resp: 116608
Ion Ratio Lower Upper
112 100
64 56.1 45.4 68.0
63 29.9 23.2 34.8









#5
Phenol-d6
Concen: 4.953 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 150343
Ion Ratio Lower Upper
99 100
42 20.9 16.6 24.8
71 32.5 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 4.784 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion: 93 Resp: 142662
Ion Ratio Lower Upper
93 100
63 74.9 59.3 88.9
95 31.9 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	31197
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.0	13.4
54	9.9	5.2	7.8#
68	7.3	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 4.757 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:	82	Resp:	119764
Ion Ratio	Lower	Upper	
82	100		
128	44.7	34.1	51.1
54	50.9	43.5	65.3





#9

Naphthalene

Concen: 4.920 ng

RT: 10.626 min Scan# 941

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:128 Resp: 422794

Ion Ratio Lower Upper

128 100

129 10.9 9.1 13.7

127 13.0 10.5 15.7



#10

Hexachlorobutadiene

Concen: 4.748 ng

RT: 10.925 min Scan# 969

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38




Tgt Ion:225 Resp: 68839


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.6 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 5.068 ng
RT: 12.177 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 225769
Ion Ratio Lower Upper
152 100
151 20.8 16.8 25.2



#12
2-Methylnaphthalene
Concen: 5.025 ng
RT: 12.248 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:142 Resp: 287955
Ion Ratio Lower Upper
142 100
141 89.6 70.3 105.5
115 34.8 27.8 41.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.430 min Scan# 1452

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:164 Resp: 17960

Ion Ratio Lower Upper

164 100

162 106.0 83.0 124.4

160 52.4 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 5.881 ng

RT: 15.932 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38




Tgt Ion:330 Resp: 43520


Ion Ratio Lower Upper

330 100

332 97.0 76.6 114.8

141 41.2 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 4.667 ng
RT: 13.051 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 335695
Ion Ratio Lower Upper
172 100
171 35.8 27.8 41.6
170 24.2 18.1 27.1



#16
Acenaphthylene
Concen: 5.267 ng
RT: 14.152 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:152 Resp: 428380
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.8
153 13.0 10.2 15.2









#17
Acenaphthene
Concen: 5.117 ng
RT: 14.494 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 291166
Ion Ratio Lower Upper
154 100
153 108.1 90.0 135.0
152 55.0 47.3 70.9



#18
Fluorene
Concen: 5.227 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:166 Resp: 389813
Ion Ratio Lower Upper
166 100
165 97.0 79.0 118.4
167 13.1 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:188 Resp: 33242

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.4 8.6 13.0



#21

4-Bromophenyl-phenylether

Concen: 5.239 ng

RT: 16.379 min Scan# 1623

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38




Tgt Ion:248 Resp: 114170


Ion Ratio Lower Upper

248 100

250 95.7 76.2 114.2

141 56.3 57.6 86.4#









#22
Hexachlorobenzene
Concen: 5.003 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:284 Resp: 123999
Ion Ratio Lower Upper
284 100
142 37.6 30.5 45.7
249 29.6 23.5 35.3



#23
Atrazine
Concen: 5.501 ng
RT: 16.652 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:200 Resp: 87823
Ion Ratio Lower Upper
200 100
173 23.4 21.4 32.2
215 45.1 36.7 55.1









#24
 Pentachlorophenol
 Concen: 6.188 ng
 RT: 16.838 min Scan# 1660
 Delta R.T. 0.000 min
 Lab File: BN038117.D
 Acq: 28 Oct 2025 16:38




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:	266	Resp:	66071
Ion Ratio	Lower	Upper	
266	100		
264	62.7	49.3	73.9
268	64.1	50.6	75.8



#25
 Phenanthrene
 Concen: 5.110 ng
 RT: 17.223 min Scan# 1691
 Delta R.T. 0.000 min
 Lab File: BN038117.D
 Acq: 28 Oct 2025 16:38

Tgt Ion:	178	Resp:	539294
Ion Ratio	Lower	Upper	
178	100		
176	19.2	15.5	23.3
179	15.1	12.2	18.4









#26
 Anthracene
 Concen: 5.411 ng
 RT: 17.322 min Scan# 1699
 Delta R.T. 0.000 min
 Lab File: BN038117.D
 Acq: 28 Oct 2025 16:38




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:178 Resp: 499584
 Ion Ratio Lower Upper
 178 100
 176 18.6 14.9 22.3
 179 15.3 11.8 17.8



#27
 Fluoranthene-d10
 Concen: 5.003 ng
 RT: 19.220 min Scan# 1947
 Delta R.T. 0.000 min
 Lab File: BN038117.D
 Acq: 28 Oct 2025 16:38

Tgt Ion:212 Resp: 419371
 Ion Ratio Lower Upper
 212 100
 106 15.8 12.6 19.0
 104 8.9 7.1 10.7









#28
 Fluoranthene
 Concen: 4.984 ng
 RT: 19.253 min Scan# 1954
 Delta R.T. 0.000 min
 Lab File: BN038117.D
 Acq: 28 Oct 2025 16:38

Instrument : BNA_N
 ClientSampleId : SSTDICC5.0

Tgt Ion:202 Resp: 587710




Ion	Ratio	Lower	Upper
202	100		
101	11.7	9.4	14.2
203	17.3	13.4	20.2




#29
 Chrysene-d12
 Concen: 0.400 ng
 RT: 21.375 min Scan# 2269
 Delta R.T. 0.000 min
 Lab File: BN038117.D
 Acq: 28 Oct 2025 16:38

Tgt Ion:240 Resp: 25416

Ion	Ratio	Lower	Upper
240	100		
120	11.7	10.7	16.1
236	28.2	23.1	34.7









#30
 Pyrene
 Concen: 5.112 ng
 RT: 19.610 min Scan# 2031
 Delta R.T. 0.000 min
 Lab File: BN038117.D
 Acq: 28 Oct 2025 16:38

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:202 Resp: 586526




Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.7	25.1
203	17.8	14.2	21.2




#31
 Terphenyl-d14
 Concen: 5.127 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. 0.000 min
 Lab File: BN038117.D
 Acq: 28 Oct 2025 16:38

Tgt Ion:244 Resp: 283126

Ion	Ratio	Lower	Upper
244	100		
212	8.9	7.8	11.6
122	17.3	15.2	22.8









#32
Benzo(a)anthracene
Concen: 5.100 ng
RT: 21.358 min Scan# 21358
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 462496
Ion Ratio Lower Upper
228 100
226 27.4 22.2 33.4
229 19.8 15.8 23.8



#33
Chrysene
Concen: 4.811 ng
RT: 21.411 min Scan# 2273
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:228 Resp: 473389
Ion Ratio Lower Upper
228 100
226 29.8 24.3 36.5
229 19.8 15.7 23.5









#34
Bis(2-ethylhexyl)phthalate
Concen: 4.679 ng
RT: 21.295 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:149 Resp: 223481
Ion Ratio Lower Upper
149 100
167 27.0 21.1 31.7
279 3.1 2.9 4.3



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.680 min Scan# 2820
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:264 Resp: 18172
Ion Ratio Lower Upper
264 100
260 27.2 21.8 32.8
265 82.9 68.6 102.8









#36
Indeno(1,2,3-cd)pyrene
Concen: 5.409 ng
RT: 26.034 min Scan# 3626
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:276 Resp: 403825
Ion Ratio Lower Upper
276 100
138 27.3 21.4 32.0
277 24.9 19.2 28.8



#37
Benzo(b)fluoranthene
Concen: 5.326 ng
RT: 22.981 min Scan# 2581
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:252 Resp: 413811
Ion Ratio Lower Upper
252 100
253 22.0 20.9 31.3
125 11.5 14.1 21.1#









#38
Benzo(k)fluoranthene
Concen: 5.387 ng
RT: 23.028 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 421255
Ion Ratio Lower Upper
252 100
253 21.8 21.2 31.8
125 11.4 14.6 22.0#



#39
Benzo(a)pyrene
Concen: 5.340 ng
RT: 23.578 min Scan# 2785
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38


Tgt Ion:252 Resp: 332107
Ion Ratio Lower Upper
252 100
253 21.9 23.1 34.7#
125 12.9 18.6 28.0#








#40
Dibenzo(a,h)anthracene
Concen: 5.545 ng
RT: 26.048 min Scan# 31
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0






Tgt Ion:278 Resp: 318538
Ion Ratio Lower Upper
278 100
139 18.2 17.6 26.4
279 23.3 22.6 33.8



#41
Benzo(g,h,i)perylene
Concen: 5.223 ng
RT: 26.750 min Scan# 3870
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:276 Resp: 342713
Ion Ratio Lower Upper
276 100
277 23.6 19.7 29.5
138 23.9 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038118.D
 Acq On : 28 Oct 2025 17:51
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102825

Quant Time: Oct 29 17:07:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	7698	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	23240	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	13297	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	26273	0.400	ng	0.00
29) Chrysene-d12	21.375	240	18729	0.400	ng	0.00
35) Perylene-d12	23.683	264	15660	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	6665	0.368	ng	0.00
5) Phenol-d6	6.937	99	7743	0.351	ng	0.00
8) Nitrobenzene-d5	8.929	82	6617	0.353	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	13607	0.410	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1861	0.340	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	21897	0.411	ng	0.01
27) Fluoranthene-d10	19.220	212	25766	0.389	ng	0.00
31) Terphenyl-d14	19.819	244	17430	0.428	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	3503	0.440	ng	98
3) n-Nitrosodimethylamine	3.564	42	4071	0.398	ng	# 96
6) bis(2-Chloroethyl)ether	7.197	93	9255	0.427	ng	100
9) Naphthalene	10.626	128	26051	0.407	ng	99
10) Hexachlorobutadiene	10.925	225	4702	0.435	ng	# 98
12) 2-Methylnaphthalene	12.253	142	16372	0.384	ng	98
16) Acenaphthylene	14.152	152	27388	0.455	ng	100
17) Acenaphthene	14.505	154	16777	0.398	ng	99
18) Fluorene	15.489	166	22859	0.414	ng	99
20) 4,6-Dinitro-2-methylph...	15.560	198	1138	0.411	ng	90
21) 4-Bromophenyl-phenylether	16.379	248	6886	0.400	ng	99
22) Hexachlorobenzene	16.503	284	8349	0.426	ng	98
23) Atrazine	16.652	200	4672	0.370	ng	99
24) Pentachlorophenol	16.838	266	2963	0.351	ng	97
25) Phenanthrene	17.223	178	33992	0.408	ng	99
26) Anthracene	17.322	178	29654	0.406	ng	99
28) Fluoranthene	19.253	202	35401	0.380	ng	99
30) Pyrene	19.615	202	35647	0.422	ng	100
32) Benzo(a)anthracene	21.358	228	26833	0.402	ng	100
33) Chrysene	21.411	228	31031	0.428	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	12587	0.358	ng	99
36) Indeno(1,2,3-cd)pyrene	26.037	276	26115	0.406	ng	98
37) Benzo(b)fluoranthene	22.984	252	27124	0.405	ng	97
38) Benzo(k)fluoranthene	23.031	252	29583	0.439	ng	98
39) Benzo(a)pyrene	23.581	252	23400	0.437	ng	# 95
40) Dibenzo(a,h)anthracene	26.051	278	19931	0.403	ng	97
41) Benzo(g,h,i)perylene	26.756	276	22180	0.392	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038118.D
Acq On : 28 Oct 2025 17:51
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN102825

Quant Time: Oct 29 17:07:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.782 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Instrument :
BNA_N
ClientSampleId :
ICVBN102825

Tgt Ion:152 Resp: 7698




Ion	Ratio	Lower	Upper
152	100		
150	151.3	122.3	183.5
115	57.8	47.4	71.0




#2
1,4-Dioxane
Concen: 0.440 ng
RT: 3.254 min Scan# 24
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion: 88 Resp: 3503

Ion	Ratio	Lower	Upper
88	100		
43	31.0	24.3	36.5
58	77.1	60.4	90.6









#3
 n-Nitrosodimethylamine
 Concen: 0.398 ng
 RT: 3.564 min Scan# 61
 Delta R.T. 0.007 min
 Lab File: BN038118.D
 Acq: 28 Oct 2025 17:51




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102825


Tgt Ion: 42 Resp: 4071
 Ion Ratio Lower Upper
 42 100
 74 124.8 96.8 145.2
 44 10.5 5.3 7.9#



#4
 2-Fluorophenol
 Concen: 0.368 ng
 RT: 5.348 min Scan# 314
 Delta R.T. 0.000 min
 Lab File: BN038118.D
 Acq: 28 Oct 2025 17:51

Tgt Ion: 112 Resp: 6665
 Ion Ratio Lower Upper
 112 100
 64 55.4 45.4 68.0
 63 29.1 23.2 34.8









#5
Phenol-d6
Concen: 0.351 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:	99	Resp:	7743
Ion	Ratio	Lower	Upper
99	100		
42	21.1	16.6	24.8
71	33.2	25.4	38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.427 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:	93	Resp:	9255
Ion	Ratio	Lower	Upper
93	100		
63	73.9	59.3	88.9
95	30.9	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:	136	Resp:	23240
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	6.6	5.2	7.8
68	5.3	4.1	6.1



#8
Nitrobenzene-d5
Concen: 0.353 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:	82	Resp:	6617
Ion	Ratio	Lower	Upper
82	100		
128	41.8	34.1	51.1
54	54.7	43.5	65.3









#9
Naphthalene
Concen: 0.407 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:	128	Resp:	26051
Ion Ratio	Lower	Upper	
128	100		
129	11.3	9.1	13.7
127	13.4	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.435 ng
RT: 10.925 min Scan# 969
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:	225	Resp:	4702
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.4	50.3	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.410 ng
RT: 12.177 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:152 Resp: 13607
Ion Ratio Lower Upper
152 100
151 21.0 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.384 ng
RT: 12.253 min Scan# 1170
Delta R.T. 0.005 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:142 Resp: 16372
Ion Ratio Lower Upper
142 100
141 89.5 70.3 105.5
115 35.3 27.8 41.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.441 min Scan# 1453

Delta R.T. 0.011 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51

Instrument :

BNA_N

ClientSampleId :

ICVBN102825





Tgt Ion:164 Resp: 13297

Ion Ratio Lower Upper

164 100

162 103.7 83.0 124.4

160 52.0 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.340 ng

RT: 15.932 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51




Tgt Ion:330 Resp: 1861


Ion Ratio Lower Upper

330 100

332 94.5 76.6 114.8

141 43.0 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.411 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:	172	Resp:	21897
Ion Ratio	Lower	Upper	
172	100		
171	34.5	27.8	41.6
170	22.9	18.1	27.1



#16
Acenaphthylene
Concen: 0.455 ng
RT: 14.152 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:	152	Resp:	27388
Ion Ratio	Lower	Upper	
152	100		
151	19.8	15.8	23.8
153	13.2	10.2	15.2









#17
Acenaphthene
Concen: 0.398 ng
RT: 14.505 min Scan# 1459
Delta R.T. 0.011 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:154 Resp: 16777
Ion Ratio Lower Upper
154 100
153 113.6 90.0 135.0
152 60.6 47.3 70.9



#18
Fluorene
Concen: 0.414 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:166 Resp: 22859
Ion Ratio Lower Upper
166 100
165 99.9 79.0 118.4
167 13.4 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51

Instrument :

BNA_N

ClientSampleId :

ICVBN102825





Tgt Ion:188 Resp: 26273

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.7 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.411 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51




Tgt Ion:198 Resp: 1138


Ion Ratio Lower Upper

198 100

51 125.7 88.9 133.3

105 63.1 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.400 ng
RT: 16.379 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:248 Resp: 6886
Ion Ratio Lower Upper
248 100
250 94.8 76.2 114.2
141 73.0 57.6 86.4



#22
Hexachlorobenzene
Concen: 0.426 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:284 Resp: 8349
Ion Ratio Lower Upper
284 100
142 39.2 30.5 45.7
249 30.0 23.5 35.3









#23
 Atrazine
 Concen: 0.370 ng
 RT: 16.652 min Scan# 1645
 Delta R.T. 0.000 min
 Lab File: BN038118.D
 Acq: 28 Oct 2025 17:51




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102825


Tgt Ion:	200	Resp:	4672
Ion Ratio	Lower	Upper	
200	100		
173	27.7	21.4	32.2
215	45.3	36.7	55.1



#24
 Pentachlorophenol
 Concen: 0.351 ng
 RT: 16.838 min Scan# 1660
 Delta R.T. 0.000 min
 Lab File: BN038118.D
 Acq: 28 Oct 2025 17:51

Tgt Ion:	266	Resp:	2963
Ion Ratio	Lower	Upper	
266	100		
264	62.7	49.3	73.9
268	66.0	50.6	75.8









#25
 Phenanthrene
 Concen: 0.408 ng
 RT: 17.223 min Scan# 1691
 Delta R.T. 0.000 min
 Lab File: BN038118.D
 Acq: 28 Oct 2025 17:51




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102825


Tgt Ion:	178	Resp:	33992
Ion Ratio	Lower	Upper	
178	100		
176	19.2	15.5	23.3
179	15.7	12.2	18.4



#26
 Anthracene
 Concen: 0.406 ng
 RT: 17.322 min Scan# 1699
 Delta R.T. 0.000 min
 Lab File: BN038118.D
 Acq: 28 Oct 2025 17:51

Tgt Ion:	178	Resp:	29654
Ion Ratio	Lower	Upper	
178	100		
176	18.8	14.9	22.3
179	15.3	11.8	17.8









#27
Fluoranthene-d10
Concen: 0.389 ng
RT: 19.220 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:212 Resp: 25766
Ion Ratio Lower Upper
212 100
106 16.1 12.6 19.0
104 9.0 7.1 10.7



#28
Fluoranthene
Concen: 0.380 ng
RT: 19.253 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:202 Resp: 35401
Ion Ratio Lower Upper
202 100
101 11.5 9.4 14.2
203 17.1 13.4 20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:240 Resp: 18729
Ion Ratio Lower Upper
240 100
120 14.0 10.7 16.1
236 28.8 23.1 34.7



#30
Pyrene
Concen: 0.422 ng
RT: 19.615 min Scan# 2032
Delta R.T. 0.005 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:202 Resp: 35647
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.6 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.428 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. 0.000 min
 Lab File: BN038118.D
 Acq: 28 Oct 2025 17:51




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102825


Tgt Ion:244 Resp: 17430
 Ion Ratio Lower Upper
 244 100
 212 9.6 7.8 11.6
 122 19.1 15.2 22.8



#32
 Benzo(a)anthracene
 Concen: 0.402 ng
 RT: 21.358 min Scan# 2267
 Delta R.T. 0.000 min
 Lab File: BN038118.D
 Acq: 28 Oct 2025 17:51

Tgt Ion:228 Resp: 26833
 Ion Ratio Lower Upper
 228 100
 226 28.0 22.2 33.4
 229 19.7 15.8 23.8









#33
Chrysene
Concen: 0.428 ng
RT: 21.411 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:	228	Resp:	31031
Ion Ratio	Lower	Upper	
228	100		
226	30.4	24.3	36.5
229	19.6	15.7	23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.358 ng
RT: 21.295 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:	149	Resp:	12587
Ion Ratio	Lower	Upper	
149	100		
167	27.0	21.1	31.7
279	3.4	2.9	4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.683 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:264 Resp: 15660
Ion Ratio Lower Upper
264 100
260 27.0 21.8 32.8
265 93.6 68.6 102.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.406 ng
RT: 26.037 min Scan# 3626
Delta R.T. 0.003 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:276 Resp: 26115
Ion Ratio Lower Upper
276 100
138 27.9 21.4 32.0
277 24.8 19.2 28.8





#37

Benzo(b)fluoranthene

Concen: 0.405 ng

RT: 22.984 min Scan# 2582

Delta R.T. 0.003 min

Lab File: BN038118.D


Acq: 28 Oct 2025 17:51

Instrument :

BNA_N

ClientSampleId :

ICVBN102825






Tgt Ion:252 Resp: 27124

Ion Ratio Lower Upper

252 100

253 25.3 20.9 31.3

125 19.9 14.1 21.1



#38

Benzo(k)fluoranthene

Concen: 0.439 ng

RT: 23.031 min Scan# 2598

Delta R.T. 0.003 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51




Tgt Ion:252 Resp: 29583


Ion Ratio Lower Upper

252 100

253 25.4 21.2 31.8

125 18.0 14.6 22.0





#39

Benzo(a)pyrene

Concen: 0.437 ng

RT: 23.581 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51





Instrument :

BNA_N

ClientSampleId :

ICVBN102825

Tgt Ion:	252	Resp:	23400
Ion Ratio	Lower	Upper	
252	100		
253	27.8	23.1	34.7
125	28.0	18.6	28.0



#40

Dibenzo(a,h)anthracene

Concen: 0.403 ng




RT: 26.051 min Scan# 3631


Delta R.T. 0.003 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51

Tgt Ion:	278	Resp:	19931
Ion Ratio	Lower	Upper	
278	100		
139	20.2	17.6	26.4
279	27.0	22.6	33.8





#41

Benzo(g,h,i)perylene

Concen: 0.392 ng

RT: 26.756 min Scan# 3872

Delta R.T. 0.006 min

Lab File: BN038118.D


Acq: 28 Oct 2025 17:51

Instrument :

BNA_N

ClientSampleId :

ICVBN102825





Tgt Ion:276 Resp: 22180

Ion Ratio Lower Upper

276 100

277 25.7 19.7 29.5

138 25.7 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038118.D
Acq On : 28 Oct 2025 17:51
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN102825

Quant Time: Oct 29 17:07:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.413	0.455	-10.2	90	0.00
3	n-Nitrosodimethylamine	0.532	0.529	0.6	82	0.00
4 S	2-Fluorophenol	0.942	0.866	8.1	76	0.00
5 S	Phenol-d6	1.148	1.006	12.4	74	0.00
6	bis(2-Chloroethyl)ether	1.128	1.202	-6.6	89	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	84	0.01
8 S	Nitrobenzene-d5	0.323	0.285	11.8	76	0.00
9	Naphthalene	1.102	1.121	-1.7	87	0.00
10	Hexachlorobutadiene	0.186	0.202	-8.6	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.571	0.585	-2.5	89	0.00
12	2-Methylnaphthalene	0.735	0.704	4.2	84	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.01
14 S	2,4,6-Tribromophenol	0.165	0.140	15.2	81	0.00
15 S	2-Fluorobiphenyl	1.602	1.647	-2.8	90	0.01
16	Acenaphthylene	1.811	2.060	-13.7	103	0.00
17	Acenaphthene	1.267	1.262	0.4	89	0.01
18	Fluorene	1.661	1.719	-3.5	92	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.043	21.8	86	0.00
21	4-Bromophenyl-phenylether	0.262	0.262	0.0	93	0.00
22	Hexachlorobenzene	0.298	0.318	-6.7	95	0.00
23	Atrazine	0.192	0.178	7.3	90	0.00
24	Pentachlorophenol	0.128	0.113	11.7	93	0.00
25	Phenanthrene	1.270	1.294	-1.9	94	0.00
26	Anthracene	1.111	1.129	-1.6	96	0.00
27 SURR	Fluoranthene-d10	1.009	0.981	2.8	91	0.00
28	Fluoranthene	1.419	1.347	5.1	89	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
30	Pyrene	1.806	1.903	-5.4	90	0.00
31 S	Terphenyl-d14	0.869	0.931	-7.1	93	0.00
32	Benzo(a)anthracene	1.427	1.433	-0.4	92	0.00
33	Chrysene	1.549	1.657	-7.0	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.752	0.672	10.6	81	0.00
35 I	Perylene-d12	1.000	1.000	0.0	87	0.00
36	Indeno(1,2,3-cd)pyrene	1.643	1.668	-1.5	91	0.00
37	Benzo(b)fluoranthene	1.710	1.732	-1.3	93	0.00
38	Benzo(k)fluoranthene	1.721	1.889	-9.8	102	0.00
39 C	Benzo(a)pyrene	1.369	1.494	-9.1	99	0.00
40	Dibenzo(a,h)anthracene	1.265	1.273	-0.6	93	0.00
41	Benzo(g,h,i)perylene	1.444	1.416	1.9	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038118.D
 Acq On : 28 Oct 2025 17:51
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102825

Quant Time: Oct 29 17:07:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.00
2	1,4-Dioxane	0.400	0.440	-10.0	90	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	82	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	76	0.00
5 S	Phenol-d6	0.400	0.351	12.3	74	0.00
6	bis(2-Chloroethyl)ether	0.400	0.427	-6.7	89	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	84	0.01
8 S	Nitrobenzene-d5	0.400	0.353	11.8	76	0.00
9	Naphthalene	0.400	0.407	-1.7	87	0.00
10	Hexachlorobutadiene	0.400	0.435	-8.7	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	89	0.00
12	2-Methylnaphthalene	0.400	0.384	4.0	84	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	0.01
14 S	2,4,6-Tribromophenol	0.400	0.340	15.0	81	0.00
15 S	2-Fluorobiphenyl	0.400	0.411	-2.7	90	0.01
16	Acenaphthylene	0.400	0.455	-13.7	103	0.00
17	Acenaphthene	0.400	0.398	0.5	89	0.01
18	Fluorene	0.400	0.414	-3.5	92	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.411	-2.7	86	0.00
21	4-Bromophenyl-phenylether	0.400	0.400	0.0	93	0.00
22	Hexachlorobenzene	0.400	0.426	-6.5	95	0.00
23	Atrazine	0.400	0.370	7.5	90	0.00
24	Pentachlorophenol	0.400	0.351	12.3	93	0.00
25	Phenanthrene	0.400	0.408	-2.0	94	0.00
26	Anthracene	0.400	0.406	-1.5	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.389	2.8	91	0.00
28	Fluoranthene	0.400	0.380	5.0	89	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	87	0.00
30	Pyrene	0.400	0.422	-5.5	90	0.00
31 S	Terphenyl-d14	0.400	0.428	-7.0	93	0.00
32	Benzo(a)anthracene	0.400	0.402	-0.5	92	0.00
33	Chrysene	0.400	0.428	-7.0	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.358	10.5	81	0.00
35 I	Perylene-d12	0.400	0.400	0.0	87	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.406	-1.5	91	0.00
37	Benzo(b)fluoranthene	0.400	0.405	-1.3	93	0.00
38	Benzo(k)fluoranthene	0.400	0.439	-9.7	102	0.00
39 C	Benzo(a)pyrene	0.400	0.437	-9.2	99	0.00
40	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	93	0.00
41	Benzo(g,h,i)perylene	0.400	0.392	2.0	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3526</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>11/04/2025 10:40</u>
Lab File ID:	<u>BN038130.D</u>	Init. Calib. Date(s):	<u>10/28/2025 10/28/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4</u>	Init. Calib. Time(s):	<u>13:00 16:38</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.571	0.541		-5.3	20.0
Fluoranthene-d10	1.008	0.889		-11.8	20.0
2-Fluorophenol	0.943	0.856		-9.2	20.0
Phenol-d6	1.148	1.025		-10.7	20.0
Nitrobenzene-d5	0.323	0.277		-14.2	20.0
2-Fluorobiphenyl	1.602	1.514		-5.5	20.0
2,4,6-Tribromophenol	0.165	0.134		-18.8	20.0
Terphenyl-d14	0.869	0.916		5.4	20.0
1,4-Dioxane	0.416	0.433		4.1	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038130.D
 Acq On : 04 Nov 2025 10:40
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Nov 04 11:42:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8599	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	24973	0.400	ng	# 0.00
13) Acenaphthene-d10	14.441	164	13947	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	27056	0.400	ng	0.00
29) Chrysene-d12	21.375	240	17059	0.400	ng	0.00
35) Perylene-d12	23.686	264	13607	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	7357	0.363	ng	0.00
5) Phenol-d6	6.937	99	8817	0.357	ng	0.00
8) Nitrobenzene-d5	8.928	82	6906	0.343	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	13509	0.379	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1874	0.326	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	21117	0.378	ng	0.01
27) Fluoranthene-d10	19.220	212	24042	0.352	ng	0.00
31) Terphenyl-d14	19.819	244	15620	0.421	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	3722	0.419	ng	99
3) n-Nitrosodimethylamine	3.564	42	4544	0.397	ng	# 93
6) bis(2-Chloroethyl)ether	7.197	93	9558	0.394	ng	99
9) Naphthalene	10.626	128	27095	0.394	ng	100
10) Hexachlorobutadiene	10.925	225	4747	0.409	ng	# 99
12) 2-Methylnaphthalene	12.248	142	17751	0.387	ng	98
16) Acenaphthylene	14.152	152	24464	0.387	ng	99
17) Acenaphthene	14.494	154	16888	0.382	ng	96
18) Fluorene	15.489	166	22996	0.397	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	1019	0.381	ng	# 84
21) 4-Bromophenyl-phenylether	16.379	248	6844	0.386	ng	98
22) Hexachlorobenzene	16.503	284	8492	0.421	ng	100
23) Atrazine	16.652	200	4361	0.336	ng	98
24) Pentachlorophenol	16.838	266	2516	0.290	ng	98
25) Phenanthrene	17.223	178	33589	0.391	ng	100
26) Anthracene	17.322	178	27861	0.371	ng	99
28) Fluoranthene	19.252	202	35123	0.366	ng	100
30) Pyrene	19.615	202	34116	0.443	ng	100
32) Benzo(a)anthracene	21.357	228	22233	0.365	ng	100
33) Chrysene	21.411	228	26943	0.408	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	12992	0.405	ng	100
36) Indeno(1,2,3-cd)pyrene	26.036	276	23804	0.426	ng	99
37) Benzo(b)fluoranthene	22.987	252	23288	0.400	ng	98
38) Benzo(k)fluoranthene	23.031	252	23877	0.408	ng	99
39) Benzo(a)pyrene	23.583	252	18445	0.396	ng	98
40) Dibenzo(a,h)anthracene	26.054	278	18347	0.427	ng	99
41) Benzo(g,h,i)perylene	26.753	276	21656	0.441	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038130.D
Acq On : 04 Nov 2025 10:40
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Quant Time: Nov 04 11:42:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:152 Resp: 8599
Ion Ratio Lower Upper
152 100
150 153.9 122.3 183.5
115 59.4 47.4 71.0



#2
1,4-Dioxane
Concen: 0.419 ng
RT: 3.254 min Scan# 24
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion: 88 Resp: 3722
Ion Ratio Lower Upper
88 100
43 31.1 24.3 36.5
58 76.5 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 0.397 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 4544
Ion Ratio Lower Upper
42 100
74 112.7 96.8 145.2
44 9.0 5.3 7.9#



#4
2-Fluorophenol
Concen: 0.363 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion: 112 Resp: 7357
Ion Ratio Lower Upper
112 100
64 55.3 45.4 68.0
63 29.7 23.2 34.8









#5
Phenol-d6
Concen: 0.357 ng
RT: 6.937 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	99	Resp:	8817
Ion	Ratio	Lower	Upper
99	100		
42	20.3	16.6	24.8
71	31.7	25.4	38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.394 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:	93	Resp:	9558
Ion	Ratio	Lower	Upper
93	100		
63	73.2	59.3	88.9
95	31.1	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 937
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	24973
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.0	13.4
54	9.9	5.2	7.8#
68	7.4	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 0.343 ng
RT: 8.928 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:	82	Resp:	6906
Ion Ratio	Lower	Upper	
82	100		
128	43.6	34.1	51.1
54	54.8	43.5	65.3









#9
Naphthalene
Concen: 0.394 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	27095
Ion Ratio	Lower	Upper	
128	100		
129	11.1	9.1	13.7
127	13.2	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.409 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:	225	Resp:	4747
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.8	50.3	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.379 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 13509
Ion Ratio Lower Upper
152 100
151 20.7 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.387 ng
RT: 12.248 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:142 Resp: 17751
Ion Ratio Lower Upper
142 100
141 89.9 70.3 105.5
115 36.5 27.8 41.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.441 min Scan# 1453

Delta R.T. 0.010 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:164 Resp: 13947

Ion Ratio Lower Upper

164 100

162 104.1 83.0 124.4

160 51.5 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.326 ng

RT: 15.932 min Scan# 1587

Delta R.T. -0.000 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40




Tgt Ion:330 Resp: 1874


Ion Ratio Lower Upper

330 100

332 95.6 76.6 114.8

141 42.3 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.378 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.010 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	21117
Ion Ratio	Lower	Upper	
172	100		
171	34.8	27.8	41.6
170	22.9	18.1	27.1



#16
Acenaphthylene
Concen: 0.387 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:	152	Resp:	24464
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.8	23.8
153	13.1	10.2	15.2





#17

Acenaphthene

Concen: 0.382 ng

RT: 14.494 min Scan# 1458

Delta R.T. -0.000 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:154 Resp: 16888

Ion Ratio Lower Upper

154 100

153 115.7 90.0 135.0

152 62.6 47.3 70.9



#18

Fluorene

Concen: 0.397 ng

RT: 15.489 min Scan# 1551

Delta R.T. -0.000 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40




Tgt Ion:166 Resp: 22996


Ion Ratio Lower Upper

166 100

165 98.5 79.0 118.4

167 13.0 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1688

Delta R.T. -0.000 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:188 Resp: 27056

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.6 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.381 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40




Tgt Ion:198 Resp: 1019


Ion Ratio Lower Upper

198 100

51 135.4 88.9 133.3

105 63.4 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.386 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	248	Resp:	6844
Ion Ratio	Lower	Upper	
248	100		
250	95.3	76.2	114.2
141	68.0	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.421 ng
RT: 16.503 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:	284	Resp:	8492
Ion Ratio	Lower	Upper	
284	100		
142	38.3	30.5	45.7
249	29.5	23.5	35.3









#23
 Atrazine
 Concen: 0.336 ng
 RT: 16.652 min Scan# 1645
 Delta R.T. -0.000 min
 Lab File: BN038130.D
 Acq: 04 Nov 2025 10:40




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	200	Resp:	4361
Ion Ratio	Lower	Upper	
200	100		
173	27.2	21.4	32.2
215	47.5	36.7	55.1



#24
 Pentachlorophenol
 Concen: 0.290 ng
 RT: 16.838 min Scan# 1660
 Delta R.T. -0.000 min
 Lab File: BN038130.D
 Acq: 04 Nov 2025 10:40

Tgt Ion:	266	Resp:	2516
Ion Ratio	Lower	Upper	
266	100		
264	62.4	49.3	73.9
268	65.6	50.6	75.8





#25

Phenanthrene

Concen: 0.391 ng

RT: 17.223 min Scan# 1691

Delta R.T. -0.000 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:178 Resp: 33589

Ion Ratio Lower Upper

178 100

176 19.3 15.5 23.3

179 15.5 12.2 18.4



#26

Anthracene

Concen: 0.371 ng

RT: 17.322 min Scan# 1699

Delta R.T. -0.000 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40




Tgt Ion:178 Resp: 27861


Ion Ratio Lower Upper

178 100

176 18.5 14.9 22.3

179 15.3 11.8 17.8









#27
Fluoranthene-d10
Concen: 0.352 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 24042
Ion Ratio Lower Upper
212 100
106 16.2 12.6 19.0
104 9.1 7.1 10.7



#28
Fluoranthene
Concen: 0.366 ng
RT: 19.252 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:202 Resp: 35123
Ion Ratio Lower Upper
202 100
101 11.9 9.4 14.2
203 16.9 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:240 Resp: 17059

Ion Ratio Lower Upper

240 100

120 15.0 10.7 16.1

236 29.6 23.1 34.7



#30

Pyrene

Concen: 0.443 ng

RT: 19.615 min Scan# 2032

Delta R.T. 0.004 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40




Tgt Ion:202 Resp: 34116


Ion Ratio Lower Upper

202 100

200 21.1 16.7 25.1

203 17.8 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.421 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. -0.000 min
 Lab File: BN038130.D
 Acq: 04 Nov 2025 10:40




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:244 Resp: 15620
 Ion Ratio Lower Upper
 244 100
 212 9.5 7.8 11.6
 122 19.3 15.2 22.8



#32
 Benzo(a)anthracene
 Concen: 0.365 ng
 RT: 21.357 min Scan# 2267
 Delta R.T. -0.000 min
 Lab File: BN038130.D
 Acq: 04 Nov 2025 10:40

Tgt Ion:228 Resp: 22233
 Ion Ratio Lower Upper
 228 100
 226 28.0 22.2 33.4
 229 19.8 15.8 23.8









#33
 Chrysene
 Concen: 0.408 ng
 RT: 21.411 min Scan# 2273
 Delta R.T. -0.000 min
 Lab File: BN038130.D
 Acq: 04 Nov 2025 10:40

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:228 Resp: 26943




Ion	Ratio	Lower	Upper
228	100		
226	31.0	24.3	36.5
229	19.6	15.7	23.5




#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.405 ng
 RT: 21.295 min Scan# 2260
 Delta R.T. -0.000 min
 Lab File: BN038130.D
 Acq: 04 Nov 2025 10:40

Tgt Ion:149 Resp: 12992

Ion	Ratio	Lower	Upper
149	100		
167	26.4	21.1	31.7
279	3.2	2.9	4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.686 min Scan# 2821
Delta R.T. 0.006 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:264 Resp: 13607
Ion Ratio Lower Upper
264 100
260 27.1 21.8 32.8
265 93.2 68.6 102.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.426 ng
RT: 26.036 min Scan# 3626
Delta R.T. 0.003 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:276 Resp: 23804
Ion Ratio Lower Upper
276 100
138 27.4 21.4 32.0
277 23.7 19.2 28.8









#37
Benzo(b)fluoranthene
Concen: 0.400 ng
RT: 22.987 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 23288
Ion Ratio Lower Upper
252 100
253 25.4 20.9 31.3
125 18.8 14.1 21.1



#38
Benzo(k)fluoranthene
Concen: 0.408 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.003 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:252 Resp: 23877
Ion Ratio Lower Upper
252 100
253 26.0 21.2 31.8
125 19.0 14.6 22.0





#39

Benzo(a)pyrene

Concen: 0.396 ng

RT: 23.583 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 18445

Ion Ratio Lower Upper

252 100

253 28.6 23.1 34.7

125 24.7 18.6 28.0



#40

Dibenzo(a,h)anthracene

Concen: 0.427 ng

RT: 26.054 min Scan# 3632

Delta R.T. 0.006 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40




Tgt Ion:278 Resp: 18347


Ion Ratio Lower Upper

278 100

139 21.1 17.6 26.4


279 27.8 22.6 33.8







#41
Benzo(g,h,i)perylene
Concen: 0.441 ng
RT: 26.753 min Scan# 38
Delta R.T. 0.003 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	276	Resp:	21656
Ion	Ratio	Lower	Upper
276	100		
277	24.8	19.7	29.5
138	25.4	20.2	30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038130.D
 Acq On : 04 Nov 2025 10:40
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 04 11:42:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.413	0.433	-4.8	96	0.00
3	n-Nitrosodimethylamine	0.532	0.528	0.8	92	0.00
4 S	2-Fluorophenol	0.942	0.856	9.1	84	0.00
5 S	Phenol-d6	1.148	1.025	10.7	85	0.00
6	bis(2-Chloroethyl)ether	1.128	1.112	1.4	92	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
8 S	Nitrobenzene-d5	0.323	0.277	14.2	80	0.00
9	Naphthalene	1.102	1.085	1.5	91	0.00
10	Hexachlorobutadiene	0.186	0.190	-2.2	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.571	0.541	5.3	89	0.00
12	2-Methylnaphthalene	0.735	0.711	3.3	91	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.01
14 S	2,4,6-Tribromophenol	0.165	0.134	18.8	81	0.00
15 S	2-Fluorobiphenyl	1.602	1.514	5.5	86	0.01
16	Acenaphthylene	1.811	1.754	3.1	92	0.00
17	Acenaphthene	1.267	1.211	4.4	90	0.00
18	Fluorene	1.661	1.649	0.7	93	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.038	30.9#	77	0.00
21	4-Bromophenyl-phenylether	0.262	0.253	3.4	93	0.00
22	Hexachlorobenzene	0.298	0.314	-5.4	97	0.00
23	Atrazine	0.192	0.161	16.1	84	0.00
24	Pentachlorophenol	0.128	0.093	27.3#	79	0.00
25	Phenanthrene	1.270	1.241	2.3	93	0.00
26	Anthracene	1.111	1.030	7.3	90	0.00
27 SURR	Fluoranthene-d10	1.009	0.889	11.9	85	0.00
28	Fluoranthene	1.419	1.298	8.5	89	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
30	Pyrene	1.806	2.000	-10.7	86	0.00
31 S	Terphenyl-d14	0.869	0.916	-5.4	83	0.00
32	Benzo(a)anthracene	1.427	1.303	8.7	76	0.00
33	Chrysene	1.549	1.579	-1.9	81	0.00
34	Bis(2-ethylhexyl)phthalate	0.752	0.762	-1.3	84	0.00
35 I	Perylene-d12	1.000	1.000	0.0	76	0.00
36	Indeno(1,2,3-cd)pyrene	1.643	1.749	-6.5	83	0.00
37	Benzo(b)fluoranthene	1.710	1.711	-0.1	80	0.00
38	Benzo(k)fluoranthene	1.721	1.755	-2.0	82	0.00
39 C	Benzo(a)pyrene	1.369	1.356	0.9	78	0.00
40	Dibenzo(a,h)anthracene	1.265	1.348	-6.6	85	0.00
41	Benzo(g,h,i)perylene	1.444	1.592	-10.2	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038130.D
Acq On : 04 Nov 2025 10:40
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 04 11:42:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	94	0.00
2	1,4-Dioxane	0.400	0.419	-4.7	96	0.00
3	n-Nitrosodimethylamine	0.400	0.397	0.8	92	0.00
4 S	2-Fluorophenol	0.400	0.363	9.3	84	0.00
5 S	Phenol-d6	0.400	0.357	10.8	85	0.00
6	bis(2-Chloroethyl)ether	0.400	0.394	1.5	92	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	90	0.00
8 S	Nitrobenzene-d5	0.400	0.343	14.2	80	0.00
9	Naphthalene	0.400	0.394	1.5	91	0.00
10	Hexachlorobutadiene	0.400	0.409	-2.2	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.379	5.3	89	0.00
12	2-Methylnaphthalene	0.400	0.387	3.3	91	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	90	0.01
14 S	2,4,6-Tribromophenol	0.400	0.326	18.5	81	0.00
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	86	0.01
16	Acenaphthylene	0.400	0.387	3.3	92	0.00
17	Acenaphthene	0.400	0.382	4.5	90	0.00
18	Fluorene	0.400	0.397	0.8	93	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	92	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.381	4.8	77	0.00
21	4-Bromophenyl-phenylether	0.400	0.386	3.5	93	0.00
22	Hexachlorobenzene	0.400	0.421	-5.2	97	0.00
23	Atrazine	0.400	0.336	16.0	84	0.00
24	Pentachlorophenol	0.400	0.290	27.5#	79	0.00
25	Phenanthrene	0.400	0.391	2.3	93	0.00
26	Anthracene	0.400	0.371	7.3	90	0.00
27 SURR	Fluoranthene-d10	0.400	0.352	12.0	85	0.00
28	Fluoranthene	0.400	0.366	8.5	89	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	79	0.00
30	Pyrene	0.400	0.443	-10.7	86	0.00
31 S	Terphenyl-d14	0.400	0.421	-5.2	83	0.00
32	Benzo(a)anthracene	0.400	0.365	8.8	76	0.00
33	Chrysene	0.400	0.408	-2.0	81	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.405	-1.3	84	0.00
35 I	Perylene-d12	0.400	0.400	0.0	76	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.426	-6.5	83	0.00
37	Benzo(b)fluoranthene	0.400	0.400	0.0	80	0.00
38	Benzo(k)fluoranthene	0.400	0.408	-2.0	82	0.00
39 C	Benzo(a)pyrene	0.400	0.396	1.0	78	0.00
40	Dibenzo(a,h)anthracene	0.400	0.427	-6.7	85	0.00
41	Benzo(g,h,i)perylene	0.400	0.441	-10.2	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG No.: Q3526
Instrument ID: BNA_N Calibration Date/Time: 11/04/2025 20:12
Lab File ID: BN038145.D Init. Calib. Date(s): 10/28/2025 10/28/2025
EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 13:00 16:38
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.571	0.538		-5.8	50.0
Fluoranthene-d10	1.008	0.826		-18.1	50.0
2-Fluorophenol	0.943	0.881		-6.6	50.0
Phenol-d6	1.148	1.042		-9.2	50.0
Nitrobenzene-d5	0.323	0.273		-15.5	50.0
2-Fluorobiphenyl	1.602	1.511		-5.7	50.0
2,4,6-Tribromophenol	0.165	0.126		-23.6	50.0
Terphenyl-d14	0.869	0.962		10.7	50.0
1,4-Dioxane	0.416	0.430		3.4	50.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038145.D
 Acq On : 04 Nov 2025 20:12
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Nov 04 21:16:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8830	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	26327	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	14572	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	25889	0.400	ng	0.00
29) Chrysene-d12	21.375	240	14271	0.400	ng	0.00
35) Perylene-d12	23.683	264	11979	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	7775	0.374	ng	0.00
5) Phenol-d6	6.937	99	9204	0.363	ng	0.00
8) Nitrobenzene-d5	8.929	82	7183	0.338	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	14165	0.377	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1835	0.306	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	22024	0.377	ng	0.00
27) Fluoranthene-d10	19.220	212	21378	0.327	ng	0.00
31) Terphenyl-d14	19.819	244	13728	0.443	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	3796	0.416	ng	99
3) n-Nitrosodimethylamine	3.564	42	4631	0.394	ng	# 94
6) bis(2-Chloroethyl)ether	7.197	93	9767	0.392	ng	100
9) Naphthalene	10.626	128	28555	0.394	ng	100
10) Hexachlorobutadiene	10.925	225	5138	0.420	ng	# 99
12) 2-Methylnaphthalene	12.248	142	18776	0.388	ng	99
16) Acenaphthylene	14.152	152	25035	0.379	ng	100
17) Acenaphthene	14.494	154	17860	0.387	ng	98
18) Fluorene	15.489	166	23115	0.382	ng	99
20) 4,6-Dinitro-2-methylph...	15.560	198	971	0.380	ng	94
21) 4-Bromophenyl-phenylether	16.379	248	6744	0.397	ng	96
22) Hexachlorobenzene	16.491	284	8174	0.423	ng	99
23) Atrazine	16.652	200	4191	0.337	ng	98
24) Pentachlorophenol	16.838	266	2211	0.266	ng	99
25) Phenanthrene	17.223	178	31983	0.389	ng	100
26) Anthracene	17.322	178	26920	0.374	ng	99
28) Fluoranthene	19.248	202	30682	0.334	ng	100
30) Pyrene	19.610	202	29749	0.462	ng	100
32) Benzo(a)anthracene	21.357	228	19156	0.376	ng	100
33) Chrysene	21.411	228	22690	0.411	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	10298	0.384	ng	99
36) Indeno(1,2,3-cd)pyrene	26.034	276	21759	0.442	ng	98
37) Benzo(b)fluoranthene	22.984	252	20105	0.393	ng	97
38) Benzo(k)fluoranthene	23.028	252	20945	0.406	ng	98
39) Benzo(a)pyrene	23.581	252	16258	0.397	ng	96
40) Dibenzo(a,h)anthracene	26.051	278	17101	0.452	ng	99
41) Benzo(g,h,i)perylene	26.747	276	19764	0.457	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038145.D
Acq On : 04 Nov 2025 20:12
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Quant Time: Nov 04 21:16:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.775 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN038145.D
 Acq: 04 Nov 2025 20:12


Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC





Tgt Ion:152 Resp: 8830
 Ion Ratio Lower Upper
 152 100
 150 153.2 122.3 183.5
 115 60.0 47.4 71.0




#2
 1,4-Dioxane
 Concen: 0.416 ng
 RT: 3.254 min Scan# 24
 Delta R.T. -0.000 min
 Lab File: BN038145.D
 Acq: 04 Nov 2025 20:12



Tgt Ion: 88 Resp: 3796
 Ion Ratio Lower Upper
 88 100
 43 30.9 24.3 36.5
 58 76.6 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 0.394 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 4631
Ion Ratio Lower Upper
42 100
74 114.9 96.8 145.2
44 8.9 5.3 7.9#



#4
2-Fluorophenol
Concen: 0.374 ng
RT: 5.341 min Scan# 313
Delta R.T. -0.007 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion: 112 Resp: 7775
Ion Ratio Lower Upper
112 100
64 56.9 45.4 68.0
63 29.9 23.2 34.8









#5
Phenol-d6
Concen: 0.363 ng
RT: 6.937 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	99	Resp:	9204
Ion	Ratio	Lower	Upper
99	100		
42	21.1	16.6	24.8
71	32.3	25.4	38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.392 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:	93	Resp:	9767
Ion	Ratio	Lower	Upper
93	100		
63	73.8	59.3	88.9
95	31.3	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 937
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	26327
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.0	13.4
54	9.6	5.2	7.8#
68	7.1	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 0.338 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:	82	Resp:	7183
Ion Ratio	Lower	Upper	
82	100		
128	45.7	34.1	51.1
54	52.7	43.5	65.3









#9
Naphthalene
Concen: 0.394 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	28555
Ion Ratio	Lower	Upper	
128	100		
129	11.4	9.1	13.7
127	13.5	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.420 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:	225	Resp:	5138
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.4	50.3	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.377 ng
RT: 12.172 min Scan# 1155
Delta R.T. -0.005 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 14165
Ion Ratio Lower Upper
152 100
151 21.3 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.388 ng
RT: 12.248 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:142 Resp: 18776
Ion Ratio Lower Upper
142 100
141 88.8 70.3 105.5
115 35.2 27.8 41.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.430 min Scan# 14

Delta R.T. -0.000 min

Lab File: BN038145.D

Acq: 04 Nov 2025 20:12

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:164 Resp: 14572

Ion Ratio Lower Upper

164 100

162 106.7 83.0 124.4

160 55.1 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.306 ng

RT: 15.932 min Scan# 1587

Delta R.T. -0.000 min

Lab File: BN038145.D

Acq: 04 Nov 2025 20:12




Tgt Ion:330 Resp: 1835


Ion Ratio Lower Upper

330 100

332 99.6 76.6 114.8

141 43.1 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.377 ng
RT: 13.051 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	22024
Ion Ratio	Lower	Upper	
172	100		
171	35.5	27.8	41.6
170	24.5	18.1	27.1



#16
Acenaphthylene
Concen: 0.379 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:	152	Resp:	25035
Ion Ratio	Lower	Upper	
152	100		
151	19.8	15.8	23.8
153	13.0	10.2	15.2









#17
Acenaphthene
Concen: 0.387 ng
RT: 14.494 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 17860
Ion Ratio Lower Upper
154 100
153 114.0 90.0 135.0
152 61.4 47.3 70.9



#18
Fluorene
Concen: 0.382 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:166 Resp: 23115
Ion Ratio Lower Upper
166 100
165 99.8 79.0 118.4
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 25889
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.9 8.6 13.0



#20
4,6-Dinitro-2-methylphenol
Concen: 0.380 ng
RT: 15.560 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:198 Resp: 971
Ion Ratio Lower Upper
198 100
51 118.1 88.9 133.3
105 58.5 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.397 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	248	Resp:	6744
Ion Ratio	Lower	Upper	
248	100		
250	96.4	76.2	114.2
141	64.8	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.423 ng
RT: 16.491 min Scan# 1632
Delta R.T. -0.013 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:	284	Resp:	8174
Ion Ratio	Lower	Upper	
284	100		
142	39.1	30.5	45.7
249	30.0	23.5	35.3









#23
Atrazine
Concen: 0.337 ng
RT: 16.652 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 4191
Ion Ratio Lower Upper
200 100
173 24.5 21.4 32.2
215 46.5 36.7 55.1



#24
Pentachlorophenol
Concen: 0.266 ng
RT: 16.838 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:266 Resp: 2211
Ion Ratio Lower Upper
266 100
264 61.3 49.3 73.9
268 65.2 50.6 75.8









#25
 Phenanthrene
 Concen: 0.389 ng
 RT: 17.223 min Scan# 1691
 Delta R.T. -0.000 min
 Lab File: BN038145.D
 Acq: 04 Nov 2025 20:12

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:178 Resp: 31983




Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.5	23.3
179	15.3	12.2	18.4




#26
 Anthracene
 Concen: 0.374 ng
 RT: 17.322 min Scan# 1699
 Delta R.T. -0.000 min
 Lab File: BN038145.D
 Acq: 04 Nov 2025 20:12

Tgt Ion:178 Resp: 26920

Ion	Ratio	Lower	Upper
178	100		
176	18.7	14.9	22.3
179	15.3	11.8	17.8









#27
Fluoranthene-d10
Concen: 0.327 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 21378
Ion Ratio Lower Upper
212 100
106 16.1 12.6 19.0
104 9.5 7.1 10.7



#28
Fluoranthene
Concen: 0.334 ng
RT: 19.248 min Scan# 1953
Delta R.T. -0.005 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12


Tgt Ion:202 Resp: 30682
Ion Ratio Lower Upper
202 100
101 12.1 9.4 14.2
203 16.9 13.4 20.2








#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC





Tgt Ion:240 Resp: 14271
Ion Ratio Lower Upper
240 100
120 13.8 10.7 16.1
236 29.3 23.1 34.7




#30
Pyrene
Concen: 0.462 ng
RT: 19.610 min Scan# 2031
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12



Tgt Ion:202 Resp: 29749
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.6 14.2 21.2





#31

Terphenyl-d14

Concen: 0.443 ng

RT: 19.819 min Scan# 2076

Delta R.T. -0.000 min

Lab File: BN038145.D

Acq: 04 Nov 2025 20:12

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:244 Resp: 13728

Ion Ratio Lower Upper

244 100

212 9.7 7.8 11.6

122 18.8 15.2 22.8



#32

Benzo(a)anthracene

Concen: 0.376 ng

RT: 21.357 min Scan# 2267

Delta R.T. -0.000 min

Lab File: BN038145.D

Acq: 04 Nov 2025 20:12




Tgt Ion:228 Resp: 19156


Ion Ratio Lower Upper

228 100

226 27.5 22.2 33.4

229 19.7 15.8 23.8









#33
 Chrysene
 Concen: 0.411 ng
 RT: 21.411 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN038145.D
 Acq: 04 Nov 2025 20:12




Instrument : BNA_N
 ClientSampleId : SSTDCCC0.4EC


Tgt Ion:	228	Resp:	22690
Ion Ratio	Lower	Upper	
228	100		
226	29.8	24.3	36.5
229	19.8	15.7	23.5



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.384 ng
 RT: 21.295 min Scan# 2260
 Delta R.T. -0.000 min
 Lab File: BN038145.D
 Acq: 04 Nov 2025 20:12


Tgt Ion:	149	Resp:	10298
Ion Ratio	Lower	Upper	
149	100		
167	26.6	21.1	31.7
279	3.3	2.9	4.3








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.683 min Scan# 2821
Delta R.T. 0.003 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 264 Resp: 11979




Ion	Ratio	Lower	Upper
264	100		
260	27.3	21.8	32.8
265	102.8	68.6	102.8




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.442 ng
RT: 26.034 min Scan# 3625
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion: 276 Resp: 21759


Ion	Ratio	Lower	Upper
276	100		
138	27.9	21.4	32.0
277	24.5	19.2	28.8








#37
Benzo(b)fluoranthene
Concen: 0.393 ng
RT: 22.984 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:252 Resp: 20105
Ion Ratio Lower Upper
252 100
253 26.5 20.9 31.3
125 20.3 14.1 21.1



#38
Benzo(k)fluoranthene
Concen: 0.406 ng
RT: 23.028 min Scan# 2597
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12


Tgt Ion:252 Resp: 20945
Ion Ratio Lower Upper
252 100
253 26.9 21.2 31.8
125 20.0 14.6 22.0








#39
Benzo(a)pyrene
Concen: 0.397 ng
RT: 23.581 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC





Tgt Ion:252 Resp: 16258
Ion Ratio Lower Upper
252 100
253 29.4 23.1 34.7
125 27.3 18.6 28.0




#40
Dibenzo(a,h)anthracene
Concen: 0.452 ng
RT: 26.051 min Scan# 3631
Delta R.T. 0.003 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Tgt Ion:278 Resp: 17101
Ion Ratio Lower Upper
278 100
139 21.0 17.6 26.4
279 28.4 22.6 33.8







#41
Benzo(g,h,i)perylene
Concen: 0.457 ng
RT: 26.747 min Scan# 3869
Delta R.T. -0.003 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:	276	Resp:	19764
Ion	Ratio	Lower	Upper
276	100		
277	24.6	19.7	29.5
138	26.1	20.2	30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038145.D
 Acq On : 04 Nov 2025 20:12
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 04 21:16:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.413	0.430	-4.1	97	0.00
3	n-Nitrosodimethylamine	0.532	0.524	1.5	94	0.00
4 S	2-Fluorophenol	0.942	0.881	6.5	89	0.00
5 S	Phenol-d6	1.148	1.042	9.2	88	0.00
6	bis(2-Chloroethyl)ether	1.128	1.106	2.0	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.323	0.273	15.5	83	0.00
9	Naphthalene	1.102	1.085	1.5	96	0.00
10	Hexachlorobutadiene	0.186	0.195	-4.8	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.571	0.538	5.8	93	0.00
12	2-Methylnaphthalene	0.735	0.713	3.0	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.165	0.126	23.6	80	0.00
15 S	2-Fluorobiphenyl	1.602	1.511	5.7	90	0.00
16	Acenaphthylene	1.811	1.718	5.1	94	0.00
17	Acenaphthene	1.267	1.226	3.2	95	0.00
18	Fluorene	1.661	1.586	4.5	93	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.038	30.9#	73	0.00
21	4-Bromophenyl-phenylether	0.262	0.260	0.8	91	0.00
22	Hexachlorobenzene	0.298	0.316	-6.0	93	-0.01
23	Atrazine	0.192	0.162	15.6	81	0.00
24	Pentachlorophenol	0.128	0.085	33.6#	69	0.00
25	Phenanthrene	1.270	1.235	2.8	89	0.00
26	Anthracene	1.111	1.040	6.4	87	0.00
27 SURR	Fluoranthene-d10	1.009	0.826	18.1	76	0.00
28	Fluoranthene	1.419	1.185	16.5	77	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	66	0.00
30	Pyrene	1.806	2.085	-15.4	75	0.00
31 S	Terphenyl-d14	0.869	0.962	-10.7	73	0.00
32	Benzo(a)anthracene	1.427	1.342	6.0	66	0.00
33	Chrysene	1.549	1.590	-2.6	68	0.00
34	Bis(2-ethylhexyl)phthalate	0.752	0.722	4.0	66	0.00
35 I	Perylene-d12	1.000	1.000	0.0	67	0.00
36	Indeno(1,2,3-cd)pyrene	1.643	1.816	-10.5	76	0.00
37	Benzo(b)fluoranthene	1.710	1.678	1.9	69	0.00
38	Benzo(k)fluoranthene	1.721	1.748	-1.6	72	0.00
39 C	Benzo(a)pyrene	1.369	1.357	0.9	69	0.00
40	Dibenzo(a,h)anthracene	1.265	1.428	-12.9	80	0.00
41	Benzo(g,h,i)perylene	1.444	1.650	-14.3	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038145.D
Acq On : 04 Nov 2025 20:12
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 04 21:16:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.416	-4.0	97	0.00
3	n-Nitrosodimethylamine	0.400	0.394	1.5	94	0.00
4 S	2-Fluorophenol	0.400	0.374	6.5	89	0.00
5 S	Phenol-d6	0.400	0.363	9.3	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.392	2.0	94	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.338	15.5	83	0.00
9	Naphthalene	0.400	0.394	1.5	96	0.00
10	Hexachlorobutadiene	0.400	0.420	-5.0	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	93	0.00
12	2-Methylnaphthalene	0.400	0.388	3.0	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.400	0.306	23.5	80	0.00
15 S	2-Fluorobiphenyl	0.400	0.377	5.8	90	0.00
16	Acenaphthylene	0.400	0.379	5.3	94	0.00
17	Acenaphthene	0.400	0.387	3.3	95	0.00
18	Fluorene	0.400	0.382	4.5	93	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	88	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.380	5.0	73	0.00
21	4-Bromophenyl-phenylether	0.400	0.397	0.8	91	0.00
22	Hexachlorobenzene	0.400	0.423	-5.7	93	-0.01
23	Atrazine	0.400	0.337	15.8	81	0.00
24	Pentachlorophenol	0.400	0.266	33.5#	69	0.00
25	Phenanthrene	0.400	0.389	2.8	89	0.00
26	Anthracene	0.400	0.374	6.5	87	0.00
27 SURR	Fluoranthene-d10	0.400	0.327	18.3	76	0.00
28	Fluoranthene	0.400	0.334	16.5	77	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	66	0.00
30	Pyrene	0.400	0.462	-15.5	75	0.00
31 S	Terphenyl-d14	0.400	0.443	-10.7	73	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	66	0.00
33	Chrysene	0.400	0.411	-2.7	68	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.384	4.0	66	0.00
35 I	Perylene-d12	0.400	0.400	0.0	67	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.442	-10.5	76	0.00
37	Benzo(b)fluoranthene	0.400	0.393	1.8	69	0.00
38	Benzo(k)fluoranthene	0.400	0.406	-1.5	72	0.00
39 C	Benzo(a)pyrene	0.400	0.397	0.8	69	0.00
40	Dibenzo(a,h)anthracene	0.400	0.452	-13.0	80	0.00
41	Benzo(g,h,i)perylene	0.400	0.457	-14.2	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



QC SAMPLE DATA


- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038110.D
Acq On : 28 Oct 2025 12:05
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Oct 29 13:09:46 2025






AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2467

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
68	69	0.00	2	1.8	4477	PASS
69	69	100	100	100.0	255398	PASS
70	69	0.00	2	0.5	1371	PASS
197	198	0.00	2	0.2	1550	PASS
198	198	100	100	100.0	669440	PASS
199	198	5	9	6.5	43461	PASS
365	198	1	100	3.4	22528	PASS
441	443	0.01	150	83.9	95293	PASS
442	442	100	100	100.0	571285	PASS
443	442	15	24	19.9	113539	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038110.D
Acq On : 28 Oct 2025 12:05
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 28 18:06:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 12:54:48 2025
Response via : Initial Calibration



TIC: BN038110.D\data.ms

(70) Pentachlorophenol (C)

16.839min (-0.017) 31764.46 ng




response 993628

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.51
264.00	61.60	67.24
0.00	0.00	0.00

```
Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\  
Data File : BN038110.D  
Acq On    : 28 Oct 2025  12:05  
Operator  : RC/JU  
Sample    : DFTPP  
Misc      :  
ALS Vial  : 1    Sample Multiplier: 1
```

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 28 18:06:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 12:54:48 2025
Response via : Initial Calibration



TIC: BN038110.D\data.ms

(77) Benzidine

19.439min (-0.005) 0.00 ng

response **4602519**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.42
183.00	13.20	12.28
0.00	0.00	0.00

Instrument :
BNA_N
ClientSampleId :
DFTPP

DDT Breakdown


Date	Instrument Name	DFTPP Data File
10/28/2025	BNA_N	<u>BN038110.D</u>
Compound Name	Response	Retention Time
DDT	2477320	20.674
DDD	29265	20.233
DDE	1740	19.733
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
31005	2508325	1.24

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038129.D
Acq On : 04 Nov 2025 10:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Oct 29 13:09:46 2025






AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2467

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
68	69	0.00	2	1.8	7977	PASS
69	69	100	100	100.0	447289	PASS
70	69	0.00	2	0.5	2389	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	1179904	PASS
199	198	5	9	6.9	81427	PASS
365	198	1	100	3.8	44821	PASS
441	443	0.01	150	77.2	173248	PASS
442	442	100	100	100.0	1122112	PASS
443	442	15	24	20.0	224533	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038129.D
Acq On : 04 Nov 2025 10:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Nov 04 14:05:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 04 14:05:21 2025
Response via : Initial Calibration



TIC: BN038129.D\data.ms

(70) Pentachlorophenol (C)

16.851min (0.000) 277.93 ng




response 1813451

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.08
264.00	61.60	60.19
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038129.D
Acq On : 04 Nov 2025 10:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Nov 04 14:05:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 04 14:05:21 2025
Response via : Initial Calibration



TIC: BN038129.D\data.ms

(77) Benzidine

19.451min (0.000) 0.00 ng

response 8673581

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.80
183.00	13.20	12.79
0.00	0.00	0.00

Instrument :
BNA_N
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
11/4/2025	BNA_N	<u>BN038129.D</u>
Compound Name	Response	Retention Time
DDT	4564306	20.686
DDD	39263	20.233
DDE	2743	19.733
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
42006	4606312	0.91

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB170381BL		SDG No.:	Q3526
Lab Sample ID:	PB170381BL		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.20	U	1	0.070	0.20	0.20	ug/L	11/04/25 11:43	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.33			30 - 150		83%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.35			30 - 150		88%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.35			55 - 111		88%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.39			53 - 106		97%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.46			58 - 132		114%	SPK: 0.4		
INTERNAL STANDARDS		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	8220								
1146-65-2	Naphthalene-d8	22200								
15067-26-2	Acenaphthene-d10	11600								
1517-22-2	Phenanthrene-d10	21800								
1719-03-5	Chrysene-d12	15200								
1520-96-3	Perylene-d12	13200								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038131.D
Acq On : 04 Nov 2025 11:43
Operator : RC/JU
Sample : PB170381BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170381BL

Quant Time: Nov 04 12:07:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.789	152	8221	0.400	ng	0.01
7) Naphthalene-d8	10.594	136	22221	0.400	ng	0.02
13) Acenaphthene-d10	14.441	164	11601	0.400	ng	0.01
19) Phenanthrene-d10	17.198	188	21781	0.400	ng	0.01
29) Chrysene-d12	21.384	240	15202	0.400	ng	0.00
35) Perylene-d12	23.695	264	13241	0.400	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.363	112	6175	0.319	ng	0.01
5) Phenol-d6	6.952	99	7077	0.300	ng	0.01
8) Nitrobenzene-d5	8.939	82	6330	0.353	ng	0.01
11) 2-Methylnaphthalene-d10	12.187	152	10488	0.331	ng	0.01
14) 2,4,6-Tribromophenol	15.945	330	557	0.117	ng	0.01
15) 2-Fluorobiphenyl	13.073	172	18033	0.388	ng	0.02
27) Fluoranthene-d10	19.229	212	19315	0.352	ng	0.00
31) Terphenyl-d14	19.829	244	15022	0.455	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038131.D
Acq On : 04 Nov 2025 11:43
Operator : RC/JU
Sample : PB170381BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170381BL

Quant Time: Nov 04 12:07:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.789 min Scan# 61
Delta R.T. 0.014 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion:152 Resp: 8221
Ion Ratio Lower Upper
152 100
150 153.2 122.3 183.5
115 58.7 47.4 71.0



#4
2-Fluorophenol
Concen: 0.319 ng
RT: 5.363 min Scan# 316
Delta R.T. 0.014 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion:112 Resp: 6175
Ion Ratio Lower Upper
112 100
64 55.6 45.4 68.0
63 30.3 23.2 34.8









#5
Phenol-d6
Concen: 0.300 ng
RT: 6.952 min Scan# 534
Delta R.T. 0.014 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion:	99	Resp:	7077
Ion	Ratio	Lower	Upper
99	100		
42	20.3	16.6	24.8
71	31.7	25.4	38.2



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 938
Delta R.T. 0.021 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion:	136	Resp:	22221
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.8	5.2	7.8
68	5.3	4.1	6.1









#8
Nitrobenzene-d5
Concen: 0.353 ng
RT: 8.939 min Scan# 782
Delta R.T. 0.011 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion:	82	Resp:	6330
Ion	Ratio	Lower	Upper
82	100		
128	41.4	34.1	51.1
54	56.9	43.5	65.3



#11
2-Methylnaphthalene-d10
Concen: 0.331 ng
RT: 12.187 min Scan# 1157
Delta R.T. 0.010 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion:	152	Resp:	10488
Ion	Ratio	Lower	Upper
152	100		
151	20.9	16.8	25.2





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.441 min Scan# 14

Delta R.T. 0.011 min

Lab File: BN038131.D

Acq: 04 Nov 2025 11:43

Instrument :

BNA_N

ClientSampleId :

PB170381BL





Tgt Ion:164 Resp: 11601

Ion Ratio Lower Upper

164 100

162 106.2 83.0 124.4

160 55.8 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.117 ng

RT: 15.945 min Scan# 1588

Delta R.T. 0.013 min

Lab File: BN038131.D

Acq: 04 Nov 2025 11:43




Tgt Ion:330 Resp: 557


Ion Ratio Lower Upper

330 100

332 93.7 76.6 114.8

141 49.6 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.388 ng
RT: 13.073 min Scan# 1325
Delta R.T. 0.021 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion:	172	Resp:	18033
Ion Ratio	Lower	Upper	
172	100		
171	35.0	27.8	41.6
170	22.9	18.1	27.1



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.198 min Scan# 1689
Delta R.T. 0.012 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion:	188	Resp:	21781
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	9.6	8.6	13.0









#27
Fluoranthene-d10
Concen: 0.352 ng
RT: 19.229 min Scan# 1949
Delta R.T. 0.009 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion:212 Resp: 19315
Ion Ratio Lower Upper
212 100
106 16.1 12.6 19.0
104 9.3 7.1 10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.384 min Scan# 2270
Delta R.T. 0.009 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion:240 Resp: 15202
Ion Ratio Lower Upper
240 100
120 14.8 10.7 16.1
236 29.1 23.1 34.7









#31
Terphenyl-d14
Concen: 0.455 ng
RT: 19.829 min Scan# 2076
Delta R.T. 0.009 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL

Tgt Ion:244 Resp: 15022
Ion Ratio Lower Upper
244 100
212 9.5 7.8 11.6
122 17.5 15.2 22.8



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.695 min Scan# 2825
Delta R.T. 0.015 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion:264 Resp: 13241
Ion Ratio Lower Upper
264 100
260 27.8 21.8 32.8
265 100.3 68.6 102.8



Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB170381BS		SDG No.:	Q3526
Lab Sample ID:	PB170381BS		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.33		1	0.070	0.20	0.20	ug/L	11/04/25 12:19	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.35			30 - 150		88%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.30			30 - 150		76%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.37			55 - 111		92%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.42			53 - 106		104%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.48			58 - 132		121%	SPK: 0.4		
INTERNAL STANDARDS										
		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	11500								
1146-65-2	Naphthalene-d8	31800								
15067-26-2	Acenaphthene-d10	16200								
1517-22-2	Phenanthrene-d10	30400								
1719-03-5	Chrysene-d12	16500								
1520-96-3	Perylene-d12	13200								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038132.D
 Acq On : 04 Nov 2025 12:19
 Operator : RC/JU
 Sample : PB170381BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170381BS

Quant Time: Nov 04 12:44:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	11523	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	31801	0.400	ng	# 0.00
13) Acenaphthene-d10	14.441	164	16154	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	30417	0.400	ng	0.00
29) Chrysene-d12	21.375	240	16510	0.400	ng	0.00
35) Perylene-d12	23.686	264	13194	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	9248	0.341	ng	0.00
5) Phenol-d6	6.937	99	10615	0.321	ng	0.00
8) Nitrobenzene-d5	8.929	82	9450	0.368	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	15982	0.352	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1687	0.253	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	26882	0.415	ng	0.01
27) Fluoranthene-d10	19.220	212	23204	0.303	ng	0.00
31) Terphenyl-d14	19.824	244	17337	0.483	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	3965	0.333	ng	# 62
3) n-Nitrosodimethylamine	3.564	42	5606	0.366	ng	# 94
6) bis(2-Chloroethyl)ether	7.197	93	11700	0.360	ng	99
9) Naphthalene	10.626	128	31176	0.356	ng	100
10) Hexachlorobutadiene	10.925	225	5678	0.384	ng	# 99
12) 2-Methylnaphthalene	12.248	142	17976	0.308	ng	99
16) Acenaphthylene	14.152	152	28630	0.391	ng	99
17) Acenaphthene	14.494	154	17838	0.349	ng	98
18) Fluorene	15.489	166	24190	0.361	ng	99
20) 4,6-Dinitro-2-methylph...	15.560	198	1040	0.363	ng	95
21) 4-Bromophenyl-phenylether	16.379	248	7282	0.365	ng	97
22) Hexachlorobenzene	16.503	284	8817	0.389	ng	99
23) Atrazine	16.652	200	4704	0.322	ng	99
24) Pentachlorophenol	16.838	266	4384	0.449	ng	99
25) Phenanthrene	17.223	178	35147	0.364	ng	100
26) Anthracene	17.322	178	30413	0.360	ng	99
28) Fluoranthene	19.252	202	32822	0.304	ng	100
30) Pyrene	19.615	202	31678	0.425	ng	100
32) Benzo(a)anthracene	21.357	228	20604	0.350	ng	99
33) Chrysene	21.411	228	24309	0.380	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	8897	0.287	ng	98
36) Indeno(1,2,3-cd)pyrene	26.036	276	23949	0.442	ng	98
37) Benzo(b)fluoranthene	22.987	252	20712	0.367	ng	98
38) Benzo(k)fluoranthene	23.031	252	21547	0.379	ng	98
39) Benzo(a)pyrene	23.584	252	17494	0.387	ng	98
40) Dibenzo(a,h)anthracene	26.054	278	17865	0.428	ng	99
41) Benzo(g,h,i)perylene	26.756	276	21276	0.447	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038132.D
Acq On : 04 Nov 2025 12:19
Operator : RC/JU
Sample : PB170381BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170381BS


Quant Time: Nov 04 12:44:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19


Instrument :
BNA_N
ClientSampleId :
PB170381BS





Tgt Ion:152 Resp: 11523
Ion Ratio Lower Upper
152 100
150 156.9 122.3 183.5
115 60.3 47.4 71.0




#2
1,4-Dioxane
Concen: 0.333 ng
RT: 3.261 min Scan# 25
Delta R.T. 0.007 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19



Tgt Ion: 88 Resp: 3965
Ion Ratio Lower Upper
88 100
43 84.7 24.3 36.5#
58 86.6 60.4 90.6









#3
 n-Nitrosodimethylamine
 Concen: 0.366 ng
 RT: 3.564 min Scan# 61
 Delta R.T. 0.007 min
 Lab File: BN038132.D
 Acq: 04 Nov 2025 12:19




Instrument :
 BNA_N
 ClientSampleId :
 PB170381BS


Tgt Ion: 42 Resp: 5606
 Ion Ratio Lower Upper
 42 100
 74 115.6 96.8 145.2
 44 14.6 5.3 7.9#



#4
 2-Fluorophenol
 Concen: 0.341 ng
 RT: 5.348 min Scan# 314
 Delta R.T. -0.000 min
 Lab File: BN038132.D
 Acq: 04 Nov 2025 12:19

Tgt Ion: 112 Resp: 9248
 Ion Ratio Lower Upper
 112 100
 64 56.5 45.4 68.0
 63 30.1 23.2 34.8









#5
Phenol-d6
Concen: 0.321 ng
RT: 6.937 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:	99	Resp:	10615
Ion	Ratio	Lower	Upper
99	100		
42	21.8	16.6	24.8
71	31.9	25.4	38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.360 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:	93	Resp:	11700
Ion	Ratio	Lower	Upper
93	100		
63	75.6	59.3	88.9
95	31.9	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 937
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:	136	Resp:	31801
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.0	13.4
54	9.9	5.2	7.8#
68	7.4	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 0.368 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:	82	Resp:	9450
Ion Ratio	Lower	Upper	
82	100		
128	44.3	34.1	51.1
54	52.5	43.5	65.3









#9
Naphthalene
Concen: 0.356 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:	128	Resp:	31176
Ion Ratio	Lower	Upper	
128	100		
129	11.3	9.1	13.7
127	13.3	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.384 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:	225	Resp:	5678
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.5	50.3	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.352 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:152 Resp: 15982
Ion Ratio Lower Upper
152 100
151 20.6 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.308 ng
RT: 12.248 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:142 Resp: 17976
Ion Ratio Lower Upper
142 100
141 88.8 70.3 105.5
115 35.3 27.8 41.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.441 min Scan# 1453

Delta R.T. 0.011 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:164 Resp: 16154

Ion Ratio Lower Upper

164 100

162 103.7 83.0 124.4

160 52.4 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.253 ng

RT: 15.932 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19




Tgt Ion:330 Resp: 1687


Ion Ratio Lower Upper

330 100

332 95.7 76.6 114.8

141 45.4 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.415 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:	172	Resp:	26882
Ion Ratio	Lower	Upper	
172	100		
171	34.1	27.8	41.6
170	22.8	18.1	27.1



#16
Acenaphthylene
Concen: 0.391 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:	152	Resp:	28630
Ion Ratio	Lower	Upper	
152	100		
151	19.4	15.8	23.8
153	12.9	10.2	15.2









#17
Acenaphthene
Concen: 0.349 ng
RT: 14.494 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:154 Resp: 17838
Ion Ratio Lower Upper
154 100
153 114.1 90.0 135.0
152 62.2 47.3 70.9



#18
Fluorene
Concen: 0.361 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:166 Resp: 24190
Ion Ratio Lower Upper
166 100
165 99.4 79.0 118.4
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:188 Resp: 30417
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.0 8.6 13.0



#20
4,6-Dinitro-2-methylphenol
Concen: 0.363 ng
RT: 15.560 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:198 Resp: 1040
Ion Ratio Lower Upper
198 100
51 103.2 88.9 133.3
105 61.7 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.365 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:	248	Resp:	7282
Ion Ratio	Lower	Upper	
248	100		
250	93.4	76.2	114.2
141	67.9	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.389 ng
RT: 16.503 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:	284	Resp:	8817
Ion Ratio	Lower	Upper	
284	100		
142	38.4	30.5	45.7
249	30.0	23.5	35.3









#23
Atrazine
 Concen: 0.322 ng
 RT: 16.652 min Scan# 1645
 Delta R.T. -0.000 min
 Lab File: BN038132.D
 Acq: 04 Nov 2025 12:19




Instrument : BNA_N
ClientSampleId : PB170381BS


Tgt Ion:	200	Resp:	4704
Ion Ratio	Lower	Upper	
200	100		
173	28.3	21.4	32.2
215	45.7	36.7	55.1



#24
Pentachlorophenol
 Concen: 0.449 ng
 RT: 16.838 min Scan# 1660
 Delta R.T. -0.000 min
 Lab File: BN038132.D
 Acq: 04 Nov 2025 12:19

Tgt Ion:	266	Resp:	4384
Ion Ratio	Lower	Upper	
266	100		
264	61.1	49.3	73.9
268	63.9	50.6	75.8





#25

Phenanthrene

Concen: 0.364 ng

RT: 17.223 min Scan# 1691

Delta R.T. -0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:178 Resp: 35147

Ion Ratio Lower Upper

178 100

176 19.4 15.5 23.3

179 15.2 12.2 18.4



#26

Anthracene

Concen: 0.360 ng

RT: 17.322 min Scan# 1699

Delta R.T. -0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19




Tgt Ion:178 Resp: 30413


Ion Ratio Lower Upper

178 100

176 18.7 14.9 22.3

179 15.4 11.8 17.8









#27
Fluoranthene-d10
Concen: 0.303 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:212 Resp: 23204
Ion Ratio Lower Upper
212 100
106 16.4 12.6 19.0
104 9.3 7.1 10.7



#28
Fluoranthene
Concen: 0.304 ng
RT: 19.252 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:202 Resp: 32822
Ion Ratio Lower Upper
202 100
101 12.2 9.4 14.2
203 16.8 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:240 Resp: 16510

Ion Ratio Lower Upper

240 100

120 14.8 10.7 16.1

236 29.1 23.1 34.7



#30

Pyrene

Concen: 0.425 ng

RT: 19.615 min Scan# 2032

Delta R.T. 0.005 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19




Tgt Ion:202 Resp: 31678


Ion Ratio Lower Upper

202 100

200 21.1 16.7 25.1

203 17.6 14.2 21.2





#31

Terphenyl-d14

Concen: 0.483 ng

RT: 19.824 min Scan# 2076

Delta R.T. 0.005 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:244 Resp: 17337

Ion Ratio Lower Upper

244 100

212 9.5 7.8 11.6

122 17.7 15.2 22.8



#32

Benzo(a)anthracene

Concen: 0.350 ng

RT: 21.357 min Scan# 2267

Delta R.T. -0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19




Tgt Ion:228 Resp: 20604


Ion Ratio Lower Upper

228 100

226 27.8 22.2 33.4

229 19.2 15.8 23.8









#33
 Chrysene
 Concen: 0.380 ng
 RT: 21.411 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN038132.D
 Acq: 04 Nov 2025 12:19




Instrument :
 BNA_N
 ClientSampleId :
 PB170381BS


Tgt Ion:	228	Resp:	24309
Ion Ratio	Lower	Upper	
228	100		
226	30.7	24.3	36.5
229	19.7	15.7	23.5



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.287 ng
 RT: 21.295 min Scan# 2260
 Delta R.T. -0.000 min
 Lab File: BN038132.D
 Acq: 04 Nov 2025 12:19


Tgt Ion:	149	Resp:	8897
Ion Ratio	Lower	Upper	
149	100		
167	25.4	21.1	31.7
279	3.5	2.9	4.3








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.686 min Scan# 2821
Delta R.T. 0.006 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Instrument :
BNA_N
ClientSampleId :
PB170381BS







Tgt Ion:264 Resp: 13194
Ion Ratio Lower Upper
264 100
260 27.3 21.8 32.8
265 94.2 68.6 102.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.442 ng
RT: 26.036 min Scan# 3626
Delta R.T. 0.003 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:276 Resp: 23949
Ion Ratio Lower Upper
276 100
138 28.4 21.4 32.0
277 23.9 19.2 28.8





#37

Benzo(b)fluoranthene

Concen: 0.367 ng

RT: 22.987 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:252 Resp: 20712

Ion Ratio Lower Upper

252 100

253 26.2 20.9 31.3

125 19.9 14.1 21.1



#38

Benzo(k)fluoranthene

Concen: 0.379 ng

RT: 23.031 min Scan# 2598

Delta R.T. 0.003 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19




Tgt Ion:252 Resp: 21547


Ion Ratio Lower Upper

252 100

253 26.2 21.2 31.8

125 19.8 14.6 22.0





#39

Benzo(a)pyrene

Concen: 0.387 ng

RT: 23.584 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:252 Resp: 17494

Ion Ratio Lower Upper

252 100

253 28.7 23.1 34.7

125 24.8 18.6 28.0



#40

Dibenzo(a,h)anthracene

Concen: 0.428 ng

RT: 26.054 min Scan# 3632

Delta R.T. 0.006 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19




Tgt Ion:278 Resp: 17865


Ion Ratio Lower Upper

278 100

139 21.4 17.6 26.4


279 28.1 22.6 33.8







#41
Benzo(g,h,i)perylene
Concen: 0.447 ng
RT: 26.756 min Scan# 3869
Delta R.T. 0.006 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Instrument :
BNA_N
ClientSampleId :
PB170381BS



Tgt Ion:	276	Resp:	21276
Ion	Ratio	Lower	Upper
276	100		
277	24.3	19.7	29.5
138	26.5	20.2	30.4



Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	
Client Sample ID:	PB170381BSD	SDG No.:	Q3526
Lab Sample ID:	PB170381BSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 mL	Test:	SVOC-SIMGroup1
Prep Method :	3510C		
	Level: LOW		
	Final Vol: 1000 uL		
	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.26		1	0.070	0.20	0.20	ug/L	11/04/25 12:56	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.29			30 - 150		73%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.25			30 - 150		63%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.31			55 - 111		76%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.32			53 - 106		81%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.36			58 - 132		91%	SPK: 0.4		
INTERNAL STANDARDS										
		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	10800								
1146-65-2	Naphthalene-d8	30500								
15067-26-2	Acenaphthene-d10	15600								
1517-22-2	Phenanthrene-d10	28800								
1719-03-5	Chrysene-d12	18100								
1520-96-3	Perylene-d12	14900								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038133.D
 Acq On : 04 Nov 2025 12:56
 Operator : RC/JU
 Sample : PB170381BSD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170381BSD

Quant Time: Nov 04 13:20:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	10823	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	30534	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	15616	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	28756	0.400	ng	0.00
29) Chrysene-d12	21.375	240	18053	0.400	ng	0.00
35) Perylene-d12	23.683	264	14856	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	7259	0.285	ng	0.00
5) Phenol-d6	6.937	99	8491	0.273	ng	0.00
8) Nitrobenzene-d5	8.929	82	7527	0.305	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	12754	0.292	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1353	0.210	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	20205	0.323	ng	0.00
27) Fluoranthene-d10	19.220	212	18424	0.254	ng	0.00
31) Terphenyl-d14	19.824	244	14219	0.363	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	2913	0.260	ng	# 69
3) n-Nitrosodimethylamine	3.564	42	4281	0.297	ng	# 93
6) bis(2-Chloroethyl)ether	7.197	93	9060	0.297	ng	98
9) Naphthalene	10.626	128	25040	0.298	ng	100
10) Hexachlorobutadiene	10.925	225	4498	0.317	ng	# 99
12) 2-Methylnaphthalene	12.248	142	14295	0.255	ng	97
16) Acenaphthylene	14.152	152	22428	0.317	ng	99
17) Acenaphthene	14.495	154	14121	0.285	ng	97
18) Fluorene	15.489	166	18901	0.291	ng	99
20) 4,6-Dinitro-2-methylph...	15.560	198	792	0.328	ng	# 80
21) 4-Bromophenyl-phenylether	16.379	248	5427	0.288	ng	100
22) Hexachlorobenzene	16.503	284	6880	0.321	ng	99
23) Atrazine	16.652	200	3562	0.258	ng	99
24) Pentachlorophenol	16.838	266	3398	0.368	ng	99
25) Phenanthrene	17.223	178	26605	0.291	ng	100
26) Anthracene	17.322	178	23247	0.291	ng	99
28) Fluoranthene	19.253	202	25972	0.255	ng	100
30) Pyrene	19.615	202	25592	0.314	ng	100
32) Benzo(a)anthracene	21.358	228	18334	0.285	ng	99
33) Chrysene	21.411	228	21765	0.311	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	7715	0.227	ng	99
36) Indeno(1,2,3-cd)pyrene	26.037	276	19208	0.315	ng	97
37) Benzo(b)fluoranthene	22.984	252	18434	0.290	ng	97
38) Benzo(k)fluoranthene	23.031	252	19683	0.308	ng	98
39) Benzo(a)pyrene	23.581	252	15027	0.296	ng	97
40) Dibenzo(a,h)anthracene	26.054	278	14795	0.315	ng	99
41) Benzo(g,h,i)perylene	26.753	276	17207	0.321	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038133.D
Acq On : 04 Nov 2025 12:56
Operator : RC/JU
Sample : PB170381BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Quant Time: Nov 04 13:20:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.775 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN038113.D
 Acq: 04 Nov 2025 12:56


Instrument :
 BNA_N
 ClientSampleId :
 PB170381BSD





Tgt Ion:152 Resp: 10823
 Ion Ratio Lower Upper
 152 100
 150 154.8 122.3 183.5
 115 59.3 47.4 71.0




#2
 1,4-Dioxane
 Concen: 0.260 ng
 RT: 3.254 min Scan# 24
 Delta R.T. 0.000 min
 Lab File: BN038113.D
 Acq: 04 Nov 2025 12:56



Tgt Ion: 88 Resp: 2913
 Ion Ratio Lower Upper
 88 100
 43 71.2 24.3 36.5#
 58 87.3 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 0.297 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion: 42 Resp: 4281
Ion Ratio Lower Upper
42 100
74 116.1 96.8 145.2
44 25.7 5.3 7.9#



#4
2-Fluorophenol
Concen: 0.285 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion: 112 Resp: 7259
Ion Ratio Lower Upper
112 100
64 57.2 45.4 68.0
63 30.8 23.2 34.8









#5
Phenol-d6
Concen: 0.273 ng
RT: 6.937 min Scan# 534
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:	99	Resp:	8491
Ion	Ratio	Lower	Upper
99	100		
42	22.9	16.6	24.8
71	31.9	25.4	38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.297 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56

Tgt Ion:	93	Resp:	9060
Ion	Ratio	Lower	Upper
93	100		
63	75.7	59.3	88.9
95	32.3	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 937
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:	136	Resp:	30534
Ion Ratio	Lower	Upper	
136	100		
137	11.1	9.0	13.4
54	9.8	5.2	7.8#
68	7.2	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 0.305 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56

Tgt Ion:	82	Resp:	7527
Ion Ratio	Lower	Upper	
82	100		
128	44.1	34.1	51.1
54	52.6	43.5	65.3





#9

Naphthalene

Concen: 0.298 ng

RT: 10.626 min Scan# 941

Delta R.T. 0.000 min

Lab File: BN038133.D

Acq: 04 Nov 2025 12:56





Instrument :

BNA_N

ClientSampleId :

PB170381BSD

Tgt Ion:	128	Resp:	25040
Ion Ratio	Lower	Upper	
128	100		
129	11.3	9.1	13.7
127	13.1	10.5	15.7



#10

Hexachlorobutadiene

Concen: 0.317 ng




RT: 10.925 min Scan# 969


Delta R.T. 0.000 min

Lab File: BN038133.D

Acq: 04 Nov 2025 12:56

Tgt Ion:	225	Resp:	4498
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.5	50.3	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.292 ng
RT: 12.172 min Scan# 1155
Delta R.T. -0.005 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:152 Resp: 12754
Ion Ratio Lower Upper
152 100
151 20.9 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.255 ng
RT: 12.248 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56

Tgt Ion:142 Resp: 14295
Ion Ratio Lower Upper
142 100
141 90.5 70.3 105.5
115 35.9 27.8 41.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.430 min Scan# 14

Delta R.T. 0.000 min

Lab File: BN038113.D

Acq: 04 Nov 2025 12:56

Instrument :

BNA_N

ClientSampleId :

PB170381BSD





Tgt Ion:164 Resp: 15616

Ion Ratio Lower Upper

164 100

162 107.3 83.0 124.4

160 57.4 40.7 61.1



#14

2,4,6-Tribromophenol

Concen: 0.210 ng

RT: 15.932 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BN038113.D

Acq: 04 Nov 2025 12:56




Tgt Ion:330 Resp: 1353


Ion Ratio Lower Upper

330 100

332 99.6 76.6 114.8

141 45.0 34.1 51.1









#15
 2-Fluorobiphenyl
 Concen: 0.323 ng
 RT: 13.051 min Scan# 1323
 Delta R.T. 0.000 min
 Lab File: BN038133.D
 Acq: 04 Nov 2025 12:56




Instrument :
 BNA_N
 ClientSampleId :
 PB170381BSD


Tgt Ion:	172	Resp:	20205
Ion Ratio	Lower	Upper	
172	100		
171	36.0	27.8	41.6
170	24.3	18.1	27.1



#16
 Acenaphthylene
 Concen: 0.317 ng
 RT: 14.152 min Scan# 1426
 Delta R.T. 0.000 min
 Lab File: BN038133.D
 Acq: 04 Nov 2025 12:56

Tgt Ion:	152	Resp:	22428
Ion Ratio	Lower	Upper	
152	100		
151	19.6	15.8	23.8
153	13.0	10.2	15.2









#17
Acenaphthene
Concen: 0.285 ng
RT: 14.495 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:154 Resp: 14121
Ion Ratio Lower Upper
154 100
153 115.5 90.0 135.0
152 62.6 47.3 70.9



#18
Fluorene
Concen: 0.291 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56

Tgt Ion:166 Resp: 18901
Ion Ratio Lower Upper
166 100
165 99.6 79.0 118.4
167 13.4 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1688

Delta R.T. 0.000 min

Lab File: BN038113.D

Acq: 04 Nov 2025 12:56

Instrument :

BNA_N

ClientSampleId :

PB170381BSD





Tgt Ion:188 Resp: 28756

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.1 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.328 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038113.D

Acq: 04 Nov 2025 12:56




Tgt Ion:198 Resp: 792


Ion Ratio Lower Upper

198 100

51 131.7 88.9 133.3

105 76.9 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.288 ng
RT: 16.379 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:	248	Resp:	5427
Ion Ratio	Lower	Upper	
248	100		
250	94.4	76.2	114.2
141	72.0	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.321 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56

Tgt Ion:	284	Resp:	6880
Ion Ratio	Lower	Upper	
284	100		
142	38.7	30.5	45.7
249	29.9	23.5	35.3









#23
Atrazine
 Concen: 0.258 ng
 RT: 16.652 min Scan# 1645
 Delta R.T. 0.000 min
 Lab File: BN038133.D
 Acq: 04 Nov 2025 12:56




Instrument :
 BNA_N
ClientSampleId :
 PB170381BSD


Tgt Ion:	200	Resp:	3562
Ion Ratio	Lower	Upper	
200	100		
173	28.0	21.4	32.2
215	45.8	36.7	55.1



#24
Pentachlorophenol
 Concen: 0.368 ng
 RT: 16.838 min Scan# 1660
 Delta R.T. 0.000 min
 Lab File: BN038133.D
 Acq: 04 Nov 2025 12:56

Tgt Ion:	266	Resp:	3398
Ion Ratio	Lower	Upper	
266	100		
264	63.1	49.3	73.9
268	63.3	50.6	75.8





#25

Phenanthrene

Concen: 0.291 ng

RT: 17.223 min Scan# 1691

Delta R.T. 0.000 min

Lab File: BN038113.D

Acq: 04 Nov 2025 12:56

Instrument :

BNA_N

ClientSampleId :

PB170381BSD





Tgt Ion:178 Resp: 26605

Ion Ratio Lower Upper

178 100

176 19.5 15.5 23.3

179 15.5 12.2 18.4



#26

Anthracene

Concen: 0.291 ng

RT: 17.322 min Scan# 1699

Delta R.T. 0.000 min

Lab File: BN038113.D

Acq: 04 Nov 2025 12:56




Tgt Ion:178 Resp: 23247


Ion Ratio Lower Upper

178 100

176 18.5 14.9 22.3

179 15.4 11.8 17.8









#27
Fluoranthene-d10
Concen: 0.254 ng
RT: 19.220 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:212 Resp: 18424
Ion Ratio Lower Upper
212 100
106 16.8 12.6 19.0
104 9.5 7.1 10.7



#28
Fluoranthene
Concen: 0.255 ng
RT: 19.253 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion:202 Resp: 25972
Ion Ratio Lower Upper
202 100
101 11.9 9.4 14.2
203 16.9 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038113.D

Acq: 04 Nov 2025 12:56

Instrument :

BNA_N

ClientSampleId :

PB170381BSD





Tgt Ion:240 Resp: 18053

Ion Ratio Lower Upper

240 100

120 14.7 10.7 16.1

236 28.9 23.1 34.7



#30

Pyrene

Concen: 0.314 ng

RT: 19.615 min Scan# 2032

Delta R.T. 0.005 min

Lab File: BN038113.D

Acq: 04 Nov 2025 12:56




Tgt Ion:202 Resp: 25592


Ion Ratio Lower Upper

202 100

200 21.0 16.7 25.1

203 17.4 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.363 ng
 RT: 19.824 min Scan# 2076
 Delta R.T. 0.005 min
 Lab File: BN038113.D
 Acq: 04 Nov 2025 12:56




Instrument :
 BNA_N
 ClientSampleId :
 PB170381BSD


Tgt Ion:244 Resp: 14219
 Ion Ratio Lower Upper
 244 100
 212 9.6 7.8 11.6
 122 18.1 15.2 22.8



#32
 Benzo(a)anthracene
 Concen: 0.285 ng
 RT: 21.358 min Scan# 2267
 Delta R.T. 0.000 min
 Lab File: BN038113.D
 Acq: 04 Nov 2025 12:56


Tgt Ion:228 Resp: 18334
 Ion Ratio Lower Upper
 228 100
 226 28.4 22.2 33.4
 229 19.9 15.8 23.8








#33
Chrysene
Concen: 0.311 ng
RT: 21.411 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56

Instrument :
BNA_N
ClientSampleId :
PB170381BSD







Tgt Ion:228 Resp: 21765
Ion Ratio Lower Upper
228 100
226 30.6 24.3 36.5
229 19.6 15.7 23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.227 ng
RT: 21.295 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56

Tgt Ion:149 Resp: 7715
Ion Ratio Lower Upper
149 100
167 27.0 21.1 31.7
279 3.6 2.9 4.3









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.683 min Scan# 2821
 Delta R.T. 0.003 min
 Lab File: BN038113.D
 Acq: 04 Nov 2025 12:56




Instrument :
 BNA_N
 ClientSampleId :
 PB170381BSD


Tgt Ion:	264	Resp:	14856
Ion Ratio	Lower	Upper	
264	100		
260	27.5	21.8	32.8
265	85.8	68.6	102.8



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.315 ng
 RT: 26.037 min Scan# 3626
 Delta R.T. 0.003 min
 Lab File: BN038113.D
 Acq: 04 Nov 2025 12:56


Tgt Ion:	276	Resp:	19208
Ion Ratio	Lower	Upper	
276	100		
138	28.0	21.4	32.0
277	25.2	19.2	28.8








#37
Benzo(b)fluoranthene
Concen: 0.290 ng
RT: 22.984 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56

Instrument :
BNA_N
ClientSampleId :
PB170381BSD







Tgt Ion:252 Resp: 18434
Ion Ratio Lower Upper
252 100
253 26.9 20.9 31.3
125 20.3 14.1 21.1



#38
Benzo(k)fluoranthene
Concen: 0.308 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.003 min
Lab File: BN038113.D
Acq: 04 Nov 2025 12:56

Tgt Ion:252 Resp: 19683
Ion Ratio Lower Upper
252 100
253 26.4 21.2 31.8
125 20.3 14.6 22.0





#39

Benzo(a)pyrene

Concen: 0.296 ng

RT: 23.581 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN038113.D

Acq: 04 Nov 2025 12:56

Instrument :

BNA_N

ClientSampleId :

PB170381BSD





Tgt Ion:252 Resp: 15027

Ion Ratio Lower Upper

252 100

253 29.7 23.1 34.7

125 25.7 18.6 28.0



#40

Dibenzo(a,h)anthracene

Concen: 0.315 ng

RT: 26.054 min Scan# 3632

Delta R.T. 0.006 min

Lab File: BN038113.D

Acq: 04 Nov 2025 12:56




Tgt Ion:278 Resp: 14795


Ion Ratio Lower Upper

278 100

139 21.6 17.6 26.4


279 28.7 22.6 33.8







#41
Benzo(g,h,i)perylene
Concen: 0.321 ng
RT: 26.753 min Scan# 38
Delta R.T. 0.003 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Instrument :
BNA_N
ClientSampleId :
PB170381BSD



Tgt Ion:	276	Resp:	17207
Ion	Ratio	Lower	Upper
276	100		
277	24.3	19.7	29.5
138	26.9	20.2	30.4



Manual Integration Report

Sequence:	BN102825	Instrument	BNA_n
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC0.8	BN038114.D	Benzo(b)fluoranthene	rahul	10/29/2025 9:43:43 AM	mohammad	10/30/2025 5:29:45 AM	Peak Integrated by Software



Manual Integration Report

Sequence:	BN110425	Instrument	BNA_n
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
-----------	---------	-----------	-----------	-----------	---------------	---------------	--------

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN102825

Review By	rahul	Review On	10/28/2025 3:41:07 PM
Supervise By	mohammad	Supervise On	10/30/2025 5:29:45 AM
SubDirectory	BN102825	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method BN102825
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904		
CCC	SP6902		
Internal Standard/PEM	SP6830,1ul/100ul sample		
ICV/I.BLK	SP6906		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN038110.D	28 Oct 2025 12:05	RC/JU	Ok
2	SSTDICC0.1	BN038111.D	28 Oct 2025 13:00	RC/JU	Ok
3	SSTDICC0.2	BN038112.D	28 Oct 2025 13:36	RC/JU	Ok
4	SSTDICCC0.4	BN038113.D	28 Oct 2025 14:13	RC/JU	Ok
5	SSTDICC0.8	BN038114.D	28 Oct 2025 14:49	RC/JU	Ok,M
6	SSTDICC1.6	BN038115.D	28 Oct 2025 15:25	RC/JU	Ok
7	SSTDICC3.2	BN038116.D	28 Oct 2025 16:02	RC/JU	Ok
8	SSTDICC5.0	BN038117.D	28 Oct 2025 16:38	RC/JU	Ok
9	SSTDICV0.4	BN038118.D	28 Oct 2025 17:51	RC/JU	Ok
10	PB170167BL	BN038119.D	28 Oct 2025 19:04	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN110425

Review By	raahul	Review On	11/5/2025 10:39:28 AM
Supervise By	Jagrut	Supervise On	11/5/2025 2:05:23 PM
SubDirectory	BN110425	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method BN102825
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904		
CCC	SP6902		
Internal Standard/PEM	SP6830,1ul/100ul sample		
ICV/I.BLK	SP6906		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN038129.D	04 Nov 2025 10:00	RC/JU	Ok
2	SSTDCCC0.4	BN038130.D	04 Nov 2025 10:40	RC/JU	Ok
3	PB170381BL	BN038131.D	04 Nov 2025 11:43	RC/JU	Ok
4	PB170381BS	BN038132.D	04 Nov 2025 12:19	RC/JU	Ok
5	PB170381BSD	BN038133.D	04 Nov 2025 12:56	RC/JU	Ok
6	Q3494-01	BN038134.D	04 Nov 2025 13:32	RC/JU	Ok
7	Q3525-01	BN038135.D	04 Nov 2025 14:08	RC/JU	Dilution
8	Q3525-02	BN038136.D	04 Nov 2025 14:45	RC/JU	Ok
9	Q3525-03	BN038137.D	04 Nov 2025 15:21	RC/JU	Ok
10	Q3526-01	BN038138.D	04 Nov 2025 15:57	RC/JU	Ok
11	Q3526-02	BN038139.D	04 Nov 2025 16:34	RC/JU	Ok
12	Q3526-03	BN038140.D	04 Nov 2025 17:10	RC/JU	Ok
13	Q3526-04	BN038141.D	04 Nov 2025 17:46	RC/JU	Ok
14	Q3527-01	BN038142.D	04 Nov 2025 18:22	RC/JU	Ok
15	Q3527-02	BN038143.D	04 Nov 2025 18:59	RC/JU	Ok
16	Q3525-01DL	BN038144.D	04 Nov 2025 19:35	RC/JU	Ok
17	SSTDCCC0.4	BN038145.D	04 Nov 2025 20:12	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102825

Review By	rahul	Review On	10/28/2025 3:41:07 PM
Supervise By	mohammad	Supervise On	10/30/2025 5:29:45 AM
SubDirectory	BN102825	HP Acquire Method	BNA_N, 8270_HP Processing Method BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038110.D	28 Oct 2025 12:05		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN038111.D	28 Oct 2025 13:00		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN038112.D	28 Oct 2025 13:36		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN038113.D	28 Oct 2025 14:13	compound #20 kept on LR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN038114.D	28 Oct 2025 14:49		RC/JU	Ok,M
6	SSTDICC1.6	SSTDICC1.6	BN038115.D	28 Oct 2025 15:25		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN038116.D	28 Oct 2025 16:02		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN038117.D	28 Oct 2025 16:38	Compound #20 removed from 5PPM	RC/JU	Ok
9	SSTDICV0.4	ICVBN102825	BN038118.D	28 Oct 2025 17:51		RC/JU	Ok
10	PB170167BL	PB170167BL	BN038119.D	28 Oct 2025 19:04	END CCC missing, Analyzed for Contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN110425

Review By	rahul	Review On	11/5/2025 10:39:28 AM
Supervise By	Jagrut	Supervise On	11/5/2025 2:05:23 PM
SubDirectory	BN110425	HP Acquire Method	BNA_N, 8270_HP Processing Method BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTTP	DFTTP	BN038129.D	04 Nov 2025 10:00		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038130.D	04 Nov 2025 10:40		RC/JU	Ok
3	PB170381BL	PB170381BL	BN038131.D	04 Nov 2025 11:43		RC/JU	Ok
4	PB170381BS	PB170381BS	BN038132.D	04 Nov 2025 12:19		RC/JU	Ok
5	PB170381BSD	PB170381BSD	BN038133.D	04 Nov 2025 12:56	Recovery fail low for 1,4-Dioxane.	RC/JU	Ok
6	Q3494-01	MW3	BN038134.D	04 Nov 2025 13:32		RC/JU	Ok
7	Q3525-01	RW5-SP100-20251030	BN038135.D	04 Nov 2025 14:08	Need 2X Dilution	RC/JU	Dilution
8	Q3525-02	RW5-SP201-20251030	BN038136.D	04 Nov 2025 14:45		RC/JU	Ok
9	Q3525-03	RW5-SP303-20251030	BN038137.D	04 Nov 2025 15:21		RC/JU	Ok
10	Q3526-01	RW7-SP100-20251030	BN038138.D	04 Nov 2025 15:57		RC/JU	Ok
11	Q3526-02	RW7-SP201-20251030	BN038139.D	04 Nov 2025 16:34		RC/JU	Ok
12	Q3526-03	RW7-SP302-20251030	BN038140.D	04 Nov 2025 17:10		RC/JU	Ok
13	Q3526-04	RW7-SP303-20251030	BN038141.D	04 Nov 2025 17:46		RC/JU	Ok
14	Q3527-01	RW8-SP100-20251030	BN038142.D	04 Nov 2025 18:22		RC/JU	Ok
15	Q3527-02	RW8-SP303-20251030	BN038143.D	04 Nov 2025 18:59		RC/JU	Ok
16	Q3525-01DL	RW5-SP100-20251030	BN038144.D	04 Nov 2025 19:35		RC/JU	Ok
17	SSTDCCC0.4	SSTDCCC0.4EC	BN038145.D	04 Nov 2025 20:12		RC/JU	Ok

M : Manual Integration

SOP ID: M3510C,3580A-Extraction SVOC-21

Clean Up SOP #: N/A

Matrix : Water

Weigh By: N/A **Extraction By:** RS

Balance check: N/A **Filter By:** RS

Balance ID: N/A **pH Meter ID:** N/A

pH Strip Lot#: E3953 **Hood ID:** 4,6,7

Extraction Method: ☒ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Extraction Start Date : 11/03/2025

Extraction Start Time : 11:20

Extraction End Date : 11/03/2025

Extraction End Time : 16:20

Concentration By: EH

Supervisor By : RUPESH

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	0.4 PPM	SP6876
Surrogate	1.0ML	0.4 PPM	SP6873
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3980
Baked Na2SO4	N/A	EP2655
H2SO4 1:1	N/A	EP2610
10N NaOH	N/A	EP2609
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

PH Adjusted TO <2 with 1:1 H2SO4 & >11 with 10N NAOH ,1.5ML Vial Lot #2210443.

KD Bath ID: WATER BATH-1 **Envap ID:** NEVAP-02

KD Bath Temperature: 60 °C **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/3/25	RS (Ext Lab)	Relsvoc
16:25	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction SVOC-21

Concentration Date: 11/03/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170381BL	SBLK381	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			SEP-1
PB170381BS	SLCS381	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			2
PB170381BS D	SLCSD381	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			3
Q3494-01	MW3	SVOC-SIMGrou p1	990	6	RUPESH	ritesh	1	D		4
Q3525-01	RW5-SP100-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			5
Q3525-02	RW5-SP201-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			6
Q3525-03	RW5-SP303-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			7
Q3526-01	RW7-SP100-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			8
Q3526-02	RW7-SP201-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			9
Q3526-03	RW7-SP302-20251030	SVOC-SIMGrou p1	940	6	RUPESH	ritesh	1			10
Q3526-04	RW7-SP303-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			11
Q3527-01	RW8-SP100-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1	D		12
Q3527-02	RW8-SP303-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1	D		13

Rs
11/3

* Extracts relinquished on the same date as received.

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3526 WorkList ID : 192855 Department : Extraction Date : 11-03-2025 11:15:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3494-01	MW3	Water	SVOC-SIMGroup1	Cool 4 deg C	GENV01	J31	10/29/2025	8270-Modified
Q3525-01	RW5-SP100-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3525-02	RW5-SP201-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3525-03	RW5-SP303-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3526-01	RW7-SP100-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3526-02	RW7-SP201-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3526-03	RW7-SP302-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3526-04	RW7-SP303-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3527-01	RW8-SP100-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	J32	10/30/2025	8270-Modified
Q3527-02	RW8-SP303-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	J32	10/30/2025	8270-Modified

Date/Time 11/3/25 11:15
 Raw Sample Received by: RS (Ex-66)
 Raw Sample Relinquished by: RS (Ex-66)

Date/Time 11/3/25 11:50
 Raw Sample Received by: RS (Ex-66)
 Raw Sample Relinquished by: RS (Ex-66)

Prep Standard - Chemical Standard Summary

Order ID : Q3526
Test : SVOC-SIMGroup1
Prepbatch ID : PB170381,
Sequence ID/Qc Batch ID: BN110425,

Standard ID :
EP2609,EP2610,EP2655,SP6830,SP6873,SP6875,SP6876,SP6896,SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904,SP6905,SP6906,

Chemical ID :
1ul/100ul
sample,E3657,E3875,E3942,E3972,E3973,E3980,M6157,S10105,S11497,S11652,S11828,S12116,S12201,S12273,S12560,S12563,S12577,S12670,S12905,S13058,S13097,S13120,S13122,S13245,S13249,S13299,W 3112,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1874	10 N SODIUM HYDROXIDE SOLN	EP2609	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2	None	Riteshkumar Patel

(EX-SC-2)

FROM 1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2	None	Riteshkumar Patel

(EX-SC-2)

FROM 1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2655	10/24/2025	01/28/2026	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/24/2025

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3493	Internal Standard 0.4 PPM	SP6830	06/17/2025	12/13/2025	Jagrut Upadhyay	None	None	Rahul Chavli 06/19/2025

FROM 0.10000ml of S12670 + 4.90000ml of E3942 = Final Quantity: 5.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3491	8270-SIM-Surrogate 0.4 PPM	SP6873	09/30/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								09/30/2025

FROM 0.00400ml of S12201 + 0.00800ml of S12905 + 0.02000ml of S11828 + 99.96800ml of E3972 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3895	50 ug/ml DFTPP 8270E	SP6875	09/30/2025	03/30/2026	Rahul Chavli	None	None	Jagrut Upadhyay
								09/30/2025

FROM 1.00000ml of S12577 + 19.00000ml of E3973 = Final Quantity: 20.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3492	8270-SIM-Spike 0.4 PPM	SP6876	09/30/2025	11/30/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								09/30/2025

FROM 0.00080ml of S11652 + 0.01000ml of S13120 + 0.02000ml of S12560 + 0.02000ml of S13097 + 0.02000ml of S13245 + 49.92920ml of E3972 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3339	8270 sim calibration stock 10ppm (CPI)	SP6896	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.03350ml of S10105 + 0.05000ml of S11497 + 0.12500ml of S11828 + 0.12500ml of S12116 + 0.25000ml of S12273 + 0.25000ml of S13058 + 24.16650ml of E3980 = Final Quantity: 25.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3361	8270-SIM MDL-5PPM CALIBRATION SOLUTION	SP6897	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.50000ml of E3980 + 0.01000ml of SP6830 + 0.50000ml of SP6896 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3341	8270-SIM MDL-3.2PPM CALIBRATION SOLUTION	SP6898	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.68000ml of E3980 + 0.01000ml of SP6830 + 0.32000ml of SP6896 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3344	8270-SIM MDL-1.6PPM CALIBRATION SOLUTION	SP6900	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.84000ml of E3980 + 0.01000ml of SP6830 + 0.16000ml of SP6896 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3342	8270-SIM MDL-0.8PPM CALIBRATION SOLUTION	SP6901	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.92000ml of E3980 + 0.01000ml of SP6830 + 0.08000ml of SP6896 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3343	8270-SIM MDL-0.4PPM CALIBRATION SOLUTION	SP6902	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.96000ml of E3980 + 0.01000ml of SP6830 + 0.04000ml of SP6896 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3345	8270-SIM MDL-0.2PPM CALIBRATION SOLUTION	SP6903	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.50000ml of E3980 + 0.01000ml of SP6830 + 0.50000ml of SP6902 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3346	8270-SIM MDL-0.1PPM CALIBRATION SOLUTION	SP6904	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.75000ml of E3980 + 0.01000ml of SP6830 + 0.25000ml of SP6902 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3355	8270-SIM MDL-3.2PPM CALIBRATION STOCK SOL- 2ND	SP6905	10/27/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.00630ml of S12201 + 0.01280ml of S12905 + 0.03200ml of S11828 + 0.03200ml of S13122 + 0.06400ml of S12563 + 0.06400ml of S13249 + 0.06400ml of S13299 + 19.72490ml of E3980 = Final Quantity: 20.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3356	8270-SIM MDL-0.4PPM CALIBRATION SOL ICV-2ND	SP6906	10/27/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
SOURCE FROM 0.87500ml of E3980 + 0.01000ml of SP6830 + 0.12500ml of SP6905 = Final Quantity: 1.010 ml								

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	07/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A2862010	12/13/2025	06/13/2025 / Rajesh	02/28/2025 / Rajesh	E3942

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25C1262005	04/16/2026	09/15/2025 / Riteshkumar	09/15/2025 / Riteshkumar	E3973

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25C1262005	04/16/2026	10/10/2025 / RUPESH	10/10/2025 / RUPESH	E3980

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	12/19/2025	06/19/2025 / Jagrut	12/09/2021 / Christian	S10105

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	04/22/2026	10/22/2025 / Jagrut	08/11/2023 / Yogesh	S11497

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0201728	01/30/2026	07/30/2025 / Rahul	11/09/2023 / Yogesh	S11652

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	33913 / SOM01.0 SIM Analysis Standard (Surrogate), 2000 PPM	A0201976	12/09/2025	06/09/2025 / Jagrut	11/21/2023 / rahul	S11828

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	04/22/2026	10/22/2025 / Jagrut	03/08/2024 / Rahul	S12116

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml, methanol, 5ml/ ampul	A0206206	03/22/2026	09/22/2025 / Jagrut	03/15/2024 / Rahul	S12201

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	520963	12/25/2025	06/25/2025 / Jagrut	05/24/2024 / Rahul	S12273

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	03/22/2026	09/22/2025 / Jagrut	07/23/2024 / RAHUL	S12560

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	04/16/2026	10/16/2025 / Jagrut	07/23/2024 / RAHUL	S12563

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH ₂ Cl ₂ , 1mL,	A0212955	06/30/2027	03/31/2025 / Rahul	08/01/2024 / Rahul	S12577

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH ₂ Cl ₂ , 1mL	A0212266	12/16/2025	06/16/2025 / anahy	09/20/2024 / anahy	S12670

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0216785	03/22/2026	09/22/2025 / Jagrut	12/09/2024 / anahy	S12905

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	531243	12/25/2025	06/25/2025 / Jagrut	01/16/2025 / anahy	S13058

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0221014	11/30/2025	09/03/2025 / Jagrut	05/20/2025 / Rahul	S13097

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0218894	02/25/2026	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13120

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0218894	04/16/2026	10/16/2025 / Jagrut	05/20/2025 / Rahul	S13122

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0228451	03/23/2026	09/22/2025 / Jagrut	08/06/2025 / Rahul	S13245

[CS 4978-1]

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0228451	04/27/2026	10/27/2025 / Jagrut	08/06/2025 / Rahul	S13249

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0229652	04/27/2026	10/27/2025 / Jagrut	10/15/2025 / rahul	S13299

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: **Storage:** **Solvent:** **Exp. Date:** **Description:**
Z-112090 440246 $\leq -10^{\circ}\text{C}$ Methylene Chloride 2/16/2026 CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL
-04

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d ₄	93951-73-6	99.3	248.12.7P	7487 $\pm$ 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 $\pm$ 17.26
phenol-d ₆	13127-88-3	99.9	949.120.8P	7481 $\pm$ 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 $\pm$ 17.17

Received on

02/25/21

by
CG

S9236
to

S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certified By: _____

Erica Castiglione
Chemist



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	$\leq 0.005 \%$	$< 0.005 \%$	PASS
Chloride	$\leq 0.005 \%$	0.002 %	PASS
Heavy Metals	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Iron	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Magnesium	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Mercury	$\leq 0.1 \text{ ppm}$	$< 0.1 \text{ ppm}$	PASS
Nickel	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Nitrogen Compounds	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Phosphate	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Potassium	$\leq 0.02 \%$	$< 0.02 \%$	PASS
Purity	$\geq 97.0 \%$	99.2 %	PASS
Sodium Carbonate	$\leq 1.0 \%$	0.5 %	PASS
Sulfate	$\leq 0.003 \%$	$< 0.003 \%$	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.


This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.





**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 25A2862010
Manufactured Date: 2024-12-18
Expiration Date: 2026-03-19
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	99.9 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μ eq/g)	≤ 0.3	<0.1
Chloride (Cl)	≤ 10 ppm	<5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	<0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3942

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date: 2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 25C1262005
Manufactured Date: 2025-01-15
Expiration Date: 2026-04-16
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

REC. on 10/10/25
RJ

E3980

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™



MG157
MS

Material No.: 9673-33

Batch No.: 24I1262013

Manufactured Date: 2024-08-07

Retest Date: 2029-08-06

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO ₂)	<= 2 ppm	<2 ppm
Ammonium (NH ₄)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO ₃)	<= 0.2 ppm	0.1 ppm
Phosphate (PO ₄)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium



Material No.: 9673-33
Batch No.: 24I1262013

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 1.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	1.1 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	< 1.0 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110094-02	506889	≤ -10 °C	Methylene Chloride	7/25/2028	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d ₄	2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl	321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5	4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14	1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 } Y.P.
↓
511498 } 08/11/2023

*Not a certified value

Certified By: Shane C. Tipton
Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872 **Lot No.:** A0201728

Description : Custom Pentachlorophenol Standard

Custom Pentachlorophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

511649
↓
511658
Y.P.
11/13/23

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP230530RSR	99%	25,000.0 µg/mL	+/- 777.0837

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Josh McCloskey - Operations Technician I

Date Mixed: 05-Sep-2023

Balance: B251644995

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 33913 **Lot No.:** A0201976

Description : SOM01.0 SIM Analysis Standard
SOM01.0 SIM Analysis Standard 2000µg/mL, Methylene chloride, 1mL /ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S11828
↓
S11832 } RC/
11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Methylnaphthalene-d10	7297-45-2	EF-135	98%	2,015.9 µg/mL	+/- 90.8098
2	Fluoranthene-d10	93951-69-0	PR-32557	99%	2,020.0 µg/mL	+/- 90.9963

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 13-Sep-2023

Balance Serial # B442140311


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 28-Sep-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane	123-91-1	100	223.1.3P	1997 ± 57.08

512112 } RC/
↓
912116 } 03/08/24

*Not a certified value

Certified By: _____

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206
Description : Acid Surrogate Mix (4/89 SOW)
Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : January 31, 2032 **Storage:** 10°C or colder
Ship: Ambient

512187 } RC/
↓
512206 } 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope A. Riglin

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024

Balance Serial # 1128360905

Christie Mills

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 4

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	520963	≤ -10 °C	Methylene Chloride	10/10/2028	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1010 ± 9.89
acenaphthylene	208-96-8	97.6	14.290.1P	1014 ± 9.93
aniline	62-53-3	99.97	64.1.4P	1001 ± 9.8
anthracene	120-12-7	99.5	15.7.1P	999.6 ± 9.79
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 9.8
benzo[a]anthracene	56-55-3	100	16.7.3P	1007 ± 9.86
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1011 ± 14.11
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 10.96
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 13.95
benzo[a]pyrene	50-32-8	97	20.286.2P	999.9 ± 22.24
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 9.82
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	1000 ± 14.69
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1003 ± 13.89
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.15P	999.4 ± 14.68
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 9.8
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 17.03
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1.1P	1000 ± 13.85
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 16.79
carbazole	86-74-8	99.4	239.7.2P	1000 ± 9.8

*Not a certified value

512270 } RC/
↓
512274 } 05/24/24

Kerry Kane

Certified By: _____

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.

Certificate of Analysis

Page 2 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
4-chloroaniline	106-47-8	100	66.7.1P	1000 ± 9.79
4-chlorophenylphenyl ether	7005-72-3	98	37.158.2P	1001 ± 17.07
4-chloro-3-methylphenol	59-50-7	99	102.1.2P	1006 ± 17.16
2-chloronaphthalene	91-58-7	99.9	42.7.6P	1000 ± 9.79
2-chlorophenol	95-57-8	99.8	103.7.1P	1007 ± 13.96
chrysene	218-01-9	96	21.286.2P	998.4 ± 12.85
dibenz[a,h]anthracene	53-70-3	99.44	22.286.3P	1000 ± 9.74
dibenzofuran	132-64-9	100	67.7.2.1P	1002 ± 9.77
di-n-butyl phthalate	84-74-2	99.84	40.286.1P	1007 ± 24.48
1,2-dichlorobenzene	95-50-1	99.8	43.7.1P	1000 ± 9.79
1,3-dichlorobenzene	541-73-1	99.5	44.1.3P	999.4 ± 9.79
1,4-dichlorobenzene	106-46-7	99.9	45.29.2P	1000 ± 9.79
2,4-dichlorophenol	120-83-2	99.6	104.7.1.1P	1005 ± 13.93
diethyl phthalate	84-66-2	99.8	38.7.1P	1011 ± 14
2,4-dimethylphenol	105-67-9	99.6	105.7.1.1P	1009 ± 13.98
dimethyl phthalate	131-11-3	99.9	39.9.2P	996.5 ± 13.8
1,2-dinitrobenzene	528-29-0	99.86	86.7.3.1P	999.5 ± 9.75
1,3-dinitrobenzene	99-65-0	100	313.7.2P	998 ± 9.79
1,4-dinitrobenzene	100-25-4	100	907.7.1P	999.5 ± 9.8
2,4-dinitrophenol	51-28-5	99.9	106.1.6DP	1002 ± 13.89
2,4-dinitrotoluene	121-14-2	100	87.7.3P	999.8 ± 13.85
2,6-dinitrotoluene	606-20-2	99.4	88.7.2.1P	999.6 ± 13.85
di-n-octyl phthalate	117-84-0	99.1	41.7.5P	991.6 ± 13.74
diphenylamine	122-39-4	100	78.1.6P	998 ± 13.79
2,3,5,6-tetrachlorophenol	935-95-5	97	1112.286.1P	1004 ± 14.02
fluoranthene	206-44-0	98.6	23.7.4P	999.6 ± 9.79
fluorene	86-73-7	98.4	24.7.1P	999.7 ± 9.79

*Not a certified value

Kerry Kane

Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 3 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
hexachlorobenzene	118-74-1	99	46.158.4P	999.9 ± 13.96
hexachlorobutadiene	87-68-3	97.4	47.1.4P	1000 ± 9.79
hexachlorocyclopentadiene	77-47-4	99.2	48.2.2P	1001 ± 9.8
hexachloroethane	67-72-1	99.9	49.1.4P	1003 ± 9.82
indeno[1,2,3-cd]pyrene	193-39-5	98	25.286.4P	999.4 ± 22.23
isophorone	78-59-1	98.9	90.1.4P	999.9 ± 13.85
2-methyl-4,6-dinitrophenol	534-52-1	99.6	107.421.2DP	991 ± 24.09
1-methylnaphthalene	90-12-0	97.1	249.7.5P	999.2 ± 13.95
2-methylnaphthalene	91-57-6	97.4	68.7.2P	1006 ± 22.38
2-methylphenol	95-48-7	99.6	114.7.3P	1001 ± 13.87
3-methylphenol	108-39-4	99.1	115.7.4P	499.7 ± 6.92
4-methylphenol	106-44-5	99.5	116.7.1P	501.2 ± 6.94
naphthalene	91-20-3	99.8	26.9.1P	1018 ± 9.97
2-nitroaniline	88-74-4	99.7	69.29.1P	999.6 ± 9.79
3-nitroaniline	99-09-2	100	70.7.3P	1000 ± 9.74
4-nitroaniline	100-01-6	99.7	71.29.1P	1001 ± 9.8
nitrobenzene	98-95-3	100	94.7.1P	1000 ± 13.85
2-nitrophenol	88-75-5	99.1	108.29.1P	996.5 ± 13.81
4-nitrophenol	100-02-7	100	109.7.1P	1000 ± 13.82
N-nitrosodimethylamine	62-75-9	99.5	57.3.19P	998.5 ± 14.67
N-nitrosodi-n-propylamine	621-64-7	99.8	59.286.1P	996.8 ± 17
pentachlorophenol	87-86-5	99	110.1.7P	1004 ± 13.92
phenanthrene	85-01-8	99.7	27.1.5P	999 ± 12.87
phenol	108-95-2	100	112.7.1P	998.5 ± 13.8
pyrene	129-00-0	99.2	28.9.2P	998.9 ± 9.78
pyridine	110-86-1	100	101.24.1P	999 ± 9.73
2,3,4,6-Tetrachlorophenol	58-90-2	91.8	120.421.1P	996.5 ± 13.92

*Not a certified value

Kerry Kane

Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 4 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	999.6 ± 9.79
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	999.5 ± 13.85
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	996 ± 13.8

*Not a certified value



Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 } RC/
↓
S12568 } 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31615 **Lot No.:** A0212955

Description : GC/MS Tuning Mixture
GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2027 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,004.5 µg/mL	+/- 44.8902
2	DFTPP (Decafluorotriphenylphosphine)	5074-71-5	Q117-147	99%	1,004.5 µg/mL	+/- 44.8902
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 44.9572
4	4,4'-DDT	50-29-3	S240530RSR	97%	1,000.1 µg/mL	+/- 44.6922

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12577
↓
S12579
} RC
8/2/24

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
10 ml/min.


Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Ethan Winiarski - Operations Tech I

Date Mixed: 19-Jun-2024 Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 26-Jun-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0212266

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,000.6 µg/mL	+/- 90.1075
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,000.3 µg/mL	+/- 90.0925
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,000.4 µg/mL	+/- 90.1000
4	Phenanthrene-d10	1517-22-2	PR-34099	99%	2,000.5 µg/mL	+/- 90.1037
5	Chrysene-d12	1719-03-5	PR-33506	99%	2,000.7 µg/mL	+/- 90.1112
6	Perylene-d12	1520-96-3	PR-33205	99%	2,000.6 µg/mL	+/- 90.1075

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12645
↓
S12674 } AC
10/1/24



5580 Skyline Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	531243	≤ -10 °C	Methylene Chloride	1/2/2030	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.4P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	993.8 ± 5.78
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

513057 } AC
↓
513061 } 1/16/25

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____

Melissa Workoff

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0221014

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

513088 }
↓
513117 } 5/20/25.

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBR4811	99%	1,005.2 µg/mL	+/- 36.5730
2	N-Nitrosodimethylamine	62-75-9	S241226RSR	99%	1,005.5 µg/mL	+/- 36.5848
3	Phenol	108-95-2	MKCK1120	99%	1,005.3 µg/mL	+/- 36.5780
4	Aniline	62-53-3	X22F726	99%	1,005.4 µg/mL	+/- 36.5816
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,004.7 µg/mL	+/- 36.5562
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.3 µg/mL	+/- 36.5766
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.6 µg/mL	+/- 36.5530
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,006.3 µg/mL	+/- 36.6125
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,005.5 µg/mL	+/- 36.5853
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,004.6 µg/mL	+/- 36.5507
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,004.8 µg/mL	+/- 36.5575
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,005.4 µg/mL	+/- 36.5803
13	3-Methylphenol (m-cresol)	108-39-4	STBL3873	99%	502.7 µg/mL	+/- 18.2888
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	502.6 µg/mL	+/- 18.2869
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,004.6 µg/mL	+/- 36.5502
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.6 µg/mL	+/- 36.5530
17	Nitrobenzene	98-95-3	10224044	99%	1,005.3 µg/mL	+/- 36.5780

18	Isophorone	78-59-1	MKCR3249	99%	1,004.6	µg/mL	+/- 36.5511	
19	2-Nitrophenol	88-75-5	RP230710	99%	1,005.7	µg/mL	+/- 36.5916	
20	2,4-Dimethylphenol	105-67-9	DIRAF	99%	1,004.9	µg/mL	+/- 36.5612	
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,004.3	µg/mL	+/- 36.5402	1
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,004.7	µg/mL	+/- 36.5571	2
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,004.6	µg/mL	+/- 36.5516	3
24	Naphthalene	91-20-3	STBL1057	99%	1,006.8	µg/mL	+/- 36.6335	4
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,005.9	µg/mL	+/- 36.5980	5
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,004.1	µg/mL	+/- 36.5328	6
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,005.4	µg/mL	+/- 36.5793	7
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,006.3	µg/mL	+/- 36.6144	8
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269	9
30	Hexachlorocyclopentadiene	77-47-4	099063P13G	99%	1,005.9	µg/mL	+/- 36.5975	10
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,004.7	µg/mL	+/- 36.5566	11
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.7	µg/mL	+/- 36.5571	12
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,005.3	µg/mL	+/- 36.5757	13
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.8	µg/mL	+/- 36.5584	14
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,004.5	µg/mL	+/- 36.5475	15
36	Acenaphthylene	208-96-8	214935V18H	95%	1,000.1	µg/mL	+/- 36.3888	16
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,004.9	µg/mL	+/- 36.5616	17
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,004.5	µg/mL	+/- 36.5489	18
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,005.2	µg/mL	+/- 36.5734	
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,005.9	µg/mL	+/- 36.6003	
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847	
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5525	
43	2,4-Dinitrophenol	51-28-5	D240927RSR	99%	1,005.0	µg/mL	+/- 36.5657	
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,005.4	µg/mL	+/- 36.5803	
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,005.4	µg/mL	+/- 36.5803	
46	4-Nitrophenol	100-02-7	20241120-1-AN	99%	1,005.7	µg/mL	+/- 36.5930	
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,004.9	µg/mL	+/- 36.5616	
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP240523RSR	99%	1,005.2	µg/mL	+/- 36.5739	
49	Fluorene	86-73-7	10246250	98%	1,005.8	µg/mL	+/- 36.5948	
50	4-Chlorophenyl phenyl ether	7005-72-3	002531K02D	99%	1,004.8	µg/mL	+/- 36.5584	
51	Diethylphthalate	84-66-2	STBL3611	99%	1,005.3	µg/mL	+/- 36.5762	
52	4-Nitroaniline	100-01-6	RP240830RSR	99%	1,005.5	µg/mL	+/- 36.5844	
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,004.8	µg/mL	+/- 36.5589	

54	Diphenylamine	122-39-4	MKCT1512	99%	1,005.2	µg/mL	+/- 36.5725
55	Azobenzene	103-33-3	BCCL3292	99%	1,004.7	µg/mL	+/- 36.5566
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,005.6	µg/mL	+/- 36.5875
57	Hexachlorobenzene	118-74-1	15828800	99%	1,005.3	µg/mL	+/- 36.5789
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.4	µg/mL	+/- 36.5825
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.8	µg/mL	+/- 36.5584
60	Anthracene	120-12-7	MKCW9141	99%	1,005.5	µg/mL	+/- 36.5834
61	Carbazole	86-74-8	15630800	99%	1,005.3	µg/mL	+/- 36.5789
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,005.5	µg/mL	+/- 36.5848
63	Fluoranthene	206-44-0	A0458721	99%	1,005.4	µg/mL	+/- <0.0001
64	Pyrene	129-00-0	BCCL8032	99%	1,005.2	µg/mL	+/- 36.5734
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,004.8	µg/mL	+/- 36.5584
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCR8567	99%	1,004.5	µg/mL	+/- 36.5480
67	Benz(a)anthracene	56-55-3	I60012022BAA	99%	1,005.5	µg/mL	+/- 36.5844
68	Chrysene	218-01-9	RP241212RSR	99%	1,005.5	µg/mL	+/- 36.5848
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,004.7	µg/mL	+/- 36.5548
70	Di-n-octyl phthalate	117-84-0	15817300	99%	1,004.8	µg/mL	+/- 36.5607
71	Benzo(b)fluoranthene	205-99-2	022013B	99%	1,005.1	µg/mL	+/- 36.5716
72	Benzo(k)fluoranthene	207-08-9	012022K	98%	1,004.6	µg/mL	+/- 36.5511
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,004.2	µg/mL	+/- 36.5377
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,004.1	µg/mL	+/- 36.5337
75	Dibenz(a,h)anthracene	53-70-3	712061504-1-1	99%	1,005.7	µg/mL	+/- 36.5930
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,005.4	µg/mL	+/- 36.5814

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:


100 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Penelope Riglin - Operations Tech I

Date Mixed: 12-Jan-2025

Balance Serial # 1128360905


Dillan Murphy - Operations Technician I

Date Passed: 21-Jan-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0218894
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : November 30, 2029 **Storage:** 0°C or colder
Ship: Ambient

513118 } RC/
↓
513147 } 5/20/25.

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBQ1693	99%	2,002.4 µg/mL	+/- 24.9202

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 07-Nov-2024

Balance Serial # 1128360905

Dillan Murphy - Operations Technician I

Date Passed: 11-Nov-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0228451

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2027 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

S13239 } RC/
↓
S13268 } 08/06/25

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S250226RSR	99%	1,006.0 µg/mL	+/- 23.0947
2	Atrazine	1912-24-9	5FYWL	99%	1,007.0 µg/mL	+/- 23.1176
3	Benzidine	92-87-5	S250227ECS	99%	1,003.0 µg/mL	+/- 23.0258
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,003.0 µg/mL	+/- 23.0258

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%


Tom Suckar - Mix Technician

Date Mixed: 01-Aug-2025

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0229652

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

513299 } RC/
↓ 10/17/25
513328 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBR4811	99%	1,004.9 µg/mL	+/- 36.5621
2	N-Nitrosodimethylamine	62-75-9	S250717RSR	99%	1,004.5 µg/mL	+/- 36.5484
3	Phenol	108-95-2	MKCT5446	99%	1,003.5 µg/mL	+/- 36.5120
4	Aniline	62-53-3	X22F726	99%	1,004.8 µg/mL	+/- 36.5575
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,003.6 µg/mL	+/- 36.5166
6	2-Chlorophenol	95-57-8	STBK4742	99%	1,004.5 µg/mL	+/- 36.5484
7	1,3-Dichlorobenzene	541-73-1	STBL1702	99%	1,003.8 µg/mL	+/- 36.5211
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.8 µg/mL	+/- 36.5211
9	Benzyl alcohol	100-51-6	094986W07G	99%	1,003.8 µg/mL	+/- 36.5211
10	1,2-Dichlorobenzene	95-50-1	SHBR4446	99%	1,004.9 µg/mL	+/- 36.5621
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,004.3 µg/mL	+/- 36.5393
12	2,2'-oxybis(1-chloropropane)	108-60-1	RP250904RSR	99%	1,005.0 µg/mL	+/- 36.5666
13	3-Methylphenol (m-cresol)	108-39-4	STBL3873	99%	502.4 µg/mL	+/- 18.2788
14	4-Methylphenol (p-cresol)	106-44-5	SHBQ7653	99%	502.3 µg/mL	+/- 18.2742
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,004.9 µg/mL	+/- 36.5621
16	Hexachloroethane	67-72-1	DAXRI	99%	1,003.9 µg/mL	+/- 36.5257
17	Nitrobenzene	98-95-3	10224044	99%	1,003.9 µg/mL	+/- 36.5257

18	Isophorone	78-59-1	MKCR3249	99%	1,004.8	µg/mL	+/- 36.5575
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.6	µg/mL	+/- 36.5166
20	2,4-Dimethylphenol	105-67-9	DIRAF	99%	1,004.1	µg/mL	+/- 36.5348
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,004.5	µg/mL	+/- 36.5484
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,004.8	µg/mL	+/- 36.5575
23	1,2,4-Trichlorobenzene	120-82-1	SHBR1701	99%	1,004.0	µg/mL	+/- 36.5302
24	Naphthalene	91-20-3	STBL1057	99%	1,004.0	µg/mL	+/- 36.5302
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,004.5	µg/mL	+/- 36.5484
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,002.7	µg/mL	+/- 36.4816
27	4-Chloro-3-methylphenol	59-50-7	BCCK5906	99%	1,004.3	µg/mL	+/- 36.5393
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,000.5	µg/mL	+/- 36.4029
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3847
30	Hexachlorocyclopentadiene	77-47-4	099063P13G	99%	1,004.0	µg/mL	+/- 36.5302
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,003.5	µg/mL	+/- 36.5120
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,003.6	µg/mL	+/- 36.5152
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,003.6	µg/mL	+/- 36.5166
34	2-Nitroaniline	88-74-4	RP250625RSR	99%	1,003.8	µg/mL	+/- 36.5211
35	1,4-Dinitrobenzene	100-25-4	RP250401RSR	99%	1,004.6	µg/mL	+/- 36.5530
36	Acenaphthylene	208-96-8	214935V18H	95%	1,000.4	µg/mL	+/- 36.3974
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,004.8	µg/mL	+/- 36.5575
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,003.5	µg/mL	+/- 36.5120
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,004.6	µg/mL	+/- 36.5530
40	1,2-Dinitrobenzene	528-29-0	RP250203RSR	99%	1,002.6	µg/mL	+/- 36.4802
41	Acenaphthene	83-32-9	MKCV8166	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,003.8	µg/mL	+/- 36.5211
43	2,4-Dinitrophenol	51-28-5	D240927RSR	99%	1,004.1	µg/mL	+/- 36.5348
44	Dibenzofuran	132-64-9	MKCW3845	99%	1,003.6	µg/mL	+/- 36.5166
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,003.8	µg/mL	+/- 36.5211
46	4-Nitrophenol	100-02-7	20241120-1-AN	99%	1,003.0	µg/mL	+/- 36.4938
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,003.8	µg/mL	+/- 36.5211
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP250724RSR	99%	1,003.1	µg/mL	+/- 36.4984
49	Fluorene	86-73-7	10246250	98%	1,002.9	µg/mL	+/- 36.4905
50	4-Chlorophenyl phenyl ether	7005-72-3	002531K02D	99%	1,004.3	µg/mL	+/- 36.5393
51	Diethylphthalate	84-66-2	223219R19C	99%	1,004.8	µg/mL	+/- 36.5575
52	4-Nitroaniline	100-01-6	RP240830RSR	99%	1,004.0	µg/mL	+/- 36.5302
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S250805RSR	99%	1,004.8	µg/mL	+/- 36.5575

54	Diphenylamine	122-39-4	MKCT1512	99%	1,004.9	µg/mL	+/- 36.5621
55	Azobenzene	103-33-3	BCCL3292	99%	1,003.9	µg/mL	+/- 36.5257
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.9	µg/mL	+/- 36.5257
57	Hexachlorobenzene	118-74-1	16302300	99%	1,004.5	µg/mL	+/- 36.5484
58	Pentachlorophenol	87-86-5	RP240411RSR	99%	1,004.9	µg/mL	+/- 36.5621
59	Phenanthrene	85-01-8	MKCV8193	99%	1,003.8	µg/mL	+/- 36.5211
60	Anthracene	120-12-7	MKCW9141	99%	1,003.3	µg/mL	+/- 36.5029
61	Carbazole	86-74-8	15821400	99%	1,003.1	µg/mL	+/- 36.4984
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,004.6	µg/mL	+/- 36.5530
63	Fluoranthene	206-44-0	A0458721	99%	1,004.8	µg/mL	+/- 36.5575
64	Pyrene	129-00-0	BCCL8032	99%	1,003.8	µg/mL	+/- 36.5211
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,004.9	µg/mL	+/- 36.5621
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,003.4	µg/mL	+/- 36.5075
67	Benz(a)anthracene	56-55-3	I80012022BAA	99%	1,004.3	µg/mL	+/- 36.5393
68	Chrysene	218-01-9	RP250815RSR	99%	1,003.9	µg/mL	+/- 36.5257
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,004.5	µg/mL	+/- 36.5484
70	Di-n-octyl phthalate	117-84-0	16197600	99%	1,004.4	µg/mL	+/- 36.5439
71	Benzo(b)fluoranthene	205-99-2	0225BF	99%	1,005.4	µg/mL	+/- 36.5803
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,004.5	µg/mL	+/- 36.5484
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,002.4	µg/mL	+/- 36.4726
74	Indeno(1,2,3-cd)pyrene	193-39-5	17-YMK-40-2	99%	1,004.0	µg/mL	+/- 36.5302
75	Dibenz(a,h)anthracene	53-70-3	712215450-1-1	99%	1,003.5	µg/mL	+/- 36.5120
76	Benzo(g,h,i)perylene	191-24-2	RP250501RSR	98%	1,003.0	µg/mL	+/- 36.4949

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 04-Sep-2025

Balance Serial # B345965662


Brittany Federinko - Operations Tech II

Date Passed: 15-Sep-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18



CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 78-8922

www.chemtech.net

Chemtech Project Number:

23526

COC Number:

CLIENT INFORMATION

COMPANY: Tetra Tech

ADDRESS: 4433 Corporation Ln, Suite 300

CITY: Virginia Beach STATE: VA ZIP: 23462

ATTENTION: Ernie Wu

PHONE: 757-466-4901

FAX: 757-461-4148

PROJECT INFORMATION

PROJECT NAME: NWIRP Bethpage

PROJECT #: 112G08005-WE13 LOCATION: RW7B

PROJECT MANAGER: Ernie Wu

E-MAIL: ernie.wu@tetrattech.com

PHONE: 757-466-4901

FAX: 757-461-4148

BILLING INFORMATION

BILL TO:

PO#

ADDRESS:

CITY:

STATE: ZIP:

ATTENTION:

PHONE:

ANALYSIS

1,4-Dioxane SW846 8270
SIM

1 2 3 4 5 6 7 8 9

PRESERVATIVES

COMMENTS

<-- Specify Preservatives
A-HCl B-HNO3
C-H2SO4 D-NaOH
E-ICE F-Other

FAX: 10 DAYS*

HARD COPY: 10 DAYS*

EDD 10 DAYS*

* TO BE APPROVED BY CHEMTECH

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

☐ RESEULTS ONLY☐ USEPA CLP☐ RESULTS + QC☐ New York State ASP "B"☐ New Jersey REDUCED☐ New York State ASP "A"☐ New Jersey CLP☐ Other☐ EDD FormatCHEMTECH
SAMPLE
IDPROJECT
SAMPLE IDENTIFICATIONSAMPLE
MATRIXSAMPLE
TYPE
COMP GRABSAMPLE
COLLECTION

DATE TIME

of Bottles

1 2 3 4 5 6 7 8 9

1. RW7-SP100-20251030

GW

X

10/30/25

12:15

1

X

2. RW7-SP201-20251030

GW

X

10/30/25

12:17

1

X

3. RW7-SP302-20251030

GW

X

10/30/25

12:27

1

X

4. RW7-SP303-20251030

GW

X

10/30/25

12:29

1

X

5.

6.

7.

8.

9.

10.

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER

DATE/TIME

RECEIVED BY

1.

10/31/25/1500

1.

11-3-25
0700

RELINQUISHED BY

DATE/TIME

RECEIVED BY

2.

DATE/TIME

2.

RELINQUISHED BY

DATE/TIME

RECEIVED FOR LAB BY

3.

DATE/TIME

3.

Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant☐ Cooler Temp 2-3°C

MeOH extraction requires an additional 4oz. Jar for percent solid

☐ Ice in Cooler?:

Comments:

Page of

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
CHEMTECH: ☐ Picked Up ☐ Overnight

Shipment Complete

☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT

YELLOW - CHEMTECH COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038114.D
 Acq On : 28 Oct 2025 14:49
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Oct 28 17:31:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	9110	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	27131	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	14920	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	28506	0.400	ng	0.00
29) Chrysene-d12	21.375	240	23535	0.400	ng	0.00
35) Perylene-d12	23.683	264	20307	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	16552	0.771	ng	0.00
5) Phenol-d6	6.937	99	19989	0.765	ng	0.00
8) Nitrobenzene-d5	8.929	82	16332	0.746	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	29806	0.769	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	4525	0.736	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	46843	0.784	ng	0.01
27) Fluoranthene-d10	19.220	212	55912	0.778	ng	0.00
31) Terphenyl-d14	19.819	244	37975	0.743	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	7707	0.819	ng	99
3) n-Nitrosodimethylamine	3.564	42	9598	0.792	ng	# 100
6) bis(2-Chloroethyl)ether	7.197	93	20444	0.796	ng	99
9) Naphthalene	10.626	128	58627	0.784	ng	99
10) Hexachlorobutadiene	10.925	225	9875	0.783	ng	# 98
12) 2-Methylnaphthalene	12.248	142	38674	0.776	ng	98
16) Acenaphthylene	14.152	152	52624	0.779	ng	99
17) Acenaphthene	14.494	154	36845	0.779	ng	100
18) Fluorene	15.489	166	49204	0.794	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	2898	0.689	ng	# 69
21) 4-Bromophenyl-phenylether	16.379	248	14571	0.780	ng	97
22) Hexachlorobenzene	16.503	284	16700	0.786	ng	99
23) Atrazine	16.652	200	10204	0.745	ng	99
24) Pentachlorophenol	16.838	266	6470	0.707	ng	99
25) Phenanthrene	17.223	178	70455	0.779	ng	100
26) Anthracene	17.322	178	60270	0.761	ng	100
28) Fluoranthene	19.253	202	79478	0.786	ng	100
30) Pyrene	19.615	202	78307	0.737	ng	100
32) Benzo(a)anthracene	21.357	228	64140	0.764	ng	100
33) Chrysene	21.411	228	71675	0.787	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	32963	0.745	ng	99
36) Indeno(1,2,3-cd)pyrene	26.039	276	65796	0.789	ng	99
37) Benzo(b)fluoranthene	23.031	252	67771	0.781	ng	94
38) Benzo(k)fluoranthene	23.031	252	67771	0.776	ng	# 93
39) Benzo(a)pyrene	23.578	252	53367	0.768	ng	# 90
40) Dibenzo(a,h)anthracene	26.054	278	50344	0.784	ng	95
41) Benzo(g,h,i)perylene	26.753	276	58155	0.793	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038114.D
Acq On : 28 Oct 2025 14:49
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Quant Time: Oct 28 17:31:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.775 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8







Tgt Ion:152 Resp: 9110
 Ion Ratio Lower Upper
 152 100
 150 157.1 122.3 183.5
 115 63.4 47.4 71.0



#2
 1,4-Dioxane
 Concen: 0.819 ng
 RT: 3.254 min Scan# 24
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion: 88 Resp: 7707
 Ion Ratio Lower Upper
 88 100
 43 30.6 24.3 36.5
 58 74.3 60.4 90.6









#3
 n-Nitrosodimethylamine
 Concen: 0.792 ng
 RT: 3.564 min Scan# 61
 Delta R.T. 0.007 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion: 42 Resp: 9598
 Ion Ratio Lower Upper
 42 100
 74 120.8 96.8 145.2
 44 8.8 5.3 7.9#



#4
 2-Fluorophenol
 Concen: 0.771 ng
 RT: 5.348 min Scan# 314
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion: 112 Resp: 16552
 Ion Ratio Lower Upper
 112 100
 64 55.7 45.4 68.0
 63 30.2 23.2 34.8









#5
Phenol-d6
Concen: 0.765 ng
RT: 6.937 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 19989
Ion Ratio Lower Upper
99 100
42 20.7 16.6 24.8
71 33.2 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.796 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion: 93 Resp: 20444
Ion Ratio Lower Upper
93 100
63 75.0 59.3 88.9
95 31.6 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	27131
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	6.3	5.2	7.8
68	5.2	4.1	6.1



#8
Nitrobenzene-d5
Concen: 0.746 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:	82	Resp:	16332
Ion	Ratio	Lower	Upper
82	100		
128	41.9	34.1	51.1
54	53.4	43.5	65.3









#9
Naphthalene
Concen: 0.784 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 58627
Ion Ratio Lower Upper
128 100
129 11.0 9.1 13.7
127 13.2 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.783 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:225 Resp: 9875
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.8 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.769 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 29806
Ion Ratio Lower Upper
152 100
151 21.2 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.776 ng
RT: 12.248 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:142 Resp: 38674
Ion Ratio Lower Upper
142 100
141 90.1 70.3 105.5
115 35.9 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 14920
Ion Ratio Lower Upper
164 100
162 104.0 83.0 124.4
160 51.9 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 0.736 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:330 Resp: 4525
Ion Ratio Lower Upper
330 100
332 95.3 76.6 114.8
141 41.9 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.784 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	172	Resp:	46843
Ion Ratio	Lower	Upper	
172	100		
171	34.3	27.8	41.6
170	22.4	18.1	27.1



#16
Acenaphthylene
Concen: 0.779 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:	152	Resp:	52624
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.8	23.8
153	12.9	10.2	15.2









#17
Acenaphthene
Concen: 0.779 ng
RT: 14.494 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:154 Resp: 36845
Ion Ratio Lower Upper
154 100
153 112.3 90.0 135.0
152 59.3 47.3 70.9



#18
Fluorene
Concen: 0.794 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:166 Resp: 49204
Ion Ratio Lower Upper
166 100
165 98.7 79.0 118.4
167 13.3 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:188 Resp: 28506

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.1 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.689 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49




Tgt Ion:198 Resp: 2898


Ion Ratio Lower Upper

198 100

51 69.5 88.9 133.3#

105 49.4 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.780 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	248	Resp:	14571
Ion Ratio	Lower	Upper	
248	100		
250	93.7	76.2	114.2
141	68.4	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.786 ng
RT: 16.503 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:	284	Resp:	16700
Ion Ratio	Lower	Upper	
284	100		
142	38.0	30.5	45.7
249	30.3	23.5	35.3









#23
 Atrazine
 Concen: 0.745 ng
 RT: 16.652 min Scan# 1645
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49




Instrument : BNA_N
 ClientSampleId : SSTDICC0.8


Tgt Ion:	200	Resp:	10204
Ion Ratio	Lower	Upper	
200	100		
173	26.2	21.4	32.2
215	46.0	36.7	55.1



#24
 Pentachlorophenol
 Concen: 0.707 ng
 RT: 16.838 min Scan# 1660
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion:	266	Resp:	6470
Ion Ratio	Lower	Upper	
266	100		
264	62.0	49.3	73.9
268	64.3	50.6	75.8









#25
 Phenanthrene
 Concen: 0.779 ng
 RT: 17.223 min Scan# 1691
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:	178	Resp:	70455
Ion Ratio	Lower	Upper	
178	100		
176	19.2	15.5	23.3
179	15.3	12.2	18.4



#26
 Anthracene
 Concen: 0.761 ng
 RT: 17.322 min Scan# 1699
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion:	178	Resp:	60270
Ion Ratio	Lower	Upper	
178	100		
176	18.6	14.9	22.3
179	15.2	11.8	17.8









#27
Fluoranthene-d10
Concen: 0.778 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
212	100		
106	16.0	12.6	19.0
104	9.0	7.1	10.7



#28
Fluoranthene
Concen: 0.786 ng
RT: 19.253 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion	Ratio	Lower	Upper
202	100		
101	11.8	9.4	14.2
203	17.0	13.4	20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:240 Resp: 23535
Ion Ratio Lower Upper
240 100
120 13.1 10.7 16.1
236 28.4 23.1 34.7



#30
Pyrene
Concen: 0.737 ng
RT: 19.615 min Scan# 2032
Delta R.T. 0.005 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:202 Resp: 78307
Ion Ratio Lower Upper
202 100
200 21.2 16.7 25.1
203 17.7 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.743 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244 Resp: 37975
 Ion Ratio Lower Upper
 244 100
 212 9.3 7.8 11.6
 122 18.4 15.2 22.8



#32
 Benzo(a)anthracene
 Concen: 0.764 ng
 RT: 21.357 min Scan# 2267
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion:228 Resp: 64140
 Ion Ratio Lower Upper
 228 100
 226 27.7 22.2 33.4
 229 19.6 15.8 23.8









#33
 Chrysene
 Concen: 0.787 ng
 RT: 21.411 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:	228	Resp:	71675
Ion Ratio	Lower	Upper	
228	100		
226	30.1	24.3	36.5
229	19.6	15.7	23.5



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.745 ng
 RT: 21.295 min Scan# 2260
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion:	149	Resp:	32963
Ion Ratio	Lower	Upper	
149	100		
167	26.8	21.1	31.7
279	3.1	2.9	4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.683 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:264 Resp: 20307
Ion Ratio Lower Upper
264 100
260 26.8 21.8 32.8
265 76.5 68.6 102.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.789 ng
RT: 26.039 min Scan# 3627
Delta R.T. 0.006 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:276 Resp: 65796
Ion Ratio Lower Upper
276 100
138 27.3 21.4 32.0
277 24.7 19.2 28.8









#37
Benzo(b)fluoranthene
Concen: 0.781 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.050 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 67771
Ion Ratio Lower Upper
252 100
253 23.4 20.9 31.3
125 14.2 14.1 21.1



#38
Benzo(k)fluoranthene
Concen: 0.776 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.003 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:252 Resp: 67771
Ion Ratio Lower Upper
252 100
253 23.4 21.2 31.8
125 14.2 14.6 22.0#








#39
Benzo(a)pyrene
Concen: 0.768 ng
RT: 23.578 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:252 Resp: 53367
Ion Ratio Lower Upper
252 100
253 24.4 23.1 34.7
125 17.2 18.6 28.0#



#40
Dibenzo(a,h)anthracene
Concen: 0.784 ng
RT: 26.054 min Scan# 3632
Delta R.T. 0.006 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:278 Resp: 50344
Ion Ratio Lower Upper
278 100
139 19.9 17.6 26.4
279 25.0 22.6 33.8





#41
Benzo(g,h,i)perylene
Concen: 0.793 ng
RT: 26.753 min Scan# 3869
Delta R.T. 0.003 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:276 Resp: 58155
Ion Ratio Lower Upper
276 100
277 24.0 19.7 29.5
138 24.9 20.2 30.4

