

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager : Ernie Wu

Order ID # Q3526

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

4 Water samples were received on 11/03/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-Semi Volatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for RW7-SP201-20251030 [Terphenyl-d14 - 140%], RW7-SP302-20251030 [Terphenyl-d14 - 148%] and RW7-SP303-20251030 [Terphenyl-d14 - 163%], The failure surrogates not associated with the client parameters list, therefore no corrective action was taken.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD for met criteria except for except for PB170381BSD for 1,4-Dioxane. %RPD was affected due to marginal difference in recoveries.

The Blank Spike Duplicate for {PB170381BSD} with File ID: BN038133.D met requirements for all compounds except for 1,4-Dioxane[65%], Failed marginally biased low, however, it is meeting QC criteria in the continues calibration.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____