ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

MATRIX: Water

ORDER ID: Q3526

Comments:

METH	IOD: 8270-Modified/3510			
		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			\checkmark
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			\checkmark
	The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		\checkmark	
7.	Surrogate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries were met for all analysis except for PB170381BL [2,4,6-Tribromophenol - 29%], RW7-SP201-20251030 [Terphenyl-d14 - 140%], RW7-SP302-20251030 [Terphenyl-d14 - 148%] and RW7-SP303-20251030 [Terphenyl-d14 - 163%], The failure surrogates not associated with the client parameters list, therefore no corrective action was taken.			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria The RPD for met criteria except for except for PB170381BSD for 1,4-Dioxane. %RPD was affected due to marginal difference in recoveries.		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The Blank Spike Duplicate for {PB170381BSD} with File ID: BN038133.D met requirements for all compounds except for 1,4-Dioxane[65%], Failed marginally biased low, however, it is meeting QC criteria in the continues calibration			
9.	Internal Standard Area/Retention Time Shift Meet Criteria			\checkmark

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

		NA	NO	YES
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
11				_
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
ADDIT	TIONAL COMMENTS:			
summ	poratory certifies that the all-electronic diskette deliverable exactly match the data ary forms (i.e. Form Is)." rm 6 is not included in the data package because the Initial Calibration was performed	using 7 point	2	
	t QT review data is reported in the Miscellaneous.	using / point		
QA RE	VIEW Date			