

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB170381BSD		SDG No.:	Q3526
Lab Sample ID:	PB170381BSD		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.26		1	0.070	0.20	0.20	ug/L	11/04/25 12:56	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.29			30 - 150		73%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.25			30 - 150		63%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.31			55 - 111		76%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.32			53 - 106		81%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.36			58 - 132		91%	SPK: 0.4		
INTERNAL STANDARDS		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	10800								
1146-65-2	Naphthalene-d8	30500								
15067-26-2	Acenaphthene-d10	15600								
1517-22-2	Phenanthrene-d10	28800								
1719-03-5	Chrysene-d12	18100								
1520-96-3	Perylene-d12	14900								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products