

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q3526	OrderDate:	11/3/2025 9:28:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	A12

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3526-01	RW7-SP100-2025103 0	Water			10/30/25			11/03/25
			SVOC-SIMGroup1	8270-Modified		11/03/25	11/04/25	
Q3526-02	RW7-SP201-2025103 0	Water			10/30/25			11/03/25
			SVOC-SIMGroup1	8270-Modified		11/03/25	11/04/25	
Q3526-03	RW7-SP302-2025103 0	Water			10/30/25			11/03/25
			SVOC-SIMGroup1	8270-Modified		11/03/25	11/04/25	
Q3526-04	RW7-SP303-2025103 0	Water			10/30/25			11/03/25
			SVOC-SIMGroup1	8270-Modified		11/03/25	11/04/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3526

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :	RW7-SP100-20251030							
Q3526-01	RW7-SP100-20251030	WATER	1,4-Dioxane	2.900	Q	0.07	0.2	0.2 ug/L
			Total Svoc :	2.90				
			Total Concentration:	2.90				



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q3526

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170381BL	PB170381BL	2-Methylnaphthalene-d10	0.4	0.33	83		30	150
		Fluoranthene-d10	0.4	0.35	88		30	150
		Nitrobenzene-d5	0.4	0.35	88		55	111
		2-Fluorobiphenyl	0.4	0.39	97		53	106
		Terphenyl-d14	0.4	0.46	114		58	132
PB170381BS	PB170381BS	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.30	76		30	150
		Nitrobenzene-d5	0.4	0.37	92		55	111
		2-Fluorobiphenyl	0.4	0.42	104		53	106
		Terphenyl-d14	0.4	0.48	121		58	132
PB170381BSD	PB170381BSD	2-Methylnaphthalene-d10	0.4	0.29	73		30	150
		Fluoranthene-d10	0.4	0.25	63		30	150
		Nitrobenzene-d5	0.4	0.31	76		55	111
		2-Fluorobiphenyl	0.4	0.32	81		53	106
		Terphenyl-d14	0.4	0.36	91		58	132
Q3526-01	RW7-SP100-20251030	2-Methylnaphthalene-d10	0.4	0.28	71		30	150
		Fluoranthene-d10	0.4	0.43	107		30	150
		Nitrobenzene-d5	0.4	0.28	70		55	111
		2-Fluorobiphenyl	0.4	0.31	77		53	106
		Terphenyl-d14	0.4	0.51	128		58	132
Q3526-02	RW7-SP201-20251030	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.31	76		55	111
		2-Fluorobiphenyl	0.4	0.31	77		53	106
		Terphenyl-d14	0.4	0.56	140	*	58	132
Q3526-03	RW7-SP302-20251030	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.33	81		55	111
		2-Fluorobiphenyl	0.4	0.36	90		53	106
		Terphenyl-d14	0.4	0.59	148	*	58	132
Q3526-04	RW7-SP303-20251030	2-Methylnaphthalene-d10	0.4	0.27	67		30	150
		Fluoranthene-d10	0.4	0.32	81		30	150
		Nitrobenzene-d5	0.4	0.27	68		55	111
		2-Fluorobiphenyl	0.4	0.28	69		53	106
		Terphenyl-d14	0.4	0.65	163	*	58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3526

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038132.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170381BS	1,4-Dioxane	0.4	0.33	ug/L	83				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3526

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038133.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Low	Limits	
								Qual			High	RPD
PB170381BSD	1,4-Dioxane	0.4	0.26	ug/L	65	24	*	*		70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB170381BL

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3526

Lab File ID: BN038131.D

Lab Sample ID: PB170381BL

Instrument ID: BNA_N

Date Extracted: 11/03/2025

Matrix: (soil/water) Water

Date Analyzed: 11/04/2025

Level: (low/med) LOW

Time Analyzed: 11:43

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB170381BS	PB170381BS	BN038132.D	11/04/2025
PB170381BSD	PB170381BSD	BN038133.D	11/04/2025
RW7-SP100-20251030	Q3526-01	BN038138.D	11/04/2025
RW7-SP201-20251030	Q3526-02	BN038139.D	11/04/2025
RW7-SP302-20251030	Q3526-03	BN038140.D	11/04/2025
RW7-SP303-20251030	Q3526-04	BN038141.D	11/04/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038110.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3526
DFTPP Injection Date: 10/28/2025
DFTPP Injection Time: 12:05

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	38.2
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	14.2
442	Greater than 50% of mass 198	85.3
443	15.0 - 24.0% of mass 442	17 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN038111.D	10/28/2025	13:00
SSTDICC0.2	SSTDICC0.2	BN038112.D	10/28/2025	13:36
SSTDICCC0.4	SSTDICCC0.4	BN038113.D	10/28/2025	14:13
SSTDICC0.8	SSTDICC0.8	BN038114.D	10/28/2025	14:49
SSTDICC1.6	SSTDICC1.6	BN038115.D	10/28/2025	15:25
SSTDICC3.2	SSTDICC3.2	BN038116.D	10/28/2025	16:02
SSTDICC5.0	SSTDICC5.0	BN038117.D	10/28/2025	16:38



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038129.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3526
DFTPP Injection Date: 11/04/2025
DFTPP Injection Time: 10:00

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	37.9
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	14.7
442	Greater than 50% of mass 198	95.1
443	15.0 - 24.0% of mass 442	19 (20) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038130.D	11/04/2025	10:40
PB170381BL	PB170381BL	BN038131.D	11/04/2025	11:43
PB170381BS	PB170381BS	BN038132.D	11/04/2025	12:19
PB170381BSD	PB170381BSD	BN038133.D	11/04/2025	12:56
RW7-SP100-20251030	Q3526-01	BN038138.D	11/04/2025	15:57
RW7-SP201-20251030	Q3526-02	BN038139.D	11/04/2025	16:34
RW7-SP302-20251030	Q3526-03	BN038140.D	11/04/2025	17:10
RW7-SP303-20251030	Q3526-04	BN038141.D	11/04/2025	17:46
SSTDCCC0.4EC	SSTDCCC0.4	BN038145.D	11/04/2025	20:12

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3526

Client ID : SSTDCCC0.4

Date Analyzed: 11/04/2025

Lab File ID: BN038130.D

Time Analyzed: 10:40

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8599	7.775	24973	10.57	13947	14.44
UPPER LIMIT	17198	8.275	49946	11.073	27894	14.941
LOWER LIMIT	4299.5	7.275	12486.5	10.073	6973.5	13.941
EPA SAMPLE NO.						
01 PB170381BL	8221	7.79	22221	10.59	11601	14.44
02 PB170381BS	11523	7.78	31801	10.57	16154	14.44
03 PB170381BSD	10823	7.78	30534	10.57	15616	14.43
04 RW7-SP100-20251030	8530	7.78	25854	10.57	14999	14.43
05 RW7-SP201-20251030	8635	7.78	25969	10.57	15192	14.43
06 RW7-SP302-20251030	8568	7.78	25578	10.57	14324	14.43
07 RW7-SP303-20251030	8535	7.78	25212	10.57	14257	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3526

Client ID: SSTDCCC0.4

Date Analyzed: 11/04/2025

Lab File ID: BN038130.D

Time Analyzed: 10:40

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	27056	17.186	17059	21.375	13607	23.686
UPPER LIMIT	54112	17.686	34118	21.875	27214	24.186
LOWER LIMIT	13528	16.686	8529.5	20.875	6803.5	23.186
EPA SAMPLE NO.						
01 PB170381BL	21781	17.20	15202	21.38	13241	23.70
02 PB170381BS	30417	17.19	16510	21.38	13194	23.69
03 PB170381BSD	28756	17.19	18053	21.38	14856	23.68
04 RW7-SP100-20251030	31102	17.19	25933	21.38	21372	23.68
05 RW7-SP201-20251030	28331	17.19	19063	21.38	14967	23.69
06 RW7-SP302-20251030	27505	17.19	16355	21.38	12980	23.68
07 RW7-SP303-20251030	26274	17.19	14750	21.38	12357	23.68

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/30/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	11/03/25
Client Sample ID:	RW7-SP100-20251030		SDG No.:	Q3526
Lab Sample ID:	Q3526-01		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	2.90	Q	1	0.070	0.20	0.20	ug/L	11/04/25 15:57	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.28			30 - 150		71%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.43			30 - 150		107%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.28			55 - 111		70%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.31			53 - 106		77%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.51			58 - 132		128%	SPK: 0.4		
INTERNAL STANDARDS		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	8530								
1146-65-2	Naphthalene-d8	25900								
15067-26-2	Acenaphthene-d10	15000								
1517-22-2	Phenanthrene-d10	31100								
1719-03-5	Chrysene-d12	25900								
1520-96-3	Perylene-d12	21400								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038138.D
Acq On : 04 Nov 2025 15:57
Operator : RC/JU
Sample : Q3526-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030

Quant Time: Nov 04 16:25:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8530	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	25854	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	14999	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	31102	0.400	ng	0.00
29) Chrysene-d12	21.375	240	25933	0.400	ng	# 0.00
35) Perylene-d12	23.680	264	21372	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2490	0.124	ng	0.00
5) Phenol-d6	6.937	99	1781	0.073	ng	0.00
8) Nitrobenzene-d5	8.929	82	5875	0.282	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	10503	0.284	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	2316	0.375	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	18588	0.309	ng	0.00
27) Fluoranthene-d10	19.220	212	33654	0.429	ng	0.00
31) Terphenyl-d14	19.819	244	28902	0.513	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	25922	2.941	ng	Qvalue 97
6) bis(2-Chloroethyl)ether	7.197	93	524	0.022	ng	98
34) Bis(2-ethylhexyl)phtha...	21.295	149	22629	0.464	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038138.D
Acq On : 04 Nov 2025 15:57
Operator : RC/JU
Sample : Q3526-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Quant Time: Nov 04 16:25:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030







Tgt Ion:152 Resp: 8530
Ion Ratio Lower Upper
152 100
150 155.7 122.3 183.5
115 59.4 47.4 71.0



#2
1,4-Dioxane
Concen: 2.941 ng
RT: 3.247 min Scan# 23
Delta R.T. -0.007 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion: 88 Resp: 25922
Ion Ratio Lower Upper
88 100
43 35.4 24.3 36.5
58 75.5 60.4 90.6









#4
2-Fluorophenol
Concen: 0.124 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion:	112	Resp:	2490
Ion	Ratio	Lower	Upper
112	100		
64	58.2	45.4	68.0
63	31.3	23.2	34.8



#5
Phenol-d6
Concen: 0.073 ng
RT: 6.937 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion:	99	Resp:	1781
Ion	Ratio	Lower	Upper
99	100		
42	19.5	16.6	24.8
71	40.9	25.4	38.2









#6
bis(2-Chloroethyl)ether
Concen: 0.022 ng
RT: 7.197 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion: 93 Resp: 524
Ion Ratio Lower Upper
93 100
63 71.9 59.3 88.9
95 30.7 25.2 37.8



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 936
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion: 136 Resp: 25854
Ion Ratio Lower Upper
136 100
137 11.0 9.0 13.4
54 9.9 5.2 7.8#
68 7.3 4.1 6.1#









#8
Nitrobenzene-d5
Concen: 0.282 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion:	82	Resp:	5875
Ion	Ratio	Lower	Upper
82	100		
128	45.7	34.1	51.1
54	53.9	43.5	65.3



#11
2-Methylnaphthalene-d10
Concen: 0.284 ng
RT: 12.172 min Scan# 1154
Delta R.T. -0.005 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion:	152	Resp:	10503
Ion	Ratio	Lower	Upper
152	100		
151	21.0	16.8	25.2









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.430 min Scan# 1453
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion:164 Resp: 14999
Ion Ratio Lower Upper
164 100
162 106.3 83.0 124.4
160 54.8 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 0.375 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion:330 Resp: 2316
Ion Ratio Lower Upper
330 100
332 95.3 76.6 114.8
141 50.2 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.309 ng
RT: 13.051 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.0	27.8	41.6
170	24.2	18.1	27.1



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	11.4	8.6	13.0









#27
Fluoranthene-d10
Concen: 0.429 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion:	212	Resp:	33654
Ion Ratio	Lower	Upper	
212	100		
106	18.8	12.6	19.0
104	9.7	7.1	10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion:	240	Resp:	25933
Ion Ratio	Lower	Upper	
240	100		
120	23.9	10.7	16.1
236	30.4	23.1	34.7









#31
Terphenyl-d14
Concen: 0.513 ng
RT: 19.819 min Scan# 2076
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57




Instrument :
BNA_N
ClientSampleId :
RW7-SP100-20251030


Tgt Ion	Ratio	Lower	Upper
244	100		
212	16.1	7.8	11.6#
122	39.1	15.2	22.8#



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.464 ng
RT: 21.295 min Scan# 2260
Delta R.T. -0.000 min
Lab File: BN038138.D
Acq: 04 Nov 2025 15:57

Tgt Ion	Ratio	Lower	Upper
149	100		
167	28.4	21.1	31.7
279	3.1	2.9	4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.680 min Scan# 28

Delta R.T. -0.000 min

Lab File: BN038138.D


Acq: 04 Nov 2025 15:57

Instrument :

BNA_N

ClientSampleId :

RW7-SP100-20251030





Tgt Ion:264 Resp: 21372

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

264	100		
-----	-----	--	--

260	27.6	21.8	32.8
-----	------	------	------

265	72.2	68.6	102.8
-----	------	------	-------





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/30/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	11/03/25
Client Sample ID:	RW7-SP201-20251030		SDG No.:	Q3526
Lab Sample ID:	Q3526-02		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.20	UQ	1	0.070	0.20	0.20	ug/L	11/04/25 16:34	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.30			30 - 150		76%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.38			30 - 150		94%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.31			55 - 111		76%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.31			53 - 106		77%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.56	*		58 - 132		140%	SPK: 0.4		
INTERNAL STANDARDS		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	8640								
1146-65-2	Naphthalene-d8	26000								
15067-26-2	Acenaphthene-d10	15200								
1517-22-2	Phenanthrene-d10	28300								
1719-03-5	Chrysene-d12	19100								
1520-96-3	Perylene-d12	15000								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038139.D
 Acq On : 04 Nov 2025 16:34
 Operator : RC/JU
 Sample : Q3526-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP201-20251030

Quant Time: Nov 04 16:59:16 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8635	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	25969	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	15192	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	28331	0.400	ng	0.00
29) Chrysene-d12	21.375	240	19063	0.400	ng	0.00
35) Perylene-d12	23.686	264	14967	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2921	0.144	ng	0.00
5) Phenol-d6	6.937	99	2056	0.083	ng	0.00
8) Nitrobenzene-d5	8.929	82	6392	0.305	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	11234	0.303	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1744	0.279	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	18738	0.308	ng	0.00
27) Fluoranthene-d10	19.220	212	26931	0.377	ng	0.00
31) Terphenyl-d14	19.819	244	23204	0.560	ng	0.00
Target Compounds						Qvalue
9) Naphthalene	10.626	128	1557	0.022	ng	# 76
34) Bis(2-ethylhexyl)phtha...	21.295	149	3536	0.099	ng	# 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038139.D
Acq On : 04 Nov 2025 16:34
Operator : RC/JU
Sample : Q3526-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030

Quant Time: Nov 04 16:59:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 651
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030


Tgt Ion	Ratio	Lower	Upper
152	100		
150	152.2	122.3	183.5
115	60.6	47.4	71.0



#4
2-Fluorophenol
Concen: 0.144 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Tgt Ion	Ratio	Lower	Upper
112	100		
64	57.8	45.4	68.0
63	31.5	23.2	34.8









#5
Phenol-d6
Concen: 0.083 ng
RT: 6.937 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030


Tgt Ion	Ratio	Lower	Upper
99	100		
42	45.8	16.6	24.8#
71	203.1	25.4	38.2#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 936
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.0	13.4
54	10.6	5.2	7.8#
68	7.8	4.1	6.1#









#8
Nitrobenzene-d5
Concen: 0.305 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030


Tgt Ion:	82	Resp:	6392
Ion	Ratio	Lower	Upper
82	100		
128	44.1	34.1	51.1
54	55.0	43.5	65.3



#9
Naphthalene
Concen: 0.022 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Tgt Ion:	128	Resp:	1557
Ion	Ratio	Lower	Upper
128	100		
129	23.8	9.1	13.7#
127	19.6	10.5	15.7#









#11
2-Methylnaphthalene-d10
Concen: 0.303 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030


Tgt Ion:152 Resp: 11234
Ion Ratio Lower Upper
152 100
151 20.7 16.8 25.2



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.430 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Tgt Ion:164 Resp: 15192
Ion Ratio Lower Upper
164 100
162 106.5 83.0 124.4
160 55.0 40.7 61.1









#14
2,4,6-Tribromophenol
Concen: 0.279 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030


Tgt Ion:330 Resp: 1744
Ion Ratio Lower Upper
330 100
332 95.9 76.6 114.8
141 43.8 34.1 51.1



#15
2-Fluorobiphenyl
Concen: 0.308 ng
RT: 13.051 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Tgt Ion:172 Resp: 18738
Ion Ratio Lower Upper
172 100
171 36.2 27.8 41.6
170 24.4 18.1 27.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1688

Delta R.T. -0.000 min

Lab File: BN038139.D

Acq: 04 Nov 2025 16:34

Instrument :

BNA_N

ClientSampleId :

RW7-SP201-20251030





Tgt Ion:188 Resp: 28331

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.9 8.6 13.0



#27

Fluoranthene-d10

Concen: 0.377 ng

RT: 19.220 min Scan# 1947

Delta R.T. -0.000 min

Lab File: BN038139.D

Acq: 04 Nov 2025 16:34




Tgt Ion:212 Resp: 26931


Ion Ratio Lower Upper

212 100

106 15.5 12.6 19.0

104 9.7 7.1 10.7









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030


Tgt Ion:240 Resp: 19063
Ion Ratio Lower Upper
240 100
120 14.2 10.7 16.1
236 28.6 23.1 34.7



#31
Terphenyl-d14
Concen: 0.560 ng
RT: 19.819 min Scan# 2076
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Tgt Ion:244 Resp: 23204
Ion Ratio Lower Upper
244 100
212 9.3 7.8 11.6
122 17.5 15.2 22.8









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.099 ng
RT: 21.295 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20251030

Tgt Ion:149 Resp: 3536
Ion Ratio Lower Upper
149 100
167 26.5 21.1 31.7
279 5.5 2.9 4.3#



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.686 min Scan# 2822
Delta R.T. 0.006 min
Lab File: BN038139.D
Acq: 04 Nov 2025 16:34

Tgt Ion:264 Resp: 14967
Ion Ratio Lower Upper
264 100
260 27.3 21.8 32.8
265 96.7 68.6 102.8





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/30/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	11/03/25
Client Sample ID:	RW7-SP302-20251030		SDG No.:	Q3526
Lab Sample ID:	Q3526-03		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	940 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.21	UQ	1	0.070	0.21	0.21	ug/L	11/04/25 17:10	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.32			30 - 150		79%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.36			30 - 150		90%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.33			55 - 111		81%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.36			53 - 106		90%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.59	*		58 - 132		148%	SPK: 0.4		
INTERNAL STANDARDS		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	8570								
1146-65-2	Naphthalene-d8	25600								
15067-26-2	Acenaphthene-d10	14300								
1517-22-2	Phenanthrene-d10	27500								
1719-03-5	Chrysene-d12	16400								
1520-96-3	Perylene-d12	13000								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038140.D
 Acq On : 04 Nov 2025 17:10
 Operator : RC/JU
 Sample : Q3526-03
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP302-20251030

Quant Time: Nov 04 17:34:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8568	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	25578	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	14324	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	27505	0.400	ng	0.00
29) Chrysene-d12	21.375	240	16355	0.400	ng	0.00
35) Perylene-d12	23.683	264	12980	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2964	0.147	ng	0.00
5) Phenol-d6	6.937	99	2079	0.085	ng	0.00
8) Nitrobenzene-d5	8.929	82	6714	0.325	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	11602	0.318	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1849	0.313	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	20625	0.360	ng	0.00
27) Fluoranthene-d10	19.220	212	25076	0.362	ng	0.00
31) Terphenyl-d14	19.819	244	21101	0.594	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.295	149	3579	0.116	ng	# 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038140.D
Acq On : 04 Nov 2025 17:10
Operator : RC/JU
Sample : Q3526-03
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030

Quant Time: Nov 04 17:34:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10




Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030


Tgt Ion:152 Resp: 8568
Ion Ratio Lower Upper
152 100
150 155.3 122.3 183.5
115 59.9 47.4 71.0



#4
2-Fluorophenol
Concen: 0.147 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10

Tgt Ion:112 Resp: 2964
Ion Ratio Lower Upper
112 100
64 57.7 45.4 68.0
63 30.7 23.2 34.8









#5
Phenol-d6
Concen: 0.085 ng
RT: 6.937 min Scan# 534
Delta R.T. 0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10




Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030


Tgt Ion	Ratio	Lower	Upper
99	100		
42	24.7	16.6	24.8
71	39.5	25.4	38.2



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 936
Delta R.T. -0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10

Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.0	13.4
54	10.0	5.2	7.8
68	7.3	4.1	6.1









#8
Nitrobenzene-d5
Concen: 0.325 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10




Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030


Tgt Ion:	82	Resp:	6714
Ion	Ratio	Lower	Upper
82	100		
128	43.5	34.1	51.1
54	54.0	43.5	65.3



#11
2-Methylnaphthalene-d10
Concen: 0.318 ng
RT: 12.172 min Scan# 1154
Delta R.T. -0.005 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10

Tgt Ion:	152	Resp:	11602
Ion	Ratio	Lower	Upper
152	100		
151	21.5	16.8	25.2









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.430 min Scan# 1453
Delta R.T. -0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10




Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030


Tgt Ion	Ratio	Lower	Upper
164	100		
162	107.2	83.0	124.4
160	56.2	40.7	61.1



#14
2,4,6-Tribromophenol
Concen: 0.313 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10

Tgt Ion	Ratio	Lower	Upper
330	100		
332	97.9	76.6	114.8
141	42.3	34.1	51.1









#15
2-Fluorobiphenyl
Concen: 0.360 ng
RT: 13.051 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10




Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030


Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.6	27.8	41.6
170	24.6	18.1	27.1



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	10.0	8.6	13.0









#27
Fluoranthene-d10
Concen: 0.362 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10




Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030


Tgt Ion	Ratio	Lower	Upper
212	100		
106	16.5	12.6	19.0
104	9.2	7.1	10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10

Tgt Ion	Ratio	Lower	Upper
240	100		
120	13.8	10.7	16.1
236	28.8	23.1	34.7









#31
Terphenyl-d14
Concen: 0.594 ng
RT: 19.819 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10




Instrument :
BNA_N
ClientSampleId :
RW7-SP302-20251030


Tgt Ion	Ratio	Lower	Upper
244	100		
212	9.3	7.8	11.6
122	17.6	15.2	22.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.116 ng
RT: 21.295 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BN038140.D
Acq: 04 Nov 2025 17:10

Tgt Ion	Ratio	Lower	Upper
149	100		
167	24.6	21.1	31.7
279	4.6	2.9	4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.683 min Scan# 2821

Delta R.T. 0.003 min

Lab File: BN038140.D


Acq: 04 Nov 2025 17:10

Instrument :

BNA_N

ClientSampleId :

RW7-SP302-20251030





Tgt Ion:264 Resp: 12980

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

264	100		
-----	-----	--	--

260	27.6	21.8	32.8
-----	------	------	------

265	103.2	68.6	102.8
-----	-------	------	-------





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/30/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	11/03/25
Client Sample ID:	RW7-SP303-20251030		SDG No.:	Q3526
Lab Sample ID:	Q3526-04		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.20	UQ	1	0.070	0.20	0.20	ug/L	11/04/25 17:46	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.27			30 - 150		67%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.32			30 - 150		81%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.27			55 - 111		68%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.28			53 - 106		69%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.65	*		58 - 132		163%	SPK: 0.4		
INTERNAL STANDARDS		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	8540								
1146-65-2	Naphthalene-d8	25200								
15067-26-2	Acenaphthene-d10	14300								
1517-22-2	Phenanthrene-d10	26300								
1719-03-5	Chrysene-d12	14800								
1520-96-3	Perylene-d12	12400								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038141.D
 Acq On : 04 Nov 2025 17:46
 Operator : RC/JU
 Sample : Q3526-04
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP303-20251030

Quant Time: Nov 04 18:13:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8535	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	25212	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	14257	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	26274	0.400	ng	0.00
29) Chrysene-d12	21.375	240	14750	0.400	ng	0.00
35) Perylene-d12	23.680	264	12357	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2450	0.122	ng	0.00
5) Phenol-d6	6.937	99	1641	0.067	ng	0.00
8) Nitrobenzene-d5	8.929	82	5492	0.270	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	9669	0.269	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1587	0.270	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	15752	0.276	ng	0.00
27) Fluoranthene-d10	19.220	212	21496	0.324	ng	0.00
31) Terphenyl-d14	19.819	244	20961	0.654	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.295	149	3107	0.112	ng	Qvalue 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038141.D
Acq On : 04 Nov 2025 17:46
Operator : RC/JU
Sample : Q3526-04
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030


Quant Time: Nov 04 18:13:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 651
Delta R.T. -0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030







Tgt Ion:152 Resp: 8535
Ion Ratio Lower Upper
152 100
150 156.1 122.3 183.5
115 59.7 47.4 71.0



#4
2-Fluorophenol
Concen: 0.122 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Tgt Ion:112 Resp: 2450
Ion Ratio Lower Upper
112 100
64 56.8 45.4 68.0
63 29.6 23.2 34.8









#5
Phenol-d6
Concen: 0.067 ng
RT: 6.937 min Scan# 534
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46




Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030


Tgt Ion:	99	Resp:	1641
Ion	Ratio	Lower	Upper
99	100		
42	20.5	16.6	24.8
71	41.1	25.4	38.2



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 936
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Tgt Ion:	136	Resp:	25212
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	9.2	5.2	7.8
68	7.0	4.1	6.1









#8
Nitrobenzene-d5
Concen: 0.270 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46




Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030


Tgt Ion:	82	Resp:	5492
Ion	Ratio	Lower	Upper
82	100		
128	45.6	34.1	51.1
54	55.1	43.5	65.3



#11
2-Methylnaphthalene-d10
Concen: 0.269 ng
RT: 12.172 min Scan# 1154
Delta R.T. -0.005 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Tgt Ion:	152	Resp:	9669
Ion	Ratio	Lower	Upper
152	100		
151	20.9	16.8	25.2









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.430 min Scan# 1453
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46




Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030


Tgt Ion:164 Resp: 14257
Ion Ratio Lower Upper
164 100
162 107.4 83.0 124.4
160 55.2 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 0.270 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Tgt Ion:330 Resp: 1587
Ion Ratio Lower Upper
330 100
332 94.8 76.6 114.8
141 44.3 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.276 ng
RT: 13.051 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46




Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.0	27.8	41.6
170	24.4	18.1	27.1



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	10.0	8.6	13.0





#27

Fluoranthene-d10

Concen: 0.324 ng

RT: 19.220 min Scan# 1947

Delta R.T. 0.000 min

Lab File: BN038141.D

Acq: 04 Nov 2025 17:46

Instrument :

BNA_N

ClientSampleId :

RW7-SP303-20251030





Tgt Ion:212 Resp: 21496

Ion Ratio Lower Upper

212 100

106 16.5 12.6 19.0

104 9.1 7.1 10.7



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 2269

Delta R.T. 0.000 min

Lab File: BN038141.D

Acq: 04 Nov 2025 17:46




Tgt Ion:240 Resp: 14750


Ion Ratio Lower Upper

240 100

120 13.0 10.7 16.1

236 28.8 23.1 34.7









#31
Terphenyl-d14
Concen: 0.654 ng
RT: 19.819 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46




Instrument :
BNA_N
ClientSampleId :
RW7-SP303-20251030


Tgt Ion	Ratio	Lower	Upper
244	100		
212	9.2	7.8	11.6
122	17.8	15.2	22.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.112 ng
RT: 21.295 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BN038141.D
Acq: 04 Nov 2025 17:46

Tgt Ion	Ratio	Lower	Upper
149	100		
167	25.4	21.1	31.7
279	3.9	2.9	4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.680 min Scan# 28

Delta R.T. 0.000 min

Lab File: BN038141.D


Acq: 04 Nov 2025 17:46

Instrument :

BNA_N

ClientSampleId :

RW7-SP303-20251030





Tgt Ion:264 Resp: 12357

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

264	100		
-----	-----	--	--

260	27.3	21.8	32.8
-----	------	------	------

265	103.3	68.6	102.8#
-----	-------	------	--------





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: AllianceContract: TETR06Lab Code: ACESDG No.: Q3526Instrument ID: BNA_NCalibration Date(s): 10/28/2025 10/28/2025Calibration Time(s): 13:00 16:38

LAB FILE ID:		RRF0.1 = BN038111.D		RRF0.2 = BN038112.D		RRF0.4 = BN038113.D		
		RRF0.8 = BN038114.D		RRF1.6 = BN038115.D		RRF3.2 = BN038116.D		
COMPOUND	RRF0.1	RRF0.2	RRF0.4	RRF0.8	RRF1.6	RRF3.2	RRF	% RSD
2-Methylnaphthalene-d10	0.581	0.569	0.551	0.549	0.586	0.584	0.571	2.7
Fluoranthene-d10	1.040	0.968	0.955	0.981	1.050	1.055	1.008	4.1
2-Fluorophenol	1.066	0.973	0.954	0.908	0.919	0.895	0.943	6.7
Phenol-d6	1.266	1.134	1.134	1.097	1.131	1.137	1.148	4.7
Nitrobenzene-d5	0.403	0.312	0.313	0.301	0.316	0.307	0.323	11.1
2-Fluorobiphenyl	1.794	1.649	1.577	1.570	1.615	1.514	1.602	6.2
2,4,6-Tribromophenol	0.157	0.151	0.149	0.152	0.171	0.182	0.165	10.7
Terphenyl-d14	0.878	0.881	0.872	0.807	0.892	0.862	0.869	3.4
1,4-Dioxane		0.446	0.424	0.423	0.416	0.400	0.416	4.8

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN102825.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Oct 29 13:09:46 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN038111.D 0.2 =BN038112.D 0.4 =BN038113.D 0.8 =BN038114.D 1.6 =BN038115.D 3.2 =BN038116.D 5 =BN038117.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.446	0.424	0.405	0.416	0.400	0.389	0.413	4.87	
3)	n-Nitrosodimet...	0.528	0.537	0.527	0.547	0.534	0.518	0.532	1.85	
4) S	2-Fluorophenol	1.066	0.973	0.954	0.907	0.918	0.895	0.882	0.942	6.73
5) S	Phenol-d6	1.266	1.134	1.134	1.097	1.131	1.135	1.137	1.148	4.71
6)	bis(2-Chloroet...	1.164	1.140	1.126	1.122	1.148	1.114	1.079	1.128	2.42
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.403	0.312	0.313	0.301	0.316	0.307	0.307	0.323	11.11
9)	Naphthalene	1.139	1.097	1.081	1.080	1.125	1.107	1.084	1.102	2.08
10)	Hexachlorobuta...	0.198	0.187	0.187	0.182	0.188	0.183	0.177	0.186	3.49
11) SURR2	Methylnaphth...	0.581	0.569	0.551	0.549	0.586	0.584	0.579	0.571	2.69
12)	2-Methylnaphth...	0.765	0.712	0.709	0.713	0.757	0.749	0.738	0.735	3.22
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.156	0.151	0.149	0.152	0.171	0.182	0.194	0.165	10.70
15) S	2-Fluorobiphenyl	1.794	1.649	1.577	1.570	1.615	1.514	1.495	1.602	6.24
16)	Acenaphthylene	1.768	1.728	1.711	1.764	1.891	1.912	1.908	1.811	4.88
17)	Acenaphthene	1.260	1.242	1.211	1.234	1.318	1.310	1.297	1.267	3.25
18)	Fluorene	1.549	1.618	1.596	1.635	1.768	1.731	1.729	1.661	4.93
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.041	0.045	0.051	0.064	0.073		0.055	24.55	
21)	4-Bromophenyl-...	0.260	0.251	0.250	0.256	0.274	0.271	0.275	0.262	4.06
22)	Hexachlorobenzene	0.304	0.292	0.298	0.293	0.306	0.297	0.298	0.298	1.76
23)	Atrazine	0.190	0.176	0.176	0.179	0.203	0.209	0.211	0.192	8.19
24)	Pentachlorophenol	0.112	0.108	0.113	0.132	0.146	0.159	0.128	16.16	
25)	Phenanthrene	1.281	1.231	1.224	1.236	1.327	1.293	1.298	1.270	3.12
26)	Anthracene	1.076	1.029	1.047	1.057	1.182	1.183	1.202	1.111	6.72
27) SURR	Fluoranthene-d10	1.040	0.969	0.957	0.981	1.050	1.055	1.009	1.009	4.01
28)	Fluoranthene	1.448	1.349	1.345	1.394	1.492	1.491	1.414	1.419	4.30
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.834	1.853	1.842	1.664	1.835	1.766	1.846	1.806	3.82
31) S	Terphenyl-d14	0.878	0.881	0.872	0.807	0.892	0.862	0.891	0.869	3.37
32)	Benzo(a)anthra...	1.491	1.379	1.358	1.363	1.469	1.474	1.456	1.427	4.06
33)	Chrysene	1.623	1.586	1.543	1.523	1.558	1.518	1.490	1.549	2.91
34)	Bis(2-ethylhex...	0.910	0.722	0.700	0.721	0.754	0.703	0.752	10.60	
35) I	Perylene-d12	-----ISTD-----								


Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN102825.M

36)	Indeno(1,2,3-c...	1.521	1.512	1.604	1.620	1.730	1.739	1.778	1.643	6.53
37)	Benzo(b)fluora...	1.603	1.543	1.630	1.729	1.806	1.839	1.822	1.710	6.94
38)	Benzo(k)fluora...	1.652	1.590	1.619	1.669	1.831	1.835	1.855	1.721	6.62
39) C	Benzo(a)pyrene	1.344	1.276	1.316	1.314	1.420	1.451	1.462	1.369	5.44
40)	Dibenzo(a,h)an...	1.149	1.150	1.196	1.240	1.352	1.363	1.402	1.265	8.42
41)	Benzo(g,h,i)pe...	1.402	1.365	1.426	1.432	1.499	1.477	1.509	1.444	3.66

(#) = Out of Range

4,6-Dinitro-2-methylphenol



Response = $7.609 \times 10^{-2} \times \text{Amt} - 3.487 \times 10^{-2}$
 Coef of Det (r^2) = 0.995885 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA N\Methods\8270-SIM-BN102825.M
 Calibration Table Last Updated: Wed Oct 29 13:09:46 2025

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038111.D
 Acq On : 28 Oct 2025 13:00
 Operator : RC/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Oct 28 17:31:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.789	152	8239	0.400	ng	0.01
7) Naphthalene-d8	10.584	136	24762	0.400	ng	# 0.01
13) Acenaphthene-d10	14.441	164	14233	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	28114	0.400	ng	# 0.00
29) Chrysene-d12	21.384	240	22622	0.400	ng	0.00
35) Perylene-d12	23.692	264	20615	0.400	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.356	112	2196	0.113	ng	0.00
5) Phenol-d6	6.952	99	2608	0.110	ng	0.01
8) Nitrobenzene-d5	8.939	82	2497	0.125	ng	0.01
11) 2-Methylnaphthalene-d10	12.182	152	3595	0.102	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	558	0.095	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	6383	0.112	ng	0.01
27) Fluoranthene-d10	19.225	212	7307	0.103	ng	0.00
31) Terphenyl-d14	19.829	244	4967	0.101	ng	0.00
Target Compounds						Qvalue
6) bis(2-Chloroethyl)ether	7.204	93	2398	0.103	ng	95
9) Naphthalene	10.637	128	7049	0.103	ng	95
10) Hexachlorobutadiene	10.936	225	1223	0.106	ng	# 98
12) 2-Methylnaphthalene	12.258	142	4736	0.104	ng	96
16) Acenaphthylene	14.163	152	6290	0.098	ng	100
17) Acenaphthene	14.505	154	4482	0.099	ng	98
18) Fluorene	15.499	166	5513	0.093	ng	97
21) 4-Bromophenyl-phenylether	16.391	248	1824	0.099	ng	# 82
22) Hexachlorobenzene	16.503	284	2137	0.102	ng	98
23) Atrazine	16.652	200	1338	0.099	ng	# 90
25) Phenanthrene	17.236	178	9006	0.101	ng	99
26) Anthracene	17.322	178	7562	0.097	ng	98
28) Fluoranthene	19.257	202	10174	0.102	ng	99
30) Pyrene	19.620	202	10370	0.102	ng	99
32) Benzo(a)anthracene	21.367	228	8433	0.104	ng	99
33) Chrysene	21.420	228	9181	0.105	ng	98
36) Indeno(1,2,3-cd)pyrene	26.048	276	7839	0.093	ng	98
37) Benzo(b)fluoranthene	22.993	252	8261	0.094	ng	# 67
38) Benzo(k)fluoranthene	23.040	252	8515	0.096	ng	# 67
39) Benzo(a)pyrene	23.590	252	6927	0.098	ng	# 58
40) Dibenzo(a,h)anthracene	26.072	278	5922	0.091	ng	# 71
41) Benzo(g,h,i)perylene	26.765	276	7227	0.097	ng	# 88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038111.D
Acq On : 28 Oct 2025 13:00
Operator : RC/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Oct 28 17:31:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.789 min Scan# 61
Delta R.T. 0.014 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 8239
Ion Ratio Lower Upper
152 100
150 151.7 122.3 183.5
115 60.8 47.4 71.0



#4
2-Fluorophenol
Concen: 0.113 ng
RT: 5.356 min Scan# 315
Delta R.T. 0.007 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:112 Resp: 2196
Ion Ratio Lower Upper
112 100
64 57.2 45.4 68.0
63 32.8 23.2 34.8









#5
Phenol-d6
Concen: 0.110 ng
RT: 6.952 min Scan# 511
Delta R.T. 0.015 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 99 Resp: 2608
Ion Ratio Lower Upper
99 100
42 23.4 16.6 24.8
71 34.0 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.103 ng
RT: 7.204 min Scan# 571
Delta R.T. 0.007 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion: 93 Resp: 2398
Ion Ratio Lower Upper
93 100
63 78.2 59.3 88.9
95 34.1 25.2 37.8





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.584 min Scan# 91

Delta R.T. 0.011 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1

Tgt Ion:136 Resp: 24762





Ion Ratio Lower Upper

136 100

137 11.3 9.0 13.4

54 10.0 5.2 7.8#

68 7.5 4.1 6.1#



#8

Nitrobenzene-d5

Concen: 0.125 ng

RT: 8.939 min Scan# 783

Delta R.T. 0.011 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00




Tgt Ion: 82 Resp: 2497


Ion Ratio Lower Upper

82 100

128 46.2 34.1 51.1

54 57.2 43.5 65.3









#9
Naphthalene
Concen: 0.103 ng
RT: 10.637 min Scan# 941
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:128 Resp: 7049
Ion Ratio Lower Upper
128 100
129 13.5 9.1 13.7
127 14.7 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.106 ng
RT: 10.936 min Scan# 970
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:225 Resp: 1223
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.3 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.102 ng
RT: 12.182 min Scan# 1156
Delta R.T. 0.005 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:152 Resp: 3595
Ion Ratio Lower Upper
152 100
151 21.4 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.104 ng
RT: 12.258 min Scan# 1171
Delta R.T. 0.010 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:142 Resp: 4736
Ion Ratio Lower Upper
142 100
141 90.0 70.3 105.5
115 39.2 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 14233
Ion Ratio Lower Upper
164 100
162 106.3 83.0 124.4
160 52.7 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 0.095 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:330 Resp: 558
Ion Ratio Lower Upper
330 100
332 95.9 76.6 114.8
141 41.0 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.112 ng
RT: 13.062 min Scan# 11
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:172 Resp: 6383
Ion Ratio Lower Upper
172 100
171 36.2 27.8 41.6
170 25.2 18.1 27.1



#16
Acenaphthylene
Concen: 0.098 ng
RT: 14.163 min Scan# 1427
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:152 Resp: 6290
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.8
153 12.9 10.2 15.2









#17
Acenaphthene
Concen: 0.099 ng
RT: 14.505 min Scan# 1459
Delta R.T. 0.011 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:154 Resp: 4482
Ion Ratio Lower Upper
154 100
153 114.1 90.0 135.0
152 60.7 47.3 70.9



#18
Fluorene
Concen: 0.093 ng
RT: 15.499 min Scan# 1552
Delta R.T. 0.010 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:166 Resp: 5513
Ion Ratio Lower Upper
166 100
165 102.4 79.0 118.4
167 13.3 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:188 Resp: 28114

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.8 8.6 13.0#



#21

4-Bromophenyl-phenylether

Concen: 0.099 ng

RT: 16.391 min Scan# 1624

Delta R.T. 0.012 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00




Tgt Ion:248 Resp: 1824


Ion Ratio Lower Upper

248 100

250 103.9 76.2 114.2

141 48.0 57.6 86.4#





#22

Hexachlorobenzene

Concen: 0.102 ng

RT: 16.503 min Scan# 1633

Delta R.T. 0.000 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:284 Resp: 2137

Ion Ratio Lower Upper

284 100

142 38.9 30.5 45.7

249 31.2 23.5 35.3



#23

Atrazine

Concen: 0.099 ng

RT: 16.652 min Scan# 1645

Delta R.T. 0.000 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00




Tgt Ion:200 Resp: 1338


Ion Ratio Lower Upper

200 100

173 38.1 21.4 32.2#

215 47.8 36.7 55.1









#25
Phenanthrene
Concen: 0.101 ng
RT: 17.236 min Scan# 1692
Delta R.T. 0.012 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:178 Resp: 9006
Ion Ratio Lower Upper
178 100
176 19.0 15.5 23.3
179 15.8 12.2 18.4



#26
Anthracene
Concen: 0.097 ng
RT: 17.322 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:178 Resp: 7562
Ion Ratio Lower Upper
178 100
176 18.4 14.9 22.3
179 16.0 11.8 17.8





#27

Fluoranthene-d10

Concen: 0.103 ng

RT: 19.225 min Scan# 1948

Delta R.T. 0.005 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:212 Resp: 7307

Ion Ratio Lower Upper

212 100

106 15.9 12.6 19.0

104 9.3 7.1 10.7



#28

Fluoranthene

Concen: 0.102 ng

RT: 19.257 min Scan# 1955

Delta R.T. 0.005 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00




Tgt Ion:202 Resp: 10174


Ion Ratio Lower Upper

202 100

101 11.4 9.4 14.2

203 17.2 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.384 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:240 Resp: 22622

Ion Ratio Lower Upper

240 100

120 12.1 10.7 16.1

236 28.6 23.1 34.7



#30

Pyrene

Concen: 0.102 ng

RT: 19.620 min Scan# 2033

Delta R.T. 0.009 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00




Tgt Ion:202 Resp: 10370


Ion Ratio Lower Upper

202 100

200 21.3 16.7 25.1

203 18.0 14.2 21.2









#31
Terphenyl-d14
Concen: 0.101 ng
RT: 19.829 min Scan# 2076
Delta R.T. 0.009 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:244 Resp: 4967
Ion Ratio Lower Upper
244 100
212 10.8 7.8 11.6
122 18.9 15.2 22.8



#32
Benzo(a)anthracene
Concen: 0.104 ng
RT: 21.367 min Scan# 2268
Delta R.T. 0.009 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:228 Resp: 8433
Ion Ratio Lower Upper
228 100
226 27.7 22.2 33.4
229 20.5 15.8 23.8









#33
Chrysene
Concen: 0.105 ng
RT: 21.420 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:228 Resp: 9181
Ion Ratio Lower Upper
228 100
226 30.5 24.3 36.5
229 21.3 15.7 23.5



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.692 min Scan# 2824
Delta R.T. 0.012 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00


Tgt Ion:264 Resp: 20615
Ion Ratio Lower Upper
264 100
260 27.0 21.8 32.8
265 75.0 68.6 102.8








#36
Indeno(1,2,3-cd)pyrene
Concen: 0.093 ng
RT: 26.048 min Scan# 3630
Delta R.T. 0.015 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1







Tgt Ion:276 Resp: 7839
Ion Ratio Lower Upper
276 100
138 25.8 21.4 32.0
277 24.8 19.2 28.8



#37
Benzo(b)fluoranthene
Concen: 0.094 ng
RT: 22.993 min Scan# 2585
Delta R.T. 0.012 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:252 Resp: 8261
Ion Ratio Lower Upper
252 100
253 37.3 20.9 31.3#
125 39.7 14.1 21.1#









#38
Benzo(k)fluoranthene
Concen: 0.096 ng
RT: 23.040 min Scan# 20
Delta R.T. 0.012 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 8515
Ion Ratio Lower Upper
252 100
253 38.3 21.2 31.8#
125 39.7 14.6 22.0#



#39
Benzo(a)pyrene
Concen: 0.098 ng
RT: 23.590 min Scan# 2789
Delta R.T. 0.012 min
Lab File: BN038111.D
Acq: 28 Oct 2025 13:00

Tgt Ion:252 Resp: 6927
Ion Ratio Lower Upper
252 100
253 44.2 23.1 34.7#
125 51.6 18.6 28.0#





#40

Dibenzo(a,h)anthracene

Concen: 0.091 ng

RT: 26.072 min Scan# 3

Delta R.T. 0.023 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:278 Resp: 5922

Ion Ratio Lower Upper

278 100

139 38.9 17.6 26.4#

279 41.0 22.6 33.8#



#41

Benzo(g,h,i)perylene

Concen: 0.097 ng

RT: 26.765 min Scan# 3875

Delta R.T. 0.015 min

Lab File: BN038111.D

Acq: 28 Oct 2025 13:00




Tgt Ion:276 Resp: 7227

Ion Ratio Lower Upper

276 100

277 27.6 19.7 29.5

138 34.5 20.2 30.4#



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038112.D
 Acq On : 28 Oct 2025 13:36
 Operator : RC/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Oct 28 17:31:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	9322	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	28183	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	15941	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	31512	0.400	ng	0.00
29) Chrysene-d12	21.375	240	23080	0.400	ng	0.00
35) Perylene-d12	23.683	264	20598	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	4536	0.207	ng	0.00
5) Phenol-d6	6.937	99	5284	0.198	ng	0.00
8) Nitrobenzene-d5	8.929	82	4403	0.194	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	8012	0.199	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1201	0.183	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	13145	0.206	ng	0.01
27) Fluoranthene-d10	19.225	212	15247	0.192	ng	0.00
31) Terphenyl-d14	19.824	244	10170	0.203	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.254	88	2078	0.216	ng	95
3) n-Nitrosodimethylamine	3.564	42	2463	0.199	ng	# 98
6) bis(2-Chloroethyl)ether	7.197	93	5312	0.202	ng	99
9) Naphthalene	10.626	128	15455	0.199	ng	98
10) Hexachlorobutadiene	10.925	225	2632	0.201	ng	# 97
12) 2-Methylnaphthalene	12.253	142	10027	0.194	ng	99
16) Acenaphthylene	14.152	152	13787	0.191	ng	99
17) Acenaphthene	14.505	154	9898	0.196	ng	100
18) Fluorene	15.489	166	13019	0.197	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	639	0.238	ng	# 40
21) 4-Bromophenyl-phenylether	16.379	248	3948	0.191	ng	98
22) Hexachlorobenzene	16.503	284	4594	0.196	ng	99
23) Atrazine	16.652	200	2774	0.183	ng	98
24) Pentachlorophenol	16.838	266	1769	0.175	ng	98
25) Phenanthrene	17.223	178	19374	0.194	ng	100
26) Anthracene	17.322	178	16201	0.185	ng	99
28) Fluoranthene	19.253	202	21232	0.190	ng	99
30) Pyrene	19.615	202	21387	0.205	ng	100
32) Benzo(a)anthracene	21.358	228	15917	0.193	ng	99
33) Chrysene	21.411	228	18301	0.205	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	10496	0.242	ng	100
36) Indeno(1,2,3-cd)pyrene	26.039	276	15575	0.184	ng	99
37) Benzo(b)fluoranthene	22.984	252	15888	0.180	ng	# 90
38) Benzo(k)fluoranthene	23.031	252	16371	0.185	ng	# 91
39) Benzo(a)pyrene	23.584	252	13137	0.186	ng	# 84
40) Dibenzo(a,h)anthracene	26.057	278	11845	0.182	ng	95
41) Benzo(g,h,i)perylene	26.753	276	14055	0.189	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038112.D
Acq On : 28 Oct 2025 13:36
Operator : RC/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Oct 28 17:31:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.782 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 9322
Ion Ratio Lower Upper
152 100
150 150.4 122.3 183.5
115 59.1 47.4 71.0



#2
1,4-Dioxane
Concen: 0.216 ng
RT: 3.254 min Scan# 24
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion: 88 Resp: 2078
Ion Ratio Lower Upper
88 100
43 36.3 24.3 36.5
58 73.3 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 0.199 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 2463
Ion Ratio Lower Upper
42 100
74 121.8 96.8 145.2
44 12.3 5.3 7.9#



#4
2-Fluorophenol
Concen: 0.207 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion: 112 Resp: 4536
Ion Ratio Lower Upper
112 100
64 56.6 45.4 68.0
63 30.2 23.2 34.8









#5
Phenol-d6
Concen: 0.198 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 99 Resp: 5284
Ion Ratio Lower Upper
99 100
42 22.1 16.6 24.8
71 32.7 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.202 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion: 93 Resp: 5312
Ion Ratio Lower Upper
93 100
63 74.3 59.3 88.9
95 32.2 25.2 37.8





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.584 min Scan# 91

Delta R.T. 0.011 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2

Tgt Ion:136 Resp: 28183





Ion Ratio Lower Upper

136 100

137 11.2 9.0 13.4

54 6.6 5.2 7.8

68 5.2 4.1 6.1



#8

Nitrobenzene-d5

Concen: 0.194 ng

RT: 8.929 min Scan# 782

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion: 82 Resp: 4403


Ion Ratio Lower Upper

82 100

128 44.4 34.1 51.1

54 57.3 43.5 65.3





#9

Naphthalene

Concen: 0.199 ng

RT: 10.626 min Scan# 941

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:128 Resp: 15455

Ion Ratio Lower Upper

128 100

129 12.2 9.1 13.7

127 13.8 10.5 15.7



#10

Hexachlorobutadiene

Concen: 0.201 ng

RT: 10.925 min Scan# 969

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion:225 Resp: 2632


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 65.1 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.199 ng
RT: 12.177 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 8012
Ion Ratio Lower Upper
152 100
151 20.7 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.194 ng
RT: 12.253 min Scan# 1170
Delta R.T. 0.005 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:142 Resp: 10027
Ion Ratio Lower Upper
142 100
141 88.4 70.3 105.5
115 36.8 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	164	162	160
Resp:	15941		
Ion Ratio	100	103.1	49.6
Lower		83.0	40.7
Upper		124.4	61.1



#14
2,4,6-Tribromophenol
Concen: 0.183 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion	330	332	141
Resp:	1201		
Ion Ratio	100	96.6	41.5
Lower		76.6	34.1
Upper		114.8	51.1









#15
2-Fluorobiphenyl
Concen: 0.206 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.0	27.8	41.6
170	23.6	18.1	27.1



#16
Acenaphthylene
Concen: 0.191 ng
RT: 14.152 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.8	23.8
153	13.0	10.2	15.2









#17
Acenaphthene
Concen: 0.196 ng
RT: 14.505 min Scan# 1459
Delta R.T. 0.011 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:154 Resp: 9898
Ion Ratio Lower Upper
154 100
153 112.7 90.0 135.0
152 58.9 47.3 70.9



#18
Fluorene
Concen: 0.197 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:166 Resp: 13019
Ion Ratio Lower Upper
166 100
165 98.6 79.0 118.4
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:188 Resp: 31512
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.0 8.6 13.0



#20
4,6-Dinitro-2-methylphenol
Concen: 0.238 ng
RT: 15.560 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:198 Resp: 639
Ion Ratio Lower Upper
198 100
51 192.9 88.9 133.3#
105 84.5 49.2 73.8#









#21
4-Bromophenyl-phenylether
Concen: 0.191 ng
RT: 16.379 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
248	100		
250	92.8	76.2	114.2
141	73.9	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.196 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion	Ratio	Lower	Upper
284	100		
142	38.0	30.5	45.7
249	30.1	23.5	35.3









#23
Atrazine
Concen: 0.183 ng
RT: 16.652 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	200	Resp:	2774
Ion Ratio	Lower	Upper	
200	100		
173	28.7	21.4	32.2
215	45.0	36.7	55.1



#24
Pentachlorophenol
Concen: 0.175 ng
RT: 16.838 min Scan# 1660
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:	266	Resp:	1769
Ion Ratio	Lower	Upper	
266	100		
264	61.0	49.3	73.9
268	65.9	50.6	75.8









#25
Phenanthrene
Concen: 0.194 ng
RT: 17.223 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:178 Resp: 19374
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.3 12.2 18.4



#26
Anthracene
Concen: 0.185 ng
RT: 17.322 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:178 Resp: 16201
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.5 11.8 17.8





#27

Fluoranthene-d10

Concen: 0.192 ng

RT: 19.225 min Scan# 1947

Delta R.T. 0.005 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:212 Resp: 15247

Ion Ratio Lower Upper

212 100

106 15.8 12.6 19.0

104 9.0 7.1 10.7



#28

Fluoranthene

Concen: 0.190 ng

RT: 19.253 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion:202 Resp: 21232


Ion Ratio Lower Upper

202 100

101 11.6 9.4 14.2

203 17.2 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:240 Resp: 23080

Ion Ratio Lower Upper

240 100

120 14.9 10.7 16.1

236 28.8 23.1 34.7



#30

Pyrene

Concen: 0.205 ng

RT: 19.615 min Scan# 2032

Delta R.T. 0.005 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion:202 Resp: 21387


Ion Ratio Lower Upper

202 100

200 21.0 16.7 25.1

203 17.7 14.2 21.2









#31
Terphenyl-d14
Concen: 0.203 ng
RT: 19.824 min Scan# 2076
Delta R.T. 0.005 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:244 Resp: 10170
Ion Ratio Lower Upper
244 100
212 9.6 7.8 11.6
122 17.6 15.2 22.8



#32
Benzo(a)anthracene
Concen: 0.193 ng
RT: 21.358 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BN038112.D
Acq: 28 Oct 2025 13:36

Tgt Ion:228 Resp: 15917
Ion Ratio Lower Upper
228 100
226 28.5 22.2 33.4
229 19.5 15.8 23.8





#33

Chrysene

Concen: 0.205 ng

RT: 21.411 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:228 Resp: 18301

Ion Ratio Lower Upper

228 100

226 31.1 24.3 36.5

229 19.9 15.7 23.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.242 ng

RT: 21.295 min Scan# 2260

Delta R.T. 0.000 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion:149 Resp: 10496


Ion Ratio Lower Upper

149 100

167 26.3 21.1 31.7

279 3.1 2.9 4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.683 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:264 Resp: 20598

Ion Ratio Lower Upper

264 100

260 26.6 21.8 32.8

265 75.6 68.6 102.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.184 ng

RT: 26.039 min Scan# 3627

Delta R.T. 0.006 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion:276 Resp: 15575


Ion Ratio Lower Upper

276 100

138 25.6 21.4 32.0

277 24.2 19.2 28.8





#37

Benzo(b)fluoranthene

Concen: 0.180 ng

RT: 22.984 min Scan# 2582

Delta R.T. 0.003 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 15888

Ion Ratio Lower Upper

252 100

253 29.8 20.9 31.3

125 24.3 14.1 21.1#



#38

Benzo(k)fluoranthene

Concen: 0.185 ng

RT: 23.031 min Scan# 2598

Delta R.T. 0.003 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion:252 Resp: 16371


Ion Ratio Lower Upper

252 100

253 29.7 21.2 31.8

125 23.6 14.6 22.0#





#39

Benzo(a)pyrene

Concen: 0.186 ng

RT: 23.584 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN038112.D


Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 13137

Ion Ratio Lower Upper

252 100

253 35.6 23.1 34.7#

125 33.1 18.6 28.0#



#40

Dibenzo(a,h)anthracene

Concen: 0.182 ng

RT: 26.057 min Scan# 3633

Delta R.T. 0.009 min

Lab File: BN038112.D

Acq: 28 Oct 2025 13:36




Tgt Ion:278 Resp: 11845


Ion Ratio Lower Upper

278 100

139 24.7 17.6 26.4

279 30.8 22.6 33.8





#41

Benzo(g,h,i)perylene

Concen: 0.189 ng

RT: 26.753 min Scan# 3871

Delta R.T. 0.003 min

Lab File: BN038112.D


Acq: 28 Oct 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:276 Resp: 14055

Ion Ratio Lower Upper

276 100

277 26.2 19.7 29.5

138 28.5 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038113.D
 Acq On : 28 Oct 2025 14:13
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Oct 28 17:31:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	9189	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	27629	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	15496	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	29547	0.400	ng	0.00
29) Chrysene-d12	21.375	240	21529	0.400	ng	0.00
35) Perylene-d12	23.683	264	17966	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	8768	0.405	ng	0.00
5) Phenol-d6	6.937	99	10420	0.395	ng	0.00
8) Nitrobenzene-d5	8.929	82	8652	0.388	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	15231	0.386	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	2303	0.361	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	24443	0.394	ng	0.01
27) Fluoranthene-d10	19.220	212	28227	0.379	ng	0.00
31) Terphenyl-d14	19.819	244	18767	0.401	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.254	88	3895	0.410	ng	100
3) n-Nitrosodimethylamine	3.564	42	4910	0.402	ng	100
6) bis(2-Chloroethyl)ether	7.197	93	10402	0.402	ng	100
9) Naphthalene	10.626	128	29858	0.392	ng	100
10) Hexachlorobutadiene	10.925	225	5174	0.403	ng	# 100
12) 2-Methylnaphthalene	12.253	142	19577	0.386	ng	100
16) Acenaphthylene	14.152	152	26507	0.378	ng	100
17) Acenaphthene	14.494	154	18774	0.382	ng	100
18) Fluorene	15.489	166	24733	0.384	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	1329	0.375	ng	100
21) 4-Bromophenyl-phenylether	16.379	248	7383	0.381	ng	100
22) Hexachlorobenzene	16.503	284	8778	0.398	ng	100
23) Atrazine	16.652	200	5175	0.365	ng	100
24) Pentachlorophenol	16.838	266	3182	0.335	ng	100
25) Phenanthrene	17.223	178	36093	0.385	ng	100
26) Anthracene	17.322	178	30869	0.376	ng	100
28) Fluoranthene	19.253	202	39664	0.378	ng	100
30) Pyrene	19.615	202	39644	0.408	ng	100
32) Benzo(a)anthracene	21.357	228	29230	0.381	ng	100
33) Chrysene	21.411	228	33214	0.398	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	15540	0.384	ng	100
36) Indeno(1,2,3-cd)pyrene	26.037	276	28812	0.390	ng	100
37) Benzo(b)fluoranthene	22.984	252	29283	0.381	ng	100
38) Benzo(k)fluoranthene	23.028	252	29088	0.376	ng	100
39) Benzo(a)pyrene	23.581	252	23635	0.384	ng	100
40) Dibenzo(a,h)anthracene	26.054	278	21487	0.378	ng	100
41) Benzo(g,h,i)perylene	26.747	276	25622	0.395	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038113.D
 Acq On : 28 Oct 2025 14:13
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4


Quant Time: Oct 28 17:31:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.782 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4







Tgt Ion:152 Resp: 9189
Ion Ratio Lower Upper
152 100
150 152.9 122.3 183.5
115 59.2 47.4 71.0



#2
1,4-Dioxane
Concen: 0.410 ng
RT: 3.254 min Scan# 24
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion: 88 Resp: 3895
Ion Ratio Lower Upper
88 100
43 30.4 24.3 36.5
58 75.5 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 0.402 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 4910
Ion Ratio Lower Upper
42 100
74 121.1 96.8 145.2
44 6.7 5.3 7.9



#4
2-Fluorophenol
Concen: 0.405 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion: 112 Resp: 8768
Ion Ratio Lower Upper
112 100
64 56.7 45.4 68.0
63 29.0 23.2 34.8









#5
Phenol-d6
Concen: 0.395 ng
RT: 6.937 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 10420
Ion Ratio Lower Upper
99 100
42 20.7 16.6 24.8
71 31.8 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.402 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion: 93 Resp: 10402
Ion Ratio Lower Upper
93 100
63 73.7 59.3 88.9
95 31.4 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	136	Resp	27629
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.5	5.2	7.8
68	5.1	4.1	6.1



#8
Nitrobenzene-d5
Concen: 0.388 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion	82	Resp	8652
Ion	Ratio	Lower	Upper
82	100		
128	42.6	34.1	51.1
54	54.4	43.5	65.3









#9
Naphthalene
Concen: 0.392 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	128	129	127
Resp:	29858		
Ion Ratio	100	11.4	13.1
Lower		9.1	10.5
Upper		13.7	15.7



#10
Hexachlorobutadiene
Concen: 0.403 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion	225	223	227
Resp:	5174		
Ion Ratio	100	0.0	62.9
Lower		0.0	50.3
Upper		0.0	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.386 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 15231
Ion Ratio Lower Upper
152 100
151 21.0 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.386 ng
RT: 12.253 min Scan# 1170
Delta R.T. 0.005 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:142 Resp: 19577
Ion Ratio Lower Upper
142 100
141 87.9 70.3 105.5
115 34.7 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 15496
Ion Ratio Lower Upper
164 100
162 103.7 83.0 124.4
160 50.9 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 0.361 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:330 Resp: 2303
Ion Ratio Lower Upper
330 100
332 96.1 76.6 114.8
141 42.6 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.394 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.7	27.8	41.6
170	22.6	18.1	27.1



#16
Acenaphthylene
Concen: 0.378 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.8	23.8
153	12.7	10.2	15.2









#17
Acenaphthene
Concen: 0.382 ng
RT: 14.494 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 18774
Ion Ratio Lower Upper
154 100
153 112.5 90.0 135.0
152 59.1 47.3 70.9



#18
Fluorene
Concen: 0.384 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:166 Resp: 24733
Ion Ratio Lower Upper
166 100
165 98.7 79.0 118.4
167 13.4 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1688

Delta R.T. -0.000 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:188 Resp: 29547

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.8 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.375 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13




Tgt Ion:198 Resp: 1329


Ion Ratio Lower Upper

198 100

51 111.1 88.9 133.3

105 61.5 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.381 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
248	100		
250	95.2	76.2	114.2
141	72.0	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.398 ng
RT: 16.503 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion	Ratio	Lower	Upper
284	100		
142	38.1	30.5	45.7
249	29.4	23.5	35.3









#23
Atrazine
Concen: 0.365 ng
RT: 16.652 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.8	21.4	32.2
215	45.9	36.7	55.1



#24
Pentachlorophenol
Concen: 0.335 ng
RT: 16.838 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion	Ratio	Lower	Upper
266	100		
264	61.6	49.3	73.9
268	63.2	50.6	75.8









#25
Phenanthrene
Concen: 0.385 ng
RT: 17.223 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 36093
Ion Ratio Lower Upper
178 100
176 19.4 15.5 23.3
179 15.3 12.2 18.4



#26
Anthracene
Concen: 0.376 ng
RT: 17.322 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:178 Resp: 30869
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 14.8 11.8 17.8





#27

Fluoranthene-d10

Concen: 0.379 ng

RT: 19.220 min Scan# 1947

Delta R.T. -0.000 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:212 Resp: 28227

Ion Ratio Lower Upper

212 100

106 15.8 12.6 19.0

104 8.9 7.1 10.7



#28

Fluoranthene

Concen: 0.378 ng

RT: 19.253 min Scan# 1954

Delta R.T. -0.000 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13




Tgt Ion:202 Resp: 39664


Ion Ratio Lower Upper

202 100

101 11.8 9.4 14.2

203 16.8 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:240 Resp: 21529

Ion Ratio Lower Upper

240 100

120 13.4 10.7 16.1

236 28.9 23.1 34.7



#30

Pyrene

Concen: 0.408 ng

RT: 19.615 min Scan# 2032

Delta R.T. 0.005 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13




Tgt Ion:202 Resp: 39644


Ion Ratio Lower Upper

202 100

200 20.9 16.7 25.1

203 17.7 14.2 21.2









#31
Terphenyl-d14
Concen: 0.401 ng
RT: 19.819 min Scan# 2076
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:244 Resp: 18767
Ion Ratio Lower Upper
244 100
212 9.7 7.8 11.6
122 19.0 15.2 22.8



#32
Benzo(a)anthracene
Concen: 0.381 ng
RT: 21.357 min Scan# 2267
Delta R.T. -0.000 min
Lab File: BN038113.D
Acq: 28 Oct 2025 14:13

Tgt Ion:228 Resp: 29230
Ion Ratio Lower Upper
228 100
226 27.8 22.2 33.4
229 19.8 15.8 23.8





#33

Chrysene

Concen: 0.398 ng

RT: 21.411 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:228 Resp: 33214

Ion Ratio Lower Upper

228 100

226 30.4 24.3 36.5

229 19.6 15.7 23.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.384 ng

RT: 21.295 min Scan# 2260

Delta R.T. -0.000 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13




Tgt Ion:149 Resp: 15540


Ion Ratio Lower Upper

149 100

167 26.4 21.1 31.7

279 3.6 2.9 4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.683 min Scan# 2821

Delta R.T. 0.003 min

Lab File: BN038113.D


Acq: 28 Oct 2025 14:13

Instrument :



BNA_N

ClientSampleId :

SSTDICCC0.4



Tgt Ion:	264	Resp:	17966
Ion Ratio	Lower	Upper	
264	100		
260	27.3	21.8	32.8
265	85.7	68.6	102.8



#36

Indeno(1,2,3-cd)pyrene


Concen: 0.390 ng

RT: 26.037 min Scan# 3626



Delta R.T. 0.003 min


Lab File: BN038113.D

Acq: 28 Oct 2025 14:13



Tgt Ion:	276	Resp:	28812
Ion Ratio	Lower	Upper	
276	100		
138	26.7	21.4	32.0
277	24.0	19.2	28.8





#37

Benzo(b)fluoranthene

Concen: 0.381 ng

RT: 22.984 min Scan# 2582

Delta R.T. 0.003 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 29283

Ion Ratio Lower Upper

252 100

253 26.1 20.9 31.3

125 17.6 14.1 21.1



#38

Benzo(k)fluoranthene

Concen: 0.376 ng

RT: 23.028 min Scan# 2597

Delta R.T. -0.000 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13




Tgt Ion:252 Resp: 29088


Ion Ratio Lower Upper

252 100

253 26.5 21.2 31.8

125 18.3 14.6 22.0





#39

Benzo(a)pyrene

Concen: 0.384 ng

RT: 23.581 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 23635

Ion Ratio Lower Upper

252 100

253 28.9 23.1 34.7

125 23.3 18.6 28.0



#40

Dibenzo(a,h)anthracene

Concen: 0.378 ng

RT: 26.054 min Scan# 3632

Delta R.T. 0.006 min

Lab File: BN038113.D

Acq: 28 Oct 2025 14:13




Tgt Ion:278 Resp: 21487


Ion Ratio Lower Upper

278 100

139 22.0 17.6 26.4

279 28.2 22.6 33.8





#41

Benzo(g,h,i)perylene

Concen: 0.395 ng

RT: 26.747 min Scan# 3869

Delta R.T. -0.003 min

Lab File: BN038113.D


Acq: 28 Oct 2025 14:13

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:276 Resp: 25622

Ion Ratio Lower Upper

276 100

277 24.6 19.7 29.5

138 25.3 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038114.D
 Acq On : 28 Oct 2025 14:49
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 10/29/2025
 Supervised By :mohammad ahmed 10/30/2025

Quant Time: Oct 28 17:31:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	9110	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	27131	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	14920	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	28506	0.400	ng	0.00
29) Chrysene-d12	21.375	240	23535	0.400	ng	0.00
35) Perylene-d12	23.683	264	20307	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	16552	0.771	ng	0.00
5) Phenol-d6	6.937	99	19989	0.765	ng	0.00
8) Nitrobenzene-d5	8.929	82	16332	0.746	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	29806	0.769	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	4525	0.736	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	46843	0.784	ng	0.01
27) Fluoranthene-d10	19.220	212	55912	0.778	ng	0.00
31) Terphenyl-d14	19.819	244	37975	0.743	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.254	88	7707	0.819	ng	99
3) n-Nitrosodimethylamine	3.564	42	9598	0.792	ng	# 100
6) bis(2-Chloroethyl)ether	7.197	93	20444	0.796	ng	99
9) Naphthalene	10.626	128	58627	0.784	ng	99
10) Hexachlorobutadiene	10.925	225	9875	0.783	ng	# 98
12) 2-Methylnaphthalene	12.248	142	38674	0.776	ng	98
16) Acenaphthylene	14.152	152	52624	0.779	ng	99
17) Acenaphthene	14.494	154	36845	0.779	ng	100
18) Fluorene	15.489	166	49204	0.794	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	2898	0.689	ng	# 69
21) 4-Bromophenyl-phenylether	16.379	248	14571	0.780	ng	97
22) Hexachlorobenzene	16.503	284	16700	0.786	ng	99
23) Atrazine	16.652	200	10204	0.745	ng	99
24) Pentachlorophenol	16.838	266	6470	0.707	ng	99
25) Phenanthrene	17.223	178	70455	0.779	ng	100
26) Anthracene	17.322	178	60270	0.761	ng	100
28) Fluoranthene	19.253	202	79478	0.786	ng	100
30) Pyrene	19.615	202	78307	0.737	ng	100
32) Benzo(a)anthracene	21.357	228	64140	0.764	ng	100
33) Chrysene	21.411	228	71675	0.787	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	32963	0.745	ng	99
36) Indeno(1,2,3-cd)pyrene	26.039	276	65796	0.789	ng	99
37) Benzo(b)fluoranthene	22.984	252	69912m	0.805	ng	
38) Benzo(k)fluoranthene	23.031	252	67771	0.776	ng	# 93
39) Benzo(a)pyrene	23.578	252	53367	0.768	ng	# 90
40) Dibenzo(a,h)anthracene	26.054	278	50344	0.784	ng	95
41) Benzo(g,h,i)perylene	26.753	276	58155	0.793	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038114.D
 Acq On : 28 Oct 2025 14:49
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/29/2025

Supervised By :mohammad ahmed 10/30/2025


Quant Time: Oct 28 17:31:46 2025


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Oct 28 17:15:13 2025

Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8






Tgt Ion:152 Resp: 9110
Ion Ratio Lower Upper
152 100
150 157.1 122.3 183.5
115 63.4 47.4 71.0

Manual Integrations




APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#2
1,4-Dioxane
Concen: 0.819 ng
RT: 3.254 min Scan# 24
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion: 88 Resp: 7707
Ion Ratio Lower Upper
88 100
43 30.6 24.3 36.5
58 74.3 60.4 90.6





#3
n-Nitrosodimethylamine
Concen: 0.792 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion: 42 Resp: 9598
Ion Ratio Lower Upper
42 100
74 120.8 96.8 145.2
44 8.8 5.3 7.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#4
2-Fluorophenol
Concen: 0.771 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:112 Resp: 16552
Ion Ratio Lower Upper
112 100
64 55.7 45.4 68.0
63 30.2 23.2 34.8





#5
Phenol-d6
Concen: 0.765 ng
RT: 6.937 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion: 99 Resp: 19989
Ion Ratio Lower Upper
99 100
42 20.7 16.6 24.8
71 33.2 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.796 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion: 93 Resp: 20444
Ion Ratio Lower Upper
93 100
63 75.0 59.3 88.9
95 31.6 25.2 37.8





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:136 Resp: 2713
Ion Ratio Lower Upper
136 100
137 11.1 9.0 13.4
54 6.3 5.2 7.8
68 5.2 4.1 6.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#8
Nitrobenzene-d5
Concen: 0.746 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion: 82 Resp: 16332
Ion Ratio Lower Upper
82 100
128 41.9 34.1 51.1
54 53.4 43.5 65.3









#9
Naphthalene
Concen: 0.784 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:128 Resp: 5862
Ion Ratio Lower Upper
128 100
129 11.0 9.1 13.7
127 13.2 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#10
Hexachlorobutadiene
Concen: 0.783 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:225 Resp: 9875
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.8 50.3 75.5





#11
2-Methylnaphthalene-d10
Concen: 0.769 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:152 Resp: 29800
Ion Ratio Lower Upper
152 100
151 21.2 16.8 25.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#12
2-Methylnaphthalene
Concen: 0.776 ng
RT: 12.248 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:142 Resp: 38674
Ion Ratio Lower Upper
142 100
141 90.1 70.3 105.5
115 35.9 27.8 41.6





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8




Tgt Ion:164 Resp: 14920
Ion Ratio Lower Upper
164 100
162 104.0 83.0 124.4
160 51.9 40.7 61.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025




#14
2,4,6-Tribromophenol
Concen: 0.736 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Tgt Ion:330 Resp: 4525
Ion Ratio Lower Upper
330 100
332 95.3 76.6 114.8
141 41.9 34.1 51.1





#15
2-Fluorobiphenyl
Concen: 0.784 ng
RT: 13.062 min Scan# 11
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:172 Resp: 4684
Ion Ratio Lower Upper
172 100
171 34.3 27.8 41.6
170 22.4 18.1 27.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#16
Acenaphthylene
Concen: 0.779 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:152 Resp: 52624
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.8
153 12.9 10.2 15.2





#17
Acenaphthene
Concen: 0.779 ng
RT: 14.494 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:154 Resp: 3684
Ion Ratio Lower Upper
154 100
153 112.3 90.0 135.0
152 59.3 47.3 70.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#18
Fluorene
Concen: 0.794 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:166 Resp: 49204
Ion Ratio Lower Upper
166 100
165 98.7 79.0 118.4
167 13.3 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1688

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :





SSTDICC0.8

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	11.1	8.6	13.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#20

4,6-Dinitro-2-methylphenol

Concen: 0.689 ng




RT: 15.560 min Scan# 1557


Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Tgt Ion	Ratio	Lower	Upper
198	100		
51	69.5	88.9	133.3
105	49.4	49.2	73.8









#21
4-Bromophenyl-phenylether
Concen: 0.780 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:248 Resp: 1457
Ion Ratio Lower Upper
248 100
250 93.7 76.2 114.2
141 68.4 57.6 86.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#22
Hexachlorobenzene
Concen: 0.786 ng
RT: 16.503 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:284 Resp: 16700
Ion Ratio Lower Upper
284 100
142 38.0 30.5 45.7
249 30.3 23.5 35.3





#23
Atrazine
Concen: 0.745 ng
RT: 16.652 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:200 Resp: 10204
Ion Ratio Lower Upper
200 100
173 26.2 21.4 32.2
215 46.0 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#24
Pentachlorophenol
Concen: 0.707 ng
RT: 16.838 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49


Tgt Ion:266 Resp: 6470
Ion Ratio Lower Upper
266 100
264 62.0 49.3 73.9
268 64.3 50.6 75.8





#25
Phenanthrene
Concen: 0.779 ng
RT: 17.223 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:178 Resp: 7045
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.3 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#26
Anthracene
Concen: 0.761 ng
RT: 17.322 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:178 Resp: 60270
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.2 11.8 17.8





#27

Fluoranthene-d10

Concen: 0.778 ng

RT: 19.220 min Scan# 1947

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:	212	Resp:	5591
Ion Ratio	Lower	Upper	
212	100		
106	16.0	12.6	19.0
104	9.0	7.1	10.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/29/2025

Supervised By :mohammad ahmed 10/30/2025



#28

Fluoranthene

Concen: 0.786 ng




RT: 19.253 min Scan# 1954


Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Tgt Ion:	202	Resp:	79478
Ion Ratio	Lower	Upper	
202	100		
101	11.8	9.4	14.2
203	17.0	13.4	20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240	Resp:	2353
Ion Ratio	Lower	Upper
240	100	
120	13.1	10.7 16.1
236	28.4	23.1 34.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/29/2025

Supervised By :mohammad ahmed 10/30/2025



#30

Pyrene

Concen: 0.737 ng




RT: 19.615 min Scan# 2032


Delta R.T. 0.005 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49


Tgt Ion:202	Resp:	78307
Ion Ratio	Lower	Upper
202	100	
200	21.2	16.7 25.1
203	17.7	14.2 21.2





#31
Terphenyl-d14
Concen: 0.743 ng
RT: 19.819 min Scan# 2076
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:244 Resp: 37975
Ion Ratio Lower Upper
244 100
212 9.3 7.8 11.6
122 18.4 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#32
Benzo(a)anthracene
Concen: 0.764 ng
RT: 21.357 min Scan# 2267
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:228 Resp: 64140
Ion Ratio Lower Upper
228 100
226 27.7 22.2 33.4
229 19.6 15.8 23.8





#33

Chrysene

Concen: 0.787 ng

RT: 21.411 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :





SSTDICC0.8

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.3	36.5
229	19.6	15.7	23.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.745 ng




RT: 21.295 min Scan# 2260


Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.8	21.1	31.7
279	3.1	2.9	4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.683 min Scan# 2821

Delta R.T. 0.003 min

Lab File: BN038114.D


Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8






Tgt Ion:264 Resp: 20307

Ion	Ratio	Lower	Upper
264	100		
260	26.8	21.8	32.8
265	76.5	68.6	102.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/29/2025
Supervised By :mohammad ahmed 10/30/2025



#36

Indeno(1,2,3-cd)pyrene


Concen: 0.789 ng

RT: 26.039 min Scan# 3627

Delta R.T. 0.006 min



Lab File: BN038114.D


Acq: 28 Oct 2025 14:49



Tgt Ion:276 Resp: 65796

Ion	Ratio	Lower	Upper
276	100		
138	27.3	21.4	32.0
277	24.7	19.2	28.8





#37

Benzo(b)fluoranthene

Concen: 0.805 ng m

RT: 22.984 min Scan# 2582

Delta R.T. 0.003 min

Lab File: BN038114.D


Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8



Tgt Ion:252 Resp: 69912

Ion Ratio Lower Upper

252 100

253 23.4 20.9 31.3




125 14.0 14.1 21.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/29/2025

Supervised By :mohammad ahmed 10/30/2025



#38

Benzo(k)fluoranthene


Concen: 0.776 ng

RT: 23.031 min Scan# 2598

Delta R.T. 0.003 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49





Tgt Ion:252 Resp: 67771


Ion Ratio Lower Upper

252 100

253 23.4 21.2 31.8

125 14.2 14.6 22.0#





#39

Benzo(a)pyrene

Concen: 0.768 ng

RT: 23.578 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038114.D


Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8






Tgt Ion:	252	Resp:	5336
Ion Ratio	Lower	Upper	
252	100		
253	24.4	23.1	34.7
125	17.2	18.6	28.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/29/2025

Supervised By :mohammad ahmed 10/30/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.784 ng




RT: 26.054 min Scan# 3632


Delta R.T. 0.006 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Tgt Ion:	278	Resp:	50344
Ion Ratio	Lower	Upper	
278	100		
139	19.9	17.6	26.4
279	25.0	22.6	33.8





#41

Benzo(g,h,i)perylene

Concen: 0.793 ng

RT: 26.753 min Scan# 3871

Delta R.T. 0.003 min

Lab File: BN038114.D


Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8



Tgt Ion:276 Resp: 5815

Ion Ratio Lower Upper

276 100

277 24.0 19.7 29.5



138 24.9 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 10/29/2025

Supervised By :mohammad ahmed 10/30/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038115.D
 Acq On : 28 Oct 2025 15:25
 Operator : RC/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Oct 28 17:31:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	9872	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	29279	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	16624	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	31469	0.400	ng	0.00
29) Chrysene-d12	21.375	240	25173	0.400	ng	0.00
35) Perylene-d12	23.680	264	19529	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	36286	1.560	ng	0.00
5) Phenol-d6	6.937	99	44651	1.576	ng	0.00
8) Nitrobenzene-d5	8.929	82	36991	1.566	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	68673	1.642	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	11390	1.663	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	107386	1.613	ng	0.00
27) Fluoranthene-d10	19.220	212	132146	1.665	ng	0.00
31) Terphenyl-d14	19.819	244	89776	1.642	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.254	88	16443	1.612	ng	99
3) n-Nitrosodimethylamine	3.557	42	21596	1.645	ng	96
6) bis(2-Chloroethyl)ether	7.197	93	45315	1.628	ng	99
9) Naphthalene	10.626	128	131755	1.634	ng	99
10) Hexachlorobutadiene	10.925	225	22006	1.617	ng	# 99
12) 2-Methylnaphthalene	12.248	142	88687	1.649	ng	99
16) Acenaphthylene	14.152	152	125798	1.671	ng	99
17) Acenaphthene	14.495	154	87659	1.664	ng	98
18) Fluorene	15.489	166	117574	1.703	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	7997	1.534	ng	# 54
21) 4-Bromophenyl-phenylether	16.379	248	34434	1.669	ng	95
22) Hexachlorobenzene	16.503	284	38518	1.642	ng	99
23) Atrazine	16.652	200	25585	1.693	ng	97
24) Pentachlorophenol	16.838	266	16638	1.646	ng	99
25) Phenanthrene	17.223	178	166989	1.671	ng	100
26) Anthracene	17.322	178	148754	1.702	ng	100
28) Fluoranthene	19.253	202	187774	1.682	ng	99
30) Pyrene	19.610	202	184735	1.626	ng	100
32) Benzo(a)anthracene	21.358	228	147889	1.647	ng	99
33) Chrysene	21.411	228	156927	1.610	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	72607	1.535	ng	99
36) Indeno(1,2,3-cd)pyrene	26.031	276	135169	1.685	ng	99
37) Benzo(b)fluoranthene	22.981	252	141041	1.689	ng	# 91
38) Benzo(k)fluoranthene	23.028	252	143019	1.702	ng	# 90
39) Benzo(a)pyrene	23.578	252	110949	1.660	ng	# 85
40) Dibenzo(a,h)anthracene	26.048	278	105589	1.710	ng	92
41) Benzo(g,h,i)perylene	26.747	276	117071	1.660	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038115.D
Acq On : 28 Oct 2025 15:25
Operator : RC/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Quant Time: Oct 28 17:31:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:152 Resp: 9872
Ion Ratio Lower Upper
152 100
150 157.2 122.3 183.5
115 60.8 47.4 71.0



#2
1,4-Dioxane
Concen: 1.612 ng
RT: 3.254 min Scan# 24
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion: 88 Resp: 16443
Ion Ratio Lower Upper
88 100
43 30.7 24.3 36.5
58 76.6 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 1.645 ng
RT: 3.557 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 21596
Ion Ratio Lower Upper
42 100
74 115.8 96.8 145.2
44 6.3 5.3 7.9



#4
2-Fluorophenol
Concen: 1.560 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion: 112 Resp: 36286
Ion Ratio Lower Upper
112 100
64 56.2 45.4 68.0
63 29.9 23.2 34.8









#5
Phenol-d6
Concen: 1.576 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 44651
Ion Ratio Lower Upper
99 100
42 20.8 16.6 24.8
71 32.5 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 1.628 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion: 93 Resp: 45315
Ion Ratio Lower Upper
93 100
63 74.8 59.3 88.9
95 31.9 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	29279
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.4
54	6.1	5.2	7.8
68	5.0	4.1	6.1



#8
Nitrobenzene-d5
Concen: 1.566 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:	82	Resp:	36991
Ion	Ratio	Lower	Upper
82	100		
128	43.5	34.1	51.1
54	51.3	43.5	65.3



#9
Naphthalene

Concen: 1.634 ng

RT: 10.626 min Scan# 941

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:128 Resp: 131755

Ion Ratio Lower Upper

128 100

129 10.8 9.1 13.7

127 13.1 10.5 15.7



#10

Hexachlorobutadiene

Concen: 1.617 ng

RT: 10.925 min Scan# 969

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25




Tgt Ion:225 Resp: 22006


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.8 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 1.642 ng
RT: 12.177 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 68673
Ion Ratio Lower Upper
152 100
151 20.9 16.8 25.2



#12
2-Methylnaphthalene
Concen: 1.649 ng
RT: 12.248 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:142 Resp: 88687
Ion Ratio Lower Upper
142 100
141 88.5 70.3 105.5
115 34.6 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
164	100		
162	103.5	83.0	124.4
160	50.5	40.7	61.1



#14
2,4,6-Tribromophenol
Concen: 1.663 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion	Ratio	Lower	Upper
330	100		
332	97.4	76.6	114.8
141	41.9	34.1	51.1









#15
2-Fluorobiphenyl
Concen: 1.613 ng
RT: 13.051 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 107386
Ion Ratio Lower Upper
172 100
171 35.9 27.8 41.6
170 24.3 18.1 27.1



#16
Acenaphthylene
Concen: 1.671 ng
RT: 14.152 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:152 Resp: 125798
Ion Ratio Lower Upper
152 100
151 19.4 15.8 23.8
153 13.0 10.2 15.2









#17
Acenaphthene
Concen: 1.664 ng
RT: 14.495 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 87659
Ion Ratio Lower Upper
154 100
153 110.3 90.0 135.0
152 57.5 47.3 70.9



#18
Fluorene
Concen: 1.703 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:166 Resp: 117574
Ion Ratio Lower Upper
166 100
165 98.4 79.0 118.4
167 13.2 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:188 Resp: 31469

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.6 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 1.534 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25




Tgt Ion:198 Resp: 7997


Ion Ratio Lower Upper

198 100

51 49.4 88.9 133.3#

105 43.5 49.2 73.8#









#21
4-Bromophenyl-phenylether
Concen: 1.669 ng
RT: 16.379 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 34434
Ion Ratio Lower Upper
248 100
250 95.6 76.2 114.2
141 62.3 57.6 86.4



#22
Hexachlorobenzene
Concen: 1.642 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:284 Resp: 38518
Ion Ratio Lower Upper
284 100
142 38.3 30.5 45.7
249 30.0 23.5 35.3









#23
Atrazine
Concen: 1.693 ng
RT: 16.652 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
200	100		
173	23.2	21.4	32.2
215	45.1	36.7	55.1



#24
Pentachlorophenol
Concen: 1.646 ng
RT: 16.838 min Scan# 1660
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.5	49.3	73.9
268	63.5	50.6	75.8









#25
Phenanthrene
Concen: 1.671 ng
RT: 17.223 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 166989
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.2 12.2 18.4



#26
Anthracene
Concen: 1.702 ng
RT: 17.322 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:178 Resp: 148754
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.2 11.8 17.8





#27

Fluoranthene-d10

Concen: 1.665 ng

RT: 19.220 min Scan# 1947

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:212 Resp: 132146

Ion Ratio Lower Upper

212 100

106 16.0 12.6 19.0

104 8.9 7.1 10.7



#28

Fluoranthene

Concen: 1.682 ng

RT: 19.253 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25




Tgt Ion:202 Resp: 187774


Ion Ratio Lower Upper

202 100

101 11.9 9.4 14.2

203 17.2 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 25173

Ion Ratio Lower Upper

240 100

120 11.5 10.7 16.1

236 28.0 23.1 34.7



#30

Pyrene

Concen: 1.626 ng

RT: 19.610 min Scan# 2031

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25




Tgt Ion:202 Resp: 184735


Ion Ratio Lower Upper

202 100

200 21.0 16.7 25.1

203 17.7 14.2 21.2









#31
Terphenyl-d14
Concen: 1.642 ng
RT: 19.819 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 89776
Ion Ratio Lower Upper
244 100
212 9.1 7.8 11.6
122 17.8 15.2 22.8



#32
Benzo(a)anthracene
Concen: 1.647 ng
RT: 21.358 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:228 Resp: 147889
Ion Ratio Lower Upper
228 100
226 26.8 22.2 33.4
229 19.6 15.8 23.8





#33

Chrysene

Concen: 1.610 ng

RT: 21.411 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 156927

Ion Ratio Lower Upper

228 100

226 29.8 24.3 36.5

229 19.7 15.7 23.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.535 ng

RT: 21.295 min Scan# 2260

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25




Tgt Ion:149 Resp: 72607


Ion Ratio Lower Upper

149 100

167 26.6 21.1 31.7

279 3.0 2.9 4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.680 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:264 Resp: 19529

Ion Ratio Lower Upper

264 100

260 27.3 21.8 32.8

265 78.3 68.6 102.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 1.685 ng

RT: 26.031 min Scan# 3624

Delta R.T. -0.003 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25




Tgt Ion:276 Resp: 135169


Ion Ratio Lower Upper

276 100

138 27.6 21.4 32.0

277 24.5 19.2 28.8









#37
Benzo(b)fluoranthene
Concen: 1.689 ng
RT: 22.981 min Scan# 2581
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 141041
Ion Ratio Lower Upper
252 100
253 22.5 20.9 31.3
125 12.6 14.1 21.1#



#38
Benzo(k)fluoranthene
Concen: 1.702 ng
RT: 23.028 min Scan# 2597
Delta R.T. 0.000 min
Lab File: BN038115.D
Acq: 28 Oct 2025 15:25

Tgt Ion:252 Resp: 143019
Ion Ratio Lower Upper
252 100
253 22.2 21.2 31.8
125 12.5 14.6 22.0#





#39

Benzo(a)pyrene

Concen: 1.660 ng

RT: 23.578 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038115.D


Acq: 28 Oct 2025 15:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 110949

Ion Ratio Lower Upper

252 100

253 22.7 23.1 34.7#

125 14.4 18.6 28.0#



#40

Dibenzo(a,h)anthracene

Concen: 1.710 ng

RT: 26.048 min Scan# 3630

Delta R.T. 0.000 min

Lab File: BN038115.D

Acq: 28 Oct 2025 15:25




Tgt Ion:278 Resp: 105589


Ion Ratio Lower Upper

278 100

139 18.7 17.6 26.4

279 23.9 22.6 33.8





#41

Benzo(g,h,i)perylene

Concen: 1.660 ng

RT: 26.747 min Scan# 3869

Delta R.T. -0.003 min

Lab File: BN038115.D


Acq: 28 Oct 2025 15:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:276 Resp: 117071

Ion Ratio Lower Upper

276 100

277 23.9 19.7 29.5

138 24.2 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038116.D
 Acq On : 28 Oct 2025 16:02
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Oct 28 17:31:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	9743	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	28864	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	16547	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	31206	0.400	ng	0.00
29) Chrysene-d12	21.375	240	26115	0.400	ng	0.00
35) Perylene-d12	23.683	264	19904	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	69778	3.040	ng	0.00
5) Phenol-d6	6.937	99	88646	3.171	ng	0.00
8) Nitrobenzene-d5	8.929	82	70830	3.041	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	134773	3.270	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	24045	3.527	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	200437	3.024	ng	0.01
27) Fluoranthene-d10	19.220	212	263418	3.348	ng	0.00
31) Terphenyl-d14	19.819	244	180175	3.176	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.254	88	31207	3.100	ng	99
3) n-Nitrosodimethylamine	3.557	42	41593	3.211	ng	96
6) bis(2-Chloroethyl)ether	7.197	93	86904	3.164	ng	99
9) Naphthalene	10.626	128	255606	3.215	ng	99
10) Hexachlorobutadiene	10.925	225	42325	3.155	ng	# 99
12) 2-Methylnaphthalene	12.248	142	172992	3.263	ng	98
16) Acenaphthylene	14.152	152	253121	3.378	ng	99
17) Acenaphthene	14.495	154	173433	3.308	ng	97
18) Fluorene	15.489	166	230067	3.348	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	18221	3.363	ng	# 47
21) 4-Bromophenyl-phenylether	16.379	248	67616	3.305	ng	99
22) Hexachlorobenzene	16.503	284	74101	3.185	ng	99
23) Atrazine	16.652	200	52181	3.482	ng	97
24) Pentachlorophenol	16.838	266	36424	3.634	ng	99
25) Phenanthrene	17.223	178	322719	3.257	ng	100
26) Anthracene	17.322	178	295333	3.408	ng	99
28) Fluoranthene	19.253	202	372238	3.363	ng	99
30) Pyrene	19.615	202	368946	3.130	ng	100
32) Benzo(a)anthracene	21.358	228	308037	3.306	ng	99
33) Chrysene	21.411	228	317065	3.136	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	157480	3.209	ng	99
36) Indeno(1,2,3-cd)pyrene	26.031	276	276870	3.386	ng	98
37) Benzo(b)fluoranthene	22.984	252	292856	3.441	ng	# 90
38) Benzo(k)fluoranthene	23.031	252	292143	3.411	ng	# 89
39) Benzo(a)pyrene	23.578	252	231067	3.392	ng	# 84
40) Dibenzo(a,h)anthracene	26.051	278	217022	3.449	ng	92
41) Benzo(g,h,i)perylene	26.750	276	235264	3.274	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038116.D
Acq On : 28 Oct 2025 16:02
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Quant Time: Oct 28 17:31:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:152 Resp: 9743
Ion Ratio Lower Upper
152 100
150 157.0 122.3 183.5
115 61.3 47.4 71.0



#2
1,4-Dioxane
Concen: 3.100 ng
RT: 3.254 min Scan# 24
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion: 88 Resp: 31207
Ion Ratio Lower Upper
88 100
43 30.6 24.3 36.5
58 76.1 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 3.211 ng
RT: 3.557 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 41593
Ion Ratio Lower Upper
42 100
74 116.1 96.8 145.2
44 6.8 5.3 7.9



#4
2-Fluorophenol
Concen: 3.040 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion: 112 Resp: 69778
Ion Ratio Lower Upper
112 100
64 56.1 45.4 68.0
63 30.0 23.2 34.8









#5
Phenol-d6
Concen: 3.171 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 88646
Ion Ratio Lower Upper
99 100
42 20.7 16.6 24.8
71 32.4 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 3.164 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion: 93 Resp: 86904
Ion Ratio Lower Upper
93 100
63 75.1 59.3 88.9
95 32.0 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	28864
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.5	5.2	7.8
68	4.9	4.1	6.1



#8
Nitrobenzene-d5
Concen: 3.041 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:	82	Resp:	70830
Ion	Ratio	Lower	Upper
82	100		
128	42.8	34.1	51.1
54	51.5	43.5	65.3









#9
Naphthalene
Concen: 3.215 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 255606
Ion Ratio Lower Upper
128 100
129 10.9 9.1 13.7
127 13.0 10.5 15.7



#10
Hexachlorobutadiene
Concen: 3.155 ng
RT: 10.925 min Scan# 969
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:225 Resp: 42325
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 3.270 ng
RT: 12.177 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 134773
Ion Ratio Lower Upper
152 100
151 20.9 16.8 25.2



#12
2-Methylnaphthalene
Concen: 3.263 ng
RT: 12.248 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:142 Resp: 172992
Ion Ratio Lower Upper
142 100
141 90.0 70.3 105.5
115 35.2 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 16547
Ion Ratio Lower Upper
164 100
162 102.2 83.0 124.4
160 50.8 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 3.527 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:330 Resp: 24045
Ion Ratio Lower Upper
330 100
332 97.1 76.6 114.8
141 42.2 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 3.024 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 200437
Ion Ratio Lower Upper
172 100
171 34.2 27.8 41.6
170 22.3 18.1 27.1



#16
Acenaphthylene
Concen: 3.378 ng
RT: 14.152 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:152 Resp: 253121
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.8
153 12.9 10.2 15.2









#17
Acenaphthene
Concen: 3.308 ng
RT: 14.495 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 173433
Ion Ratio Lower Upper
154 100
153 109.4 90.0 135.0
152 56.0 47.3 70.9



#18
Fluorene
Concen: 3.348 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:166 Resp: 230067
Ion Ratio Lower Upper
166 100
165 99.1 79.0 118.4
167 13.4 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 31206

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.6 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 3.363 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02




Tgt Ion:198 Resp: 18221


Ion Ratio Lower Upper

198 100

51 40.8 88.9 133.3#

105 38.5 49.2 73.8#









#21
4-Bromophenyl-phenylether
Concen: 3.305 ng
RT: 16.379 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 67616
Ion Ratio Lower Upper
248 100
250 95.3 76.2 114.2
141 69.9 57.6 86.4



#22
Hexachlorobenzene
Concen: 3.185 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:284 Resp: 74101
Ion Ratio Lower Upper
284 100
142 38.7 30.5 45.7
249 29.9 23.5 35.3









#23
Atrazine
Concen: 3.482 ng
RT: 16.652 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 52181
Ion Ratio Lower Upper
200 100
173 24.0 21.4 32.2
215 45.3 36.7 55.1



#24
Pentachlorophenol
Concen: 3.634 ng
RT: 16.838 min Scan# 1660
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:266 Resp: 36424
Ion Ratio Lower Upper
266 100
264 62.5 49.3 73.9
268 63.4 50.6 75.8









#25
Phenanthrene
Concen: 3.257 ng
RT: 17.223 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 322719
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.3 12.2 18.4



#26
Anthracene
Concen: 3.408 ng
RT: 17.322 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:178 Resp: 295333
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.2 11.8 17.8





#27

Fluoranthene-d10

Concen: 3.348 ng

RT: 19.220 min Scan# 1947

Delta R.T. 0.000 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:212 Resp: 263418

Ion Ratio Lower Upper

212 100

106 15.8 12.6 19.0

104 8.9 7.1 10.7



#28

Fluoranthene

Concen: 3.363 ng

RT: 19.253 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02




Tgt Ion:202 Resp: 372238


Ion Ratio Lower Upper

202 100

101 11.8 9.4 14.2

203 17.2 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038116.D


Acq: 28 Oct 2025 16:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2






Tgt Ion:240 Resp: 26115

Ion Ratio Lower Upper

240 100

120 12.3 10.7 16.1

236 28.4 23.1 34.7



#30

Pyrene


Concen: 3.130 ng

RT: 19.615 min Scan# 2032

Delta R.T. 0.005 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02





Tgt Ion:202 Resp: 368946


Ion Ratio Lower Upper

202 100

200 21.1 16.7 25.1

203 17.7 14.2 21.2









#31
Terphenyl-d14
Concen: 3.176 ng
RT: 19.819 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 180175
Ion Ratio Lower Upper
244 100
212 9.0 7.8 11.6
122 17.3 15.2 22.8



#32
Benzo(a)anthracene
Concen: 3.306 ng
RT: 21.358 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BN038116.D
Acq: 28 Oct 2025 16:02

Tgt Ion:228 Resp: 308037
Ion Ratio Lower Upper
228 100
226 27.3 22.2 33.4
229 19.5 15.8 23.8





#33

Chrysene

Concen: 3.136 ng

RT: 21.411 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 317065

Ion Ratio Lower Upper

228 100

226 29.9 24.3 36.5

229 19.5 15.7 23.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 3.209 ng

RT: 21.295 min Scan# 2260

Delta R.T. 0.000 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02




Tgt Ion:149 Resp: 157480


Ion Ratio Lower Upper

149 100

167 26.9 21.1 31.7

279 3.0 2.9 4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.683 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:264 Resp: 19904

Ion Ratio Lower Upper

264 100

260 26.5 21.8 32.8

265 76.0 68.6 102.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.386 ng

RT: 26.031 min Scan# 3624

Delta R.T. -0.003 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02




Tgt Ion:276 Resp: 276870


Ion Ratio Lower Upper

276 100

138 27.4 21.4 32.0

277 24.8 19.2 28.8





#37

Benzo(b)fluoranthene

Concen: 3.441 ng

RT: 22.984 min Scan# 2582

Delta R.T. 0.003 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 292856

Ion Ratio Lower Upper

252 100

253 22.3 20.9 31.3

125 11.7 14.1 21.1#



#38

Benzo(k)fluoranthene

Concen: 3.411 ng

RT: 23.031 min Scan# 2598

Delta R.T. 0.003 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02




Tgt Ion:252 Resp: 292143


Ion Ratio Lower Upper

252 100

253 22.1 21.2 31.8

125 11.6 14.6 22.0#





#39

Benzo(a)pyrene

Concen: 3.392 ng

RT: 23.578 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038116.D


Acq: 28 Oct 2025 16:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 231067

Ion Ratio Lower Upper

252 100

253 22.1 23.1 34.7#

125 13.4 18.6 28.0#



#40

Dibenzo(a,h)anthracene

Concen: 3.449 ng

RT: 26.051 min Scan# 3631

Delta R.T. 0.003 min

Lab File: BN038116.D

Acq: 28 Oct 2025 16:02




Tgt Ion:278 Resp: 217022


Ion Ratio Lower Upper

278 100

139 18.4 17.6 26.4

279 23.7 22.6 33.8





#41

Benzo(g,h,i)perylene

Concen: 3.274 ng

RT: 26.750 min Scan# 3869

Delta R.T. 0.000 min

Lab File: BN038116.D


Acq: 28 Oct 2025 16:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:276 Resp: 235264

Ion Ratio Lower Upper

276 100

277 23.7 19.7 29.5

138 24.1 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038117.D
 Acq On : 28 Oct 2025 16:38
 Operator : RC/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Oct 28 17:31:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	10579	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	31197	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	17960	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	33242	0.400	ng	0.00
29) Chrysene-d12	21.375	240	25416	0.400	ng	0.00
35) Perylene-d12	23.680	264	18172	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	116608	4.679	ng	0.00
5) Phenol-d6	6.937	99	150343	4.953	ng	0.00
8) Nitrobenzene-d5	8.929	82	119764	4.757	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	225769	5.068	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	43520	5.881	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	335695	4.667	ng	0.00
27) Fluoranthene-d10	19.220	212	419371	5.003	ng	0.00
31) Terphenyl-d14	19.819	244	283126	5.127	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.254	88	51426	4.705	ng	100
3) n-Nitrosodimethylamine	3.557	42	68522	4.871	ng	96
6) bis(2-Chloroethyl)ether	7.197	93	142662	4.784	ng	99
9) Naphthalene	10.626	128	422794	4.920	ng	99
10) Hexachlorobutadiene	10.925	225	68839	4.748	ng	# 99
12) 2-Methylnaphthalene	12.248	142	287955	5.025	ng	99
16) Acenaphthylene	14.152	152	428380	5.267	ng	99
17) Acenaphthene	14.494	154	291166	5.117	ng	95
18) Fluorene	15.489	166	389813	5.227	ng	98
21) 4-Bromophenyl-phenylether	16.379	248	114170	5.239	ng	# 92
22) Hexachlorobenzene	16.503	284	123999	5.003	ng	99
23) Atrazine	16.652	200	87823	5.501	ng	97
24) Pentachlorophenol	16.838	266	66071	6.188	ng	99
25) Phenanthrene	17.223	178	539294	5.110	ng	100
26) Anthracene	17.322	178	499584	5.411	ng	99
28) Fluoranthene	19.253	202	587710	4.984	ng	99
30) Pyrene	19.610	202	586526	5.112	ng	100
32) Benzo(a)anthracene	21.358	228	462496	5.100	ng	100
33) Chrysene	21.411	228	473389	4.811	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	223481	4.679	ng	99
36) Indeno(1,2,3-cd)pyrene	26.034	276	403825	5.409	ng	99
37) Benzo(b)fluoranthene	22.981	252	413811	5.326	ng	# 90
38) Benzo(k)fluoranthene	23.028	252	421255	5.387	ng	# 88
39) Benzo(a)pyrene	23.578	252	332107	5.340	ng	# 83
40) Dibenzo(a,h)anthracene	26.048	278	318538	5.545	ng	91
41) Benzo(g,h,i)perylene	26.750	276	342713	5.223	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038117.D
 Acq On : 28 Oct 2025 16:38
 Operator : RC/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Quant Time: Oct 28 17:31:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0







Tgt Ion:152 Resp: 10579
Ion Ratio Lower Upper
152 100
150 155.6 122.3 183.5
115 60.1 47.4 71.0



#2
1,4-Dioxane
Concen: 4.705 ng
RT: 3.254 min Scan# 24
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion: 88 Resp: 51426
Ion Ratio Lower Upper
88 100
43 30.5 24.3 36.5
58 75.4 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 4.871 ng
RT: 3.557 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 68522
Ion Ratio Lower Upper
42 100
74 116.1 96.8 145.2
44 6.5 5.3 7.9



#4
2-Fluorophenol
Concen: 4.679 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion: 112 Resp: 116608
Ion Ratio Lower Upper
112 100
64 56.1 45.4 68.0
63 29.9 23.2 34.8









#5
Phenol-d6
Concen: 4.953 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 150343
Ion Ratio Lower Upper
99 100
42 20.9 16.6 24.8
71 32.5 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 4.784 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion: 93 Resp: 142662
Ion Ratio Lower Upper
93 100
63 74.9 59.3 88.9
95 31.9 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:136 Resp: 31197
Ion Ratio Lower Upper
136 100
137 11.4 9.0 13.4
54 9.9 5.2 7.8#
68 7.3 4.1 6.1#



#8
Nitrobenzene-d5
Concen: 4.757 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion: 82 Resp: 119764
Ion Ratio Lower Upper
82 100
128 44.7 34.1 51.1
54 50.9 43.5 65.3









#9
Naphthalene
Concen: 4.920 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 422794
Ion Ratio Lower Upper
128 100
129 10.9 9.1 13.7
127 13.0 10.5 15.7



#10
Hexachlorobutadiene
Concen: 4.748 ng
RT: 10.925 min Scan# 969
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:225 Resp: 68839
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 5.068 ng
RT: 12.177 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 225769
Ion Ratio Lower Upper
152 100
151 20.8 16.8 25.2



#12
2-Methylnaphthalene
Concen: 5.025 ng
RT: 12.248 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:142 Resp: 287955
Ion Ratio Lower Upper
142 100
141 89.6 70.3 105.5
115 34.8 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.430 min Scan# 1453
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 17960
Ion Ratio Lower Upper
164 100
162 106.0 83.0 124.4
160 52.4 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 5.881 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:330 Resp: 43520
Ion Ratio Lower Upper
330 100
332 97.0 76.6 114.8
141 41.2 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 4.667 ng
RT: 13.051 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 335695
Ion Ratio Lower Upper
172 100
171 35.8 27.8 41.6
170 24.2 18.1 27.1



#16
Acenaphthylene
Concen: 5.267 ng
RT: 14.152 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:152 Resp: 428380
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.8
153 13.0 10.2 15.2









#17
Acenaphthene
Concen: 5.117 ng
RT: 14.494 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 291166
Ion Ratio Lower Upper
154 100
153 108.1 90.0 135.0
152 55.0 47.3 70.9



#18
Fluorene
Concen: 5.227 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:166 Resp: 389813
Ion Ratio Lower Upper
166 100
165 97.0 79.0 118.4
167 13.1 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:188 Resp: 33242

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.4 8.6 13.0



#21

4-Bromophenyl-phenylether

Concen: 5.239 ng

RT: 16.379 min Scan# 1623

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38




Tgt Ion:248 Resp: 114170


Ion Ratio Lower Upper

248 100

250 95.7 76.2 114.2

141 56.3 57.6 86.4#









#22
Hexachlorobenzene
Concen: 5.003 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:284 Resp: 123999
Ion Ratio Lower Upper
284 100
142 37.6 30.5 45.7
249 29.6 23.5 35.3



#23
Atrazine
Concen: 5.501 ng
RT: 16.652 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:200 Resp: 87823
Ion Ratio Lower Upper
200 100
173 23.4 21.4 32.2
215 45.1 36.7 55.1









#24
Pentachlorophenol
Concen: 6.188 ng
RT: 16.838 min Scan# 1660
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion	266	Resp	66071
Ion Ratio	Lower	Upper	
266	100		
264	62.7	49.3	73.9
268	64.1	50.6	75.8



#25
Phenanthrene
Concen: 5.110 ng
RT: 17.223 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion	178	Resp	539294
Ion Ratio	Lower	Upper	
178	100		
176	19.2	15.5	23.3
179	15.1	12.2	18.4









#26
Anthracene
Concen: 5.411 ng
RT: 17.322 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 499584
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.3 11.8 17.8



#27
Fluoranthene-d10
Concen: 5.003 ng
RT: 19.220 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:212 Resp: 419371
Ion Ratio Lower Upper
212 100
106 15.8 12.6 19.0
104 8.9 7.1 10.7









#28
Fluoranthene
Concen: 4.984 ng
RT: 19.253 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 587710
Ion Ratio Lower Upper
202 100
101 11.7 9.4 14.2
203 17.3 13.4 20.2



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:240 Resp: 25416
Ion Ratio Lower Upper
240 100
120 11.7 10.7 16.1
236 28.2 23.1 34.7









#30
Pyrene
Concen: 5.112 ng
RT: 19.610 min Scan# 2032
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 586526
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.8 14.2 21.2



#31
Terphenyl-d14
Concen: 5.127 ng
RT: 19.819 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:244 Resp: 283126
Ion Ratio Lower Upper
244 100
212 8.9 7.8 11.6
122 17.3 15.2 22.8









#32
Benzo(a)anthracene
Concen: 5.100 ng
RT: 21.358 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 462496
Ion Ratio Lower Upper
228 100
226 27.4 22.2 33.4
229 19.8 15.8 23.8



#33
Chrysene
Concen: 4.811 ng
RT: 21.411 min Scan# 2273
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:228 Resp: 473389
Ion Ratio Lower Upper
228 100
226 29.8 24.3 36.5
229 19.8 15.7 23.5









#34
Bis(2-ethylhexyl)phthalate
Concen: 4.679 ng
RT: 21.295 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:149 Resp: 223481
Ion Ratio Lower Upper
149 100
167 27.0 21.1 31.7
279 3.1 2.9 4.3



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.680 min Scan# 2820
Delta R.T. 0.000 min
Lab File: BN038117.D
Acq: 28 Oct 2025 16:38

Tgt Ion:264 Resp: 18172
Ion Ratio Lower Upper
264 100
260 27.2 21.8 32.8
265 82.9 68.6 102.8





#36

Indeno(1,2,3-cd)pyrene

Concen: 5.409 ng

RT: 26.034 min Scan# 3626

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:276 Resp: 403825

Ion Ratio Lower Upper

276 100

138 27.3 21.4 32.0

277 24.9 19.2 28.8



#37

Benzo(b)fluoranthene

Concen: 5.326 ng

RT: 22.981 min Scan# 2581

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38




Tgt Ion:252 Resp: 413811


Ion Ratio Lower Upper

252 100

253 22.0 20.9 31.3

125 11.5 14.1 21.1#





#38

Benzo(k)fluoranthene

Concen: 5.387 ng

RT: 23.028 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:252 Resp: 421255

Ion Ratio Lower Upper

252 100

253 21.8 21.2 31.8

125 11.4 14.6 22.0#



#39

Benzo(a)pyrene

Concen: 5.340 ng

RT: 23.578 min Scan# 2785

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38




Tgt Ion:252 Resp: 332107


Ion Ratio Lower Upper

252 100

253 21.9 23.1 34.7#

125 12.9 18.6 28.0#





#40

Dibenzo(a,h)anthracene

Concen: 5.545 ng

RT: 26.048 min Scan# 30

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:278 Resp: 318538

Ion Ratio Lower Upper

278 100

139 18.2 17.6 26.4

279 23.3 22.6 33.8



#41

Benzo(g,h,i)perylene

Concen: 5.223 ng

RT: 26.750 min Scan# 3870

Delta R.T. 0.000 min

Lab File: BN038117.D

Acq: 28 Oct 2025 16:38




Tgt Ion:276 Resp: 342713

Ion Ratio Lower Upper

276 100

277 23.6 19.7 29.5

138 23.9 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038118.D
 Acq On : 28 Oct 2025 17:51
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102825

Quant Time: Oct 29 17:07:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	7698	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	23240	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	13297	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	26273	0.400	ng	0.00
29) Chrysene-d12	21.375	240	18729	0.400	ng	0.00
35) Perylene-d12	23.683	264	15660	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	6665	0.368	ng	0.00
5) Phenol-d6	6.937	99	7743	0.351	ng	0.00
8) Nitrobenzene-d5	8.929	82	6617	0.353	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	13607	0.410	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1861	0.340	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	21897	0.411	ng	0.01
27) Fluoranthene-d10	19.220	212	25766	0.389	ng	0.00
31) Terphenyl-d14	19.819	244	17430	0.428	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.254	88	3503	0.440	ng	98
3) n-Nitrosodimethylamine	3.564	42	4071	0.398	ng	# 96
6) bis(2-Chloroethyl)ether	7.197	93	9255	0.427	ng	100
9) Naphthalene	10.626	128	26051	0.407	ng	99
10) Hexachlorobutadiene	10.925	225	4702	0.435	ng	# 98
12) 2-Methylnaphthalene	12.253	142	16372	0.384	ng	98
16) Acenaphthylene	14.152	152	27388	0.455	ng	100
17) Acenaphthene	14.505	154	16777	0.398	ng	99
18) Fluorene	15.489	166	22859	0.414	ng	99
20) 4,6-Dinitro-2-methylph...	15.560	198	1138	0.411	ng	90
21) 4-Bromophenyl-phenylether	16.379	248	6886	0.400	ng	99
22) Hexachlorobenzene	16.503	284	8349	0.426	ng	98
23) Atrazine	16.652	200	4672	0.370	ng	99
24) Pentachlorophenol	16.838	266	2963	0.351	ng	97
25) Phenanthrene	17.223	178	33992	0.408	ng	99
26) Anthracene	17.322	178	29654	0.406	ng	99
28) Fluoranthene	19.253	202	35401	0.380	ng	99
30) Pyrene	19.615	202	35647	0.422	ng	100
32) Benzo(a)anthracene	21.358	228	26833	0.402	ng	100
33) Chrysene	21.411	228	31031	0.428	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	12587	0.358	ng	99
36) Indeno(1,2,3-cd)pyrene	26.037	276	26115	0.406	ng	98
37) Benzo(b)fluoranthene	22.984	252	27124	0.405	ng	97
38) Benzo(k)fluoranthene	23.031	252	29583	0.439	ng	98
39) Benzo(a)pyrene	23.581	252	23400	0.437	ng	# 95
40) Dibenzo(a,h)anthracene	26.051	278	19931	0.403	ng	97
41) Benzo(g,h,i)perylene	26.756	276	22180	0.392	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038118.D
Acq On : 28 Oct 2025 17:51
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Quant Time: Oct 29 17:07:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.782 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51


Instrument :
BNA_N
ClientSampleId :
ICVBN102825





Tgt Ion:152 Resp: 7698
Ion Ratio Lower Upper
152 100
150 151.3 122.3 183.5
115 57.8 47.4 71.0




#2
1,4-Dioxane
Concen: 0.440 ng
RT: 3.254 min Scan# 24
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51



Tgt Ion: 88 Resp: 3503
Ion Ratio Lower Upper
88 100
43 31.0 24.3 36.5
58 77.1 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 0.398 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion: 42 Resp: 4071
Ion Ratio Lower Upper
42 100
74 124.8 96.8 145.2
44 10.5 5.3 7.9#



#4
2-Fluorophenol
Concen: 0.368 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion: 112 Resp: 6665
Ion Ratio Lower Upper
112 100
64 55.4 45.4 68.0
63 29.1 23.2 34.8









#5
Phenol-d6
Concen: 0.351 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion: 99 Resp: 7743
Ion Ratio Lower Upper
99 100
42 21.1 16.6 24.8
71 33.2 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.427 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion: 93 Resp: 9255
Ion Ratio Lower Upper
93 100
63 73.9 59.3 88.9
95 30.9 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:	136	Resp:	23240
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	6.6	5.2	7.8
68	5.3	4.1	6.1



#8
Nitrobenzene-d5
Concen: 0.353 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:	82	Resp:	6617
Ion	Ratio	Lower	Upper
82	100		
128	41.8	34.1	51.1
54	54.7	43.5	65.3









#9
Naphthalene
Concen: 0.407 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion	128	129	127
Resp:	26051		
Ion Ratio	100	11.3	13.4
Lower		9.1	10.5
Upper		13.7	15.7



#10
Hexachlorobutadiene
Concen: 0.435 ng
RT: 10.925 min Scan# 969
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion	225	223	227
Resp:	4702		
Ion Ratio	100	0.0	64.4
Lower		0.0	50.3
Upper		0.0	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.410 ng
RT: 12.177 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:152 Resp: 13607
Ion Ratio Lower Upper
152 100
151 21.0 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.384 ng
RT: 12.253 min Scan# 1170
Delta R.T. 0.005 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:142 Resp: 16372
Ion Ratio Lower Upper
142 100
141 89.5 70.3 105.5
115 35.3 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:164 Resp: 13297
Ion Ratio Lower Upper
164 100
162 103.7 83.0 124.4
160 52.0 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 0.340 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:330 Resp: 1861
Ion Ratio Lower Upper
330 100
332 94.5 76.6 114.8
141 43.0 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.411 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.5	27.8	41.6
170	22.9	18.1	27.1



#16
Acenaphthylene
Concen: 0.455 ng
RT: 14.152 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.8	23.8
153	13.2	10.2	15.2









#17
Acenaphthene
Concen: 0.398 ng
RT: 14.505 min Scan# 1459
Delta R.T. 0.011 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:154 Resp: 16777
Ion Ratio Lower Upper
154 100
153 113.6 90.0 135.0
152 60.6 47.3 70.9



#18
Fluorene
Concen: 0.414 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:166 Resp: 22859
Ion Ratio Lower Upper
166 100
165 99.9 79.0 118.4
167 13.4 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1688

Delta R.T. -0.000 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51

Instrument :

BNA_N

ClientSampleId :

ICVBN102825





Tgt Ion:188 Resp: 26273

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.7 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.411 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51




Tgt Ion:198 Resp: 1138


Ion Ratio Lower Upper

198 100

51 125.7 88.9 133.3

105 63.1 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.400 ng
RT: 16.379 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion	Ratio	Lower	Upper
248	100		
250	94.8	76.2	114.2
141	73.0	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.426 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion	Ratio	Lower	Upper
284	100		
142	39.2	30.5	45.7
249	30.0	23.5	35.3









#23
Atrazine
Concen: 0.370 ng
RT: 16.652 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:200 Resp: 4672
Ion Ratio Lower Upper
200 100
173 27.7 21.4 32.2
215 45.3 36.7 55.1



#24
Pentachlorophenol
Concen: 0.351 ng
RT: 16.838 min Scan# 1660
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:266 Resp: 2963
Ion Ratio Lower Upper
266 100
264 62.7 49.3 73.9
268 66.0 50.6 75.8









#25
Phenanthrene
Concen: 0.408 ng
RT: 17.223 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:178 Resp: 33992
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.7 12.2 18.4



#26
Anthracene
Concen: 0.406 ng
RT: 17.322 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:178 Resp: 29654
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.3 11.8 17.8





#27

Fluoranthene-d10

Concen: 0.389 ng

RT: 19.220 min Scan# 1947

Delta R.T. 0.000 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51

Instrument :

BNA_N

ClientSampleId :

ICVBN102825





Tgt Ion:212 Resp: 25766

Ion Ratio Lower Upper

212 100

106 16.1 12.6 19.0

104 9.0 7.1 10.7



#28

Fluoranthene

Concen: 0.380 ng

RT: 19.253 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51




Tgt Ion:202 Resp: 35401


Ion Ratio Lower Upper

202 100

101 11.5 9.4 14.2

203 17.1 13.4 20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion	Ratio	Lower	Upper
240	100		
120	14.0	10.7	16.1
236	28.8	23.1	34.7



#30
Pyrene
Concen: 0.422 ng
RT: 19.615 min Scan# 2032
Delta R.T. 0.005 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.7	25.1
203	17.6	14.2	21.2









#31
Terphenyl-d14
Concen: 0.428 ng
RT: 19.819 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51




Instrument :
BNA_N
ClientSampleId :
ICVBN102825


Tgt Ion:244 Resp: 17430
Ion Ratio Lower Upper
244 100
212 9.6 7.8 11.6
122 19.1 15.2 22.8



#32
Benzo(a)anthracene
Concen: 0.402 ng
RT: 21.358 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BN038118.D
Acq: 28 Oct 2025 17:51

Tgt Ion:228 Resp: 26833
Ion Ratio Lower Upper
228 100
226 28.0 22.2 33.4
229 19.7 15.8 23.8





#33

Chrysene

Concen: 0.428 ng

RT: 21.411 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51

Instrument :

BNA_N

ClientSampleId :

ICVBN102825





Tgt Ion:228 Resp: 31031

Ion Ratio Lower Upper

228 100

226 30.4 24.3 36.5

229 19.6 15.7 23.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.358 ng

RT: 21.295 min Scan# 2260

Delta R.T. 0.000 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51




Tgt Ion:149 Resp: 12587


Ion Ratio Lower Upper

149 100

167 27.0 21.1 31.7

279 3.4 2.9 4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.683 min Scan# 2821

Delta R.T. 0.003 min

Lab File: BN038118.D


Acq: 28 Oct 2025 17:51

Instrument :

BNA_N

ClientSampleId :

ICVBN102825





Tgt Ion:264 Resp: 15660

Ion Ratio Lower Upper

264 100

260 27.0 21.8 32.8

265 93.6 68.6 102.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.406 ng

RT: 26.037 min Scan# 3626

Delta R.T. 0.003 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51




Tgt Ion:276 Resp: 26115


Ion Ratio Lower Upper

276 100

138 27.9 21.4 32.0

277 24.8 19.2 28.8





#37

Benzo(b)fluoranthene

Concen: 0.405 ng

RT: 22.984 min Scan# 2582

Delta R.T. 0.003 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51

Instrument :

BNA_N

ClientSampleId :

ICVBN102825





Tgt Ion:252 Resp: 27124

Ion Ratio Lower Upper

252 100

253 25.3 20.9 31.3

125 19.9 14.1 21.1



#38

Benzo(k)fluoranthene

Concen: 0.439 ng

RT: 23.031 min Scan# 2598

Delta R.T. 0.003 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51




Tgt Ion:252 Resp: 29583


Ion Ratio Lower Upper

252 100

253 25.4 21.2 31.8

125 18.0 14.6 22.0





#39

Benzo(a)pyrene

Concen: 0.437 ng

RT: 23.581 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN038118.D


Acq: 28 Oct 2025 17:51

Instrument :



BNA_N

ClientSampleId :

ICVBN102825



Tgt Ion:	252	Resp:	23400
Ion Ratio	Lower	Upper	
252	100		
253	27.8	23.1	34.7
125	28.0	18.6	28.0



#40

Dibenzo(a,h)anthracene

Concen: 0.403 ng




RT: 26.051 min Scan# 3631


Delta R.T. 0.003 min

Lab File: BN038118.D

Acq: 28 Oct 2025 17:51

Tgt Ion:	278	Resp:	19931
Ion Ratio	Lower	Upper	
278	100		
139	20.2	17.6	26.4
279	27.0	22.6	33.8





#41

Benzo(g,h,i)perylene

Concen: 0.392 ng

RT: 26.756 min Scan# 3872

Delta R.T. 0.006 min

Lab File: BN038118.D


Acq: 28 Oct 2025 17:51

Instrument :

BNA_N

ClientSampleId :

ICVBN102825





Tgt Ion:276 Resp: 22180

Ion Ratio Lower Upper

276 100

277 25.7 19.7 29.5

138 25.7 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038118.D
 Acq On : 28 Oct 2025 17:51
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102825

Quant Time: Oct 29 17:07:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.413	0.455	-10.2	90	0.00
3	n-Nitrosodimethylamine	0.532	0.529	0.6	82	0.00
4 S	2-Fluorophenol	0.942	0.866	8.1	76	0.00
5 S	Phenol-d6	1.148	1.006	12.4	74	0.00
6	bis(2-Chloroethyl)ether	1.128	1.202	-6.6	89	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	84	0.01
8 S	Nitrobenzene-d5	0.323	0.285	11.8	76	0.00
9	Naphthalene	1.102	1.121	-1.7	87	0.00
10	Hexachlorobutadiene	0.186	0.202	-8.6	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.571	0.585	-2.5	89	0.00
12	2-Methylnaphthalene	0.735	0.704	4.2	84	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.01
14 S	2,4,6-Tribromophenol	0.165	0.140	15.2	81	0.00
15 S	2-Fluorobiphenyl	1.602	1.647	-2.8	90	0.01
16	Acenaphthylene	1.811	2.060	-13.7	103	0.00
17	Acenaphthene	1.267	1.262	0.4	89	0.01
18	Fluorene	1.661	1.719	-3.5	92	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.043	21.8	86	0.00
21	4-Bromophenyl-phenylether	0.262	0.262	0.0	93	0.00
22	Hexachlorobenzene	0.298	0.318	-6.7	95	0.00
23	Atrazine	0.192	0.178	7.3	90	0.00
24	Pentachlorophenol	0.128	0.113	11.7	93	0.00
25	Phenanthrene	1.270	1.294	-1.9	94	0.00
26	Anthracene	1.111	1.129	-1.6	96	0.00
27 SURR	Fluoranthene-d10	1.009	0.981	2.8	91	0.00
28	Fluoranthene	1.419	1.347	5.1	89	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
30	Pyrene	1.806	1.903	-5.4	90	0.00
31 S	Terphenyl-d14	0.869	0.931	-7.1	93	0.00
32	Benzo(a)anthracene	1.427	1.433	-0.4	92	0.00
33	Chrysene	1.549	1.657	-7.0	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.752	0.672	10.6	81	0.00
35 I	Perylene-d12	1.000	1.000	0.0	87	0.00
36	Indeno(1,2,3-cd)pyrene	1.643	1.668	-1.5	91	0.00
37	Benzo(b)fluoranthene	1.710	1.732	-1.3	93	0.00
38	Benzo(k)fluoranthene	1.721	1.889	-9.8	102	0.00
39 C	Benzo(a)pyrene	1.369	1.494	-9.1	99	0.00
40	Dibenzo(a,h)anthracene	1.265	1.273	-0.6	93	0.00
41	Benzo(g,h,i)perylene	1.444	1.416	1.9	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN102825

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038118.D
 Acq On : 28 Oct 2025 17:51
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102825

Quant Time: Oct 29 17:07:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.00
2	1,4-Dioxane	0.400	0.440	-10.0	90	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	82	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	76	0.00
5 S	Phenol-d6	0.400	0.351	12.3	74	0.00
6	bis(2-Chloroethyl)ether	0.400	0.427	-6.7	89	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	84	0.01
8 S	Nitrobenzene-d5	0.400	0.353	11.8	76	0.00
9	Naphthalene	0.400	0.407	-1.7	87	0.00
10	Hexachlorobutadiene	0.400	0.435	-8.7	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	89	0.00
12	2-Methylnaphthalene	0.400	0.384	4.0	84	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	0.01
14 S	2,4,6-Tribromophenol	0.400	0.340	15.0	81	0.00
15 S	2-Fluorobiphenyl	0.400	0.411	-2.7	90	0.01
16	Acenaphthylene	0.400	0.455	-13.7	103	0.00
17	Acenaphthene	0.400	0.398	0.5	89	0.01
18	Fluorene	0.400	0.414	-3.5	92	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.411	-2.7	86	0.00
21	4-Bromophenyl-phenylether	0.400	0.400	0.0	93	0.00
22	Hexachlorobenzene	0.400	0.426	-6.5	95	0.00
23	Atrazine	0.400	0.370	7.5	90	0.00
24	Pentachlorophenol	0.400	0.351	12.3	93	0.00
25	Phenanthrene	0.400	0.408	-2.0	94	0.00
26	Anthracene	0.400	0.406	-1.5	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.389	2.8	91	0.00
28	Fluoranthene	0.400	0.380	5.0	89	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	87	0.00
30	Pyrene	0.400	0.422	-5.5	90	0.00
31 S	Terphenyl-d14	0.400	0.428	-7.0	93	0.00
32	Benzo(a)anthracene	0.400	0.402	-0.5	92	0.00
33	Chrysene	0.400	0.428	-7.0	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.358	10.5	81	0.00
35 I	Perylene-d12	0.400	0.400	0.0	87	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.406	-1.5	91	0.00
37	Benzo(b)fluoranthene	0.400	0.405	-1.3	93	0.00
38	Benzo(k)fluoranthene	0.400	0.439	-9.7	102	0.00
39 C	Benzo(a)pyrene	0.400	0.437	-9.2	99	0.00
40	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	93	0.00
41	Benzo(g,h,i)perylene	0.400	0.392	2.0	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN102825



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3526</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>11/04/2025 10:40</u>
Lab File ID:	<u>BN038130.D</u>	Init. Calib. Date(s):	<u>10/28/2025 10/28/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4</u>	Init. Calib. Time(s):	<u>13:00 16:38</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.571	0.541		-5.3	20.0
Fluoranthene-d10	1.008	0.889		-11.8	20.0
2-Fluorophenol	0.943	0.856		-9.2	20.0
Phenol-d6	1.148	1.025		-10.7	20.0
Nitrobenzene-d5	0.323	0.277		-14.2	20.0
2-Fluorobiphenyl	1.602	1.514		-5.5	20.0
2,4,6-Tribromophenol	0.165	0.134		-18.8	20.0
Terphenyl-d14	0.869	0.916		5.4	20.0
1,4-Dioxane	0.416	0.433		4.1	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038130.D
 Acq On : 04 Nov 2025 10:40
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Nov 04 11:42:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8599	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	24973	0.400	ng	# 0.00
13) Acenaphthene-d10	14.441	164	13947	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	27056	0.400	ng	0.00
29) Chrysene-d12	21.375	240	17059	0.400	ng	0.00
35) Perylene-d12	23.686	264	13607	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	7357	0.363	ng	0.00
5) Phenol-d6	6.937	99	8817	0.357	ng	0.00
8) Nitrobenzene-d5	8.928	82	6906	0.343	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	13509	0.379	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1874	0.326	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	21117	0.378	ng	0.01
27) Fluoranthene-d10	19.220	212	24042	0.352	ng	0.00
31) Terphenyl-d14	19.819	244	15620	0.421	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.254	88	3722	0.419	ng	99
3) n-Nitrosodimethylamine	3.564	42	4544	0.397	ng	# 93
6) bis(2-Chloroethyl)ether	7.197	93	9558	0.394	ng	99
9) Naphthalene	10.626	128	27095	0.394	ng	100
10) Hexachlorobutadiene	10.925	225	4747	0.409	ng	# 99
12) 2-Methylnaphthalene	12.248	142	17751	0.387	ng	98
16) Acenaphthylene	14.152	152	24464	0.387	ng	99
17) Acenaphthene	14.494	154	16888	0.382	ng	96
18) Fluorene	15.489	166	22996	0.397	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	1019	0.381	ng	# 84
21) 4-Bromophenyl-phenylether	16.379	248	6844	0.386	ng	98
22) Hexachlorobenzene	16.503	284	8492	0.421	ng	100
23) Atrazine	16.652	200	4361	0.336	ng	98
24) Pentachlorophenol	16.838	266	2516	0.290	ng	98
25) Phenanthrene	17.223	178	33589	0.391	ng	100
26) Anthracene	17.322	178	27861	0.371	ng	99
28) Fluoranthene	19.252	202	35123	0.366	ng	100
30) Pyrene	19.615	202	34116	0.443	ng	100
32) Benzo(a)anthracene	21.357	228	22233	0.365	ng	100
33) Chrysene	21.411	228	26943	0.408	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	12992	0.405	ng	100
36) Indeno(1,2,3-cd)pyrene	26.036	276	23804	0.426	ng	99
37) Benzo(b)fluoranthene	22.987	252	23288	0.400	ng	98
38) Benzo(k)fluoranthene	23.031	252	23877	0.408	ng	99
39) Benzo(a)pyrene	23.583	252	18445	0.396	ng	98
40) Dibenzo(a,h)anthracene	26.054	278	18347	0.427	ng	99
41) Benzo(g,h,i)perylene	26.753	276	21656	0.441	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038130.D
Acq On : 04 Nov 2025 10:40
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Quant Time: Nov 04 11:42:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:152 Resp: 8599
Ion Ratio Lower Upper
152 100
150 153.9 122.3 183.5
115 59.4 47.4 71.0



#2
1,4-Dioxane
Concen: 0.419 ng
RT: 3.254 min Scan# 24
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion: 88 Resp: 3722
Ion Ratio Lower Upper
88 100
43 31.1 24.3 36.5
58 76.5 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 0.397 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 4544
Ion Ratio Lower Upper
42 100
74 112.7 96.8 145.2
44 9.0 5.3 7.9#



#4
2-Fluorophenol
Concen: 0.363 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion: 112 Resp: 7357
Ion Ratio Lower Upper
112 100
64 55.3 45.4 68.0
63 29.7 23.2 34.8









#5
Phenol-d6
Concen: 0.357 ng
RT: 6.937 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
99	100		
42	20.3	16.6	24.8
71	31.7	25.4	38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.394 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion	Ratio	Lower	Upper
93	100		
63	73.2	59.3	88.9
95	31.1	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 937
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	24973
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.0	13.4
54	9.9	5.2	7.8#
68	7.4	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 0.343 ng
RT: 8.928 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:	82	Resp:	6906
Ion Ratio	Lower	Upper	
82	100		
128	43.6	34.1	51.1
54	54.8	43.5	65.3









#9
Naphthalene
Concen: 0.394 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 27095
Ion Ratio Lower Upper
128 100
129 11.1 9.1 13.7
127 13.2 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.409 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:225 Resp: 4747
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.379 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 13509
Ion Ratio Lower Upper
152 100
151 20.7 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.387 ng
RT: 12.248 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:142 Resp: 17751
Ion Ratio Lower Upper
142 100
141 89.9 70.3 105.5
115 36.5 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.010 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
164	100		
162	104.1	83.0	124.4
160	51.5	40.7	61.1



#14
2,4,6-Tribromophenol
Concen: 0.326 ng
RT: 15.932 min Scan# 1587
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.6	76.6	114.8
141	42.3	34.1	51.1









#15
2-Fluorobiphenyl
Concen: 0.378 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.010 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.8	27.8	41.6
170	22.9	18.1	27.1



#16
Acenaphthylene
Concen: 0.387 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.8	23.8
153	13.1	10.2	15.2









#17
Acenaphthene
Concen: 0.382 ng
RT: 14.494 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 16888
Ion Ratio Lower Upper
154 100
153 115.7 90.0 135.0
152 62.6 47.3 70.9



#18
Fluorene
Concen: 0.397 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:166 Resp: 22996
Ion Ratio Lower Upper
166 100
165 98.5 79.0 118.4
167 13.0 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	10.6	8.6	13.0



#20
4,6-Dinitro-2-methylphenol
Concen: 0.381 ng
RT: 15.560 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion	Ratio	Lower	Upper
198	100		
51	135.4	88.9	133.3
105	63.4	49.2	73.8









#21
4-Bromophenyl-phenylether
Concen: 0.386 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
248	100		
250	95.3	76.2	114.2
141	68.0	57.6	86.4



#22
Hexachlorobenzene
Concen: 0.421 ng
RT: 16.503 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion	Ratio	Lower	Upper
284	100		
142	38.3	30.5	45.7
249	29.5	23.5	35.3









#23
Atrazine
Concen: 0.336 ng
RT: 16.652 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 4361
Ion Ratio Lower Upper
200 100
173 27.2 21.4 32.2
215 47.5 36.7 55.1



#24
Pentachlorophenol
Concen: 0.290 ng
RT: 16.838 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:266 Resp: 2516
Ion Ratio Lower Upper
266 100
264 62.4 49.3 73.9
268 65.6 50.6 75.8









#25
Phenanthrene
Concen: 0.391 ng
RT: 17.223 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 33589
Ion Ratio Lower Upper
178 100
176 19.3 15.5 23.3
179 15.5 12.2 18.4



#26
Anthracene
Concen: 0.371 ng
RT: 17.322 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:178 Resp: 27861
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.3 11.8 17.8









#27
Fluoranthene-d10
Concen: 0.352 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
212	100		
106	16.2	12.6	19.0
104	9.1	7.1	10.7



#28
Fluoranthene
Concen: 0.366 ng
RT: 19.252 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion	Ratio	Lower	Upper
202	100		
101	11.9	9.4	14.2
203	16.9	13.4	20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	240	Resp:	17059
Ion Ratio	Lower	Upper	
240	100		
120	15.0	10.7	16.1
236	29.6	23.1	34.7



#30
Pyrene
Concen: 0.443 ng
RT: 19.615 min Scan# 2032
Delta R.T. 0.004 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:	202	Resp:	34116
Ion Ratio	Lower	Upper	
202	100		
200	21.1	16.7	25.1
203	17.8	14.2	21.2









#31
Terphenyl-d14
Concen: 0.421 ng
RT: 19.819 min Scan# 2076
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 15620
Ion Ratio Lower Upper
244 100
212 9.5 7.8 11.6
122 19.3 15.2 22.8



#32
Benzo(a)anthracene
Concen: 0.365 ng
RT: 21.357 min Scan# 2267
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:228 Resp: 22233
Ion Ratio Lower Upper
228 100
226 28.0 22.2 33.4
229 19.8 15.8 23.8









#33
Chrysene
Concen: 0.408 ng
RT: 21.411 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.0	24.3	36.5
229	19.6	15.7	23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.405 ng
RT: 21.295 min Scan# 2260
Delta R.T. -0.000 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.4	21.1	31.7
279	3.2	2.9	4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.686 min Scan# 2821
Delta R.T. 0.006 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:264 Resp: 13607
Ion Ratio Lower Upper
264 100
260 27.1 21.8 32.8
265 93.2 68.6 102.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.426 ng
RT: 26.036 min Scan# 3626
Delta R.T. 0.003 min
Lab File: BN038130.D
Acq: 04 Nov 2025 10:40

Tgt Ion:276 Resp: 23804
Ion Ratio Lower Upper
276 100
138 27.4 21.4 32.0
277 23.7 19.2 28.8





#37

Benzo(b)fluoranthene

Concen: 0.400 ng

RT: 22.987 min Scan# 2582

Delta R.T. 0.006 min

Lab File: BN038130.D


Acq: 04 Nov 2025 10:40

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4






Tgt Ion:252 Resp: 23288

Ion Ratio Lower Upper

252 100

253 25.4 20.9 31.3

125 18.8 14.1 21.1



#38

Benzo(k)fluoranthene

Concen: 0.408 ng

RT: 23.031 min Scan# 2598

Delta R.T. 0.003 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40




Tgt Ion:252 Resp: 23877


Ion Ratio Lower Upper

252 100

253 26.0 21.2 31.8

125 19.0 14.6 22.0





#39

Benzo(a)pyrene

Concen: 0.396 ng

RT: 23.583 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 18445

Ion Ratio Lower Upper

252 100

253 28.6 23.1 34.7

125 24.7 18.6 28.0



#40

Dibenzo(a,h)anthracene

Concen: 0.427 ng

RT: 26.054 min Scan# 3632

Delta R.T. 0.006 min

Lab File: BN038130.D

Acq: 04 Nov 2025 10:40




Tgt Ion:278 Resp: 18347


Ion Ratio Lower Upper

278 100

139 21.1 17.6 26.4

279 27.8 22.6 33.8





#41

Benzo(g,h,i)perylene

Concen: 0.441 ng

RT: 26.753 min Scan# 38

Delta R.T. 0.003 min

Lab File: BN038130.D


Acq: 04 Nov 2025 10:40

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:276 Resp: 21656

Ion Ratio Lower Upper

276 100

277 24.8 19.7 29.5

138 25.4 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038130.D
 Acq On : 04 Nov 2025 10:40
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 04 11:42:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.413	0.433	-4.8	96	0.00
3	n-Nitrosodimethylamine	0.532	0.528	0.8	92	0.00
4 S	2-Fluorophenol	0.942	0.856	9.1	84	0.00
5 S	Phenol-d6	1.148	1.025	10.7	85	0.00
6	bis(2-Chloroethyl)ether	1.128	1.112	1.4	92	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
8 S	Nitrobenzene-d5	0.323	0.277	14.2	80	0.00
9	Naphthalene	1.102	1.085	1.5	91	0.00
10	Hexachlorobutadiene	0.186	0.190	-2.2	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.571	0.541	5.3	89	0.00
12	2-Methylnaphthalene	0.735	0.711	3.3	91	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.01
14 S	2,4,6-Tribromophenol	0.165	0.134	18.8	81	0.00
15 S	2-Fluorobiphenyl	1.602	1.514	5.5	86	0.01
16	Acenaphthylene	1.811	1.754	3.1	92	0.00
17	Acenaphthene	1.267	1.211	4.4	90	0.00
18	Fluorene	1.661	1.649	0.7	93	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.038	30.9#	77	0.00
21	4-Bromophenyl-phenylether	0.262	0.253	3.4	93	0.00
22	Hexachlorobenzene	0.298	0.314	-5.4	97	0.00
23	Atrazine	0.192	0.161	16.1	84	0.00
24	Pentachlorophenol	0.128	0.093	27.3#	79	0.00
25	Phenanthrene	1.270	1.241	2.3	93	0.00
26	Anthracene	1.111	1.030	7.3	90	0.00
27 SURR	Fluoranthene-d10	1.009	0.889	11.9	85	0.00
28	Fluoranthene	1.419	1.298	8.5	89	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
30	Pyrene	1.806	2.000	-10.7	86	0.00
31 S	Terphenyl-d14	0.869	0.916	-5.4	83	0.00
32	Benzo(a)anthracene	1.427	1.303	8.7	76	0.00
33	Chrysene	1.549	1.579	-1.9	81	0.00
34	Bis(2-ethylhexyl)phthalate	0.752	0.762	-1.3	84	0.00
35 I	Perylene-d12	1.000	1.000	0.0	76	0.00
36	Indeno(1,2,3-cd)pyrene	1.643	1.749	-6.5	83	0.00
37	Benzo(b)fluoranthene	1.710	1.711	-0.1	80	0.00
38	Benzo(k)fluoranthene	1.721	1.755	-2.0	82	0.00
39 C	Benzo(a)pyrene	1.369	1.356	0.9	78	0.00
40	Dibenzo(a,h)anthracene	1.265	1.348	-6.6	85	0.00
41	Benzo(g,h,i)perylene	1.444	1.592	-10.2	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038130.D
 Acq On : 04 Nov 2025 10:40
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 04 11:42:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	94	0.00
2	1,4-Dioxane	0.400	0.419	-4.7	96	0.00
3	n-Nitrosodimethylamine	0.400	0.397	0.8	92	0.00
4 S	2-Fluorophenol	0.400	0.363	9.3	84	0.00
5 S	Phenol-d6	0.400	0.357	10.8	85	0.00
6	bis(2-Chloroethyl)ether	0.400	0.394	1.5	92	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	90	0.00
8 S	Nitrobenzene-d5	0.400	0.343	14.2	80	0.00
9	Naphthalene	0.400	0.394	1.5	91	0.00
10	Hexachlorobutadiene	0.400	0.409	-2.2	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.379	5.3	89	0.00
12	2-Methylnaphthalene	0.400	0.387	3.3	91	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	90	0.01
14 S	2,4,6-Tribromophenol	0.400	0.326	18.5	81	0.00
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	86	0.01
16	Acenaphthylene	0.400	0.387	3.3	92	0.00
17	Acenaphthene	0.400	0.382	4.5	90	0.00
18	Fluorene	0.400	0.397	0.8	93	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	92	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.381	4.8	77	0.00
21	4-Bromophenyl-phenylether	0.400	0.386	3.5	93	0.00
22	Hexachlorobenzene	0.400	0.421	-5.2	97	0.00
23	Atrazine	0.400	0.336	16.0	84	0.00
24	Pentachlorophenol	0.400	0.290	27.5#	79	0.00
25	Phenanthrene	0.400	0.391	2.3	93	0.00
26	Anthracene	0.400	0.371	7.3	90	0.00
27 SURR	Fluoranthene-d10	0.400	0.352	12.0	85	0.00
28	Fluoranthene	0.400	0.366	8.5	89	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	79	0.00
30	Pyrene	0.400	0.443	-10.7	86	0.00
31 S	Terphenyl-d14	0.400	0.421	-5.2	83	0.00
32	Benzo(a)anthracene	0.400	0.365	8.8	76	0.00
33	Chrysene	0.400	0.408	-2.0	81	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.405	-1.3	84	0.00
35 I	Perylene-d12	0.400	0.400	0.0	76	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.426	-6.5	83	0.00
37	Benzo(b)fluoranthene	0.400	0.400	0.0	80	0.00
38	Benzo(k)fluoranthene	0.400	0.408	-2.0	82	0.00
39 C	Benzo(a)pyrene	0.400	0.396	1.0	78	0.00
40	Dibenzo(a,h)anthracene	0.400	0.427	-6.7	85	0.00
41	Benzo(g,h,i)perylene	0.400	0.441	-10.2	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3526</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>11/04/2025 20:12</u>
Lab File ID:	<u>BN038145.D</u>	Init. Calib. Date(s):	<u>10/28/2025 10/28/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4EC</u>	Init. Calib. Time(s):	<u>13:00 16:38</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.571	0.538		-5.8	50.0
Fluoranthene-d10	1.008	0.826		-18.1	50.0
2-Fluorophenol	0.943	0.881		-6.6	50.0
Phenol-d6	1.148	1.042		-9.2	50.0
Nitrobenzene-d5	0.323	0.273		-15.5	50.0
2-Fluorobiphenyl	1.602	1.511		-5.7	50.0
2,4,6-Tribromophenol	0.165	0.126		-23.6	50.0
Terphenyl-d14	0.869	0.962		10.7	50.0
1,4-Dioxane	0.416	0.430		3.4	50.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038145.D
 Acq On : 04 Nov 2025 20:12
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Nov 04 21:16:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8830	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	26327	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	14572	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	25889	0.400	ng	0.00
29) Chrysene-d12	21.375	240	14271	0.400	ng	0.00
35) Perylene-d12	23.683	264	11979	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	7775	0.374	ng	0.00
5) Phenol-d6	6.937	99	9204	0.363	ng	0.00
8) Nitrobenzene-d5	8.929	82	7183	0.338	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	14165	0.377	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1835	0.306	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	22024	0.377	ng	0.00
27) Fluoranthene-d10	19.220	212	21378	0.327	ng	0.00
31) Terphenyl-d14	19.819	244	13728	0.443	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.254	88	3796	0.416	ng	99
3) n-Nitrosodimethylamine	3.564	42	4631	0.394	ng	# 94
6) bis(2-Chloroethyl)ether	7.197	93	9767	0.392	ng	100
9) Naphthalene	10.626	128	28555	0.394	ng	100
10) Hexachlorobutadiene	10.925	225	5138	0.420	ng	# 99
12) 2-Methylnaphthalene	12.248	142	18776	0.388	ng	99
16) Acenaphthylene	14.152	152	25035	0.379	ng	100
17) Acenaphthene	14.494	154	17860	0.387	ng	98
18) Fluorene	15.489	166	23115	0.382	ng	99
20) 4,6-Dinitro-2-methylph...	15.560	198	971	0.380	ng	94
21) 4-Bromophenyl-phenylether	16.379	248	6744	0.397	ng	96
22) Hexachlorobenzene	16.491	284	8174	0.423	ng	99
23) Atrazine	16.652	200	4191	0.337	ng	98
24) Pentachlorophenol	16.838	266	2211	0.266	ng	99
25) Phenanthrene	17.223	178	31983	0.389	ng	100
26) Anthracene	17.322	178	26920	0.374	ng	99
28) Fluoranthene	19.248	202	30682	0.334	ng	100
30) Pyrene	19.610	202	29749	0.462	ng	100
32) Benzo(a)anthracene	21.357	228	19156	0.376	ng	100
33) Chrysene	21.411	228	22690	0.411	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	10298	0.384	ng	99
36) Indeno(1,2,3-cd)pyrene	26.034	276	21759	0.442	ng	98
37) Benzo(b)fluoranthene	22.984	252	20105	0.393	ng	97
38) Benzo(k)fluoranthene	23.028	252	20945	0.406	ng	98
39) Benzo(a)pyrene	23.581	252	16258	0.397	ng	96
40) Dibenzo(a,h)anthracene	26.051	278	17101	0.452	ng	99
41) Benzo(g,h,i)perylene	26.747	276	19764	0.457	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038145.D
Acq On : 04 Nov 2025 20:12
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Nov 04 21:16:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:152 Resp: 8830
Ion Ratio Lower Upper
152 100
150 153.2 122.3 183.5
115 60.0 47.4 71.0




#2
1,4-Dioxane
Concen: 0.416 ng
RT: 3.254 min Scan# 24
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12


Tgt Ion: 88 Resp: 3796
Ion Ratio Lower Upper
88 100
43 30.9 24.3 36.5
58 76.6 60.4 90.6




Abundance Scan 67 (3.564 min): BN038113.D\data.ms (-60)



Abundance Scan 67 (3.564 min): BN038145.D\data.ms



Abundance Scan 67 (3.564 min): BN038145.D\data.ms (-45)



#3

n-Nitrosodimethylamine

Concen: 0.394 ng

RT: 3.564 min Scan# 61

Delta R.T. 0.007 min

Lab File: BN038145.D

Acq: 04 Nov 2025 20:12

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion: 42 Resp: 4631


Ion Ratio Lower Upper

42 100


74 114.9 96.8 145.2

44 8.9 5.3 7.9#


Abundance




Abundance Scan 314 (5.348 min): BN038113.D\data.ms (-30)



Abundance Scan 313 (5.341 min): BN038145.D\data.ms



Abundance Scan 313 (5.341 min): BN038145.D\data.ms (-28)



#4

2-Fluorophenol

Concen: 0.374 ng

RT: 5.341 min Scan# 313

Delta R.T. -0.007 min

Lab File: BN038145.D

Acq: 04 Nov 2025 20:12

Tgt Ion: 112 Resp: 7775


Ion Ratio Lower Upper


112 100

64 56.9 45.4 68.0

63 29.9 23.2 34.8

Abundance









#5
Phenol-d6
Concen: 0.363 ng
RT: 6.937 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 9204
Ion Ratio Lower Upper
99 100
42 21.1 16.6 24.8
71 32.3 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.392 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion: 93 Resp: 9767
Ion Ratio Lower Upper
93 100
63 73.8 59.3 88.9
95 31.3 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 937
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	26327
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	9.6	5.2	7.8#
68	7.1	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 0.338 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:	82	Resp:	7183
Ion	Ratio	Lower	Upper
82	100		
128	45.7	34.1	51.1
54	52.7	43.5	65.3









#9
Naphthalene
Concen: 0.394 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 28555
Ion Ratio Lower Upper
128 100
129 11.4 9.1 13.7
127 13.5 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.420 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:225 Resp: 5138
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.4 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.377 ng
RT: 12.172 min Scan# 1154
Delta R.T. -0.005 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 14165
Ion Ratio Lower Upper
152 100
151 21.3 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.388 ng
RT: 12.248 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:142 Resp: 18776
Ion Ratio Lower Upper
142 100
141 88.8 70.3 105.5
115 35.2 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.430 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 14572
Ion Ratio Lower Upper
164 100
162 106.7 83.0 124.4
160 55.1 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 0.306 ng
RT: 15.932 min Scan# 1587
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:330 Resp: 1835
Ion Ratio Lower Upper
330 100
332 99.6 76.6 114.8
141 43.1 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.377 ng
RT: 13.051 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172 Resp: 22024
Ion Ratio Lower Upper
172 100
171 35.5 27.8 41.6
170 24.5 18.1 27.1



#16
Acenaphthylene
Concen: 0.379 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:152 Resp: 25035
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 13.0 10.2 15.2









#17
Acenaphthene
Concen: 0.387 ng
RT: 14.494 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 17860
Ion Ratio Lower Upper
154 100
153 114.0 90.0 135.0
152 61.4 47.3 70.9



#18
Fluorene
Concen: 0.382 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:166 Resp: 23115
Ion Ratio Lower Upper
166 100
165 99.8 79.0 118.4
167 13.4 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1688

Delta R.T. -0.000 min

Lab File: BN038145.D

Acq: 04 Nov 2025 20:12

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:188 Resp: 25889

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.9 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.380 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038145.D

Acq: 04 Nov 2025 20:12




Tgt Ion:198 Resp: 971


Ion Ratio Lower Upper

198 100

51 118.1 88.9 133.3

105 58.5 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.397 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 6744
Ion Ratio Lower Upper
248 100
250 96.4 76.2 114.2
141 64.8 57.6 86.4



#22
Hexachlorobenzene
Concen: 0.423 ng
RT: 16.491 min Scan# 1632
Delta R.T. -0.013 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:284 Resp: 8174
Ion Ratio Lower Upper
284 100
142 39.1 30.5 45.7
249 30.0 23.5 35.3









#23
Atrazine
Concen: 0.337 ng
RT: 16.652 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
200	100		
173	24.5	21.4	32.2
215	46.5	36.7	55.1



#24
Pentachlorophenol
Concen: 0.266 ng
RT: 16.838 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion	Ratio	Lower	Upper
266	100		
264	61.3	49.3	73.9
268	65.2	50.6	75.8









#25
Phenanthrene
Concen: 0.389 ng
RT: 17.223 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 31983
Ion Ratio Lower Upper
178 100
176 19.4 15.5 23.3
179 15.3 12.2 18.4



#26
Anthracene
Concen: 0.374 ng
RT: 17.322 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:178 Resp: 26920
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.3 11.8 17.8









#27
Fluoranthene-d10
Concen: 0.327 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
212	100		
106	16.1	12.6	19.0
104	9.5	7.1	10.7



#28
Fluoranthene
Concen: 0.334 ng
RT: 19.248 min Scan# 1953
Delta R.T. -0.005 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion	Ratio	Lower	Upper
202	100		
101	12.1	9.4	14.2
203	16.9	13.4	20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
240	100		
120	13.8	10.7	16.1
236	29.3	23.1	34.7



#30
Pyrene
Concen: 0.462 ng
RT: 19.610 min Scan# 2031
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.7	25.1
203	17.6	14.2	21.2









#31
Terphenyl-d14
Concen: 0.443 ng
RT: 19.819 min Scan# 2076
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 13728
Ion Ratio Lower Upper
244 100
212 9.7 7.8 11.6
122 18.8 15.2 22.8



#32
Benzo(a)anthracene
Concen: 0.376 ng
RT: 21.357 min Scan# 2267
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:228 Resp: 19156
Ion Ratio Lower Upper
228 100
226 27.5 22.2 33.4
229 19.7 15.8 23.8









#33
Chrysene
Concen: 0.411 ng
RT: 21.411 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.3	36.5
229	19.8	15.7	23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.384 ng
RT: 21.295 min Scan# 2260
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.6	21.1	31.7
279	3.3	2.9	4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.683 min Scan# 2821
Delta R.T. 0.003 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:264 Resp: 11979
Ion Ratio Lower Upper
264 100
260 27.3 21.8 32.8
265 102.8 68.6 102.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.442 ng
RT: 26.034 min Scan# 3625
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:276 Resp: 21759
Ion Ratio Lower Upper
276 100
138 27.9 21.4 32.0
277 24.5 19.2 28.8









#37
Benzo(b)fluoranthene
Concen: 0.393 ng
RT: 22.984 min Scan# 2582
Delta R.T. 0.003 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 20105
Ion Ratio Lower Upper
252 100
253 26.5 20.9 31.3
125 20.3 14.1 21.1



#38
Benzo(k)fluoranthene
Concen: 0.406 ng
RT: 23.028 min Scan# 2597
Delta R.T. -0.000 min
Lab File: BN038145.D
Acq: 04 Nov 2025 20:12

Tgt Ion:252 Resp: 20945
Ion Ratio Lower Upper
252 100
253 26.9 21.2 31.8
125 20.0 14.6 22.0





#39

Benzo(a)pyrene

Concen: 0.397 ng

RT: 23.581 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN038145.D

Acq: 04 Nov 2025 20:12

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:252 Resp: 16258

Ion Ratio Lower Upper

252 100

253 29.4 23.1 34.7

125 27.3 18.6 28.0



#40

Dibenzo(a,h)anthracene

Concen: 0.452 ng

RT: 26.051 min Scan# 3631

Delta R.T. 0.003 min

Lab File: BN038145.D

Acq: 04 Nov 2025 20:12




Tgt Ion:278 Resp: 17101


Ion Ratio Lower Upper

278 100

139 21.0 17.6 26.4

279 28.4 22.6 33.8





#41

Benzo(g,h,i)perylene

Concen: 0.457 ng

RT: 26.747 min Scan# 3869

Delta R.T. -0.003 min

Lab File: BN038145.D


Acq: 04 Nov 2025 20:12

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:276 Resp: 19764

Ion Ratio Lower Upper

276 100

277 24.6 19.7 29.5

138 26.1 20.2 30.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038145.D
 Acq On : 04 Nov 2025 20:12
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 04 21:16:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.413	0.430	-4.1	97	0.00
3	n-Nitrosodimethylamine	0.532	0.524	1.5	94	0.00
4 S	2-Fluorophenol	0.942	0.881	6.5	89	0.00
5 S	Phenol-d6	1.148	1.042	9.2	88	0.00
6	bis(2-Chloroethyl)ether	1.128	1.106	2.0	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.323	0.273	15.5	83	0.00
9	Naphthalene	1.102	1.085	1.5	96	0.00
10	Hexachlorobutadiene	0.186	0.195	-4.8	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.571	0.538	5.8	93	0.00
12	2-Methylnaphthalene	0.735	0.713	3.0	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.165	0.126	23.6	80	0.00
15 S	2-Fluorobiphenyl	1.602	1.511	5.7	90	0.00
16	Acenaphthylene	1.811	1.718	5.1	94	0.00
17	Acenaphthene	1.267	1.226	3.2	95	0.00
18	Fluorene	1.661	1.586	4.5	93	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.038	30.9#	73	0.00
21	4-Bromophenyl-phenylether	0.262	0.260	0.8	91	0.00
22	Hexachlorobenzene	0.298	0.316	-6.0	93	-0.01
23	Atrazine	0.192	0.162	15.6	81	0.00
24	Pentachlorophenol	0.128	0.085	33.6#	69	0.00
25	Phenanthrene	1.270	1.235	2.8	89	0.00
26	Anthracene	1.111	1.040	6.4	87	0.00
27 SURR	Fluoranthene-d10	1.009	0.826	18.1	76	0.00
28	Fluoranthene	1.419	1.185	16.5	77	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	66	0.00
30	Pyrene	1.806	2.085	-15.4	75	0.00
31 S	Terphenyl-d14	0.869	0.962	-10.7	73	0.00
32	Benzo(a)anthracene	1.427	1.342	6.0	66	0.00
33	Chrysene	1.549	1.590	-2.6	68	0.00
34	Bis(2-ethylhexyl)phthalate	0.752	0.722	4.0	66	0.00
35 I	Perylene-d12	1.000	1.000	0.0	67	0.00
36	Indeno(1,2,3-cd)pyrene	1.643	1.816	-10.5	76	0.00
37	Benzo(b)fluoranthene	1.710	1.678	1.9	69	0.00
38	Benzo(k)fluoranthene	1.721	1.748	-1.6	72	0.00
39 C	Benzo(a)pyrene	1.369	1.357	0.9	69	0.00
40	Dibenzo(a,h)anthracene	1.265	1.428	-12.9	80	0.00
41	Benzo(g,h,i)perylene	1.444	1.650	-14.3	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038145.D
 Acq On : 04 Nov 2025 20:12
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 04 21:16:11 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.416	-4.0	97	0.00
3	n-Nitrosodimethylamine	0.400	0.394	1.5	94	0.00
4 S	2-Fluorophenol	0.400	0.374	6.5	89	0.00
5 S	Phenol-d6	0.400	0.363	9.3	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.392	2.0	94	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.338	15.5	83	0.00
9	Naphthalene	0.400	0.394	1.5	96	0.00
10	Hexachlorobutadiene	0.400	0.420	-5.0	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	93	0.00
12	2-Methylnaphthalene	0.400	0.388	3.0	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.400	0.306	23.5	80	0.00
15 S	2-Fluorobiphenyl	0.400	0.377	5.8	90	0.00
16	Acenaphthylene	0.400	0.379	5.3	94	0.00
17	Acenaphthene	0.400	0.387	3.3	95	0.00
18	Fluorene	0.400	0.382	4.5	93	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	88	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.380	5.0	73	0.00
21	4-Bromophenyl-phenylether	0.400	0.397	0.8	91	0.00
22	Hexachlorobenzene	0.400	0.423	-5.7	93	-0.01
23	Atrazine	0.400	0.337	15.8	81	0.00
24	Pentachlorophenol	0.400	0.266	33.5#	69	0.00
25	Phenanthrene	0.400	0.389	2.8	89	0.00
26	Anthracene	0.400	0.374	6.5	87	0.00
27 SURR	Fluoranthene-d10	0.400	0.327	18.3	76	0.00
28	Fluoranthene	0.400	0.334	16.5	77	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	66	0.00
30	Pyrene	0.400	0.462	-15.5	75	0.00
31 S	Terphenyl-d14	0.400	0.443	-10.7	73	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	66	0.00
33	Chrysene	0.400	0.411	-2.7	68	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.384	4.0	66	0.00
35 I	Perylene-d12	0.400	0.400	0.0	67	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.442	-10.5	76	0.00
37	Benzo(b)fluoranthene	0.400	0.393	1.8	69	0.00
38	Benzo(k)fluoranthene	0.400	0.406	-1.5	72	0.00
39 C	Benzo(a)pyrene	0.400	0.397	0.8	69	0.00
40	Dibenzo(a,h)anthracene	0.400	0.452	-13.0	80	0.00
41	Benzo(g,h,i)perylene	0.400	0.457	-14.2	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4




QC SAMPLE DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038110.D
Acq On : 28 Oct 2025 12:05
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Oct 29 13:09:46 2025






AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
68	69	0.00	2	1.8	4477	PASS
69	69	100	100	100.0	255398	PASS
70	69	0.00	2	0.5	1371	PASS
197	198	0.00	2	0.2	1550	PASS
198	198	100	100	100.0	669440	PASS
199	198	5	9	6.5	43461	PASS
365	198	1	100	3.4	22528	PASS
441	443	0.01	150	83.9	95293	PASS
442	442	100	100	100.0	571285	PASS
443	442	15	24	19.9	113539	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038110.D
Acq On : 28 Oct 2025 12:05
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 28 18:06:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 12:54:48 2025
Response via : Initial Calibration



TIC: BN038110.D\data.ms

(70) Pentachlorophenol (C)

16.839min (-0.017) 31764.46 ng




response 993628

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.51
264.00	61.60	67.24
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038110.D
Acq On : 28 Oct 2025 12:05
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 28 18:06:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 12:54:48 2025
Response via : Initial Calibration



TIC: BN038110.D\data.ms

(77) Benzidine

19.439min (-0.005) 0.00 ng

response 4602519

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.42
183.00	13.20	12.28
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/28/2025	BNA_N	<u>BN038110.D</u>
Compound Name	Response	Retention Time
DDT	2477320	20.674
DDD	29265	20.233
DDE	1740	19.733
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
31005	2508325	1.24



Instrument :
BNA_N
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038129.D
Acq On : 04 Nov 2025 10:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Oct 29 13:09:46 2025






AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2467

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
68	69	0.00	2	1.8	7977	PASS
69	69	100	100	100.0	447289	PASS
70	69	0.00	2	0.5	2389	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	1179904	PASS
199	198	5	9	6.9	81427	PASS
365	198	1	100	3.8	44821	PASS
441	443	0.01	150	77.2	173248	PASS
442	442	100	100	100.0	1122112	PASS
443	442	15	24	20.0	224533	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038129.D
Acq On : 04 Nov 2025 10:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Nov 04 14:05:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 04 14:05:21 2025
Response via : Initial Calibration



TIC: BN038129.D\data.ms

(70) Pentachlorophenol (C)

16.851min (0.000) 277.93 ng




response 1813451

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.08
264.00	61.60	60.19
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038129.D
Acq On : 04 Nov 2025 10:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Nov 04 14:05:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 04 14:05:21 2025
Response via : Initial Calibration



TIC: BN038129.D\data.ms

(77) Benzidine

19.451min (0.000) 0.00 ng

response 8673581

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.80
183.00	13.20	12.79
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
11/4/2025	BNA_N	<u>BN038129.D</u>
Compound Name	Response	Retention Time
DDT	4564306	20.686
DDD	39263	20.233
DDE	2743	19.733
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
42006	4606312	0.91

Instrument :
BNA_N
ClientSampleId :
DFTPP



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB170381BL		SDG No.:	Q3526
Lab Sample ID:	PB170381BL		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.20	U	1	0.070	0.20	0.20	ug/L	11/04/25 11:43	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.33			30 - 150		83%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.35			30 - 150		88%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.35			55 - 111		88%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.39			53 - 106		97%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.46			58 - 132		114%	SPK: 0.4		
INTERNAL STANDARDS										
		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	8220								
1146-65-2	Naphthalene-d8	22200								
15067-26-2	Acenaphthene-d10	11600								
1517-22-2	Phenanthrene-d10	21800								
1719-03-5	Chrysene-d12	15200								
1520-96-3	Perylene-d12	13200								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038131.D
 Acq On : 04 Nov 2025 11:43
 Operator : RC/JU
 Sample : PB170381BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170381BL

Quant Time: Nov 04 12:07:36 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.789	152	8221	0.400	ng	0.01
7) Naphthalene-d8	10.594	136	22221	0.400	ng	0.02
13) Acenaphthene-d10	14.441	164	11601	0.400	ng	0.01
19) Phenanthrene-d10	17.198	188	21781	0.400	ng	0.01
29) Chrysene-d12	21.384	240	15202	0.400	ng	0.00
35) Perylene-d12	23.695	264	13241	0.400	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.363	112	6175	0.319	ng	0.01
5) Phenol-d6	6.952	99	7077	0.300	ng	0.01
8) Nitrobenzene-d5	8.939	82	6330	0.353	ng	0.01
11) 2-Methylnaphthalene-d10	12.187	152	10488	0.331	ng	0.01
14) 2,4,6-Tribromophenol	15.945	330	557	0.117	ng	0.01
15) 2-Fluorobiphenyl	13.073	172	18033	0.388	ng	0.02
27) Fluoranthene-d10	19.229	212	19315	0.352	ng	0.00
31) Terphenyl-d14	19.829	244	15022	0.455	ng	0.00


Target Compounds	Qvalue
-----	-----


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038131.D
Acq On : 04 Nov 2025 11:43
Operator : RC/JU
Sample : PB170381BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170381BL

Quant Time: Nov 04 12:07:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.789 min Scan# 61
Delta R.T. 0.014 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion:152 Resp: 8221
Ion Ratio Lower Upper
152 100
150 153.2 122.3 183.5
115 58.7 47.4 71.0



#4
2-Fluorophenol
Concen: 0.319 ng
RT: 5.363 min Scan# 316
Delta R.T. 0.014 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion:112 Resp: 6175
Ion Ratio Lower Upper
112 100
64 55.6 45.4 68.0
63 30.3 23.2 34.8









#5
Phenol-d6
Concen: 0.300 ng
RT: 6.952 min Scan# 534
Delta R.T. 0.014 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion: 99 Resp: 7077
Ion Ratio Lower Upper
99 100
42 20.3 16.6 24.8
71 31.7 25.4 38.2



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 938
Delta R.T. 0.021 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion: 136 Resp: 22221
Ion Ratio Lower Upper
136 100
137 11.2 9.0 13.4
54 6.8 5.2 7.8
68 5.3 4.1 6.1









#8
Nitrobenzene-d5
Concen: 0.353 ng
RT: 8.939 min Scan# 782
Delta R.T. 0.011 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion:	82	Resp:	6330
Ion Ratio	Lower	Upper	
82	100		
128	41.4	34.1	51.1
54	56.9	43.5	65.3



#11
2-Methylnaphthalene-d10
Concen: 0.331 ng
RT: 12.187 min Scan# 1157
Delta R.T. 0.010 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion:	152	Resp:	10488
Ion Ratio	Lower	Upper	
152	100		
151	20.9	16.8	25.2









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion	Ratio	Lower	Upper
164	100		
162	106.2	83.0	124.4
160	55.8	40.7	61.1



#14
2,4,6-Tribromophenol
Concen: 0.117 ng
RT: 15.945 min Scan# 1588
Delta R.T. 0.013 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion	Ratio	Lower	Upper
330	100		
332	93.7	76.6	114.8
141	49.6	34.1	51.1









#15
2-Fluorobiphenyl
Concen: 0.388 ng
RT: 13.073 min Scan# 1324
Delta R.T. 0.021 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion:	172	Resp:	18033
Ion Ratio	Lower	Upper	
172	100		
171	35.0	27.8	41.6
170	22.9	18.1	27.1



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.198 min Scan# 1689
Delta R.T. 0.012 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion:	188	Resp:	21781
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	9.6	8.6	13.0









#27
Fluoranthene-d10
Concen: 0.352 ng
RT: 19.229 min Scan# 19
Delta R.T. 0.009 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL


Tgt Ion	Ratio	Lower	Upper
212	100		
106	16.1	12.6	19.0
104	9.3	7.1	10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.384 min Scan# 2270
Delta R.T. 0.009 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion	Ratio	Lower	Upper
240	100		
120	14.8	10.7	16.1
236	29.1	23.1	34.7









#31
Terphenyl-d14
Concen: 0.455 ng
RT: 19.829 min Scan# 2076
Delta R.T. 0.009 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43




Instrument :
BNA_N
ClientSampleId :
PB170381BL

Tgt Ion:244 Resp: 15022
Ion Ratio Lower Upper
244 100
212 9.5 7.8 11.6
122 17.5 15.2 22.8



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.695 min Scan# 2825
Delta R.T. 0.015 min
Lab File: BN038131.D
Acq: 04 Nov 2025 11:43

Tgt Ion:264 Resp: 13241
Ion Ratio Lower Upper
264 100
260 27.8 21.8 32.8
265 100.3 68.6 102.8





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB170381BS		SDG No.:	Q3526
Lab Sample ID:	PB170381BS		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.33		1	0.070	0.20	0.20	ug/L	11/04/25 12:19	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.35			30 - 150		88%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.30			30 - 150		76%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.37			55 - 111		92%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.42			53 - 106		104%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.48			58 - 132		121%	SPK: 0.4		
INTERNAL STANDARDS		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	11500								
1146-65-2	Naphthalene-d8	31800								
15067-26-2	Acenaphthene-d10	16200								
1517-22-2	Phenanthrene-d10	30400								
1719-03-5	Chrysene-d12	16500								
1520-96-3	Perylene-d12	13200								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038132.D
 Acq On : 04 Nov 2025 12:19
 Operator : RC/JU
 Sample : PB170381BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170381BS

Quant Time: Nov 04 12:44:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	11523	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	31801	0.400	ng	# 0.00
13) Acenaphthene-d10	14.441	164	16154	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	30417	0.400	ng	0.00
29) Chrysene-d12	21.375	240	16510	0.400	ng	0.00
35) Perylene-d12	23.686	264	13194	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	9248	0.341	ng	0.00
5) Phenol-d6	6.937	99	10615	0.321	ng	0.00
8) Nitrobenzene-d5	8.929	82	9450	0.368	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	15982	0.352	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1687	0.253	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	26882	0.415	ng	0.01
27) Fluoranthene-d10	19.220	212	23204	0.303	ng	0.00
31) Terphenyl-d14	19.824	244	17337	0.483	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.261	88	3965	0.333	ng	# 62
3) n-Nitrosodimethylamine	3.564	42	5606	0.366	ng	# 94
6) bis(2-Chloroethyl)ether	7.197	93	11700	0.360	ng	99
9) Naphthalene	10.626	128	31176	0.356	ng	100
10) Hexachlorobutadiene	10.925	225	5678	0.384	ng	# 99
12) 2-Methylnaphthalene	12.248	142	17976	0.308	ng	99
16) Acenaphthylene	14.152	152	28630	0.391	ng	99
17) Acenaphthene	14.494	154	17838	0.349	ng	98
18) Fluorene	15.489	166	24190	0.361	ng	99
20) 4,6-Dinitro-2-methylph...	15.560	198	1040	0.363	ng	95
21) 4-Bromophenyl-phenylether	16.379	248	7282	0.365	ng	97
22) Hexachlorobenzene	16.503	284	8817	0.389	ng	99
23) Atrazine	16.652	200	4704	0.322	ng	99
24) Pentachlorophenol	16.838	266	4384	0.449	ng	99
25) Phenanthrene	17.223	178	35147	0.364	ng	100
26) Anthracene	17.322	178	30413	0.360	ng	99
28) Fluoranthene	19.252	202	32822	0.304	ng	100
30) Pyrene	19.615	202	31678	0.425	ng	100
32) Benzo(a)anthracene	21.357	228	20604	0.350	ng	99
33) Chrysene	21.411	228	24309	0.380	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	8897	0.287	ng	98
36) Indeno(1,2,3-cd)pyrene	26.036	276	23949	0.442	ng	98
37) Benzo(b)fluoranthene	22.987	252	20712	0.367	ng	98
38) Benzo(k)fluoranthene	23.031	252	21547	0.379	ng	98
39) Benzo(a)pyrene	23.584	252	17494	0.387	ng	98
40) Dibenzo(a,h)anthracene	26.054	278	17865	0.428	ng	99
41) Benzo(g,h,i)perylene	26.756	276	21276	0.447	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038132.D
Acq On : 04 Nov 2025 12:19
Operator : RC/JU
Sample : PB170381BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170381BS


Quant Time: Nov 04 12:44:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Instrument :
BNA_N
ClientSampleId :
PB170381BS







Tgt Ion: 152 Resp: 11523
Ion Ratio Lower Upper
152 100
150 156.9 122.3 183.5
115 60.3 47.4 71.0



#2
1,4-Dioxane
Concen: 0.333 ng
RT: 3.261 min Scan# 25
Delta R.T. 0.007 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion: 88 Resp: 3965
Ion Ratio Lower Upper
88 100
43 84.7 24.3 36.5
58 86.6 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 0.366 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion: 42 Resp: 5606
Ion Ratio Lower Upper
42 100
74 115.6 96.8 145.2
44 14.6 5.3 7.9#



#4
2-Fluorophenol
Concen: 0.341 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion: 112 Resp: 9248
Ion Ratio Lower Upper
112 100
64 56.5 45.4 68.0
63 30.1 23.2 34.8









#5
Phenol-d6
Concen: 0.321 ng
RT: 6.937 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion: 99 Resp: 10615
Ion Ratio Lower Upper
99 100
42 21.8 16.6 24.8
71 31.9 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.360 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion: 93 Resp: 11700
Ion Ratio Lower Upper
93 100
63 75.6 59.3 88.9
95 31.9 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 936
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:136 Resp: 31801
Ion Ratio Lower Upper
136 100
137 11.4 9.0 13.4
54 9.9 5.2 7.8#
68 7.4 4.1 6.1#



#8
Nitrobenzene-d5
Concen: 0.368 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion: 82 Resp: 9450
Ion Ratio Lower Upper
82 100
128 44.3 34.1 51.1
54 52.5 43.5 65.3









#9
Naphthalene
Concen: 0.356 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:	128	Resp:	31176
Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.1	13.7
127	13.3	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.384 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:	225	Resp:	5678
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.5	50.3	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.352 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:152 Resp: 15982
Ion Ratio Lower Upper
152 100
151 20.6 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.308 ng
RT: 12.248 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:142 Resp: 17976
Ion Ratio Lower Upper
142 100
141 88.8 70.3 105.5
115 35.3 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion	164	162	160
Resp:	16154	103.7	52.4
Lower		83.0	40.7
Upper		124.4	61.1



#14
2,4,6-Tribromophenol
Concen: 0.253 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion	330	332	141
Resp:	1687	95.7	45.4
Lower		76.6	34.1
Upper		114.8	51.1









#15
2-Fluorobiphenyl
Concen: 0.415 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.1	27.8	41.6
170	22.8	18.1	27.1



#16
Acenaphthylene
Concen: 0.391 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.8	23.8
153	12.9	10.2	15.2









#17
Acenaphthene
Concen: 0.349 ng
RT: 14.494 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:154 Resp: 17838
Ion Ratio Lower Upper
154 100
153 114.1 90.0 135.0
152 62.2 47.3 70.9



#18
Fluorene
Concen: 0.361 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:166 Resp: 24190
Ion Ratio Lower Upper
166 100
165 99.4 79.0 118.4
167 13.5 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1688

Delta R.T. -0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:188 Resp: 30417

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.0 8.6 13.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.363 ng

RT: 15.560 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19




Tgt Ion:198 Resp: 1040


Ion Ratio Lower Upper

198 100

51 103.2 88.9 133.3

105 61.7 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.365 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:248 Resp: 7282
Ion Ratio Lower Upper
248 100
250 93.4 76.2 114.2
141 67.9 57.6 86.4



#22
Hexachlorobenzene
Concen: 0.389 ng
RT: 16.503 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:284 Resp: 8817
Ion Ratio Lower Upper
284 100
142 38.4 30.5 45.7
249 30.0 23.5 35.3









#23
Atrazine
Concen: 0.322 ng
RT: 16.652 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:200 Resp: 4704
Ion Ratio Lower Upper
200 100
173 28.3 21.4 32.2
215 45.7 36.7 55.1



#24
Pentachlorophenol
Concen: 0.449 ng
RT: 16.838 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:266 Resp: 4384
Ion Ratio Lower Upper
266 100
264 61.1 49.3 73.9
268 63.9 50.6 75.8









#25
Phenanthrene
Concen: 0.364 ng
RT: 17.223 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:178 Resp: 35147
Ion Ratio Lower Upper
178 100
176 19.4 15.5 23.3
179 15.2 12.2 18.4



#26
Anthracene
Concen: 0.360 ng
RT: 17.322 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:178 Resp: 30413
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.4 11.8 17.8





#27

Fluoranthene-d10

Concen: 0.303 ng

RT: 19.220 min Scan# 1947

Delta R.T. -0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:212 Resp: 23204

Ion Ratio Lower Upper

212 100

106 16.4 12.6 19.0

104 9.3 7.1 10.7



#28

Fluoranthene

Concen: 0.304 ng

RT: 19.252 min Scan# 1954

Delta R.T. -0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19




Tgt Ion:202 Resp: 32822


Ion Ratio Lower Upper

202 100

101 12.2 9.4 14.2

203 16.8 13.4 20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:240 Resp: 16510
Ion Ratio Lower Upper
240 100
120 14.8 10.7 16.1
236 29.1 23.1 34.7



#30
Pyrene
Concen: 0.425 ng
RT: 19.615 min Scan# 2032
Delta R.T. 0.005 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:202 Resp: 31678
Ion Ratio Lower Upper
202 100
200 21.1 16.7 25.1
203 17.6 14.2 21.2









#31
Terphenyl-d14
Concen: 0.483 ng
RT: 19.824 min Scan# 2076
Delta R.T. 0.005 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:244 Resp: 17337
Ion Ratio Lower Upper
244 100
212 9.5 7.8 11.6
122 17.7 15.2 22.8



#32
Benzo(a)anthracene
Concen: 0.350 ng
RT: 21.357 min Scan# 2267
Delta R.T. -0.000 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:228 Resp: 20604
Ion Ratio Lower Upper
228 100
226 27.8 22.2 33.4
229 19.2 15.8 23.8





#33

Chrysene

Concen: 0.380 ng

RT: 21.411 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:228 Resp: 24309

Ion Ratio Lower Upper

228 100

226 30.7 24.3 36.5

229 19.7 15.7 23.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.287 ng

RT: 21.295 min Scan# 2260

Delta R.T. -0.000 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19




Tgt Ion:149 Resp: 8897


Ion Ratio Lower Upper

149 100

167 25.4 21.1 31.7

279 3.5 2.9 4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.686 min Scan# 2821
Delta R.T. 0.006 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:264 Resp: 13194
Ion Ratio Lower Upper
264 100
260 27.3 21.8 32.8
265 94.2 68.6 102.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.442 ng
RT: 26.036 min Scan# 3626
Delta R.T. 0.003 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:276 Resp: 23949
Ion Ratio Lower Upper
276 100
138 28.4 21.4 32.0
277 23.9 19.2 28.8









#37
Benzo(b)fluoranthene
Concen: 0.367 ng
RT: 22.987 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19




Instrument :
BNA_N
ClientSampleId :
PB170381BS


Tgt Ion:252 Resp: 20712
Ion Ratio Lower Upper
252 100
253 26.2 20.9 31.3
125 19.9 14.1 21.1



#38
Benzo(k)fluoranthene
Concen: 0.379 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.003 min
Lab File: BN038132.D
Acq: 04 Nov 2025 12:19

Tgt Ion:252 Resp: 21547
Ion Ratio Lower Upper
252 100
253 26.2 21.2 31.8
125 19.8 14.6 22.0





#39

Benzo(a)pyrene

Concen: 0.387 ng

RT: 23.584 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:252 Resp: 17494

Ion Ratio Lower Upper

252 100

253 28.7 23.1 34.7

125 24.8 18.6 28.0



#40

Dibenzo(a,h)anthracene

Concen: 0.428 ng

RT: 26.054 min Scan# 3632

Delta R.T. 0.006 min

Lab File: BN038132.D

Acq: 04 Nov 2025 12:19




Tgt Ion:278 Resp: 17865


Ion Ratio Lower Upper

278 100

139 21.4 17.6 26.4

279 28.1 22.6 33.8





#41

Benzo(g,h,i)perylene

Concen: 0.447 ng

RT: 26.756 min Scan# 38

Delta R.T. 0.006 min

Lab File: BN038132.D


Acq: 04 Nov 2025 12:19

Instrument :

BNA_N

ClientSampleId :

PB170381BS





Tgt Ion:276 Resp: 21276

Ion Ratio Lower Upper

276 100

277 24.3 19.7 29.5

138 26.5 20.2 30.4





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB170381BSD		SDG No.:	Q3526
Lab Sample ID:	PB170381BSD		Matrix:	Water
Analytical Method:	SW8270ESIM	Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	3510C	Prep Date: 11/03/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.26		1	0.070	0.20	0.20	ug/L	11/04/25 12:56	PB170381
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.29			30 - 150		73%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.25			30 - 150		63%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.31			55 - 111		76%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.32			53 - 106		81%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.36			58 - 132		91%	SPK: 0.4		
INTERNAL STANDARDS		Area Count								
3855-82-1	1,4-Dichlorobenzene-d4	10800								
1146-65-2	Naphthalene-d8	30500								
15067-26-2	Acenaphthene-d10	15600								
1517-22-2	Phenanthrene-d10	28800								
1719-03-5	Chrysene-d12	18100								
1520-96-3	Perylene-d12	14900								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038133.D
 Acq On : 04 Nov 2025 12:56
 Operator : RC/JU
 Sample : PB170381BSD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170381BSD

Quant Time: Nov 04 13:20:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	10823	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	30534	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	15616	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	28756	0.400	ng	0.00
29) Chrysene-d12	21.375	240	18053	0.400	ng	0.00
35) Perylene-d12	23.683	264	14856	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	7259	0.285	ng	0.00
5) Phenol-d6	6.937	99	8491	0.273	ng	0.00
8) Nitrobenzene-d5	8.929	82	7527	0.305	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	12754	0.292	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1353	0.210	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	20205	0.323	ng	0.00
27) Fluoranthene-d10	19.220	212	18424	0.254	ng	0.00
31) Terphenyl-d14	19.824	244	14219	0.363	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.254	88	2913	0.260	ng	# 69
3) n-Nitrosodimethylamine	3.564	42	4281	0.297	ng	# 93
6) bis(2-Chloroethyl)ether	7.197	93	9060	0.297	ng	98
9) Naphthalene	10.626	128	25040	0.298	ng	100
10) Hexachlorobutadiene	10.925	225	4498	0.317	ng	# 99
12) 2-Methylnaphthalene	12.248	142	14295	0.255	ng	97
16) Acenaphthylene	14.152	152	22428	0.317	ng	99
17) Acenaphthene	14.495	154	14121	0.285	ng	97
18) Fluorene	15.489	166	18901	0.291	ng	99
20) 4,6-Dinitro-2-methylph...	15.560	198	792	0.328	ng	# 80
21) 4-Bromophenyl-phenylether	16.379	248	5427	0.288	ng	100
22) Hexachlorobenzene	16.503	284	6880	0.321	ng	99
23) Atrazine	16.652	200	3562	0.258	ng	99
24) Pentachlorophenol	16.838	266	3398	0.368	ng	99
25) Phenanthrene	17.223	178	26605	0.291	ng	100
26) Anthracene	17.322	178	23247	0.291	ng	99
28) Fluoranthene	19.253	202	25972	0.255	ng	100
30) Pyrene	19.615	202	25592	0.314	ng	100
32) Benzo(a)anthracene	21.358	228	18334	0.285	ng	99
33) Chrysene	21.411	228	21765	0.311	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	7715	0.227	ng	99
36) Indeno(1,2,3-cd)pyrene	26.037	276	19208	0.315	ng	97
37) Benzo(b)fluoranthene	22.984	252	18434	0.290	ng	97
38) Benzo(k)fluoranthene	23.031	252	19683	0.308	ng	98
39) Benzo(a)pyrene	23.581	252	15027	0.296	ng	97
40) Dibenzo(a,h)anthracene	26.054	278	14795	0.315	ng	99
41) Benzo(g,h,i)perylene	26.753	276	17207	0.321	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038133.D
 Acq On : 04 Nov 2025 12:56
 Operator : RC/JU
 Sample : PB170381BSD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170381BSD

Quant Time: Nov 04 13:20:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:152 Resp: 10823
Ion Ratio Lower Upper
152 100
150 154.8 122.3 183.5
115 59.3 47.4 71.0



#2
1,4-Dioxane
Concen: 0.260 ng
RT: 3.254 min Scan# 24
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion: 88 Resp: 2913
Ion Ratio Lower Upper
88 100
43 71.2 24.3 36.5#
58 87.3 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 0.297 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion: 42 Resp: 4281
Ion Ratio Lower Upper
42 100
74 116.1 96.8 145.2
44 25.7 5.3 7.9#



#4
2-Fluorophenol
Concen: 0.285 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion: 112 Resp: 7259
Ion Ratio Lower Upper
112 100
64 57.2 45.4 68.0
63 30.8 23.2 34.8









#5
Phenol-d6
Concen: 0.273 ng
RT: 6.937 min Scan# 534
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion	Ratio	Lower	Upper
99	100		
42	22.9	16.6	24.8
71	31.9	25.4	38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.297 ng
RT: 7.197 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion	Ratio	Lower	Upper
93	100		
63	75.7	59.3	88.9
95	32.3	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 937
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:	136	Resp:	30534
Ion Ratio	Lower	Upper	
136	100		
137	11.1	9.0	13.4
54	9.8	5.2	7.8#
68	7.2	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 0.305 ng
RT: 8.929 min Scan# 782
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion:	82	Resp:	7527
Ion Ratio	Lower	Upper	
82	100		
128	44.1	34.1	51.1
54	52.6	43.5	65.3









#9
Naphthalene
Concen: 0.298 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.1	13.7
127	13.1	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.317 ng
RT: 10.925 min Scan# 969
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.5	50.3	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.292 ng
RT: 12.172 min Scan# 1155
Delta R.T. -0.005 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:152 Resp: 12754
Ion Ratio Lower Upper
152 100
151 20.9 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.255 ng
RT: 12.248 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion:142 Resp: 14295
Ion Ratio Lower Upper
142 100
141 90.5 70.3 105.5
115 35.9 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.430 min Scan# 1453
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion	Ratio	Lower	Upper
164	100		
162	107.3	83.0	124.4
160	57.4	40.7	61.1



#14
2,4,6-Tribromophenol
Concen: 0.210 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion	Ratio	Lower	Upper
330	100		
332	99.6	76.6	114.8
141	45.0	34.1	51.1









#15
2-Fluorobiphenyl
Concen: 0.323 ng
RT: 13.051 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.0	27.8	41.6
170	24.3	18.1	27.1



#16
Acenaphthylene
Concen: 0.317 ng
RT: 14.152 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.8	23.8
153	13.0	10.2	15.2









#17
Acenaphthene
Concen: 0.285 ng
RT: 14.495 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion	154	153	152
Resp:	14121	115.5	62.6
Ratio	100	115.5	62.6
Lower		90.0	47.3
Upper		135.0	70.9



#18
Fluorene
Concen: 0.291 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion	166	165	167
Resp:	18901	99.6	13.4
Ratio	100	99.6	13.4
Lower		79.0	10.7
Upper		118.4	16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	11.1	8.6	13.0



#20
4,6-Dinitro-2-methylphenol
Concen: 0.328 ng
RT: 15.560 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion	Ratio	Lower	Upper
198	100		
51	131.7	88.9	133.3
105	76.9	49.2	73.8









#21
4-Bromophenyl-phenylether
Concen: 0.288 ng
RT: 16.379 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:248 Resp: 5427
Ion Ratio Lower Upper
248 100
250 94.4 76.2 114.2
141 72.0 57.6 86.4



#22
Hexachlorobenzene
Concen: 0.321 ng
RT: 16.503 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion:284 Resp: 6880
Ion Ratio Lower Upper
284 100
142 38.7 30.5 45.7
249 29.9 23.5 35.3









#23
Atrazine
Concen: 0.258 ng
RT: 16.652 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion	Ratio	Lower	Upper
200	100		
173	28.0	21.4	32.2
215	45.8	36.7	55.1



#24
Pentachlorophenol
Concen: 0.368 ng
RT: 16.838 min Scan# 1660
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.1	49.3	73.9
268	63.3	50.6	75.8









#25
Phenanthrene
Concen: 0.291 ng
RT: 17.223 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:178 Resp: 26605
Ion Ratio Lower Upper
178 100
176 19.5 15.5 23.3
179 15.5 12.2 18.4



#26
Anthracene
Concen: 0.291 ng
RT: 17.322 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion:178 Resp: 23247
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.4 11.8 17.8









#27
Fluoranthene-d10
Concen: 0.254 ng
RT: 19.220 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion	Ratio	Lower	Upper
212	100		
106	16.8	12.6	19.0
104	9.5	7.1	10.7



#28
Fluoranthene
Concen: 0.255 ng
RT: 19.253 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion	Ratio	Lower	Upper
202	100		
101	11.9	9.4	14.2
203	16.9	13.4	20.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:240 Resp: 18053
Ion Ratio Lower Upper
240 100
120 14.7 10.7 16.1
236 28.9 23.1 34.7



#30
Pyrene
Concen: 0.314 ng
RT: 19.615 min Scan# 2032
Delta R.T. 0.005 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion:202 Resp: 25592
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.4 14.2 21.2









#31
Terphenyl-d14
Concen: 0.363 ng
RT: 19.824 min Scan# 2076
Delta R.T. 0.005 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion:244 Resp: 14219
Ion Ratio Lower Upper
244 100
212 9.6 7.8 11.6
122 18.1 15.2 22.8



#32
Benzo(a)anthracene
Concen: 0.285 ng
RT: 21.358 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion:228 Resp: 18334
Ion Ratio Lower Upper
228 100
226 28.4 22.2 33.4
229 19.9 15.8 23.8









#33
Chrysene
Concen: 0.311 ng
RT: 21.411 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56




Instrument :
BNA_N
ClientSampleId :
PB170381BSD


Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.6	24.3	36.5
229	19.6	15.7	23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.227 ng
RT: 21.295 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56


Tgt Ion	Ratio	Lower	Upper
149	100		
167	27.0	21.1	31.7
279	3.6	2.9	4.3








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.683 min Scan# 2821
Delta R.T. 0.003 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Instrument :
BNA_N
ClientSampleId :
PB170381BSD







Tgt Ion:264 Resp: 14856
Ion Ratio Lower Upper
264 100
260 27.5 21.8 32.8
265 85.8 68.6 102.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.315 ng
RT: 26.037 min Scan# 3626
Delta R.T. 0.003 min
Lab File: BN038133.D
Acq: 04 Nov 2025 12:56

Tgt Ion:276 Resp: 19208
Ion Ratio Lower Upper
276 100
138 28.0 21.4 32.0
277 25.2 19.2 28.8





#37

Benzo(b)fluoranthene

Concen: 0.290 ng

RT: 22.984 min Scan# 2582

Delta R.T. 0.003 min

Lab File: BN038133.D

Acq: 04 Nov 2025 12:56

Instrument :

BNA_N

ClientSampleId :

PB170381BSD





Tgt Ion:252 Resp: 18434

Ion Ratio Lower Upper

252 100

253 26.9 20.9 31.3

125 20.3 14.1 21.1



#38

Benzo(k)fluoranthene

Concen: 0.308 ng

RT: 23.031 min Scan# 2598

Delta R.T. 0.003 min

Lab File: BN038133.D

Acq: 04 Nov 2025 12:56




Tgt Ion:252 Resp: 19683


Ion Ratio Lower Upper

252 100

253 26.4 21.2 31.8

125 20.3 14.6 22.0





#39

Benzo(a)pyrene

Concen: 0.296 ng

RT: 23.581 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN038133.D

Acq: 04 Nov 2025 12:56

Instrument :

BNA_N

ClientSampleId :

PB170381BSD





Tgt Ion:252 Resp: 15027

Ion Ratio Lower Upper

252 100

253 29.7 23.1 34.7

125 25.7 18.6 28.0



#40

Dibenzo(a,h)anthracene

Concen: 0.315 ng

RT: 26.054 min Scan# 3632

Delta R.T. 0.006 min

Lab File: BN038133.D

Acq: 04 Nov 2025 12:56




Tgt Ion:278 Resp: 14795


Ion Ratio Lower Upper

278 100

139 21.6 17.6 26.4

279 28.7 22.6 33.8





#41

Benzo(g,h,i)perylene

Concen: 0.321 ng

RT: 26.753 min Scan# 38

Delta R.T. 0.003 min

Lab File: BN038133.D


Acq: 04 Nov 2025 12:56

Instrument :

BNA_N

ClientSampleId :

PB170381BSD





Tgt Ion:276 Resp: 17207

Ion Ratio Lower Upper

276 100

277 24.3 19.7 29.5

138 26.9 20.2 30.4





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BN102825	Instrument	BNA_n
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC0.8	BN038114.D	Benzo(b)fluoranthene	rahul	10/29/2025 9:43:43 AM	mohammad	10/30/2025 5:29:45 AM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

BN110425

Instrument

BNA_n

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN102825

Review By	rahul	Review On	10/28/2025 3:41:07 PM
Supervise By	mohammad	Supervise On	10/30/2025 5:29:45 AM
SubDirectory	BN102825	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method BN102825
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904		
CCC	SP6902		
Internal Standard/PEM	SP6830,1ul/100ul sample		
ICV/I.BLK	SP6906		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN038110.D	28 Oct 2025 12:05	RC/JU	Ok
2	SSTDICC0.1	BN038111.D	28 Oct 2025 13:00	RC/JU	Ok
3	SSTDICC0.2	BN038112.D	28 Oct 2025 13:36	RC/JU	Ok
4	SSTDICCC0.4	BN038113.D	28 Oct 2025 14:13	RC/JU	Ok
5	SSTDICC0.8	BN038114.D	28 Oct 2025 14:49	RC/JU	Ok,M
6	SSTDICC1.6	BN038115.D	28 Oct 2025 15:25	RC/JU	Ok
7	SSTDICC3.2	BN038116.D	28 Oct 2025 16:02	RC/JU	Ok
8	SSTDICC5.0	BN038117.D	28 Oct 2025 16:38	RC/JU	Ok
9	SSTDICV0.4	BN038118.D	28 Oct 2025 17:51	RC/JU	Ok
10	PB170167BL	BN038119.D	28 Oct 2025 19:04	RC/JU	Not Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN110425

Review By	rahul	Review On	11/5/2025 10:39:28 AM
Supervise By	Jagrut	Supervise On	11/5/2025 2:05:23 PM
SubDirectory	BN110425	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method BN102825
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904		
CCC	SP6902		
Internal Standard/PEM	SP6830,1ul/100ul sample		
ICV/I.BLK	SP6906		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN038129.D	04 Nov 2025 10:00	RC/JU	Ok
2	SSTDCCC0.4	BN038130.D	04 Nov 2025 10:40	RC/JU	Ok
3	PB170381BL	BN038131.D	04 Nov 2025 11:43	RC/JU	Ok
4	PB170381BS	BN038132.D	04 Nov 2025 12:19	RC/JU	Ok
5	PB170381BSD	BN038133.D	04 Nov 2025 12:56	RC/JU	Ok
6	Q3494-01	BN038134.D	04 Nov 2025 13:32	RC/JU	Ok
7	Q3525-01	BN038135.D	04 Nov 2025 14:08	RC/JU	Dilution
8	Q3525-02	BN038136.D	04 Nov 2025 14:45	RC/JU	Ok
9	Q3525-03	BN038137.D	04 Nov 2025 15:21	RC/JU	Ok
10	Q3526-01	BN038138.D	04 Nov 2025 15:57	RC/JU	Ok
11	Q3526-02	BN038139.D	04 Nov 2025 16:34	RC/JU	Ok
12	Q3526-03	BN038140.D	04 Nov 2025 17:10	RC/JU	Ok
13	Q3526-04	BN038141.D	04 Nov 2025 17:46	RC/JU	Ok
14	Q3527-01	BN038142.D	04 Nov 2025 18:22	RC/JU	Ok
15	Q3527-02	BN038143.D	04 Nov 2025 18:59	RC/JU	Ok
16	Q3525-01DL	BN038144.D	04 Nov 2025 19:35	RC/JU	Ok
17	SSTDCCC0.4	BN038145.D	04 Nov 2025 20:12	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102825

Review By	rahul	Review On	10/28/2025 3:41:07 PM
Supervise By	mohammad	Supervise On	10/30/2025 5:29:45 AM
SubDirectory	BN102825	HP Acquire Method	BNA_N, 8270_HP Processing Method BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038110.D	28 Oct 2025 12:05		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN038111.D	28 Oct 2025 13:00		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN038112.D	28 Oct 2025 13:36		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN038113.D	28 Oct 2025 14:13	compound #20 kept on LR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN038114.D	28 Oct 2025 14:49		RC/JU	Ok,M
6	SSTDICC1.6	SSTDICC1.6	BN038115.D	28 Oct 2025 15:25		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN038116.D	28 Oct 2025 16:02		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN038117.D	28 Oct 2025 16:38	Compound #20 removed from 5PPM	RC/JU	Ok
9	SSTDICV0.4	ICVBN102825	BN038118.D	28 Oct 2025 17:51		RC/JU	Ok
10	PB170167BL	PB170167BL	BN038119.D	28 Oct 2025 19:04	END CCC missing, Analyzed for Contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN110425

Review By	rahul	Review On	11/5/2025 10:39:28 AM
Supervise By	Jagrut	Supervise On	11/5/2025 2:05:23 PM
SubDirectory	BN110425	HP Acquire Method	BNA_N, 8270_HP Processing Method BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038129.D	04 Nov 2025 10:00		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038130.D	04 Nov 2025 10:40		RC/JU	Ok
3	PB170381BL	PB170381BL	BN038131.D	04 Nov 2025 11:43		RC/JU	Ok
4	PB170381BS	PB170381BS	BN038132.D	04 Nov 2025 12:19		RC/JU	Ok
5	PB170381BSD	PB170381BSD	BN038133.D	04 Nov 2025 12:56	Recovery fail low for 1,4-Dioxane.	RC/JU	Ok
6	Q3494-01	MW3	BN038134.D	04 Nov 2025 13:32		RC/JU	Ok
7	Q3525-01	RW5-SP100-20251030	BN038135.D	04 Nov 2025 14:08	Need 2X Dilution	RC/JU	Dilution
8	Q3525-02	RW5-SP201-20251030	BN038136.D	04 Nov 2025 14:45		RC/JU	Ok
9	Q3525-03	RW5-SP303-20251030	BN038137.D	04 Nov 2025 15:21		RC/JU	Ok
10	Q3526-01	RW7-SP100-20251030	BN038138.D	04 Nov 2025 15:57		RC/JU	Ok
11	Q3526-02	RW7-SP201-20251030	BN038139.D	04 Nov 2025 16:34		RC/JU	Ok
12	Q3526-03	RW7-SP302-20251030	BN038140.D	04 Nov 2025 17:10		RC/JU	Ok
13	Q3526-04	RW7-SP303-20251030	BN038141.D	04 Nov 2025 17:46		RC/JU	Ok
14	Q3527-01	RW8-SP100-20251030	BN038142.D	04 Nov 2025 18:22		RC/JU	Ok
15	Q3527-02	RW8-SP303-20251030	BN038143.D	04 Nov 2025 18:59		RC/JU	Ok
16	Q3525-01DL	RW5-SP100-20251030	BN038144.D	04 Nov 2025 19:35		RC/JU	Ok
17	SSTDCCC0.4	SSTDCCC0.4EC	BN038145.D	04 Nov 2025 20:12		RC/JU	Ok

M : Manual Integration

SOP ID: M3510C,3580A-Extraction SVOC-21

Clean Up SOP #: N/A

Extraction Start Date : 11/03/2025

Matrix : Water

Extraction Start Time : 11:20

Weigh By: N/A

Extraction By: RS

Extraction End Date : 11/03/2025

Balance check: N/A

Filter By: RS

Extraction End Time : 16:20

Balance ID: N/A

pH Meter ID: N/A

Concentration By: EH

pH Strip Lot#: E3953

Hood ID: 4,6,7

Supervisor By : RUPESH

Extraction Method: ☒ Separatory Funnel

☐ Continuous Liquid/Liquid

☐ Sonication

☐ Waste Dilution

☐ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	0.4 PPM	SP6876
Surrogate	1.0ML	0.4 PPM	SP6873
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3980
Baked Na2SO4	N/A	EP2655
H2SO4 1:1	N/A	EP2610
10N NaOH	N/A	EP2609
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

PH Adjusted TO <2 with 1:1 H2SO4 & >11with 10N NAOH ,1.5ML Vial Lot #2210443.

KD Bath ID: WATER BATH-1

Envap ID: NEVAP-02

KD Bath Temperature: 60 °C


Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/3/25	RS (Ext Lab)	Relsvoc
16:25	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction SVOC-21

Concentration Date: 11/03/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170381BL	SBLK381	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			SEP-1
PB170381BS	SLCS381	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			2
PB170381BS D	SLCSD381	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			3
Q3494-01	MW3	SVOC-SIMGrou p1	990	6	RUPESH	ritesh	1	D		4
Q3525-01	RW5-SP100-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			5
Q3525-02	RW5-SP201-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			6
Q3525-03	RW5-SP303-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			7
Q3526-01	RW7-SP100-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			8
Q3526-02	RW7-SP201-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			9
Q3526-03	RW7-SP302-20251030	SVOC-SIMGrou p1	940	6	RUPESH	ritesh	1			10
Q3526-04	RW7-SP303-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			11
Q3527-01	RW8-SP100-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1	D		12
Q3527-02	RW8-SP303-20251030	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1	D		13



170381
11:20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3526 WorkList ID : 192855 Department : Extraction Date : 11-03-2025 11:15:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3494-01	MW3	Water	SVOC-SIMGroup1	Cool 4 deg C	GENV01	J31	10/29/2025	8270-Modified
Q3525-01	RW5-SP100-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3525-02	RW5-SP201-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3525-03	RW5-SP303-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3526-01	RW7-SP100-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3526-02	RW7-SP201-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3526-03	RW7-SP302-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3526-04	RW7-SP303-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	A12	10/30/2025	8270-Modified
Q3527-01	RW8-SP100-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	J32	10/30/2025	8270-Modified
Q3527-02	RW8-SP303-20251030	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	J32	10/30/2025	8270-Modified

Date/Time 11/3/25 11:15
Raw Sample Received by: RS (Ex-66)
Raw Sample Relinquished by: RS (Ex-66)

Date/Time 11/3/25 11:50
Raw Sample Received by: RS (Ex-66)
Raw Sample Relinquished by: RS (Ex-66)

Prep Standard - Chemical Standard Summary

Order ID : Q3526

Test : SVOC-SIMGroup1

Prepbatch ID : PB170381,

Sequence ID/Qc Batch ID: BN110425,

Standard ID :

EP2609,EP2610,EP2655,SP6830,SP6873,SP6875,SP6876,SP6896,SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904,SP6905,SP6906,

Chemical ID :

1ul/100ul

sample,E3657,E3875,E3942,E3972,E3973,E3980,M6157,S10105,S11497,S11652,S11828,S12116,S12201,S12273,S12560,S12563,S12577,S12670,S12905,S13058,S13097,S13120,S13122,S13245,S13249,S13299,W 3112,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1874	10 N SODIUM HYDROXIDE SOLN	EP2609	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025
<u>FROM</u>	1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025
<u>FROM</u>	1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2655	10/24/2025	01/28/2026	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/24/2025
<u>FROM</u> 4000.00000gram of E3875 = Final Quantity: 4000.000 gram								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3493	Internal Standard 0.4 PPM	SP6830	06/17/2025	12/13/2025	Jagrut Upadhyay	None	None	Rahul Chavli 06/19/2025
<u>FROM</u>	0.10000ml of S12670 + 4.90000ml of E3942 = Final Quantity: 5.000 ml							

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3491	8270-SIM-Surrogate 0.4 PPM	SP6873	09/30/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								09/30/2025

FROM 0.00400ml of S12201 + 0.00800ml of S12905 + 0.02000ml of S11828 + 99.96800ml of E3972 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3895	50 ug/ml DFTPP 8270E	SP6875	09/30/2025	03/30/2026	Rahul Chavli	None	None	Jagrut Upadhyay
								09/30/2025

FROM 1.00000ml of S12577 + 19.00000ml of E3973 = Final Quantity: 20.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3492	8270-SIM-Spike 0.4 PPM	SP6876	09/30/2025	11/30/2025	Jagrut Upadhyay	None	None	Rahul Chavli 09/30/2025
<u>FROM</u> 0.00080ml of S11652 + 0.01000ml of S13120 + 0.02000ml of S12560 + 0.02000ml of S13097 + 0.02000ml of S13245 + 49.92920ml of E3972 = Final Quantity: 50.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3339	8270 sim calibration stock 10ppm (CPI)	SP6896	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli 11/06/2025
<u>FROM</u>	0.03350ml of S10105 + 0.05000ml of S11497 + 0.12500ml of S11828 + 0.12500ml of S12116 + 0.25000ml of S12273 + 0.25000ml of S13058 + 24.16650ml of E3980 = Final Quantity: 25.000 ml							

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3361	8270-SIM MDL-5PPM CALIBRATION SOLUTION	SP6897	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.50000ml of E3980 + 0.01000ml of SP6830 + 0.50000ml of SP6896 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3341	8270-SIM MDL-3.2PPM CALIBRATION SOLUTION	SP6898	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.68000ml of E3980 + 0.01000ml of SP6830 + 0.32000ml of SP6896 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3344	8270-SIM MDL-1.6PPM CALIBRATION SOLUTION	SP6900	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.84000ml of E3980 + 0.01000ml of SP6830 + 0.16000ml of SP6896 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3342	8270-SIM MDL-0.8PPM CALIBRATION SOLUTION	SP6901	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.92000ml of E3980 + 0.01000ml of SP6830 + 0.08000ml of SP6896 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3343	8270-SIM MDL-0.4PPM CALIBRATION SOLUTION	SP6902	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.96000ml of E3980 + 0.01000ml of SP6830 + 0.04000ml of SP6896 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3345	8270-SIM MDL-0.2PPM CALIBRATION SOLUTION	SP6903	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								11/06/2025

FROM 0.50000ml of E3980 + 0.01000ml of SP6830 + 0.50000ml of SP6902 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3346	8270-SIM MDL-0.1PPM CALIBRATION SOLUTION	SP6904	10/22/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli 11/06/2025

FROM 0.75000ml of E3980 + 0.01000ml of SP6830 + 0.25000ml of SP6902 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3355	8270-SIM MDL-3.2PPM CALIBRATION STOCK SOL- 2ND	SP6905	10/27/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli 11/06/2025

FROM SOURCE
0.00630ml of S12201 + 0.01280ml of S12905 + 0.03200ml of S11828 + 0.03200ml of S13122 + 0.06400ml of S12563 +
0.06400ml of S13249 + 0.06400ml of S13299 + 19.72490ml of E3980 = Final Quantity: 20.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3356	8270-SIM MDL-0.4PPM CALIBRATION SOL ICV-2ND	SP6906	10/27/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
SOURCE 0.87500ml of E3980 + 0.01000ml of SP6830 + 0.12500ml of SP6905 = Final Quantity: 1.010 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	07/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A2862010	12/13/2025	06/13/2025 / Rajesh	02/28/2025 / Rajesh	E3942

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25C1262005	04/16/2026	09/15/2025 / Riteshkumar	09/15/2025 / Riteshkumar	E3973

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25C1262005	04/16/2026	10/10/2025 / RUPESH	10/10/2025 / RUPESH	E3980

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	12/19/2025	06/19/2025 / Jagrut	12/09/2021 / Christian	S10105

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	04/22/2026	10/22/2025 / Jagrut	08/11/2023 / Yogesh	S11497

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0201728	01/30/2026	07/30/2025 / Rahul	11/09/2023 / Yogesh	S11652

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	33913 / SOM01.0 SIM Analysis Standard (Surrogate), 2000 PPM	A0201976	12/09/2025	06/09/2025 / Jagrut	11/21/2023 / rahul	S11828

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	04/22/2026	10/22/2025 / Jagrut	03/08/2024 / Rahul	S12116

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml, methanol, 5ml/ ampul	A0206206	03/22/2026	09/22/2025 / Jagrut	03/15/2024 / Rahul	S12201

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	520963	12/25/2025	06/25/2025 / Jagrut	05/24/2024 / Rahul	S12273

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	03/22/2026	09/22/2025 / Jagrut	07/23/2024 / RAHUL	S12560

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	04/16/2026	10/16/2025 / Jagrut	07/23/2024 / RAHUL	S12563

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH ₂ Cl ₂ , 1mL,	A0212955	06/30/2027	03/31/2025 / Rahul	08/01/2024 / Rahul	S12577

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH ₂ Cl ₂ , 1mL	A0212266	12/16/2025	06/16/2025 / anahy	09/20/2024 / anahy	S12670

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml, CH ₂ Cl ₂ , 5ml	A0216785	03/22/2026	09/22/2025 / Jagrut	12/09/2024 / anahy	S12905

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	531243	12/25/2025	06/25/2025 / Jagrut	01/16/2025 / anahy	S13058

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0221014	11/30/2025	09/03/2025 / Jagrut	05/20/2025 / Rahul	S13097

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0218894	02/25/2026	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13120

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0218894	04/16/2026	10/16/2025 / Jagrut	05/20/2025 / Rahul	S13122

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0228451	03/23/2026	09/22/2025 / Jagrut	08/06/2025 / Rahul	S13245

[CS 4978-1]

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0228451	04/27/2026	10/27/2025 / Jagrut	08/06/2025 / Rahul	S13249

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0229652	04/27/2026	10/27/2025 / Jagrut	10/15/2025 / rahul	S13299

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: **Storage:** **Solvent:** **Exp. Date:** **Description:**
Z-112090 440246 $\leq -10^{\circ}\text{C}$ Methylene Chloride 2/16/2026 CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL
-04

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d ₄	93951-73-6	99.3	248.12.7P	7487 $\pm$ 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 $\pm$ 17.26
phenol-d ₆	13127-88-3	99.9	949.120.8P	7481 $\pm$ 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 $\pm$ 17.17

Received on

02/25/21

by
CG

S9236
to

S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Erica Castiglione

Certified By:

Erica Castiglione
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	$\leq 0.005 \%$	$< 0.005 \%$	PASS
Chloride	$\leq 0.005 \%$	0.002 %	PASS
Heavy Metals	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Iron	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Magnesium	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Mercury	$\leq 0.1 \text{ ppm}$	$< 0.1 \text{ ppm}$	PASS
Nickel	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Nitrogen Compounds	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Phosphate	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Potassium	$\leq 0.02 \%$	$< 0.02 \%$	PASS
Purity	$\geq 97.0 \%$	99.2 %	PASS
Sodium Carbonate	$\leq 1.0 \%$	0.5 %	PASS
Sulfate	$\leq 0.003 \%$	$< 0.003 \%$	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.


This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.





**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 25A2862010
Manufactured Date: 2024-12-18
Expiration Date: 2026-03-19
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	99.9 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μ eq/g)	≤ 0.3	<0.1
Chloride (Cl)	≤ 10 ppm	<5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	<0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3942

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 25C1262005
Manufactured Date: 2025-01-15
Expiration Date: 2026-04-16
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.1 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	$\leq 10 \text{ ppm}$	$< 5 \text{ ppm}$
Water (by KF, coulometric)	$\leq 0.02 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

REC. on 10/10/25
RJ

E3980

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087 U.S.A. Phone 610.386.1700

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™



MG157
MS

Material No.: 9673-33

Batch No.: 24I1262013

Manufactured Date: 2024-08-07

Retest Date: 2029-08-06

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO ₂)	<= 2 ppm	<2 ppm
Ammonium (NH ₄)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO ₃)	<= 0.2 ppm	0.1 ppm
Phosphate (PO ₄)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

 **avantors**™



Material No.: 9673-33

Batch No.: 2411262013

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 1.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	1.1 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	< 1.0 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak

Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



5580 Skyline Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Page 1 of 1

Catalog No.: Lot No.:	Storage:	Solvent:	Exp. Date:	Rev	Description:
Z-110094-02 506889	≤ -10 °C	Methylene Chloride	7/25/2028	0	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml
Compound		CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d ₄		2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl		321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5		4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14		1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 } Y.P.
↓ 08/11/2023
511498

*Not a certified value

Certified By: _____
Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555872 Lot No.: A0201728

Description: Custom Pentachlorophenol Standard

Custom Pentachlorophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size: 2 mL Pkg Amt: > 1 mL

Expiration Date: September 30, 2026 Storage: 10°C or colder

Ship: Ambient

511649 } Y.P.
↓ 11/13/23
511658 }

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP230530RSR	99%	25,000.0 µg/mL	+/- 777.0837

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Josh McCloskey - Operations Technician I

Date Mixed: 05-Sep-2023 Balance: B251644995

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 33913 **Lot No.:** A0201976
Description : SOM01.0 SIM Analysis Standard
SOM01.0 SIM Analysis Standard 2000µg/mL, Methylene chloride, 1mL /ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S11828
↓
S11832 } RC/
11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Methylnaphthalene-d10	7297-45-2	EF-135	98%	2,015.9 µg/mL	+/- 90.8098
2	Fluoranthene-d10	93951-69-0	PR-32557	99%	2,020.0 µg/mL	+/- 90.9963

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 13-Sep-2023

Balance Serial # B442140311


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 28-Sep-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:	
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL	
Compound			CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane			123-91-1	100	223.1.3P	1997 ± 57.08

512112 } RC/
↓
912116 } 03/08/24

*Not a certified value

Certified By: _____

Melissa Workoff

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206

Description : Acid Surrogate Mix (4/89 SOW)
Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : January 31, 2032 **Storage:** 10°C or colder

Ship: Ambient

512187 } RC/
↓ } 03/18/24
512206 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope S. Riglin

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024

Balance Serial # 1128360905

Christie Mills

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 4

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	520963	≤ -10 °C	Methylene Chloride	10/10/2028	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1010 ± 9.89
acenaphthylene	208-96-8	97.6	14.290.1P	1014 ± 9.93
aniline	62-53-3	99.97	64.1.4P	1001 ± 9.8
anthracene	120-12-7	99.5	15.7.1P	999.6 ± 9.79
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 9.8
benzo[a]anthracene	56-55-3	100	16.7.3P	1007 ± 9.86
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1011 ± 14.11
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 10.96
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 13.95
benzo[a]pyrene	50-32-8	97	20.286.2P	999.9 ± 22.24
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 9.82
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	1000 ± 14.69
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1003 ± 13.89
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.15P	999.4 ± 14.68
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 9.8
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 17.03
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1.1P	1000 ± 13.85
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 16.79
carbazole	86-74-8	99.4	239.7.2P	1000 ± 9.8

*Not a certified value

512270 } RC/
↓
512274 } 05/24/24

Kerry Kane

Certified By: _____

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.

Certificate of Analysis

Page 2 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
4-chloroaniline	106-47-8	100	66.7.1P	1000 ± 9.79
4-chlorophenylphenyl ether	7005-72-3	98	37.158.2P	1001 ± 17.07
4-chloro-3-methylphenol	59-50-7	99	102.1.2P	1006 ± 17.16
2-chloronaphthalene	91-58-7	99.9	42.7.6P	1000 ± 9.79
2-chlorophenol	95-57-8	99.8	103.7.1P	1007 ± 13.96
chrysene	218-01-9	96	21.286.2P	998.4 ± 12.85
dibenz[a,h]anthracene	53-70-3	99.44	22.286.3P	1000 ± 9.74
dibenzofuran	132-64-9	100	67.7.2.1P	1002 ± 9.77
di-n-butyl phthalate	84-74-2	99.84	40.286.1P	1007 ± 24.48
1,2-dichlorobenzene	95-50-1	99.8	43.7.1P	1000 ± 9.79
1,3-dichlorobenzene	541-73-1	99.5	44.1.3P	999.4 ± 9.79
1,4-dichlorobenzene	106-46-7	99.9	45.29.2P	1000 ± 9.79
2,4-dichlorophenol	120-83-2	99.6	104.7.1.1P	1005 ± 13.93
diethyl phthalate	84-66-2	99.8	38.7.1P	1011 ± 14
2,4-dimethylphenol	105-67-9	99.6	105.7.1.1P	1009 ± 13.98
dimethyl phthalate	131-11-3	99.9	39.9.2P	996.5 ± 13.8
1,2-dinitrobenzene	528-29-0	99.86	86.7.3.1P	999.5 ± 9.75
1,3-dinitrobenzene	99-65-0	100	313.7.2P	998 ± 9.79
1,4-dinitrobenzene	100-25-4	100	907.7.1P	999.5 ± 9.8
2,4-dinitrophenol	51-28-5	99.9	106.1.6DP	1002 ± 13.89
2,4-dinitrotoluene	121-14-2	100	87.7.3P	999.8 ± 13.85
2,6-dinitrotoluene	606-20-2	99.4	88.7.2.1P	999.6 ± 13.85
di-n-octyl phthalate	117-84-0	99.1	41.7.5P	991.6 ± 13.74
diphenylamine	122-39-4	100	78.1.6P	998 ± 13.79
2,3,5,6-tetrachlorophenol	935-95-5	97	1112.286.1P	1004 ± 14.02
fluoranthene	206-44-0	98.6	23.7.4P	999.6 ± 9.79
fluorene	86-73-7	98.4	24.7.1P	999.7 ± 9.79

*Not a certified value

Kerry Kane

Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 3 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
hexachlorobenzene	118-74-1	99	46.158.4P	999.9 ± 13.96
hexachlorobutadiene	87-68-3	97.4	47.1.4P	1000 ± 9.79
hexachlorocyclopentadiene	77-47-4	99.2	48.2.2P	1001 ± 9.8
hexachloroethane	67-72-1	99.9	49.1.4P	1003 ± 9.82
indeno[1,2,3-cd]pyrene	193-39-5	98	25.286.4P	999.4 ± 22.23
isophorone	78-59-1	98.9	90.1.4P	999.9 ± 13.85
2-methyl-4,6-dinitrophenol	534-52-1	99.6	107.421.2DP	991 ± 24.09
1-methylnaphthalene	90-12-0	97.1	249.7.5P	999.2 ± 13.95
2-methylnaphthalene	91-57-6	97.4	68.7.2P	1006 ± 22.38
2-methylphenol	95-48-7	99.6	114.7.3P	1001 ± 13.87
3-methylphenol	108-39-4	99.1	115.7.4P	499.7 ± 6.92
4-methylphenol	106-44-5	99.5	116.7.1P	501.2 ± 6.94
naphthalene	91-20-3	99.8	26.9.1P	1018 ± 9.97
2-nitroaniline	88-74-4	99.7	69.29.1P	999.6 ± 9.79
3-nitroaniline	99-09-2	100	70.7.3P	1000 ± 9.74
4-nitroaniline	100-01-6	99.7	71.29.1P	1001 ± 9.8
nitrobenzene	98-95-3	100	94.7.1P	1000 ± 13.85
2-nitrophenol	88-75-5	99.1	108.29.1P	996.5 ± 13.81
4-nitrophenol	100-02-7	100	109.7.1P	1000 ± 13.82
N-nitrosodimethylamine	62-75-9	99.5	57.3.19P	998.5 ± 14.67
N-nitrosodi-n-propylamine	621-64-7	99.8	59.286.1P	996.8 ± 17
pentachlorophenol	87-86-5	99	110.1.7P	1004 ± 13.92
phenanthrene	85-01-8	99.7	27.1.5P	999 ± 12.87
phenol	108-95-2	100	112.7.1P	998.5 ± 13.8
pyrene	129-00-0	99.2	28.9.2P	998.9 ± 9.78
pyridine	110-86-1	100	101.24.1P	999 ± 9.73
2,3,4,6-Tetrachlorophenol	58-90-2	91.8	120.421.1P	996.5 ± 13.92

*Not a certified value

Kerry Kane

Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 4 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

<u>Compound</u>	<u>CAS No.</u>	<u>Purity (%)</u>	<u>Compound Lot No.</u>	<u>Concentration, mg/L</u>
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	999.6 ± 9.79
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	999.5 ± 13.85
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	996 ± 13.8

*Not a certified value



Certified By: _____

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 } RC/
↓
S12568 } 7/24/24


Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 } RC/
↓
S12568 } 7/24/24


Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31615 **Lot No.:** A0212955

Description : GC/MS Tuning Mixture
GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2027 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,004.5 µg/mL	+/- 44.8902
2	DFTPP (Decafluorotriphenylphosphine)	5074-71-5	Q117-147	99%	1,004.5 µg/mL	+/- 44.8902
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 44.9572
4	4,4'-DDT	50-29-3	S240530RSR	97%	1,000.1 µg/mL	+/- 44.6922

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12577 } RC
↓
S12579 } 8/2/24

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Ethan Winiarski - Operations Tech I

Date Mixed: 19-Jun-2024

Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 26-Jun-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0212266

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,000.6 µg/mL	+/- 90.1075
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,000.3 µg/mL	+/- 90.0925
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,000.4 µg/mL	+/- 90.1000
4	Phenanthrene-d10	1517-22-2	PR-34099	99%	2,000.5 µg/mL	+/- 90.1037
5	Chrysene-d12	1719-03-5	PR-33506	99%	2,000.7 µg/mL	+/- 90.1112
6	Perylene-d12	1520-96-3	PR-33205	99%	2,000.6 µg/mL	+/- 90.1075

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12645
↓
S12674 } AC
10/1/24



5580 Skyline Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	531243	$\leq -10^{\circ}\text{C}$	Methylene Chloride	1/2/2030	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.4P	997 $\pm$ 5.81
benzidine	92-87-5	99.9	124.18.6.2P	993.8 $\pm$ 5.78
caprolactam	105-60-2	99.9	271.1.6P	999 $\pm$ 5.82

513057 } AC
↓
513061 } 1/16/25

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0221014

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : November 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

513088 }
↓
513117 } 24
5/20/25.

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBR4811	99%	1,005.2 µg/mL	+/- 36.5730
2	N-Nitrosodimethylamine	62-75-9	S241226RSR	99%	1,005.5 µg/mL	+/- 36.5848
3	Phenol	108-95-2	MKCK1120	99%	1,005.3 µg/mL	+/- 36.5780
4	Aniline	62-53-3	X22F726	99%	1,005.4 µg/mL	+/- 36.5816
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,004.7 µg/mL	+/- 36.5562
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.3 µg/mL	+/- 36.5766
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.6 µg/mL	+/- 36.5530
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,006.3 µg/mL	+/- 36.6125
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,005.5 µg/mL	+/- 36.5853
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,004.6 µg/mL	+/- 36.5507
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,004.8 µg/mL	+/- 36.5575
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,005.4 µg/mL	+/- 36.5803
13	3-Methylphenol (m-cresol)	108-39-4	STBL3873	99%	502.7 µg/mL	+/- 18.2888
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	502.6 µg/mL	+/- 18.2869
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,004.6 µg/mL	+/- 36.5502
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.6 µg/mL	+/- 36.5530
17	Nitrobenzene	98-95-3	10224044	99%	1,005.3 µg/mL	+/- 36.5780

18	Isophorone	78-59-1	MKCR3249	99%	1,004.6	µg/mL	+/- 36.5511
19	2-Nitrophenol	88-75-5	RP230710	99%	1,005.7	µg/mL	+/- 36.5916
20	2,4-Dimethylphenol	105-67-9	DIRAF	99%	1,004.9	µg/mL	+/- 36.5612
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,004.3	µg/mL	+/- 36.5402
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,004.7	µg/mL	+/- 36.5571
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,004.6	µg/mL	+/- 36.5516
24	Naphthalene	91-20-3	STBL1057	99%	1,006.8	µg/mL	+/- 36.6335
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,005.9	µg/mL	+/- 36.5980
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,004.1	µg/mL	+/- 36.5328
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,005.4	µg/mL	+/- 36.5793
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,006.3	µg/mL	+/- 36.6144
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063P13G	99%	1,005.9	µg/mL	+/- 36.5975
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,004.7	µg/mL	+/- 36.5566
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.7	µg/mL	+/- 36.5571
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,005.3	µg/mL	+/- 36.5757
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.8	µg/mL	+/- 36.5584
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,004.5	µg/mL	+/- 36.5475
36	Acenaphthylene	208-96-8	214935V18H	95%	1,000.1	µg/mL	+/- 36.3888
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,004.9	µg/mL	+/- 36.5616
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,004.5	µg/mL	+/- 36.5489
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,005.2	µg/mL	+/- 36.5734
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,005.9	µg/mL	+/- 36.6003
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5525
43	2,4-Dinitrophenol	51-28-5	D240927RSR	99%	1,005.0	µg/mL	+/- 36.5657
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,005.4	µg/mL	+/- 36.5803
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,005.4	µg/mL	+/- 36.5803
46	4-Nitrophenol	100-02-7	20241120-1-AN	99%	1,005.7	µg/mL	+/- 36.5930
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,004.9	µg/mL	+/- 36.5616
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP240523RSR	99%	1,005.2	µg/mL	+/- 36.5739
49	Fluorene	86-73-7	10246250	98%	1,005.8	µg/mL	+/- 36.5948
50	4-Chlorophenyl phenyl ether	7005-72-3	002531K02D	99%	1,004.8	µg/mL	+/- 36.5584
51	Diethylphthalate	84-66-2	STBL3611	99%	1,005.3	µg/mL	+/- 36.5762
52	4-Nitroaniline	100-01-6	RP240830RSR	99%	1,005.5	µg/mL	+/- 36.5844
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,004.8	µg/mL	+/- 36.5589

54	Diphenylamine	122-39-4	MKCT1512	99%	1,005.2	µg/mL	+/- 36.5725
55	Azobenzene	103-33-3	BCCL3292	99%	1,004.7	µg/mL	+/- 36.5566
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,005.6	µg/mL	+/- 36.5875
57	Hexachlorobenzene	118-74-1	15828800	99%	1,005.3	µg/mL	+/- 36.5789
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.4	µg/mL	+/- 36.5825
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.8	µg/mL	+/- 36.5584
60	Anthracene	120-12-7	MKCW9141	99%	1,005.5	µg/mL	+/- 36.5834
61	Carbazole	86-74-8	15630800	99%	1,005.3	µg/mL	+/- 36.5789
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,005.5	µg/mL	+/- 36.5848
63	Fluoranthene	206-44-0	A0458721	99%	1,005.4	µg/mL	+/- <0.0001
64	Pyrene	129-00-0	BCCL8032	99%	1,005.2	µg/mL	+/- 36.5734
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,004.8	µg/mL	+/- 36.5584
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCR8567	99%	1,004.5	µg/mL	+/- 36.5480
67	Benz(a)anthracene	56-55-3	I60012022BAA	99%	1,005.5	µg/mL	+/- 36.5844
68	Chrysene	218-01-9	RP241212RSR	99%	1,005.5	µg/mL	+/- 36.5848
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,004.7	µg/mL	+/- 36.5548
70	Di-n-octyl phthalate	117-84-0	15817300	99%	1,004.8	µg/mL	+/- 36.5607
71	Benzo(b)fluoranthene	205-99-2	022013B	99%	1,005.1	µg/mL	+/- 36.5716
72	Benzo(k)fluoranthene	207-08-9	012022K	98%	1,004.6	µg/mL	+/- 36.5511
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,004.2	µg/mL	+/- 36.5377
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,004.1	µg/mL	+/- 36.5337
75	Dibenz(a,h)anthracene	53-70-3	712061504-1-1	99%	1,005.7	µg/mL	+/- 36.5930
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,005.4	µg/mL	+/- 36.5814

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:


100 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Penelope Riglin - Operations Tech I

Date Mixed: 12-Jan-2025

Balance Serial # 1128360905


Dillan Murphy - Operations Technician I

Date Passed: 21-Jan-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0218894
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : November 30, 2029 **Storage:** 0°C or colder
Ship: Ambient

513118 } RC/
↓
513147 } 5/20/25.

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBQ1693	99%	2,002.4 µg/mL	+/- 24.9202

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Righlin - Operations Tech I

Date Mixed: 07-Nov-2024

Balance Serial # 1128360905

Dillan Murphy - Operations Technician I

Date Passed: 11-Nov-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0218894
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : November 30, 2029 **Storage:** 0°C or colder
Ship: Ambient

513118 } RC/
↓
513147 } 5/20/25.

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBQ1693	99%	2,002.4 µg/mL	+/- 24.9202

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 07-Nov-2024

Balance Serial # 1128360905

Dillan Murphy - Operations Technician I

Date Passed: 11-Nov-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0228451

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2027 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

S13239 } RC/
↓
S13268 } 08/06/25

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S250226RSR	99%	1,006.0 µg/mL	+/- 23.0947
2	Atrazine	1912-24-9	5FYWL	99%	1,007.0 µg/mL	+/- 23.1176
3	Benzidine	92-87-5	S250227ECS	99%	1,003.0 µg/mL	+/- 23.0258
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,003.0 µg/mL	+/- 23.0258

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tom Sucka - Mix Technician

Date Mixed: 01-Aug-2025

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0228451

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2027 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

S13239 } RC/
↓
S13268 } 08/06/25

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S250226RSR	99%	1,006.0 µg/mL	+/- 23.0947
2	Atrazine	1912-24-9	5FYWL	99%	1,007.0 µg/mL	+/- 23.1176
3	Benzidine	92-87-5	S250227ECS	99%	1,003.0 µg/mL	+/- 23.0258
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,003.0 µg/mL	+/- 23.0258

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%


Tom Sucka - Mix Technician

Date Mixed: 01-Aug-2025

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0229652

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

513299 } RC/
↓ 10/17/25
513328 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBR4811	99%	1,004.9 µg/mL	+/- 36.5621
2	N-Nitrosodimethylamine	62-75-9	S250717RSR	99%	1,004.5 µg/mL	+/- 36.5484
3	Phenol	108-95-2	MKCT5446	99%	1,003.5 µg/mL	+/- 36.5120
4	Aniline	62-53-3	X22F726	99%	1,004.8 µg/mL	+/- 36.5575
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,003.6 µg/mL	+/- 36.5166
6	2-Chlorophenol	95-57-8	STBK4742	99%	1,004.5 µg/mL	+/- 36.5484
7	1,3-Dichlorobenzene	541-73-1	STBL1702	99%	1,003.8 µg/mL	+/- 36.5211
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.8 µg/mL	+/- 36.5211
9	Benzyl alcohol	100-51-6	094986W07G	99%	1,003.8 µg/mL	+/- 36.5211
10	1,2-Dichlorobenzene	95-50-1	SHBR4446	99%	1,004.9 µg/mL	+/- 36.5621
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,004.3 µg/mL	+/- 36.5393
12	2,2'-oxybis(1-chloropropane)	108-60-1	RP250904RSR	99%	1,005.0 µg/mL	+/- 36.5666
13	3-Methylphenol (m-cresol)	108-39-4	STBL3873	99%	502.4 µg/mL	+/- 18.2788
14	4-Methylphenol (p-cresol)	106-44-5	SHBQ7653	99%	502.3 µg/mL	+/- 18.2742
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,004.9 µg/mL	+/- 36.5621
16	Hexachloroethane	67-72-1	DAXRI	99%	1,003.9 µg/mL	+/- 36.5257
17	Nitrobenzene	98-95-3	10224044	99%	1,003.9 µg/mL	+/- 36.5257

18	Isophorone	78-59-1	MKCR3249	99%	1,004.8	µg/mL	+/- 36.5575
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.6	µg/mL	+/- 36.5166
20	2,4-Dimethylphenol	105-67-9	DIRAF	99%	1,004.1	µg/mL	+/- 36.5348
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,004.5	µg/mL	+/- 36.5484
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,004.8	µg/mL	+/- 36.5575
23	1,2,4-Trichlorobenzene	120-82-1	SHBR1701	99%	1,004.0	µg/mL	+/- 36.5302
24	Naphthalene	91-20-3	STBL1057	99%	1,004.0	µg/mL	+/- 36.5302
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,004.5	µg/mL	+/- 36.5484
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,002.7	µg/mL	+/- 36.4816
27	4-Chloro-3-methylphenol	59-50-7	BCCK5906	99%	1,004.3	µg/mL	+/- 36.5393
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,000.5	µg/mL	+/- 36.4029
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3847
30	Hexachlorocyclopentadiene	77-47-4	099063P13G	99%	1,004.0	µg/mL	+/- 36.5302
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,003.5	µg/mL	+/- 36.5120
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,003.6	µg/mL	+/- 36.5152
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,003.6	µg/mL	+/- 36.5166
34	2-Nitroaniline	88-74-4	RP250625RSR	99%	1,003.8	µg/mL	+/- 36.5211
35	1,4-Dinitrobenzene	100-25-4	RP250401RSR	99%	1,004.6	µg/mL	+/- 36.5530
36	Acenaphthylene	208-96-8	214935V18H	95%	1,000.4	µg/mL	+/- 36.3974
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,004.8	µg/mL	+/- 36.5575
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,003.5	µg/mL	+/- 36.5120
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,004.6	µg/mL	+/- 36.5530
40	1,2-Dinitrobenzene	528-29-0	RP250203RSR	99%	1,002.6	µg/mL	+/- 36.4802
41	Acenaphthene	83-32-9	MKCV8166	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,003.8	µg/mL	+/- 36.5211
43	2,4-Dinitrophenol	51-28-5	D240927RSR	99%	1,004.1	µg/mL	+/- 36.5348
44	Dibenzofuran	132-64-9	MKCW3845	99%	1,003.6	µg/mL	+/- 36.5166
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,003.8	µg/mL	+/- 36.5211
46	4-Nitrophenol	100-02-7	20241120-1-AN	99%	1,003.0	µg/mL	+/- 36.4938
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,003.8	µg/mL	+/- 36.5211
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP250724RSR	99%	1,003.1	µg/mL	+/- 36.4984
49	Fluorene	86-73-7	10246250	98%	1,002.9	µg/mL	+/- 36.4905
50	4-Chlorophenyl phenyl ether	7005-72-3	002531K02D	99%	1,004.3	µg/mL	+/- 36.5393
51	Diethylphthalate	84-66-2	223219R19C	99%	1,004.8	µg/mL	+/- 36.5575
52	4-Nitroaniline	100-01-6	RP240830RSR	99%	1,004.0	µg/mL	+/- 36.5302
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S250805RSR	99%	1,004.8	µg/mL	+/- 36.5575

54	Diphenylamine	122-39-4	MKCT1512	99%	1,004.9	µg/mL	+/- 36.5621
55	Azobenzene	103-33-3	BCCL3292	99%	1,003.9	µg/mL	+/- 36.5257
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.9	µg/mL	+/- 36.5257
57	Hexachlorobenzene	118-74-1	16302300	99%	1,004.5	µg/mL	+/- 36.5484
58	Pentachlorophenol	87-86-5	RP240411RSR	99%	1,004.9	µg/mL	+/- 36.5621
59	Phenanthrene	85-01-8	MKCV8193	99%	1,003.8	µg/mL	+/- 36.5211
60	Anthracene	120-12-7	MKCW9141	99%	1,003.3	µg/mL	+/- 36.5029
61	Carbazole	86-74-8	15821400	99%	1,003.1	µg/mL	+/- 36.4984
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,004.6	µg/mL	+/- 36.5530
63	Fluoranthene	206-44-0	A0458721	99%	1,004.8	µg/mL	+/- 36.5575
64	Pyrene	129-00-0	BCCL8032	99%	1,003.8	µg/mL	+/- 36.5211
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,004.9	µg/mL	+/- 36.5621
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,003.4	µg/mL	+/- 36.5075
67	Benz(a)anthracene	56-55-3	I80012022BAA	99%	1,004.3	µg/mL	+/- 36.5393
68	Chrysene	218-01-9	RP250815RSR	99%	1,003.9	µg/mL	+/- 36.5257
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,004.5	µg/mL	+/- 36.5484
70	Di-n-octyl phthalate	117-84-0	16197600	99%	1,004.4	µg/mL	+/- 36.5439
71	Benzo(b)fluoranthene	205-99-2	0225BF	99%	1,005.4	µg/mL	+/- 36.5803
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,004.5	µg/mL	+/- 36.5484
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,002.4	µg/mL	+/- 36.4726
74	Indeno(1,2,3-cd)pyrene	193-39-5	17-YMK-40-2	99%	1,004.0	µg/mL	+/- 36.5302
75	Dibenz(a,h)anthracene	53-70-3	712215450-1-1	99%	1,003.5	µg/mL	+/- 36.5120
76	Benzo(g,h,i)perylene	191-24-2	RP250501RSR	98%	1,003.0	µg/mL	+/- 36.4949

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 04-Sep-2025

Balance Serial # B345965662


Brittany Federinko - Operations Tech II

Date Passed: 15-Sep-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



SHIPPING DOCUMENTS



CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 78-8922

www.chemtech.net

Chemtech Project Number:

23526

COC Number:

CLIENT INFORMATION

COMPANY: Tetra Tech

ADDRESS: 4433 Corporation Ln, Suite 300

CITY: Virginia Beach STATE: VA ZIP: 23462

ATTENTION: Ernie Wu

PHONE: 757-466-4901

FAX: 757-461-4148

PROJECT INFORMATION

PROJECT NAME: NWIRP Bethpage

PROJECT #: 112G08005-WE13 LOCATION: RW7B

PROJECT MANAGER: Ernie Wu

E-MAIL: ernie.wu@tetrattech.com

PHONE: 757-466-4901

FAX: 757-461-4148

BILLING INFORMATION

BILL TO:

PO#

ADDRESS:

CITY:

STATE: ZIP:

ATTENTION:

PHONE:

DATA TURNAROUND INFORMATION

FAX: 10 DAYS*

HARD COPY: 10 DAYS*

EDD 10 DAYS*

* TO BE APPROVED BY CHEMTECH

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ RESEULTS ONLY☐ RESULTS + QC☐ New Jersey REDUCED☐ New Jersey CLP☐ EDD Format☐ USEPA CLP☐ New York State ASP "B"☐ New York State ASP "A"☐ Other

ANALYSIS

1,4-Dioxane SW846 8270
SIM

1

2

3

4

5

6

7

8

9

PRESERVATIVES

COMMENTS

<-- Specify Preservatives

A-HCl B-HNO3

C-H2SO4 D-NaOH

E-ICE F-Other

CHEMTECH
SAMPLE
IDPROJECT
SAMPLE IDENTIFICATIONSAMPLE
MATRIXSAMPLE
TYPE
COMP GRABSAMPLE
COLLECTION

DATE TIME

of Bottles

1

2

3

4

5

6

7

8

9

1.

RW7-SP100-20251030

GW

X

10/30/25

12:15

1

X

2.

RW7-SP201-20251030

GW

X

10/30/25

12:17

1

X

3.

RW7-SP302-20251030

GW

X

10/30/25

12:27

1

X

4.

RW7-SP303-20251030

GW

X

10/30/25

12:29

1

X

5.

6.

7.

8.

9.

10.

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER

DATE/TIME

RECEIVED BY

1.

10/31/25/1500

1.

11-3-25
0700

Conditions of bottles or coolers at receipt:

☐ Compliant☐ Non Compliant☐ Cooler Temp 2-3°C

MeOH extraction requires an additional 4oz. Jar for percent solid

☐ Ice in Cooler?:

RELINQUISHED BY

DATE/TIME

RECEIVED BY

2.

2.

RELINQUISHED BY

DATE/TIME

RECEIVED FOR LAB BY

3.

3.

Page ____ of ____

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ OvernightCHEMTECH: ☐ Picked Up ☐ Overnight

Shipment Complete

☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT

YELLOW - CHEMTECH COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
 Data File : BN038114.D
 Acq On : 28 Oct 2025 14:49
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Oct 28 17:31:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 28 17:15:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	9110	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	27131	0.400	ng	0.01
13) Acenaphthene-d10	14.441	164	14920	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	28506	0.400	ng	0.00
29) Chrysene-d12	21.375	240	23535	0.400	ng	0.00
35) Perylene-d12	23.683	264	20307	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	16552	0.771	ng	0.00
5) Phenol-d6	6.937	99	19989	0.765	ng	0.00
8) Nitrobenzene-d5	8.929	82	16332	0.746	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	29806	0.769	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	4525	0.736	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	46843	0.784	ng	0.01
27) Fluoranthene-d10	19.220	212	55912	0.778	ng	0.00
31) Terphenyl-d14	19.819	244	37975	0.743	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.254	88	7707	0.819	ng	99
3) n-Nitrosodimethylamine	3.564	42	9598	0.792	ng	# 100
6) bis(2-Chloroethyl)ether	7.197	93	20444	0.796	ng	99
9) Naphthalene	10.626	128	58627	0.784	ng	99
10) Hexachlorobutadiene	10.925	225	9875	0.783	ng	# 98
12) 2-Methylnaphthalene	12.248	142	38674	0.776	ng	98
16) Acenaphthylene	14.152	152	52624	0.779	ng	99
17) Acenaphthene	14.494	154	36845	0.779	ng	100
18) Fluorene	15.489	166	49204	0.794	ng	100
20) 4,6-Dinitro-2-methylph...	15.560	198	2898	0.689	ng	# 69
21) 4-Bromophenyl-phenylether	16.379	248	14571	0.780	ng	97
22) Hexachlorobenzene	16.503	284	16700	0.786	ng	99
23) Atrazine	16.652	200	10204	0.745	ng	99
24) Pentachlorophenol	16.838	266	6470	0.707	ng	99
25) Phenanthrene	17.223	178	70455	0.779	ng	100
26) Anthracene	17.322	178	60270	0.761	ng	100
28) Fluoranthene	19.253	202	79478	0.786	ng	100
30) Pyrene	19.615	202	78307	0.737	ng	100
32) Benzo(a)anthracene	21.357	228	64140	0.764	ng	100
33) Chrysene	21.411	228	71675	0.787	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	32963	0.745	ng	99
36) Indeno(1,2,3-cd)pyrene	26.039	276	65796	0.789	ng	99
37) Benzo(b)fluoranthene	23.031	252	67771	0.781	ng	94
38) Benzo(k)fluoranthene	23.031	252	67771	0.776	ng	# 93
39) Benzo(a)pyrene	23.578	252	53367	0.768	ng	# 90
40) Dibenzo(a,h)anthracene	26.054	278	50344	0.784	ng	95
41) Benzo(g,h,i)perylene	26.753	276	58155	0.793	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102825\
Data File : BN038114.D
Acq On : 28 Oct 2025 14:49
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Quant Time: Oct 28 17:31:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 28 17:15:13 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.775 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8







Tgt Ion:152 Resp: 9110
 Ion Ratio Lower Upper
 152 100
 150 157.1 122.3 183.5
 115 63.4 47.4 71.0



#2
 1,4-Dioxane
 Concen: 0.819 ng
 RT: 3.254 min Scan# 24
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion: 88 Resp: 7707
 Ion Ratio Lower Upper
 88 100
 43 30.6 24.3 36.5
 58 74.3 60.4 90.6









#3
 n-Nitrosodimethylamine
 Concen: 0.792 ng
 RT: 3.564 min Scan# 61
 Delta R.T. 0.007 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion: 42 Resp: 9598
 Ion Ratio Lower Upper
 42 100
 74 120.8 96.8 145.2
 44 8.8 5.3 7.9#



#4
 2-Fluorophenol
 Concen: 0.771 ng
 RT: 5.348 min Scan# 314
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion: 112 Resp: 16552
 Ion Ratio Lower Upper
 112 100
 64 55.7 45.4 68.0
 63 30.2 23.2 34.8









#5
Phenol-d6
Concen: 0.765 ng
RT: 6.937 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 19989
Ion Ratio Lower Upper
99 100
42 20.7 16.6 24.8
71 33.2 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.796 ng
RT: 7.197 min Scan# 570
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion: 93 Resp: 20444
Ion Ratio Lower Upper
93 100
63 75.0 59.3 88.9
95 31.6 25.2 37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	27131
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	6.3	5.2	7.8
68	5.2	4.1	6.1



#8
Nitrobenzene-d5
Concen: 0.746 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:	82	Resp:	16332
Ion	Ratio	Lower	Upper
82	100		
128	41.9	34.1	51.1
54	53.4	43.5	65.3









#9
Naphthalene
Concen: 0.784 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 58627
Ion Ratio Lower Upper
128 100
129 11.0 9.1 13.7
127 13.2 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.783 ng
RT: 10.925 min Scan# 969
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:225 Resp: 9875
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.8 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.769 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 29806
Ion Ratio Lower Upper
152 100
151 21.2 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.776 ng
RT: 12.248 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:142 Resp: 38674
Ion Ratio Lower Upper
142 100
141 90.1 70.3 105.5
115 35.9 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.441 min Scan# 1453
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 14920
Ion Ratio Lower Upper
164 100
162 104.0 83.0 124.4
160 51.9 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 0.736 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:330 Resp: 4525
Ion Ratio Lower Upper
330 100
332 95.3 76.6 114.8
141 41.9 34.1 51.1









#15
2-Fluorobiphenyl
Concen: 0.784 ng
RT: 13.062 min Scan# 1324
Delta R.T. 0.011 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 46843
Ion Ratio Lower Upper
172 100
171 34.3 27.8 41.6
170 22.4 18.1 27.1



#16
Acenaphthylene
Concen: 0.779 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:152 Resp: 52624
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.8
153 12.9 10.2 15.2





#17

Acenaphthene

Concen: 0.779 ng

RT: 14.494 min Scan# 1458

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:154 Resp: 36845

Ion Ratio Lower Upper

154 100

153 112.3 90.0 135.0

152 59.3 47.3 70.9



#18

Fluorene

Concen: 0.794 ng

RT: 15.489 min Scan# 1551

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49




Tgt Ion:166 Resp: 49204


Ion Ratio Lower Upper

166 100

165 98.7 79.0 118.4

167 13.3 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:188 Resp: 28506
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.1 8.6 13.0



#20
4,6-Dinitro-2-methylphenol
Concen: 0.689 ng
RT: 15.560 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:198 Resp: 2898
Ion Ratio Lower Upper
198 100
51 69.5 88.9 133.3#
105 49.4 49.2 73.8









#21
4-Bromophenyl-phenylether
Concen: 0.780 ng
RT: 16.379 min Scan# 1623
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 14571
Ion Ratio Lower Upper
248 100
250 93.7 76.2 114.2
141 68.4 57.6 86.4



#22
Hexachlorobenzene
Concen: 0.786 ng
RT: 16.503 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:284 Resp: 16700
Ion Ratio Lower Upper
284 100
142 38.0 30.5 45.7
249 30.3 23.5 35.3









#23
Atrazine
Concen: 0.745 ng
RT: 16.652 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:200 Resp: 10204
Ion Ratio Lower Upper
200 100
173 26.2 21.4 32.2
215 46.0 36.7 55.1



#24
Pentachlorophenol
Concen: 0.707 ng
RT: 16.838 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:266 Resp: 6470
Ion Ratio Lower Upper
266 100
264 62.0 49.3 73.9
268 64.3 50.6 75.8









#25
Phenanthrene
Concen: 0.779 ng
RT: 17.223 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 70455
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.3 12.2 18.4



#26
Anthracene
Concen: 0.761 ng
RT: 17.322 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:178 Resp: 60270
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.2 11.8 17.8









#27
Fluoranthene-d10
Concen: 0.778 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 55912
Ion Ratio Lower Upper
212 100
106 16.0 12.6 19.0
104 9.0 7.1 10.7



#28
Fluoranthene
Concen: 0.786 ng
RT: 19.253 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:202 Resp: 79478
Ion Ratio Lower Upper
202 100
101 11.8 9.4 14.2
203 17.0 13.4 20.2





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.375 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 23535

Ion Ratio Lower Upper

240 100

120 13.1 10.7 16.1

236 28.4 23.1 34.7



#30

Pyrene

Concen: 0.737 ng

RT: 19.615 min Scan# 2032

Delta R.T. 0.005 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49




Tgt Ion:202 Resp: 78307


Ion Ratio Lower Upper

202 100

200 21.2 16.7 25.1

203 17.7 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.743 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244 Resp: 37975
 Ion Ratio Lower Upper
 244 100
 212 9.3 7.8 11.6
 122 18.4 15.2 22.8



#32
 Benzo(a)anthracene
 Concen: 0.764 ng
 RT: 21.357 min Scan# 2267
 Delta R.T. -0.000 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49

Tgt Ion:228 Resp: 64140
 Ion Ratio Lower Upper
 228 100
 226 27.7 22.2 33.4
 229 19.6 15.8 23.8





#33

Chrysene

Concen: 0.787 ng

RT: 21.411 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:228 Resp: 71675

Ion Ratio Lower Upper

228 100

226 30.1 24.3 36.5

229 19.6 15.7 23.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.745 ng

RT: 21.295 min Scan# 2260

Delta R.T. -0.000 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49




Tgt Ion:149 Resp: 32963


Ion Ratio Lower Upper

149 100

167 26.8 21.1 31.7

279 3.1 2.9 4.3









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.683 min Scan# 21
 Delta R.T. 0.003 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:264 Resp: 20307
 Ion Ratio Lower Upper
 264 100
 260 26.8 21.8 32.8
 265 76.5 68.6 102.8



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.789 ng
 RT: 26.039 min Scan# 3627
 Delta R.T. 0.006 min
 Lab File: BN038114.D
 Acq: 28 Oct 2025 14:49


Tgt Ion:276 Resp: 65796
 Ion Ratio Lower Upper
 276 100
 138 27.3 21.4 32.0
 277 24.7 19.2 28.8








#37
Benzo(b)fluoranthene
Concen: 0.781 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.050 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8







Tgt Ion:252 Resp: 67771
Ion Ratio Lower Upper
252 100
253 23.4 20.9 31.3
125 14.2 14.1 21.1



#38
Benzo(k)fluoranthene
Concen: 0.776 ng
RT: 23.031 min Scan# 2598
Delta R.T. 0.003 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Tgt Ion:252 Resp: 67771
Ion Ratio Lower Upper
252 100
253 23.4 21.2 31.8
125 14.2 14.6 22.0#





#39

Benzo(a)pyrene

Concen: 0.768 ng

RT: 23.578 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN038114.D


Acq: 28 Oct 2025 14:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 53367

Ion Ratio Lower Upper

252 100

253 24.4 23.1 34.7

125 17.2 18.6 28.0#



#40

Dibenzo(a,h)anthracene

Concen: 0.784 ng

RT: 26.054 min Scan# 3632

Delta R.T. 0.006 min

Lab File: BN038114.D

Acq: 28 Oct 2025 14:49




Tgt Ion:278 Resp: 50344


Ion Ratio Lower Upper

278 100

139 19.9 17.6 26.4


279 25.0 22.6 33.8





#41
Benzo(g,h,i)perylene
Concen: 0.793 ng
RT: 26.753 min Scan# 3871
Delta R.T. 0.003 min
Lab File: BN038114.D
Acq: 28 Oct 2025 14:49

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:	276	Resp:	58155
Ion Ratio	Lower	Upper	
276	100		
277	24.0	19.7	29.5
138	24.9	20.2	30.4

