

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH QC BATCHID: LB137761

	NA	NO	YES
1. Calibration Summary met criteria.	_____	_____	_____✓
2. ICP Interference Check Sample Results Summary Submitted.	_____	_____	_____✓
3. Serial Dilution Summary (if applicable) Submitted.	_____	_____	_____✓
4. Laboratory Control Sample Summary (if applicable) Submitted.:	_____	_____	_____✓
5. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓	_____
6. Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
7. Sample Duplicate Analysis Met QC Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
8. Digestion Holding Time Met If not met, list number of days exceeded for each sample:	_____	_____	_____✓
9. Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓

ADDITIONAL COMMENTS:

Q3520-01 (Sb,As,Ba,Be,Cr,Se,Ag,Na,V) MS/ MSD FAILS : this fails due to unknown interference during digestion as sample is having black tiny particles and stones with soil and this could be possible cause of failure.

Analyst

Date

QA REVIEW

Date