

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3527

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4030	4000	101	90 - 110	P	11/04/2025	11:58	LB137761

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3527

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	101	100	101	80 - 120	P	11/04/2025	12:16	LB137761

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3527

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5010	5000	100	90 - 110	P	11/04/2025	12:56	LB137761
CCV02	Iron	5090	5000	102	90 - 110	P	11/04/2025	13:55	LB137761
CCV03	Iron	5000	5000	100	90 - 110	P	11/04/2025	15:06	LB137761
CCV04	Iron	5000	5000	100	90 - 110	P	11/04/2025	16:10	LB137761
CCV05	Iron	5030	5000	101	90 - 110	P	11/04/2025	16:55	LB137761