

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc.

SDG No.: Q3527

Contract: TETR06

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	99300	101000	98	85600	116500	11/04/2025	12:34	LB137761
ICSAB01	Iron	99600	99300	100	84400	114500	11/04/2025	12:38	LB137761