

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q3527
contract: TETR06 **lab code:** ACE
matrix: Water **sample id:** Q3527-02 **client id:** RW8-SP303-20251030MS
Percent Solids for Sample: NA **Spiked ID:** Q3527-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1460		40.2	J	1500	95		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q3527
contract: TETR06 **lab code:** ACE
matrix: Water **sample id:** Q3527-02 **client id:** RW8-SP303-20251030MSD
Percent Solids for Sample: NA **Spiked ID:** Q3527-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1500		40.2	J	1500	98		P