

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3527

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038133.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Low	Limits	
								Qual			High	RPD
PB170381BSD	1,4-Dioxane	0.4	0.26	ug/L	65	24	*	*		70	130	20