

Continuing Calibration Report for SequenceID : FE121125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057308.D

Aliphatic C9-C12	11793375.000	60.000	3.187	6.813	196556.250	200655.681	2.043
Aliphatic C12-C16	8076264.000	40.000	6.814	10.258	201906.600	203177.346	0.625
Aliphatic C16-C21	11796188.000	60.000	10.259	13.632	196603.133	198315.768	0.864
Aliphatic C21-C28	14853828.000	80.000	13.633	17.300	185672.850	198747.045	6.578
Aliphatic C28-C40	21042467.000	120.000	17.301	22.225	175353.892	199899.890	12.279
Aliphatic EPH	67562122.000	360.000	3.187	22.225	187672.561	199869.809	6.103