

Continuing Calibration Report for SequenceID : FE121125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057319.D

Aliphatic C9-C12	12457633.000	60.000	3.187	6.813	207627.217	200655.681	-3.474
Aliphatic C12-C16	8535955.000	40.000	6.814	10.258	213398.875	203177.346	-5.031
Aliphatic C16-C21	12376951.000	60.000	10.259	13.632	206282.517	198315.768	-4.017
Aliphatic C21-C28	15642691.000	80.000	13.633	17.300	195533.638	198747.045	1.617
Aliphatic C28-C40	23195701.000	120.000	17.301	22.225	193297.508	199899.890	3.303
Aliphatic EPH	72208931.000	360.000	3.187	22.225	200580.364	199869.809	-0.356