

Continuing Calibration Report for SequenceID : FE121125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057329.D

Aliphatic C9-C12	12386777.000	60.000	3.187	6.813	206446.283	200655.681	-2.886
Aliphatic C12-C16	8503241.000	40.000	6.814	10.258	212581.025	203177.346	-4.628
Aliphatic C16-C21	12423919.000	60.000	10.259	13.632	207065.317	198315.768	-4.412
Aliphatic C21-C28	15716092.000	80.000	13.633	17.300	196451.150	198747.045	1.155
Aliphatic C28-C40	23115468.000	120.000	17.301	22.225	192628.900	199899.890	3.637
Aliphatic EPH	72145497.000	360.000	3.187	22.225	200404.158	199869.809	-0.267