

Continuing Calibration Report for SequenceID : FF120825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016859.D

Aliphatic C9-C12	7397810.000	60.000	3.203	6.839	123296.833	121796.865	-1.232
Aliphatic C12-C16	4980633.000	40.000	6.840	10.282	124515.825	124246.420	-0.217
Aliphatic C16-C21	7781541.000	60.000	10.283	13.656	129692.350	131451.873	1.339
Aliphatic C21-C28	10282376.000	80.000	13.657	17.329	128529.700	137206.394	6.324
Aliphatic C28-C40	13787320.000	120.000	17.330	22.276	114894.333	110903.455	-3.599
Aliphatic EPH	44229680.000	360.000	3.203	22.276	122860.222	123471.409	0.495