

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

ProjectID: LOD-LOQ-MDL Study

Lab Code: ACE

SDG No.: Q3530

Calibration Sequence : FF102925			Test : TPH GC	
Concentration	(PPM)	Area Count	Reference Factor	File ID
85		9472915	111446	FF016603.D
170		18165984	106859	FF016604.D
340		37023074	108891	FF016605.D
850		89698209	105527	FF016606.D
1700		169764046	99861	FF016607.D
AVG RF : 106517		% RSD : 4.074		AVG RT : 15.059