

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: ALLI03
ProjectID: LOD-LOQ-MDL Study
Lab Code: ACE SDG No.: Q3530
DataFile: FF016774.D Analyst Name: YP\AJ Analyst Date: 11-21-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	99553872	117122	106517	9.956