

Manual Integration Report

Sample ID	ClientID ID	File ID	Sequence ID	Parameter	Supervised By	Supervised On	Reason
50 PPM TRPH STD		FF016759.D	FF112025	N-OCTACOSANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
50 PPM TRPH STD		FF016759.D	FF112025	N-TETRACONTANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
RT MARKER		FF016760.D	FF112025	N-OCTACOSANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
PB170659BS		FF016762.D	FF112025	N-OCTACOSANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
PB170659BS		FF016762.D	FF112025	N-TETRACONTANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
Q3530-04		FF016764.D	FF112025	N-TETRACONTANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
Q3530-05		FF016765.D	FF112025	N-TETRACONTANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
Q3661-01MS		FF016767.D	FF112025	N-OCTATRIACONTANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
Q3661-01MS		FF016767.D	FF112025	N-TETRACONTANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
Q3661-01MSD		FF016768.D	FF112025	N-TETRACONTANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
Q3530-10		FF016771.D	FF112025	N-TETRACONTANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
Q3530-11		FF016772.D	FF112025	N-TETRACONTANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly
50 PPM TRPH STD		FF016774.D	FF112025	N-TETRACONTANE	mohammad	11/22/2025 12:08:27 AM	Peak Integrated by Software incorrectly