

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Elegant Jewelers Mfg. Co. Inc.</u>	level: <u>low</u>	sdg no.: <u>Q3531</u>
contract: <u>ELEG01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3575-02</u>	client id: <u>002 35th Ave (Nov)MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3575-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	159		2.48	J	150	104		P
Zinc	ug/L	75 - 125	113		33.9		100	80		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Elegant Jewelers Mfg. Co. Inc.</u>	level:	<u>low</u>	sdg no.:	<u>Q3531</u>
contract:	<u>ELEG01</u>			lab code:	<u>ACE</u>
matrix:	<u>Water</u>	sample id:	<u>Q3575-02</u>	client id:	<u>002 35th Ave (Nov)MSD</u>
Percent Solids for Sample:	NA	Spiked ID:	Q3575-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	161		2.48	J	150	106		P
Zinc	ug/L	75 - 125	115		33.9		100	81		P