

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q3531

Contract: ELEG01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170461BS							
Copper	ug/L	150	153		102	85 - 115	P
Zinc	ug/L	100	104		104	85 - 115	P