

## Analytical Summary Report

Analysis Method: 9040C  
Parameter: pH  
Run Number: LB137776

Analyst By : jignesh  
Supervisor Review By : Iwona  
Slope : 98.6  
pH Meter ID : WC PH METER-1

| Calibration Standards           | Chemtech Log# |
|---------------------------------|---------------|
| PH 4 BUFFER SOLUTION            | W3178         |
| BUFFER PH 7.00 GREEN 1PINT PK6  | W3093         |
| PH 10.01 BUFFER, COLOR CD 475ML | W3191         |
| buffer solution pH 7 yellow     | W3217         |
| Buffer Solution, PH2 (500ml)    | W3161         |
| pH 12.00 Buffer                 | W3200         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

True Value of CCV3 = 2.00 Control Limits[+/- 0.05].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 11/05/2025 | 10:10     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.3             | 7.00        | 11/05/2025 | 10:11     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.4             | 10.03       | 11/05/2025 | 10:14     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.1             | 6.98        | 11/05/2025 | 10:15     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.3             | 2.02        | 11/05/2025 | 10:19     |
| 6   | Q3534-01    | 1  | Water  | NA          | NA          | 20.6             | 6.20        | 11/05/2025 | 10:25     |
| 7   | Q3534-02    | 1  | Water  | NA          | NA          | 20.4             | 6.52        | 11/05/2025 | 10:26     |
| 8   | Q3534-03    | 1  | Water  | NA          | NA          | 20.5             | 6.64        | 11/05/2025 | 10:33     |
| 9   | Q3534-04    | 1  | Water  | NA          | NA          | 20.9             | 6.56        | 11/05/2025 | 10:37     |
| 10  | Q3534-05    | 1  | Water  | NA          | NA          | 20.1             | 6.32        | 11/05/2025 | 10:40     |
| 11  | Q3534-06    | 1  | Water  | NA          | NA          | 20.8             | 6.56        | 11/05/2025 | 10:47     |
| 12  | Q3534-07    | 1  | Water  | NA          | NA          | 20.8             | 6.70        | 11/05/2025 | 10:50     |
| 13  | Q3537-03    | 1  | Water  | NA          | NA          | 20.6             | 6.80        | 11/05/2025 | 11:00     |
| 14  | Q3541-01    | 1  | Water  | NA          | NA          | 20.7             | 8.49        | 11/05/2025 | 11:10     |
| 15  | Q3542-01    | 1  | Water  | NA          | NA          | 20.6             | 9.70        | 11/05/2025 | 11:20     |
| 16  | CCV2        | 1  | Water  | NA          | NA          | 20.3             | 12.02       | 11/05/2025 | 11:25     |
| 17  | Q3542-01DUP | 1  | Water  | NA          | NA          | 20.8             | 9.72        | 11/05/2025 | 11:27     |
| 18  | CCV3        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 11/05/2025 | 11:30     |

# WORKLIST(Hardcopy Internal Chain)

137776

WorkList Name : ph q w3542      WorkList ID : 192887      Department : Wet-Chemistry      Date : 11-05-2025 08:39:39

| Sample   | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|------|--------------|----------|-----------------------------|--------------|--------|
| Q3534-01 | MW-6            | Water  | pH   | Cool 4 deg C | REMI02   | D31                         | 11/03/2025   | 9040C  |
| Q3534-02 | MW-1            | Water  | pH   | Cool 4 deg C | REMI02   | D31                         | 11/03/2025   | 9040C  |
| Q3534-03 | MW-11           | Water  | pH   | Cool 4 deg C | REMI02   | D31                         | 11/03/2025   | 9040C  |
| Q3534-04 | MW-5            | Water  | pH   | Cool 4 deg C | REMI02   | D31                         | 11/03/2025   | 9040C  |
| Q3534-05 | MW-EB-1         | Water  | pH   | Cool 4 deg C | REMI02   | D31                         | 11/04/2025   | 9040C  |
| Q3534-06 | MW-EB-2         | Water  | pH   | Cool 4 deg C | REMI02   | D31                         | 11/03/2025   | 9040C  |
| Q3534-07 | MW-EB-3         | Water  | pH   | Cool 4 deg C | REMI02   | D31                         | 11/04/2025   | 9040C  |
| Q3537-03 | MW-17-WATER     | Water  | pH   | Cool 4 deg C | REMI02   | D31                         | 11/03/2025   | 9040C  |
| Q3541-01 | N48969          | Water  | pH   | Cool 4 deg C | PSEG03   | D41                         | 11/04/2025   | 9040C  |
| Q3542-01 | AUD-25-182-194  | Water  | pH   | Cool 4 deg C | PSEG03   | D41                         | 11/04/2025   | 9040C  |

Date/Time 11/05/25 08:45

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 11/05/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]